



PHILIPS

Philips Medical Systems

Federal Communications Commission
445 12th Street, SW
Washington, DC 20554

January 13, 2006

Re: FCC ID: PQC-WMTS-MODULE

To Whom It May Concern:

Philips Medical Systems of Andover, Massachusetts is submitting documentation for a Class II Permissive Change pursuant to 47 CFR 2.1043 for FCC Identifier PQC-WMTS-MODULE.

The modification is only the removal of the aluminum housing to enable the incorporation of the module into a medical device (i.e. space constraints). There are no changes to the electrical design of the radio, i.e., no changes to the basic frequency determining and stabilizing circuitry, clocks, data rates, RF shielding, frequency multiplication stages, modulator circuitry or maximum power or field strength ratings. The aluminum bracket is intended for mounting purposes only, and its removal does not compromise RF shielding integrity, which is provided by a separate shielded enclosure within the module.

The supplied test report to 47 CFR Part 95 includes the test data of the module without the aluminum housing. Due to an improvement in the set-up for the test procedures, test results are not identical to the originally submitted data (27-Apr-05), specifically as pertains to radiated emissions, however all test results are well within the allowable margins and fully meet the technical requirements as defined in 47 CFR Part 95.1115.

Sincerely,

Denise Haley
Quality & Regulatory Engineer
3000 Minuteman Road, MS 0240
Andover, MA 01810
Tel: 978-659-4358
e-mail: denise.haley@philips.com

3000 Minuteman Road
Andover, Massachusetts 01810
Tel: (978) 687 1501, Fax: (978) 794 7646
www.medical.philips.com