

## Philips Medical Systems

Federal Communications Commission Washington DC 20554

21 February 2002,

Dear Sir or Madam,

Philips Medical Systems (formerly Agilent Technologies) of Andover Massachusetts currently desires to submit an application for Equipment Authorization under Part 95, Subpart H for a wireless medical telemetry system (the M2601A telemetry transmitter).

The M2601A currently has a Grant of Equipment Authorization under Part 15 for the range of 590-632 MHz (FCC Identifier PQCM2601A). This Form 731 application will be for the exact same device but specifically for the WMTS band of 608-614 MHz under Part 95.

Upon receipt of this grant, additional labeling will be applied to our devices per 47 CFR 95.1109.

In addition, I wish to modify the address and contact name for Grantee Code PQC to Denise Haley, as stated below. The previous contact name (T. Sweeney) and address (Bothell WA) were temporary as Agilent transferred to Philips Medical. Under FRN 0005142203, TIN 133429115, my name and address have already been modified in the CORES system.

Thank you for your consideration of this application, and please do not hesitate to contact me should you require additional information.

Sincerely,

Denise Haley Quality and Regulatory Engineer Philips Medical Systems 3000 Minuteman Road Mail Stop 0490 Andover, MA 01810-1099 Tel: 978-659-2701 Fax. 978-685-5624 Denise\_haley@hsgmed.com

formerly: Agilent Technologies Healthcare Solutions Group

3000 Minuteman. Road Andover, Massachusetts 01810 Tel: (978) 687 1501, Fax: (978) 794 7646 www.medical.philips.com