

**Philips Medical Systems**



To: Federal Communications Commission, Washington, DC 20554  
Attention: Ms. Diane Poole  
Date: October 25, 2004  
From: Barry Wyshogrod, Philips Medical Systems, Andover, MA  
Subject: **Reply to Correspondences, References #27853, #27854  
731 Application #EA408759**

Dear Ms. Poole and FCC Reviewers,

On 22-Oct-04, you e-mailed me the following request:

Please add the FCC statement on the label as in 15.19 (a) of the FCC rules to the product.

Philips Medical Systems' response:

I apologize for my oversight and will implement your request. We have added the appropriate text to the product's FCC ID Labeling, as per my phone conversation with you, Diane. We will add the text as per 47 CFR Part 15.19(a)(1):

"This device complies with part 15 of the FCC Rules. Operation is subject to the condition that this device does not cause harmful interference."

I hereby attach a revised FCC ID Label for the device. This attachment supercedes the previous label submitted on 19-Oct-04 as part of my reply to correspondences #27768, #27769, #27770.

Diane, thank you for your help. For any further issues, I can be reached at **(978) 659-7383** or by e-mail to: [barry.wyshogrod@philips.com](mailto:barry.wyshogrod@philips.com). If I am not available, please contact Ms. Denise Haley at **(978) 659-4358** or by e-mail to: [denise.haley@philips.com](mailto:denise.haley@philips.com).

Sincerely yours,

Barry Wyshogrod, Regulatory Engineer  
*Signature already on-file at the FCC*