

Test Laboratory: Compliance Certification Services Inc.

D2450V2 SN-728 MSL 2450

Communication System: CW; Frequency: 2450 MHz; Duty Cycle: 1:1
Medium parameters used: $f = 2450$ MHz; $\sigma = 1.971$ mho/m; $\epsilon_r = 53.694$; $\rho = 1000$ kg/m³
Phantom section: Flat Section
Air Temperature: 24.2 deg C; Liquid Temperature: 23.2 deg C
Area Scan Find Secondary Maximum Within 2dB and with a peak SAR value greater than 0.0012W/kg

DASY Configuration:

- Probe: EX3DV4 - SN3665; ConvF(7.11, 7.11, 7.11); Calibrated: 4/27/2012;
- Sensor-Surface: 4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn877; Calibrated: 3/16/2012
- Phantom: ELI 4.0; Type: QDOVA001BA; Serial: 1056
- Measurement SW: DASY52, Version 52.8 (1)SEMCAD X Version 14.6.5 (6469)

Pin=250mW,d=10mm/Area Scan (7x9x1): Measurement grid: dx=15mm, dy=15mm
Maximum value of SAR (measured) = 14.2 mW/g

Pin=250mW,d=10mm/Zoom Scan (7x7x7)/Cube 0: Measurement grid: dx=5mm, dy=5mm, dz=5mm
Reference Value = 84.464 V/m; Power Drift = 0.24 dB
Peak SAR (extrapolated) = 27.686 mW/g
SAR(1 g) = 13.7 mW/g; SAR(10 g) = 6.43 mW/g
Maximum value of SAR (measured) = 15.7 mW/g

