

12 April 2022

Element Materials Technology

6775 NE Evergreen Pkwy

#400

Hillsboro, OR 97124

FCC ID: PX2-GMIPG1

To whom it may concern,

We, Abbott Laboratories (Neuromodulation Division 6901 Preston Road, Plano, TX 75024), are submitting an equipment authorization for an Original certification for the following model which is part of a neurostimulation therapy system. Details of the model name variations, system compatibility and system functionality can be found within the operational description and manual (IFU).

IPG Model 32400

Rule part 15.247 is applicable since it uses Bluetooth Low Energy technology as described in KDB 558074 Section 8. The product operates at 0dBm or less when factoring in low duty cycle operation. The IPG model is an implantable device that operate within the torso.

Manasi Khare

Director, Regulatory Affairs