

# Label

Product Name : Medical Tablet PC  
Model No : M101MDT  
Rating : +19V  $\Rightarrow$  3.15A  
For use only Power Supply :  
GlobTek (GTM96600-6019-T3)  
Input : 100-240Vac, 50-60Hz, 1.5A,  
Output : 19VDC 3.15A Max.  
Tablet PC Input Rating : 19 VDC 3.15A

This device complies with part 15 of the FCC Rules. Operation is subject to the following two conditions:  
(1) This device may not cause harmful interference, and  
(2) This device must accept any interference received, including interference that may cause undesired operation. This Class B digital apparatus meets all requirements of the Canadian Interference-Causing Equipment Regulations. Cet appareil numérique de la classe B respecte toutes les exigences du Règlement sur le matériel brouilleur du Canada.



ISO 7000-1641  
Operating instructions



E489236

Complies with 21CFR 1040.10 and 1040.11 except for deviations pursuant to Laser Notice No.50, dated June 24 2007

MEDICAL - GENERAL MEDICAL EQUIPMENT AS TO ELECTRICAL SHOCK, FIRE AND MECHANICAL HAZARDS ONLY IN ACCORDANCE WITH ANSI/AAMI ES 60601-1 (2005) + AMD (2012) and CAN/CSA-C22.2 No.6060-1 (2014)



Contains  
FCC ID : PX9-AC9260NGW  
IC : 9911A-AC9260NGW

Winmate Inc.  
9F, No.111-6, Shing-De Rd., San-Chung Dist.,  
New Taipei City 24158, Taiwan

Manufactured Date :

Part number of this label 91D00000344H

MADE IN TAIWAN

# Label location

