

Please see below data with RF scan data centered on the T-coil for the worst-case configurations. The RT T-coil emission category is M3.



Peak E-field in V/m

Grid 1	Grid 2	Grid 3
57.2	50.2	44.2
Grid 4	Grid 5	Grid 6
60.1	52.7	45.2
Grid 7	Grid 8	Grid 9
58.5	50.4	43.5



Cellular E-field – Max 60.1 V/m (M4)

Peak H-field in A/m

Grid 1	Grid 2	Grid 3
0.115	0.086	0.062
Grid 4	Grid 5	Grid 6
0.112	0.083	0.062
Grid 7	Grid 8	Grid 9
0.095	0.069	0.055



Cellular H-field – Max 0.115 A/m (M4)

Peak E-field in V/m

Grid 1	Grid 2	Grid 3
49.6	39.2	36.9
Grid 4	Grid 5	Grid 6
61.4	43.8	37.4
Grid 7	Grid 8	Grid 9
61.2	42.3	37.1



PCS E-field – Max 61.4 V/m (M3)

Peak H-field in A/m

Grid 1	Grid 2	Grid 3
0.051	0.057	0.066
Grid 4	Grid 5	Grid 6
0.065	0.075	0.076
Grid 7	Grid 8	Grid 9
0.106	0.095	0.087



PCS H-field – Max 0.106 A/m (M4)