

Variant model list

Date: February 18, 2019

UL Japan, Inc. 4383-326 Asama-cho, Ise-shi, Mie 516-0021 Japan

FCC ID : Q6ZHEM7340T

Applicant: OMRON HEALTHCARE Co., Ltd

To whom it may concern:

We hereby declare that the following models are essentially the same devices except presence or absence of LCD backlight.

| Model No. | HEM-7342T | HEM-7340T |
|-----------|------------|-----------|
| | (presence) | (absence) |

We hereby declare that the following models are essentially the same devices except sales destination.

| | | |
|-----------|-----------|--------------|
| Model No. | HEM-7342T | HEM-7343T |

We hereby declare that the following models are essentially the same devices except sales destination.

| Model No. | HEM-7340T | HEM-7341T |
|-----------|-----------|-----------|

These differences cause no influence to radio specification.

There was no degradation of EMC characteristic.

We can consider them electrically identical.

Thank you for your attention to this matter.

Sincerely,

Toshiaki Yuasa

OMRON HEALTHCARE Co., Ltd.