

RE: RF Module Model Number 0702-014-028 FCC statements
1/24/07

User Manual Statement

The product listed below is not marketed to a consumer as such. There is no user manual or advertising literature. Stryker Instruments will install all modules manufactured.

Furthermore, we understand that all products with this unit installed will contain the following information on the label:


Contains Stryker® RF Module Model Number 0702-014-028
FCC ID: Q9R-0702014028
IC: 4919A-014028

"This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation."

Where a device label is too small for the above statements, they will be placed in a user manual for the product with this module installed.

The user manuals for the products with this module installed will also have statements cautioning the user that changes or modifications not expressly approved by Stryker Instruments could void the user's authority to operate the equipment. Also, the following note will be conspicuously placed in the users manual: *"The term "IC:" before the radio certification number only signifies that Industry Canada technical specifications were met."* Lastly, the applicable RF Exposure Statement will also be included.

Greg L. Lawford
Manager R&D Power Tools



1/25/08

Paul A. Welch
Supervisor Product Approvals



1/25/08