Little Anne QCPR Sensor

Important Product Information



English Little Anne QCPR Sensor

A Warning states a condition, hazard, or unsafe practice that can result in serious personal injury or death.

A Caution states a condition, hazard, or unsafe practice that can result in minor personal injury or damage to the manikin.

Care, Maintenance and Cleaning

Clean with a dry cloth only.

↑ Caution

During use and assembly, do not touch the uncovered section of circuit board.

Battery

Consumer grade, non-rechargeable, carbon zinc or alkaline batteries, LR6 1.5V AA only to be used. Battery must be replaced with correct type.

♠ Caution

Risk of explosion if battery is replaced by an incorrect type. Dispose of used batteries according to the instructions.

Federal Communications Commission Statement and Industry Canada Statements

This device complies with part 15 of the FCC rules. Operation is subject to the following two conditions:

- 1. This device may not cause harmful interference
- This device must accept any interference received, including interference that may cause undesired operation.

△Caution

Changes or modifications not expressly approved by Laerdal Medical could void the user's authority to operate the equipment.

This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:

- · Reorient or relocate the receiving antenna.
- Increase the separation between the equipment and receiver.
- Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
- Consult the dealer or an experienced radio/TV technician for help.

Contains FCC ID: QHQ-120-60750 IC: 20263-12060750


Industry Canada ICES-003 Compliance: CAN ICES-3 (B)/NMB-3(B)

Korea Certification:

MSIP-CMM-Lm1-120-60750

English Little Anne OCPR Sensor

Symbol Glossary



Waste Handling

The Little Anne QCPR Sensor contains electronic components. Dispose of it at the applicable recycling facility in accordance with local regulations.

European directive 2012/19/EU (WEEE)
WEEE: this appliance is marked according to the
European directive 2012/19/EU on Waste Electrical
and Electronic Equipment (WEEE). By ensuring this
product is disposed of correctly, you will help prevent
potential negative consequences for the environment
and human health, which could otherwise be caused
by inappropriate waste handling of this product.

The symbol on the product, or on the documents accompanying the product, indicates that this appliance may not be treated as household waste. Instead it shall be handed over to the applicable collection point for the recycling of electrial and electronic equipment. Disposal must be carried out in accordance with local environmental regulations for waste disposal. For more detailed information about treatment, recovery and recycling of this product, please contact your local city office, your household waste disposal service or the Laerdal representative where you purchased the product.

Warranty

Refer to the Laerdal Global Warranty for terms and conditions. For more information visit www.laerdal.com

Specifications

-	
Operating temperature	15°C to 30°C (59 °F to 86 °F)
Storage temperature	-10 °C to 50 °C (14 °F to 122 °F)
Humidity	5% - 95% relative humidity
Weight	200 g (0.44 lbs)
Dimensions	165 × 220 × 55 mm (6.5 × 8.7 × 2.2 in)

Do not touch

© 2015 Laerdal Medical AS, All rights reserved.

Manufactured in China for: Laerdal Medical AS P.O. Box 377 Tanke Svilandsgate 30, 4002 Stavanger, Norway T: (+47) 51 51 17 00

Printed in China

.0-08939 Rev C

