



Laerdal
helping save lives

Tanke Svilandsgate 30, 377, Stavanger, Norway 4002

Date: February 22, 2016

Declaration

We, **Laerdal Medical AS**, declare that the device, FCC ID: **QHQ-20430250**,
Model Name: Link Box PLUS, does not support any non-US channels in all the
operational mode(s). All non-US frequencies, US 2.4G channel 12-13 and Country
code selection are disabled through proprietary software and are not user changeable.

Should you have any question or comment regarding this matter, please do not
hesitate to contact me.

Sincerely yours,

Mari Kaada / Corporate Regulatory Affairs Manager

Laerdal Medical AS

Tel: +4751511700

Fax: +4751511717

E-mail: mari.kaada@laerdal.no