



Product Service

---

**Choose certainty.  
Add value.**

# Report On

RF Exposure Assessment of the  
Laerdal Medical AS  
801-002xx CPRmeter 2 (Medical equipm.)

FCC ID: QHQ-801002  
IC: 20263-801002

Document 75932076 Report 06 Issue 1

February 2016



Product Service

TÜV SÜD Product Service, Octagon House, Concorde Way, Segensworth North,  
Fareham, Hampshire, United Kingdom, PO15 5RL  
Tel: +44 (0) 1489 558100. Website: [www.tuv-sud.co.uk](http://www.tuv-sud.co.uk)

**REPORT ON**

RF Exposure Assessment of the  
Laerdal Medical AS  
801-002xx CPRmeter 2 (Medical equipm.)

Document 75932076 Report 06 Issue 1

February 2016

**PREPARED FOR**

Laerdal Medical AS  
PO Box 377  
Tanke Svilandsgt. 30  
N-4002  
Stavanger  
Norway

**PREPARED BY**

Handwritten signature of Sara McRitchie in black ink.

**Sara McRitchie**  
Project Manager

**APPROVED BY**

Handwritten signature of Mark Jenkins in black ink.

**Mark Jenkins**  
Authorised Signatory

**DATED**

22 February 2016



Product Service

## CONTENTS

Section		Page No
<b>1</b>	<b>REPORT SUMMARY .....</b>	<b>3</b>
1.1	Introduction .....	4
1.2	Regional Requirements .....	5
1.3	Product Information .....	6
1.3.1	Technical Description .....	6
1.3.2	Supported Features .....	6
1.3.3	Antennas.....	6
1.4	Brief Summary of Results.....	7
<b>2</b>	<b>TEST DETAILS .....</b>	<b>9</b>
2.1	Rationale for Assessment of the RF Exposure .....	10
2.2	Test Result Details.....	11
<b>3</b>	<b>DISCLAIMERS AND COPYRIGHT.....</b>	<b>12</b>
3.1	Disclaimers and Copyright.....	13



Product Service

## **SECTION 1**

### **REPORT SUMMARY**

RF Exposure Assessment of the  
Laerdal Medical AS  
801-002xx CPRmeter 2 (Medical equipm.)



Product Service

## 1.1 INTRODUCTION

The information contained in this report is intended to show verification of the RF Exposure Assessment of the Laerdal Medical AS 801-002xx CPRmeter 2 (Medical equipm.) to the requirements of the applied test specifications.

Objective	To perform RF Exposure Assessment to determine the Equipment Under Test's (EUT's) compliance of the applied rules.
Applicant	Laerdal Medical AS
Manufacturer	Laerdal Medical AS
Manufacturing Description	CPRmeter 2 (Medical equipm.)
Model Number(s)	801-002xx
Test Specification/Issue/Date	EN 62311:2008 CFR 47 Pt1.1310 Health Canada Safety Code 6 ARPANSA Radiation Protection Series No.3



Product Service

## 1.2 REGIONAL REQUIREMENTS

The table below shows the regional requirements that are referenced in this test report. A full list of the requirements is shown in Annex A.

Report Reference	Regional Requirement
EU	EN 62311:2008
FCC	CFR 47 Pt1.1310
IC	Health Canada Safety Code 6
AUS	ARPANSA Radiation Protection Series No.3



Product Service

### 1.3 PRODUCT INFORMATION

#### 1.3.1 Technical Description

The Equipment under test was a Laerdal Medical AS 801-002xx CPRmeter 2 (Medical equipm.). A full technical description can be found in the manufacturer's documentation.

All reported calculations were carried out on the relevant information supplied for the 801-002xx CPRmeter 2 (Medical equipm.) to demonstrate compliance with the applied test specification(s). The sample assessed was found to comply with the requirements of the applied rules.

#### 1.3.2 Supported Features

The following radio access technologies and frequency bands are supported by the equipment under test.

Radio Access Technology	Bluetooth Low Energy
Frequency Band	2400 MHz to 2483.5 MHz

#### 1.3.3 Antennas

The following antennas are supported by the equipment under test.

No.	Model	Gain (dBi)
1	Integral Antenna	1.7



Product Service

#### 1.4 BRIEF SUMMARY OF RESULTS

The wireless device described within this report has been shown to be capable of compliance with the basic restrictions related to human exposure to electromagnetic fields for both General Public and Occupational. The calculations shown in this report were made in accordance the procedures specified in the applied test specification(s).

Required Compliance Boundary (m)	
Occupational	General Population
0.01	0.01

**Table 1 – Compliance Boundary Results**





Product Service

Regional Requirement	Calculated RF exposure level at compliance boundary of 0.01 m					
	S Field (W/m <sup>2</sup> )		E Field (V/m)		H Field (A/m)	
	Result	Limit	Result	Limit	Result	Limit
ICNIRP	2.9430	50.0000	33.3089	137.0000	0.0884	0.3630
FCC*	0.2943	5.0000	N/A	N/A	N/A	N/A
RSS	2.9430	31.6229	33.3089	109.1886	0.0884	0.2896
ARPANSA	2.9430	50.0000	33.3089	137.0000	0.0884	0.3640

\* Requirement and Result in mW/cm<sup>2</sup>

**Table 2 – Occupational Results**

The calculations show that the EUT complies with the occupational exposure levels described in the EN 62311:2008, CFR 47 Pt1.1310, Health Canada Safety Code 6 and ARPANSA Radiation Protection Series No.3 at the point of investigation, 0.01 m.

Regional Requirement	Calculated RF exposure level at compliance boundary of 0.01 m					
	S Field (W/m <sup>2</sup> )		E Field (V/m)		H Field (A/m)	
	Result	Limit	Result	Limit	Result	Limit
ICNIRP	2.9430	10.0000	33.3089	61.0000	0.0884	0.1620
FCC*	0.2943	1.0000	N/A	N/A	N/A	N/A
RSS	2.9430	5.3478	33.3089	44.8977	0.0884	0.1191
ARPANSA	2.9430	10.0000	33.3089	61.4000	0.0884	0.1630

\* Requirement and Result in mW/cm<sup>2</sup>

**Table 3 – General Population Results**

The calculations show that the EUT complies with the occupational exposure levels described in the EN 62311:2008, CFR 47 Pt1.1310, Health Canada Safety Code 6 and ARPANSA Radiation Protection Series No.3 at the point of investigation, 0.01 m.



Product Service

## **SECTION 2**

### **TEST DETAILS**



## 2.1 RATIONALE FOR ASSESSMENT OF THE RF EXPOSURE

The aim of the assessment report is to evaluate the compliance boundary for a set of given input power(s) according to the basic restrictions (directly or indirectly via compliance with reference levels) related to human exposure to radio frequency electromagnetic fields.

The chosen assessment method to establish the compliance boundary in the far-field region is the reference method as defined in the relevant specifications.

The RF exposure assessment is based upon the following criteria:

The 801-002xx CPRmeter 2 (Medical equipm.) operates with the following transmitters active on the antenna ports shown in Section 1.3.3. For each transmitter, the Radio Access Technology (RAT), EIRP inclusive of antenna gain and duty cycle, gain of the antenna and lowest frequency of operation are shown as they contribute to the calculation of S Field, E field and H field values according to the following formulas.

The power flux (S Field):

$$S = \frac{PG_{(\theta, \phi)}}{4\pi \cdot r^2}$$

The electric field strength (E Field):

$$E = \frac{\sqrt{30PG_{(\theta, \phi)}}}{r}$$

The magnetic field strength (H Field):

$$H = \frac{E}{\eta_0}$$

Where:

P = Average Power (W)

G = Antenna Gain (dBi)

r = Distance (cm) or (m)

$\eta_0 = 377$



Product Service

## 2.2 TEST RESULT DETAILS

The frequencies shown in the tables below have been chosen based on the lowest possible frequency that the EUT can transmit.

Antenna Port	Tx No.	Ant No.	RAT	EIRP (W)	Duty Cycle (%)	Gain (dBi)	Frequency (MHz)	RF Exposure Level at compliance boundary of 0.01 m		
								S Field	E Field	H Field
1	1	1	Bluetooth Low Energy	0.004	100	1.7	2400	2.9430	33.3089	0.0884

**Table 4 – Occupational Transmitter Summary**

Antenna Port	Tx No.	Ant No.	RAT	EIRP (W)	Duty Cycle (%)	Gain (dBi)	Frequency (MHz)	RF Exposure Level at compliance boundary of 0.01 m		
								S Field	E Field	H Field
1	1	1	Bluetooth Low Energy	0.004	100	1.7	2400	2.9430	33.3089	0.0884

**Table 5 – General Population Transmitter Summary**



Product Service

## **SECTION 3**

### **DISCLAIMERS AND COPYRIGHT**



Product Service

### 3.1 DISCLAIMERS AND COPYRIGHT

This report relates only to the actual item/items tested.

This report must not be reproduced, except in its entirety, without the written permission of  
TÜV SÜD Product Service

© 2016 TÜV SÜD Product Service



Product Service

## **ANNEX A**

### **REGIONAL REQUIREMENTS**



Product Service

Frequency Range (MHz)	Power Density (W/m <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
0.065 - 1	-	610/f	1.6/f
1 - 10	-	610/f	1.6/f
10 - 400	10	61	0.162
400 - 2000	f/40	3*f <sup>0.5</sup>	0.00796*f <sup>0.5</sup>
2000 - 300000	50	137	0.363

**Table A.1 – EN 62311:2008 Occupational Limits**

Frequency Range (MHz)	Power Density (W/m <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
0.003 - 0.15	-	87	5
0.15 - 1	-	87/f	0.73/f
1 - 10	-	87/f <sup>0.5</sup>	0.73/f
10 - 400	2	27	0.071
400 - 2000	f/200	1.375*f <sup>0.5</sup>	0.00364*f <sup>0.5</sup>
2000 - 300000	10	61	0.162

**Table A.2 – EN 62311:2008 General Population Limits**

Frequency Range (MHz)	S Field (mW/cm <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
0 - 0.3	-	-	-
0.3 - 3	100	614	1.63
3 - 30	900/f <sup>2</sup>	1842/f	4.89/f
30 - 300	1	61.4	0.163
300 - 1500	f/300	-	-
1500 - 100000	5	-	-

**Table A.3 – CFR 47 Pt1.1310 Occupational Limits**

Frequency Range (MHz)	S Field (mW/cm <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
0 - 0.3	-	-	-
0.3 - 3	100	614	1.63
3 - 30	180/f <sup>2</sup>	824/f	2.19/f
30 - 300	0.2	27.5	0.073
300 - 1500	f/1500	-	-
1500 - 100000	1	-	-

**Table A.4 – CFR 47 Pt1.1310 General Population Limits**

Frequency Range (MHz)	Power Density (W/m <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
10 - 20	10	61.4	0.163
20 - 48	44.72/f <sup>0.5</sup>	129.8/f <sup>0.25</sup>	0.3444/f <sup>0.25</sup>
48 - 100	6.455	49.33	0.1309
100 - 6000	0.6455*f <sup>0.5</sup>	15.60*f <sup>0.25</sup>	0.04138*f <sup>0.25</sup>
6000 - 150000	50	137	0.364

**Table A.5 – Health Canada Safety Code 6 Occupational Limits**

Frequency Range (MHz)	Power Density (W/m <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
10 - 20	2	27.46	0.0728
20 - 48	8.944/f <sup>0.5</sup>	58.07/f <sup>0.25</sup>	0.1540/f <sup>0.25</sup>
48 - 300	1.291	22.06	0.05852
300 - 6000	0.02619*f <sup>0.6834</sup>	3.142*f <sup>0.3417</sup>	0.008335*f <sup>0.3417</sup>
6000 - 15000	10	61.4	0.163

**Table A.6 – Health Canada Safety Code 6 General Population Limits**

Frequency Range (MHz)	Power Density (W/m <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
0.1 - 1	-	614	1.63/f





Product Service

1 - 10	$1000/f^2$	614	$1.63/f$
10 - 400	10	61.4	0.163
400 - 2000	$f/40$	$3.07*f^{0.5}$	$0.00814*f^{0.5}$
2000 - 300000	50	137	0.364

**Table A.7 – ARPANSA Radiation Protection Series No.3 Occupational Limits**

Frequency Range (MHz)	Power Density (W/m <sup>2</sup> )	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)
0.1 - 0.15	-	86.8	4.86
0.15 - 1	-	86.8	$0.729/f$
1 - 10	-	$86.8/f^{0.5}$	$0.729/f$
10 - 400	2	27.4	0.0729
400 - 2000	$f/200$	$1.37*f^{0.5}$	$0.00364*f^{0.5}$
2000 - 300000	10	61.4	0.163

**Table A.8 – ARPANSA Radiation Protection Series No.3 General Population Limits**