



Rayence Co.,Ltd.

14, Samsung 1-ro 1-gil, Hwaseong-si, Gyeonggi-do, Korea
Tel. +82-31-8015-6459 / Fax +82-31-8015-6394

Date: March 31, 2017

To: SIEMIC, INC.
775 Montague Expressway,
Milpitas, CA 95035
USA

Dear Sir/Madam,

Re: Letter of Similarity

We, Rayence Co.,Ltd., declare the differences for between original model and variant models.

-Product: Medical Image Processing Unit

-Model name: 1417WCC

-Variant Model name: 1417WGC

1417WCC and 1417WGC are identical electrical components except for scintillator;

- 1) C: Csl scintillator
- 2) G: Gadox scintillator

Sincerely Yours,

NAME : Namyoung, Ha
TITLE : Assistant Manager
E-MAIL: namyoung.ha@rayence.com
FAX: + 82-31-8015-6394
TEL: + 82-31-8015-6459
Rayence Co., Ltd.