

Rehabtronics Inc. #4352, 10230 Jasper Avenue Edmonton, Alberta Canada T5J4P6 **PHONE:** (+1) 780-638-2520 **FAX:** (+1) 780-428-3010 **FMAII:** inquiries@rehabtronics

EMAIL: inquiries@rehabtronics.com **WEBSITE:** www.rehabtronics.com

August 29, 2013 Federal Communications Commission Authorization and Evaluation Division

RE: Change to identification of presently authorized equipment (New FCC ID: QIQ-H200W-ICT-V02)

Dear Sir/Madam:

This is a request for a change in identification of presently authorized equipment. The details of the original grant are as follows:

Original FCC ID: QIQ-H200W-ICT-V01 Date of Original Grant: 02/17/2013

Requested New ID: QIQ-H200W-ICT-V02

It must be noted that no devices were released to market with the authorized ID under the original grant. Prior to going to market, it was decided that a small hardware change was necessary. Specifically, a power-amplifier was added to mitigate some performance issues identified in the original unit.

As a result, all tests from the previous grant application are obsolete. All the tests have been repeated following the changes. These new test-reports, along with updated schematics and block-diagrams have been supplied with the application for the new ID.

Sincerely,

Mike Pearson MR, QA & RA

August 29, 2013