

### Appendix C. Calibration Certificate

Table of contents		
DAE4-1492		
Probe EX3DV4-7381		
Dipole D835V2-4d126		
Dipole D1900V2-5d143		
Dipole D2450V2-978		
Dipole D2600V2-1119		

Schmid & Partner Engineering AG

s p e a

a

Zeughausstrasse 43, 8004 Zurich, Switzerland Phone +41 44 245 9700, Fax +41 44 245 9779 info@speag.com, http://www.speag.com

### IMPORTANT NOTICE

### USAGE OF THE DAE 4

The DAE unit is a delicate, high precision instrument and requires careful treatment by the user. There are no serviceable parts inside the DAE. Special attention shall be given to the following points:

Battery Exchange: The battery cover of the DAE4 unit is closed using a screw, over tightening the screw may cause the threads inside the DAE to wear out.

Shipping of the DAE: Before shipping the DAE to SPEAG for calibration, remove the batteries and pack the DAE in an antistatic bag. This antistatic bag shall then be packed into a larger box or container which protects the DAE from impacts during transportation. The package shall be marked to indicate that a fragile instrument is inside.

E-Stop Failures: Touch detection may be malfunctioning due to broken magnets in the E-stop. Rough handling of the E-stop may lead to damage of these magnets. Touch and collision errors are often caused by dust and dirt accumulated in the E-stop. To prevent E-stop failure, the customer shall always mount the probe to the DAE carefully and keep the DAE unit in a non-dusty environment if not used for measurements.

Repair: Minor repairs are performed at no extra cost during the annual calibration. However, SPEAG reserves the right to charge for any repair especially if rough unprofessional handling caused the defect.

DASY Configuration Files: Since the exact values of the DAE input resistances, as measured during the calibration procedure of a DAE unit, are not used by the DASY software, a nominal value of 200 MOhm is given in the corresponding configuration file.

### Important Note:

Warranty and calibration is void if the DAE unit is disassembled partly or fully by the Customer.

#### Important Note:

Never attempt to grease or oil the E-stop assembly. Cleaning and readjusting of the Estop assembly is allowed by certified SPEAG personnel only and is part of the annual calibration procedure.

#### Important Note:

To prevent damage of the DAE probe connector pins, use great care when installing the probe to the DAE. Carefully connect the probe with the connector notch oriented in the mating position. Avoid any rotational movement of the probe body versus the DAE while turning the locking nut of the connector. The same care shall be used when disconnecting the probe from the DAE.

Schmid & Partner Engineering

TN\_BR040315AD DAE4.doc

11.12.2009

### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client Huawei - SZ (Auden)

Certificate No: DAE4-1492\_Sep16

Accreditation No.: SCS 0108

A CAL-06.v29 alibration procedu eptember 28, 201 he traceability to nation es with confidence pro	04 BM - SN: 1492 ure for the data acquisition electro 16 hal standards, which realize the physical units bability are given on the following pages and a facility: environment temperature (22 ± 3)°C a	of measurements (SI). are part of the certificate.
alibration procedu eptember 28, 20 <sup>-1</sup> he traceability to nation es with confidence pro	16 hal standards, which realize the physical units bability are given on the following pages and	of measurements (SI). are part of the certificate.
eptember 28, 20 he traceability to nation es with confidence pro n the closed laboratory	16 hal standards, which realize the physical units bability are given on the following pages and	of measurements (SI). are part of the certificate.
he traceability to nation es with confidence pro n the closed laboratory	nal standards, which realize the physical units bability are given on the following pages and	are part of the certificate.
es with confidence pro	bability are given on the following pages and	are part of the certificate.
es with confidence pro	bability are given on the following pages and	are part of the certificate.
	facility: environment temperature (22 $\pm$ 3)°C a	and humidity < 70%.
ical for calibration)		
0#	Cal Date (Certificate No.)	Scheduled Calibration
N: 0810278	09-Sep-16 (No:19065)	Sep-17
) #	Check Date (in house)	Scheduled Check
		In house check: Jan-17
E UMS 006 AA 1002	05-Jan-16 (in house check)	In house check: Jan-17
lame	Function	Signature
ric Hainfeld	Technician	2011
in Bomholt	Deputy Technical Manager	i.v.Blund
		Issued: September 28, 2016
	N: 0810278 D # E UWS 053 AA 1001 E UMS 006 AA 1002 ame ric Hainfeld	N: 0810278     09-Sep-16 (No:19065)       0 #     Check Date (in house)       E UWS 053 AA 1001     05-Jan-16 (in house check)       E UMS 006 AA 1002     05-Jan-16 (in house check)       ame     Function       ric Hainfeld     Technician

### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst

- Service suisse d'étalonnage
- C Servizio svizzero di taratura
- S Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

### Glossary DAE

Connector angle

data acquisition electronics

information used in DASY system to align probe sensor X to the robot coordinate system.

### Methods Applied and Interpretation of Parameters

- DC Voltage Measurement: Calibration Factor assessed for use in DASY system by comparison with a calibrated instrument traceable to national standards. The figure given corresponds to the full scale range of the voltmeter in the respective range.
- Connector angle: The angle of the connector is assessed measuring the angle mechanically by a tool inserted. Uncertainty is not required.
- The following parameters as documented in the Appendix contain technical information as a
  result from the performance test and require no uncertainty.
  - DC Voltage Measurement Linearity: Verification of the Linearity at +10% and -10% of the nominal calibration voltage. Influence of offset voltage is included in this measurement.
  - Common mode sensitivity: Influence of a positive or negative common mode voltage on the differential measurement.
  - Channel separation: Influence of a voltage on the neighbor channels not subject to an input voltage.
  - AD Converter Values with inputs shorted: Values on the internal AD converter corresponding to zero input voltage
  - Input Offset Measurement: Output voltage and statistical results over a large number of zero voltage measurements.
  - Input Offset Current: Typical value for information; Maximum channel input offset current, not considering the input resistance.
  - Input resistance: Typical value for information: DAE input resistance at the connector, during internal auto-zeroing and during measurement.
  - Low Battery Alarm Voltage: Typical value for information. Below this voltage, a battery alarm signal is generated.
  - Power consumption: Typical value for information. Supply currents in various operating modes.

### DC Voltage Measurement

A/D - Converter Reso	lution nominal			
High Range:	1LSB =	6.1µV,	full range =	-100+300 mV
Low Range:	1LSB =	61nV ,	full range =	-1+3mV
DASY measurement	parameters: Aut	to Zero Time: 3	sec; Measuring	time: 3 sec

<b>Calibration Factors</b>	X	Y	Z
High Range	403.902 ± 0.02% (k=2)	403.818 ± 0.02% (k=2)	404.047 ± 0.02% (k=2)
Low Range	3.96999 ± 1.50% (k=2)	3.92849 ± 1.50% (k=2)	3.97726 ± 1.50% (k=2)

### **Connector Angle**

Connector Angle to be used in DAST system 350.5 ± 1	Connector Angle to be used in DASY system	358.5 ° ± 1 °
---	---	---------------

Appendix (Additional assessments outside the scope of SCS0108)

### 1. DC Voltage Linearity

High Range		Reading (µV)	Difference (µV)	Error (%)
Channel X	+ Input	200040.96	3.16	0.00
Channel X	+ Input	20007.78	2.86	0.01
Channel X	- Input	-20002.10	3.74	-0.02
Channel Y	+ Input	200037.08	-0.56	-0.00
Channel Y	+ Input	20004.16	-0.68	-0.00
Channel Y	- Input	-20004.07	1.87	-0.01
Channel Z	+ Input	200037.08	-0.92	-0.00
Channel Z	+ Input	20001.95	-2.96	-0.01
Channel Z	- Input	-20007.20	-1.17	0.01

Low Range	Reading (µV)	Difference (µV)	Error (%)
Channel X + Input	2000.72	-0.17	-0.01
Channel X + Input	201.39	0.43	0.22
Channel X - Input	-198.65	0.30	-0.15
Channel Y + Input	2000.43	-0.35	-0.02
Channel Y + Input	201.03	0.15	0.08
Channel Y - Input	-199.67	-0.58	0.29
Channel Z + Input	2001.51	0.77	0.04
Channel Z + Input	199.94	-0.84	-0.42
Channel Z - Input	-200.77	-1.62	0.82

### 2. Common mode sensitivity

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Common mode Input Voltage (mV)	High Range Average Reading (μV)	Low Range Average Reading (µV)
Channel X	200	24.30	22.82
	- 200	-22.24	-23.45
Channel Y	200	0.04	-0.11
	- 200	-1.28	-1.79
Channel Z	200	5.14	5.22
	- 200	-8.29	-8.11

3. Channel separation DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	Input Voltage (mV)	Channel X (µV)	Channel Y (µV)	Channel Z (µV)
Channel X	200		-2.94	-2.04
Channel Y	200	6.62		-0.59
Channel Z	200	10.52	3.49	

Certificate No: DAE4-1492\_Sep16

### 4. AD-Converter Values with inputs shorted

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec

	High Range (LSB)	Low Range (LSB)
Channel X	16395	16252
Channel Y	16550	15762
Channel Z	15906	16049

### 5. Input Offset Measurement

DASY measurement parameters: Auto Zero Time: 3 sec; Measuring time: 3 sec Input 10MΩ

	Average (µV)	min. Offset (µV)	max. Offset (μV)	Std. Deviation (µV)
Channel X	0.66	-2.98	5.34	0.85
Channel Y	0.06	-1.40	2.11	0.60
Channel Z	-1.10	-3.01	0.20	0.47

### 6. Input Offset Current

Nominal Input circuitry offset current on all channels: <25fA

### 7. Input Resistance (Typical values for information)

	Zeroing (kOhm)	Measuring (MOhm)
Channel X	200	200
Channel Y	200	200
Channel Z	200	200

### 8. Low Battery Alarm Voltage (Typical values for information)

Typical values	Alarm Level (VDC)	
Supply (+ Vcc)	+7.9	
Supply (- Vcc)	-7.6	

### 9. Power Consumption (Typical values for information)

Typical values	Switched off (mA)	Stand by (mA)	Transmitting (mA)
Supply (+ Vcc)	+0.01	+6	+14
Supply (- Vcc)	-0.01	-8	-9

#### Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst

- C Service suisse d'étalonnage
- S Servizio svizzero di taratura
- Swiss Calibration Service

Accreditation No.: SCS 0108

In house check: Jun-18

In house check: Jun-18

In house check: Oct-16

Certificate No: EX3-7381\_Sep16

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client Huawei-SZ (Auden)

CALIBRATION CERTIFICATE Object EX3DV4 - SN:7381 QA CAL-01.v9, QA CAL-14.v4, QA CAL-23.v5, QA CAL-25.v6 Calibration procedure(s) Calibration procedure for dosimetric E-field probes September 29, 2016 Calibration date: This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Scheduled Calibration Primary Standards ID Cal Date (Certificate No.) Power meter NRP SN: 104778 06-Apr-16 (No. 217-02288/02289) Apr-17 Power sensor NRP-Z91 SN: 103244 06-Apr-16 (No. 217-02288) Apr-17 SN: 103245 Power sensor NRP-Z91 06-Apr-16 (No. 217-02289) Apr-17 Reference 20 dB Attenuator SN: S5277 (20x) 05-Apr-16 (No. 217-02293) Apr-17 Reference Probe ES3DV2 SN: 3013 31-Dec-15 (No. ES3-3013\_Dec15) Dec-16 DAE4 SN: 660 23-Dec-15 (No. DAE4-660\_Dec15) Dec-16 Secondary Standards ID Scheduled Check Check Date (in house) Power meter E4419B SN: GB41293874 06-Apr-16 (in house check Jun-16) In house check: Jun-18 Power sensor E4412A SN: MY41498087 06-Apr-16 (in house check Jun-16) In house check: Jun-18

	Name	Function	Signature
Calibrated by:	Leif Klysner	Laboratory Technician	Selftlym
Approved by:	Katja Pokovic	Technical Manager	il the
			Issued: October 4, 2016
This calibration certificate	e shall not be reproduced except in ful	I without written approval of the laboratory	

06-Apr-16 (in house check Jun-16)

04-Aug-99 (in house check Jun-16)

18-Oct-01 (in house check Oct-15)

Power sensor E4412A

RF generator HP 8648C

Network Analyzer HP 8753E

SN: 000110210

SN: US37390585

SN: US3642U01700

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Schweizerischer Kalibrierdienst s

- Service suisse d'étalonnage C
  - Servizio svizzero di taratura s

Accreditation No.: SCS 0108

Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

### Glossary:

TSL	tissue simulating liquid
NORMx,y,z	sensitivity in free space
ConvF	sensitivity in TSL / NORMx,y,z
DCP	diode compression point
CF	crest factor (1/duty_cycle) of the RF signal
A, B, C, D	modulation dependent linearization parameters
Polarization $\phi$	φ rotation around probe axis
Polarization 9	9 rotation around an axis that is in the plane normal to probe axis (at measurement center),
	i.e., 9 = 0 is normal to probe axis
Connector Angle	information used in DASY system to align probe sensor X to the robot coordinate system

Connector Angle

### Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

### Methods Applied and Interpretation of Parameters:

- NORMx, y, z: Assessed for E-field polarization 9 = 0 (f ≤ 900 MHz in TEM-cell; f > 1800 MHz: R22 waveguide). NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not affect the E2-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z \* frequency\_response (see Frequency Response Chart). This linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z; Bx,y,z; Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f ≤ 800 MHz) and inside waveguide using analytical field distributions based on power measurements for f > 800 MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty values are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx, y, z \* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from ± 50 MHz to ± 100 MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

September 29, 2016

EX3DV4 - SN:7381

# Probe EX3DV4

## SN:7381

Repaired: Calibrated:

Manufactured: April 13, 2015 September 22, 2016 September 29, 2016

Calibrated for DASY/EASY Systems (Note: non-compatible with DASY2 system!)

Certificate No: EX3-7381\_Sep16

Page 3 of 38

### **Basic Calibration Parameters**

	Sensor X	Sensor Y	Sensor Z	Unc (k=2)
Norm (µV/(V/m) <sup>2</sup> ) <sup>A</sup>	0.41	0.43	0.42	± 10.1 %
DCP (mV) <sup>B</sup>	96.5	97.3	97.3	

### Modulation Calibration Parameters

UID	Communication System Name		A dB	B dBõV	С	D dB	VR mV	Unc <sup>E</sup> (k=2)
0	CW	X 0.0	0.0	0.0	1.0	0.00	143.0	±3.3 %
		Y	0.0	0.0	1.0		148.6	
		Z	0.0	0.0	1.0		139.1	

Note: For details on UID parameters see Appendix.

### Sensor Model Parameters

	C1 fF	C2 fF	α V <sup>-1</sup>	T1 ms.V <sup>-2</sup>	T2 ms.V <sup>-1</sup>	T3 ms	T4 V <sup>-2</sup>	T5 V <sup>-1</sup>	T6
Х	44.89	343	37.26	7.442	0.477	4.968	0.385	0.246	1.003
Y	42.6	325	37.25	2.135	0.645	4.974	0	0.313	1.004
Z	41.15	309.4	36.13	5.149	0.484	4.961	1.509	0.064	1.004

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

<sup>A</sup> The uncertainties of Norm X,Y,Z do not affect the E<sup>2</sup>-field uncertainty inside TSL (see Pages 5 and 6).

<sup>II</sup> Numerical linearization parameter: uncertainty not required. <sup>E</sup> Uncertainty is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

f (MHz) <sup>C</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	41.9	0.89	11.23	11.23	11.23	0.40	1.09	± 12.0 %
850	41.5	0.92	10.11	10.11	10.11	0.55	0.81	± 12.0 %
1750	40.1	1.37	8.68	8.68	8.68	0.35	0.81	± 12.0 %
1900	40.0	1.40	8.35	8.35	8.35	0.33	0.80	± 12.0 %
2300	39.5	1.67	7.85	7.85	7.85	0.19	1.14	± 12.0 %
2450	39.2	1.80	7.39	7.39	7.39	0.36	0.80	± 12.0 %
2600	39.0	1.96	7.27	7.27	7.27	0.20	1.23	± 12.0 %
5250	35.9	4.71	5.47	5.47	5.47	0.30	1.80	± 13.1 %
5600	35.5	5.07	4.75	4.75	4.75	0.40	1.80	± 13.1 %
5750	35.4	5.22	5.01	5.01	5.01	0.40	1.80	± 13.1 %

#### Calibration Parameter Determined in Head Tissue Simulating Media

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity calibration below 20 MHz.

<sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters (ε and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ε and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.
<sup>G</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is

<sup>o</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.


f (MHz) <sup>c</sup>	Relative Permittivity <sup>F</sup>	Conductivity (S/m) <sup>F</sup>	ConvF X	ConvF Y	ConvF Z	Alpha <sup>G</sup>	Depth <sup>G</sup> (mm)	Unc (k=2)
750	55.5	0.96	9.80	9.80	9.80	0.25	1.11	± 12.0 %
850	55.2	0.99	9.78	9.78	9.78	0.42	0.84	± 12.0 %
1750	53.4	1.49	8.08	8.08	8.08	0.30	0.99	± 12.0 %
1900	53.3	1.52	7.95	7.95	7.95	0.20	1.26	± 12.0 %
2300	52.9	1.81	7.67	7.67	7.67	0.34	0.80	± 12.0 %
2450	52.7	1.95	7.45	7.45	7.45	0.34	0.90	± 12.0 %
2600	52.5	2.16	7.19	7.19	7.19	0.31	0.95	± 12.0 %
5250	48.9	5.36	4.56	4.56	4.56	0.50	1.90	± 13.1 %
5600	48.5	5.77	3.78	3.78	3.78	0.60	1.90	± 13.1 %
5750	48.3	5.94	3.97	3.97	3.97	0.60	1.90	± 13.1 %

### Calibration Parameter Determined in Body Tissue Simulating Media

<sup>C</sup> Frequency validity above 300 MHz of ± 100 MHz only applies for DASY v4.4 and higher (see Page 2), else it is restricted to ± 50 MHz. The uncertainty is the RSS of the ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz. <sup>F</sup> At frequencies below 3 GHz, the validity of tissue parameters (c and σ) can be relaxed to ± 10% if liquid compensation formula is applied to

F At frequencies below 3 GHz, the validity of tissue parameters (c and σ) can be relaxed to ± 10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (c and σ) is restricted to ± 5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.


the ConvF uncertainty for indicated target tissue parameters. <sup>0</sup> Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.



### Frequency Response of E-Field (TEM-Cell:ifi110 EXX, Waveguide: R22)


Uncertainty of Frequency Response of E-field: ± 6.3% (k=2)

September 29, 2016



### Receiving Pattern ( $\phi$ ), $\vartheta = 0^{\circ}$


Uncertainty of Axial Isotropy Assessment: ± 0.5% (k=2)



### Dynamic Range f(SAR<sub>head</sub>) (TEM cell , f<sub>eval</sub>= 1900 MHz)

Uncertainty of Linearity Assessment: ± 0.6% (k=2)

Page 9 of 38



### **Conversion Factor Assessment**

Page 10 of 38

### **Other Probe Parameters**

Sensor Arrangement	Triangular
Connector Angle (°)	116.2
Mechanical Surface Detection Mode	enabled
Optical Surface Detection Mode	disabled
Probe Overall Length	337 mm
Probe Body Diameter	10 mm
Tip Length	9 mm
Tip Diameter	2.5 mm
Probe Tip to Sensor X Calibration Point	1 mm
Probe Tip to Sensor Y Calibration Point	1 mm
Probe Tip to Sensor Z Calibration Point	1 mm
Recommended Measurement Distance from Surface	1.4 mm

### Appendix: Modulation Calibration Parameters

UID	Communication System Name		A dB	B dBõV	с	D dB	VR mV	Max Unc <sup>E</sup> (k=2)
0	CW	X	0.00	0.00	1.00	0.00	143.0	± 3.3 %
		Y	0.00	0.00	1.00		148.6	
10010		Z	0.00	0.00	1.00		139.1	
10010- CAA	SAR Validation (Square, 100ms, 10ms)	X	1.87	63.41	8.60	10.00	20.0	± 9.6 %
		Y	2.16	64.29	9.29		20.0	
10011		Z	1.87	63.10	8.36		20.0	-
10011- CAB	UMTS-FDD (WCDMA)	×	1.26	71.18	17.68	0.00	150.0	± 9.6 %
		Y	1.51	75.12	19.67		150.0	
10012-	JEEE 000 11h WIELD 4 OUL (DODD 1	Z	1.11	69.16	16.38		150.0	
CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps)	X	1.18	64.26	15.89	0.41	150.0	±9.6 %
		Y	1.18	64.88	16.50		150.0	
10013-	IEEE 802.11g WiFi 2.4 GHz (DSSS-	X	1.14 4.81	63.73 66.53	15.30	4.10	150.0	1000
CAB	OFDM, 6 Mbps)	1.122	133300		17.07	1.46	150.0	± 9.6 %
		Y	4.78	66.68	17.20		150.0	
10021- DAB	GSM-FDD (TDMA, GMSK)	X	4.72 7.64	66.47 78.57	16.88 15.86	9.39	150.0 50.0	± 9.6 %
		Y	14.78	86.38	18.78		50.0	
		Z	5.00	73.55	13.93		50.0	
10023- DAB	GPRS-FDD (TDMA, GMSK, TN 0)	X	6.18	76.04	15.01	9.57	50.0	±9.6 %
		Y	9.47	81.04	17.14		50.0	
		Z	4.50	72.15	13.41		50.0	
10024- DAB	GPRS-FDD (TDMA, GMSK, TN 0-1)	×	31.84	93.57	18.88	6.56	60.0	± 9.6 %
		Y	100.00	107.28	22.67		60.0	
		Z	4.25	74.61	13.23		60.0	
10025- DAB	EDGE-FDD (TDMA, 8PSK, TN 0)	×	6.58	84.96	33.31	12.57	50.0	±9.6 %
		Y	4.22	70.11	25.42		50.0	
10026-	EDGE EDD (TOMA ADON THA A)	Z	5.77	80.29	30.67		50.0	
DAB	EDGE-FDD (TDMA, 8PSK, TN 0-1)	×	7.55	88.58	31.44	9.56	60.0	±9.6 %
		Y	6.07	83.03	29.17		60.0	
10027- DAB	GPRS-FDD (TDMA, GMSK, TN 0-1-2)	Z X	6.67 100.00	85.50 104.25	30.01 20.61	4.80	60.0 80.0	± 9.6 %
		Y	100.00	109.36	22.60		20.0	
		Z	53.53	97.01	22.60		80.0 80.0	
10028- DAB	GPRS-FDD (TDMA, GMSK, TN 0-1-2-3)	×	100.00	106.10	20.78	3.55	100.0	±9.6 %
		Y	100.00	114.53	23.94		100.0	
		Z	100.00	102.88	19.22		100.0	
10029- DAB	EDGE-FDD (TDMA, 8PSK, TN 0-1-2)	×	4.61	77.41	25.78	7.80	80.0	±9.6 %
		Y	3.91	73.86	24.29		80.0	
10000	IFFE ANA JE J PL	Z	4.16	75.20	24.61		80.0	
10030- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH1)	×	15.51	86.57	16.42	5.30	70.0	±9.6 %
_		Y	100.00	105.98	21.50		70.0	
10024		Z	2.43	70.53	11.23		70.0	
10031- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH3)	X	100.00	105.44	19.36	1.88	100.0	± 9.6 %
		Y	100.00	119.70	24.36	_	100.0	
		Z	100.00	98.05	16.18		100.0	

### September 29, 2016

10032- CAA	IEEE 802.15.1 Bluetooth (GFSK, DH5)	x	100.00	124.69	26.03	1.17	100.0	±9.6 %
0.01		Y	100.00	213.77	57.78		100.0	
		Z	100.00	106.22	18.67		100.0	
10033- CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH1)	X	4.94	81.54	20.44	5.30	70.0	±9.6 %
onn	bitty	Y	4.68	81.79	20.78		70.0	
		Z	3.51	76.21	17.97	-	70.0	
10034- CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH3)	X	2.27	74.32	16.92	1.88	100.0	±9.6 %
UM	Dhaj	Y	2.54	77.31	18.28		100.0	
		Z	1.66	70.14	14.62		100.0	
10035- CAA	IEEE 802.15.1 Bluetooth (PI/4-DQPSK, DH5)	X	1.87	72.93	16.30	1.17	100.0	±9.6 %
		Y	2.25	76.71	17.94		100.0	5
		Z	1.41	69.22	14.13		100.0	
10036- CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH1)	Х	6.11	84.90	21.67	5.30	70.0	±9.6 %
0.01		Y	5.95	85.65	22.20		70.0	0.000
		Z	4.06	78.47	18.89		70.0	
10037- CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH3)	X	2.10	73.44	16.54	1.88	100.0	±9.6 %
0/01		Y	2.25	75.81	17.70		100.0	
		Z	1.56	69.43	14.29		100.0	
10038- CAA	IEEE 802.15.1 Bluetooth (8-DPSK, DH5)	X	1.89	73.33	16.59	1.17	100.0	± 9.6 %
onn		Y	2.29	77.33	18.33		100.0	
		Z	1.42	69.48	14.37		100.0	
10039- CAB	CDMA2000 (1xRTT, RC1)	X	3.53	81.73	19.56	0.00	150.0	± 9.6 %
UND		Y	17.77	104.06	26.01		150.0	
		Z	2.30	75.80	16.79		150.0	
10042- CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4- DQPSK, Halfrate)	X	3.81	72.34	12.74	7.78	50.0	±9.6 %
0,10	D'al Ori, Hamaloj	Y	8.30	80.58	15.84		50.0	
		Z	2.60	68.50	11.06		50.0	
10044- CAA	IS-91/EIA/TIA-553 FDD (FDMA, FM)	X	0.00	104.39	1.95	0.00	150.0	± 9.6 %
		Y	0.00	112.21	1.28		150.0	
		Z	0.00	100.36	0.38	-	150.0	
10048- CAA	DECT (TDD, TDMA/FDM, GFSK, Full Slot, 24)	X	4.84	69.06	13.68	13.80	25.0	± 9.6 %
		Y	5.92	70.50	14.69		25.0	
		Z	4.51	67.49	12.92		25.0	
10049- CAA	DECT (TDD, TDMA/FDM, GFSK, Double Slot, 12)	×	4.80	71.85	13.65	10.79	40.0	± 9.6 %
Carl Charles		Y	5.89	73.89	14.85		40.0	
		Z	4.26	70.00	12.77		40.0	
10056- CAA	UMTS-TDD (TD-SCDMA, 1.28 Mcps)	X	11.27	87.10	22.00	9.03	50.0	± 9.6 %
		Y	11.36	86.97	22.03		50.0	
		Z	8.58	82.33	19.97		50.0	
10058- DAB	EDGE-FDD (TDMA, 8PSK, TN 0-1-2-3)	X	3.64	72.87	23.03	6.55	100.0	± 9.6 %
		Y	3.18	70.25	21.96		100.0	
		Z	3.33	71.14	22.05		100.0	
10059- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps)	X	1.18	65.02	16.27	0.61	110.0	± 9.6 %
		Y	1.17	65.53	16.86	-	110.0	
		Z	1.13	64.29	15.58		110.0	
10060- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps)	X	10.47	107.76	29.35	1.30	110.0	± 9.6 %
		Y	19.68	123.43	34.57		110.0	
		Z	2.79	87.05	22.88	_	110.0	

### September 29, 2016

10061- CAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)	X	1.98	75.54	20.50	2.04	110.0	±9.6 %
		Y	1.77	75.08	20.90		110.0	
		Z	1.60	72.06	18.71	1	110.0	
10062- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps)	X	4.66	66.71	16.67	0.49	100.0	±9.6 %
		Y	4.64	66.91	16.83		100.0	E
		Z	4.57	66.63	16.48		100.0	
10063- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps)	X	4.66	66.75	16.72	0,72	100.0	±9.6 %
		Y	4.64	66.94	16.88		100.0	
		Z	4.57	66.67	16.52		100.0	
10064- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps)	X	4.93	66.98	16.92	0.86	100.0	±9.6 %
		Y	4.91	67.14	17.06		100.0	
		Z	4.83	66.88	16.72		100.0	
10065- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps)	X	4.78	66.78	16.95	1.21	100.0	±9.6 %
		Y	4.75	66.94	17.08		100.0	
		Z	4.68	66.67	16.74		100.0	
10066- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps)	×	4.78	66.74	17.06	1.46	100.0	±9.6 %
		Y	4.75	66.87	17.19		100.0	
		Z	4.67	66.62	16.84		100.0	
10067- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps)	X	5.05	66.87	17.45	2.04	100.0	± 9.6 %
		Y	5.02	67.00	17.56		100.0	
		Z	4.95	66.80	17.25		100.0	
10068- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps)	×	5.07	66.82	17.60	2.55	100.0	± 9.6 %
		Y	5.03	66.89	17.68		100.0	
		Z	4.96	66.69	17.37		100.0	
10069- CAB	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps)	X	5.15	66.82	17.78	2.67	100.0	± 9.6 %
		Y	5.10	66.89	17.85		100.0	
		Z	5.04	66.71	17.55		100.0	
10071- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 9 Mbps)	×	4.88	66.53	17.30	1.99	100.0	±9.6 %
10/01/C		Y	4.85	66.66	17.42		100.0	
		Z	4.79	66.47	17.11		100.0	
10072- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 12 Mbps)	X	4.83	66.76	17.45	2.30	100.0	±9.6 %
		Y	4.80	66.88	17.57		100.0	
		Z	4.74	66.66	17.24		100.0	
10073- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 18 Mbps)	X	4.86	66.83	17.70	2.83	100.0	± 9.6 %
		Y	4.84	66.93	17.80		100.0	
		Z	4.77	66.74	17.49		100.0	
10074- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 24 Mbps)	X	4.83	66.67	17.80	3.30	100.0	± 9.6 %
		Y	4.81	66.76	17.89		100.0	
		Z	4.75	66.60	17.59		100.0	
10075- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 36 Mbps)	X	4.85	66.71	18.05	3.82	90.0	± 9.6 %
-		Y	4.82	66.76	18.11		90.0	
		Z	4.76	66.60	17.82		90.0	
10076- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 48 Mbps)	X	4.86	66.49	18.15	4.15	90.0	± 9.6 %
		Y	4.83	66.55	18.21		90.0	
2010/22/22/01		Z	4.78	66.43	17.94		90.0	
10077- CAB	IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps)	X	4.88	66.54	18.23	4.30	90.0	± 9.6 %
		Y	4.86	66.60	18.30		90.0	
		Z	4.81	66.49	18.03		90.0	

September 29, 2016

10081- CAB	CDMA2000 (1xRTT, RC3)	×	1.20	71.41	15.32	0.00	150.0	±9.6 %
P. 1.1.1		Y	2.16	79.85	18.48		150.0	
		Z	0.90	67.70	13.03		150.0	
10082- CAB	IS-54 / IS-136 FDD (TDMA/FDM, PI/4- DQPSK, Fullrate)	×	6.20	67.51	6.04	4.77	80.0	± 9.6 %
		Y	0.54	60.00	3.52		80.0	
		Z	0.58	60.00	3.48		80.0	
10090- DAB	GPRS-FDD (TDMA, GMSK, TN 0-4)	x	28.40	92.49	18.63	6.56	60.0	±9.6 %
		Y	100.00	107.28	22.68		60.0	
		Z	4,18	74.43	13.18		60.0	
10097- CAB	UMTS-FDD (HSDPA)	X	2.03	69.78	17.03	0.00	150.0	± 9.6 %
		Y	2.23	71.90	18.12		150.0	
		Z	1.92	69.05	16.37		150.0	
10098- CAB		X	2.00	69.77	17.02	0.00	150.0	±9.6 %
		Y	2.19	71.91	18.13		150.0	
		Z	1.89	69.01	16.35		150.0	
10099- DAB	EDGE-FDD (TDMA, 8PSK, TN 0-4)	×	7.60	88.71	31.48	9.56	60.0	± 9.6 %
		Y	6.10	83.12	29.20		60.0	
		Z	6.71	85.62	30.04		60.0	
10100- CAB	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	×	3.36	71.73	17.69	0.00	150.0	±9.6 %
		Y	3.53	72.97	18.39	-	150.0	
10101		Z	3.18	70.97	17.20		150.0	
10101- CAB	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	×	3.33	68.12	16.52	0.00	150.0	±9.6 %
		Y	3.37	68.64	16.88		150.0	
		Z	3.23	67.78	16.20		150.0	
10102- CAB	LTE-FDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	×	3.43	68.05	16.58	0.00	150.0	±9.6 %
		Y	3.47	68.56	16.94		150.0	
10100		Z	3.33	67.75	16.28		150.0	
10103- CAB	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK)	×	5.16	72.86	19.25	3.98	65.0	± 9.6 %
		Y	4.84	72.32	19.23		65.0	
	1 1000 1000 10 10 0	Z	4.99	72.57	18.99		65.0	
10104- CAB	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM)	×	5.46	71.77	19.59	3.98	65.0	± 9.6 %
		Y	5.05	70.72	19.24		65.0	
10/15	A restar term, do July do resta a secondaria	Z	5.15	70.93	19.05	-	65.0	
10105- CAB	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM)	×	5.07	70.09	19.10	3.98	65.0	± 9.6 %
		Y	4.77	69.31	18.87		65.0	
10105		Z	5.04	70.28	19.05		65.0	
10108- CAC	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	×	2.93	71.05	17.58	0.00	150.0	± 9.6 %
		Y	3.08	72.40	18.34		150.0	
10100		Z	2.76	70.28	17.06	0.07	150.0	
10109- CAC	LTE-FDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	×	2.99	68.14	16.51	0.00	150.0	± 9.6 %
		Y	3.04	68.84	16.95	-	150.0	-
10115		Z	2.88	67.77	16.14		150.0	
10110- CAC	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	x	2.40	70.44	17.33	0.00	150.0	± 9.6 %
		Y	2.55	72.06	18.22		150.0	-
10/11		Z	2.24	69.57	16.70		150.0	
10111-	LTE-FDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	2.77	69.53	17.03	0.00	150.0	± 9.6 %
CAC	10-02/1WI)							
CAC	10-0,410	YZ	2.91 2.64	70.94 69.09	17.79 16.56		150.0	

September 29, 2016

CAC         64           10114- CAB         IEI ME           10115- CAB         IEI CAB           10116- CAB         IEI CAB           10117- CAB         IEI CAB           10118- CAB         IEI CAB           10118- CAB         IEI CAB           10119- CAB         IEI CAB           10119- CAB         IEI CAB           101140- CAB         LT CAB           10141- CAB         LT CAC	TE-FDD (SC-FDMA, 100% RB, 5 MHz, 4-QAM) EEE 802.11n (HT Greenfield, 13.5 lbps, BPSK) EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64- AM)	Y Z X Y Z X Y Z X Y Z X Y Z X Y Z X Y Y Z X Y	3.16 3.01 2.92 3.06 2.80 5.18 5.17 5.08 5.45 5.43 5.45 5.43 5.33 5.27 5.26 5.17 5.26 5.17 5.14 5.13 5.05	68.77 67.78 69.60 70.98 69.22 67.44 67.61 67.30 67.50 67.64 67.50 67.64 67.62 67.80 67.49 67.26 67.46	16.96 16.19 17.12 17.85 16.68 16.78 16.94 16.59 16.81 16.95 16.61 16.80 16.61 16.71	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 % ± 9.6 % ± 9.6 %
CAC         64           10114-         IEI           CAB         Mb           10115-         IEI           CAB         16           10115-         IEI           CAB         64           10116-         IEI           CAB         64           10117-         IEI           CAB         94           10117-         IEI           CAB         QA           10118-         IEI           CAB         QA           10119-         IEI           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LT           CAC         QF	4-QAM) EEE 802.11n (HT Greenfield, 13.5 lbps, BPSK) EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, 16- AM) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X Y Z X Y Z X Y Z X Y Z X Y Z X	3.01 2.92 3.06 2.80 5.18 5.17 5.08 5.45 5.43 5.45 5.43 5.33 5.27 5.26 5.17 5.26 5.17 5.14 5.13 5.05	67.78 69.60 70.98 69.22 67.44 67.61 67.30 67.50 67.64 67.50 67.64 67.62 67.62 67.80 67.49 67.26	16.19 17.12 17.85 16.68 16.78 16.94 16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 % ± 9.6 %
CAC         64           10114-         IEI           10115-         IEI           CAB         16           10115-         IEI           CAB         16           10116-         IEI           CAB         64           10117-         IEI           CAB         64           10117-         IEI           CAB         QA           10118-         IEI           CAB         QA           10119-         IEI           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LT           CAC         QF	4-QAM) EEE 802.11n (HT Greenfield, 13.5 lbps, BPSK) EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, 16- AM) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	X Y Z X Y Z X Y Z X Y Z X Y Z X	2.92 3.06 2.80 5.18 5.17 5.08 5.45 5.45 5.43 5.33 5.27 5.26 5.17 5.26 5.17 5.14 5.13 5.05	69.60 70.98 69.22 67.44 67.61 67.30 67.50 67.64 67.50 67.64 67.62 67.62 67.80 67.49 67.26	17.12 17.85 16.68 16.78 16.94 16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 % ± 9.6 %
CAB         Mb           10115-         IEI           CAB         16           10115-         IEI           CAB         16           10116-         IEI           CAB         64           10117-         IEI           CAB         BP           10117-         IEI           CAB         QA           10118-         IEI           CAB         QA           10119-         IEI           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LT           CAC         QF	EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X Y Z X Y Z X Y Z X Y Z X	2.80 5.18 5.17 5.08 5.45 5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	69.22 67.44 67.61 67.30 67.50 67.64 67.64 67.62 67.62 67.80 67.49 67.26	16.68 16.78 16.94 16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 %
CAB         Mb           10115-         IEI           CAB         16           10115-         IEI           CAB         16           10116-         IEI           CAB         64           10117-         IEI           CAB         BP           10117-         IEI           CAB         QA           10118-         IEI           CAB         QA           10119-         IEI           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LT           CAC         QF	EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	X Y Z X Y Z X Y Z X Y Z X	2.80 5.18 5.17 5.08 5.45 5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	69.22 67.44 67.61 67.30 67.50 67.64 67.64 67.62 67.62 67.80 67.49 67.26	16.68 16.78 16.94 16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 %
CAB         Mb           10115-         IEI           CAB         16           10115-         IEI           CAB         16           10116-         IEI           CAB         64           10117-         IEI           CAB         BP           10117-         IEI           CAB         QA           10118-         IEI           CAB         QA           10119-         IEI           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LT           CAC         QF	EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	X Y Z X Y Z X Y Z X Y Z X	5.18 5.17 5.08 5.45 5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	67.44 67.61 67.30 67.50 67.64 67.64 67.62 67.62 67.80 67.49 67.26	16.78 16.94 16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 %
CAB         Mt           10115-         IEI           CAB         16           10115-         IEI           CAB         64           10116-         IEI           CAB         64           10117-         IEI           CAB         94           10117-         IEI           CAB         QA           10118-         IEI           CAB         QA           10119-         IEI           CAB         QA           10140-         LTI           CAB         MH           10141-         LTI           CAB         MH           10142-         LTI           CAC         QF	EEE 802.11n (HT Greenfield, 81 Mbps, 6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Y Z X Y Z X Y Z X Y Z X	5.17 5.08 5.45 5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	67.61 67.30 67.50 67.64 67.34 67.62 67.80 67.49 67.26	16.94 16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0 150.0	± 9.6 %
CAB         16           10116-         IEE           CAB         64           10117-         IEE           CAB         BP           10117-         IEE           CAB         QA           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LT           CAB         MI-           10141-         LT           CAB         MI-           10142-         LT           CAC         QF	6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X Y Z X Y Z X Y Z X	5.08 5.45 5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	67.30 67.50 67.64 67.34 67.62 67.80 67.49 67.26	16.59 16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0 150.0	
CAB         16           10116-         IEE           CAB         64           10117-         IEE           CAB         BP           10117-         IEE           CAB         QA           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LTI           CAB         MH           10141-         LTI           CAB         MH           10142-         LTI           CAC         QF	6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	X Y Z X Y Z X Y Z X	5.45 5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	67.50 67.64 67.34 67.62 67.80 67.49 67.26	16.81 16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0 150.0	
CAB         16           10116-         IEE           CAB         64           10117-         IEE           CAB         BP           10117-         IEE           CAB         QA           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LT           CAB         MI-           10141-         LT           CAB         MI-           10142-         LT           CAC         QF	6-QAM) EEE 802.11n (HT Greenfield, 135 Mbps, 4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Y Z X Y Z X Y Z X	5.43 5.33 5.27 5.26 5.17 5.14 5.13 5.05	67.64 67.34 67.62 67.80 67.49 67.26	16.95 16.61 16.80 16.96 16.61	0.00	150.0 150.0 150.0 150.0 150.0	
CAB 64 10117- IEE CAB QA 10118- IEE CAB QA 10119- IEE CAB QA 10119- IEE CAB QA 10140- LT CAB MH 10140- LT CAB MH 10141- LT CAB MH 10141- LT CAB MH	4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X Y Z X Y Z X	5.33 5.27 5.26 5.17 5.14 5.13 5.05	67.34 67.62 67.80 67.49 67.26	16.61 16.80 16.96 16.61		150.0 150.0 150.0 150.0	± 9.6 %
CAB         64           10117-         IEE           CAB         BP           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LTI           CAC         QF	4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	X Y Z X Y Z X	5.27 5.26 5.17 5.14 5.13 5.05	67.62 67.80 67.49 67.26	16.80 16.96 16.61		150.0 150.0 150.0	± 9.6 %
CAB         64           10117-         IEE           CAB         BP           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LTI           CAC         QF	4-QAM) EEE 802.11n (HT Mixed, 13.5 Mbps, PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Y Z X Y Z X	5.26 5.17 5.14 5.13 5.05	67.80 67.49 67.26	16.96 16.61		150.0 150.0	± 9.6 %
CAB BP 10118- IEE CAB QA 10119- IEE CAB QA 10140- LT CAB MI- 10141- LT CAB MI- 10142- LT CAC QP	PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X Y Z X	5.17 5.14 5.13 5.05	67.49 67.26	16.61	0.00	150.0	
CAB BP 10118- IEE CAB QA 10119- IEE CAB QA 10140- LT CAB MI- 10141- LT CAB MI- 10142- LT CAC QP	PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X Y Z X	5.17 5.14 5.13 5.05	67.49 67.26	16.61	0.00	150.0	
CAB         BP           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LT           CAB         MH           10141-         LT           CAB         MH           10142-         LT           CAC         QF	PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	X Y Z X	5.14 5.13 5.05	67.26		0.00		
CAB         BP           10118-         IEE           CAB         QA           10119-         IEE           CAB         QA           10140-         LTI           CAB         MH           10141-         LTI           CAB         MH           10142-         LTI           CAC         QF	PSK) EEE 802.11n (HT Mixed, 81 Mbps, 16- AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Y Z X	5.13 5.05	42232423	19203		150.0	± 9.6 %
CAB QA 10119- IEE CAB QA 10140- LT CAB MI- 10141- LT CAB MI- 10142- LT CAC QF	AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X	5.05	67.46		0.00	10010	- 0.0 /0
CAB QA 10119- IEE CAB QA 10140- LT CAB MI- 10141- LT CAB MI- 10142- LT CAC QF	AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	Z X	5.05		16.88		150.0	
CAB QA 10119- IEE CAB QA 10140- LT CAB MI- 10141- LT CAB MI- 10142- LT CAC QF	AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-	x	the second s	67.19	16.55		150.0	
CAB QA 10119- IEE CAB QA 10140- LT CAB MI- 10141- LT CAB MI- 10142- LT CAC QF	AM) EEE 802.11n (HT Mixed, 135 Mbps, 64-		5.53	67.72	16.93	0.00	150.0	± 9.6 %
CAB QA 10140- LT CAB MH 10141- LT CAB MH 10142- LT CAC QF		Y I			122004	0.00	1952,235	1 0.0 %
CAB QA 10140- LT CAB MH 10141- LT CAB MH 10142- LT CAC QF			5.51	67.87	17.07	_	150.0	
CAB QA 10140- LT CAB MH 10141- LT CAB MH 10142- LT CAC QF		Z	5.41	67.55	16.72	0.00	150.0	
CAB MH 10141- LT CAB MH 10142- LT CAC QF		x	5.26	67.59	16.79	0.00	150.0	±9.6 %
CAB MH 10141- LT CAB MH 10142- LT CAC QF		Y	5.25	67.77	16.95		150.0	
CAB MH 10141- LT CAB MH 10142- LT CAC QF		Z	5.15	67.45	16.60		150.0	
CAB MI 10142- LTI CAC QF	TE-FDD (SC-FDMA, 100% RB, 15 Hz, 16-QAM)	X	3.46	68.05	16.50	0.00	150.0	±9.6 %
CAB MI 10142- LTI CAC QF		Y	3.50	68.55	16.85		150.0	
CAB MI 10142- LTI CAC QF		Z	3.36	67.76	16.20		150.0	
10142- LT CAC QF	TE-FDD (SC-FDMA, 100% RB, 15 Hz, 64-QAM)	X	3.58	68.14	16.65	0.00	150.0	± 9.6 %
CAC QF		Y	3.62	68.65	17.00		150.0	
CAC QF		z	3.48	67.88	16.37		150.0	
	TE-FDD (SC-FDMA, 100% RB, 3 MHz, PSK)	X	2.23	70.98	17.19	0.00	150.0	±9.6 %
	( SK)	Y	2.46	73.31	40.00		160.0	
10110		Z	2.46	69.88	18.33		150.0	
	TE-FDD (SC-FDMA, 100% RB, 3 MHz, 6-QAM)	X	2.04	70.97	16.37 16.97	0.00	150.0 150.0	±9.6 %
10	- see strip	Y	3.04	73.29	18.02		150.0	
		Z	2.55	70.19	16.25		150.0	
the second se	TE-FDD (SC-FDMA, 100% RB, 3 MHz, 4-QAM)	X	2.36	67.83	14.95	0.00	150.0	± 9.6 %
04	e successfully	Y	2.44	68.84	15.41		150.0	
		Z	2.44				the second design of the second s	
10145- LT	TE-EDD (SC-EDMA 100% DB 14			67.08	14.22	0.00	150.0	+0.0.01
13/2/02/02/02 02/02/02/02/02/02/02/02/02/02/02/02/02/0	TE-FDD (SC-FDMA, 100% RB, 1.4 Hz, QPSK)	×	1.37	67.00	12.74	0.00	150.0	±9.6 %
		Y	1.47	68.36	13.15		150.0	
		Z	1.09	64.53	10.89		150.0	
10 C - C - C - C - C - C - C - C - C - C	TE-FDD (SC-FDMA, 100% RB, 1.4 Hz, 16-QAM)	×	1.50	64.09	10.18	0.00	150.0	±9.6 %
		Y	1.46	64.20	10.14		150.0	
		Z	1.34	62.84	8.82		150.0	
		X	1.70	65.48	11.02	0.00	150.0	± 9.6 %
in the second second	TE-FDD (SC-FDMA, 100% RB, 1.4 Hz 64-OAM)	Y	1.70	65.97	11.15		150.0	
	TE-FDD (SC-FDMA, 100% RB, 1.4 Hz, 64-QAM)		1.46	63.74	9.40		150.0	

September 29, 2016

10149- CAB	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	x	3.00	68.21	16.56	0.00	150.0	±9.6 %
and a second		Y	3.06	68.92	17.01		150.0	
		Z	2.89	67.84	16.19		150.0	
10150- CAB	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	x	3.12	68.16	16.58	0.00	150.0	± 9.6 %
		Y	3.17	68.84	17.01		150.0	
		Z	3.01	67.84	16.23		150.0	
10151- CAB	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	x	5.47	75.55	20.46	3.98	65.0	±9.6 %
		Y	4.99	74.68	20.33	1	65.0	
		Z	5.00	74.31	19.78	1	65.0	
10152- CAB	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM)	x	4.98	71.60	19.18	3.98	65.0	±9.6 %
		Y	4.57	70.54	18.83		65.0	
		Z	4.65	70.66	18.56		65.0	and the second second
10153- CAB	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM)	x	5.31	72.52	19.96	3.98	65.0	±9.6 %
		Y	4.89	71.51	19.65		65.0	
		Z	4.98	71.62	19.37		65.0	
10154- CAC	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	x	2.47	70.94	17.63	0.00	150.0	±9.6 %
		Y	2.65	72.76	18.59		150.0	
		Z	2.29	70.03	16.97		150.0	
10155- CAC	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	x	2.77	69.55	17.05	0.00	150.0	±9.6 %
		Y	2.91	70.97	17.81		150.0	
		Z	2.65	69.12	16.59		150.0	
10156- CAC	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	×	2.12	71.56	17.17	0.00	150.0	± 9.6 %
		Y	2.45	74.60	18.56		150.0	
		Z	1.89	70.11	16.14		150.0	
10157- CAC	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	x	2,27	68.91	15.20	0.00	150.0	± 9.6 %
		Y	2.43	70.48	15.88		150.0	
		Z	2.06	67.81	14.26		150.0	
10158- CAC	LTE-FDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	×	2.92	69.69	17.17	0.00	150.0	± 9.6 %
		Y	3.07	71.09	17.92		150.0	
		Z	2.80	69.31	16.74		150.0	· · · · ·
10159- CAC	LTE-FDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	×	2.39	69.45	15.50	0.00	150.0	± 9.6 %
		Y	2.60	71.22	16.27		150.0	
		Z	2.17	68.30	14.54		150.0	· · · · · · · · · · · · · · · · · · ·
10160- CAB	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	×	2.92	70.01	17.28	0.00	150.0	± 9.6 %
		Y	3.04	71.16	17.94		150.0	-
		Z	2.77	69.39	16.80		150.0	
10161- CAB	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	×	3.02	68.16	16.53	0.00	150.0	± 9.6 %
		Y	3.08	68.92	16.99		150.0	
		Z	2.91	67.84	16.16		150.0	
10162- CAB	LTE-FDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	×	3.13	68.31	16.63	0.00	150.0	± 9.6 %
		Y	3.19	69.07	17.09		150.0	
		Z	3.02	68.02	16.28		150.0	
10166- CAC	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	X	3.31	68.86	19.00	3.01	150.0	± 9.6 %
		Y	3.31	69.26	19.38		150.0	
		Z	3.34	69.52	19.19		150.0	
10167- CAC	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	x	3.86	71.43	19.37	3.01	150.0	± 9.6 %
		Y	3.83	71.75	19.71		150.0	
		Z	4.12	73.12	19.92		150.0	

#### September 29, 2016

10168- CAC	LTE-FDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	×	4.30	73.82	20.81	3.01	150.0	± 9.6 %
		Y	4.34	74.63	21.40		150.0	
a description of the		Z	4.74	76.16	21.61		150.0	
10169- CAB	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	X	2.55	67.28	18.39	3.01	150.0	± 9.6 %
		Y	2.53	67.34	18.64		150.0	
		Z	2.66	68.48	18.81		150.0	-
10170- CAB	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	x	3.18	72.46	20.65	3.01	150.0	± 9.6 %
		Y	3.15	72.66	21.04		150.0	
		Z	3.80	76.29	22.04		150.0	
10171- AAB	LTE-FDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	x	2.66	68.73	17.89	3.01	150.0	± 9.6 %
		Y	2.60	68.52	18.02	-	150.0	
		Z	2.96	71.04	18.67		150.0	
10172- CAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK)	x	3.68	77.23	23.15	6.02	65.0	±9.6 %
		Y	3.29	75.64	22.83		65.0	-
		Z	4.07	79.96	24.07		65.0	
10173- CAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM)	x	6.37	85.05	24.09	6.02	65.0	± 9.6 %
Contra Co		Y	5.22	82.65	23.74		65.0	
		Z	6.85	86.84	24.39	-	65.0	
10174- CAB	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64-QAM)	x	4.45	78.21	21.05	6.02	65.0	± 9.6 %
		Y	4.12	77.81	21.34		65.0	
		Z	5.27	81.52	21.90		65.0	
10175- CAC	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	x	2.53	67.02	18.15	3.01	150.0	±9.6 %
		Y	2.50	67.04	18.38		150.0	
		Z	2.62	68.16	18.55		150.0	-
10176- CAC	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	×	3.18	72.49	20.66	3.01	150.0	± 9.6 %
		Y	3.15	72.68	21.05		150.0	
		Z	3.80	76.33	22.06		150.0	
10177- CAE	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	X	2.54	67.15	18.24	3.01	150.0	± 9.6 %
		Y	2,52	67.19	18.47		150.0	
		Z	2.65	68.31	18.64		150.0	
10178- CAC	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM)	x	3.16	72.30	20.55	3.01	150.0	± 9.6 %
		Y	3.13	72.47	20.93	-	150.0	
		Z	3.76	76.08	21.93		150.0	
10179- CAC	LTE-FDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	X	2.90	70.53	19.16	3.01	150.0	± 9.6 %
		Y	2.85	70.49	19.41		150.0	
		Z	3.33	73.51	20.20		150.0	
10180- CAC	LTE-FDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM)	x	2.65	68.67	17.85	3.01	150.0	± 9.6 %
		Y	2.59	68.46	17.98		150.0	
		Z	2.95	70.97	18.62		150.0	
10181- CAB	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	x	2.54	67.14	18.23	3.01	150.0	± 9.6 %
		Y	2.52	67.17	18.47		150.0	
		Z	2.64	68.29	18.63		150.0	
10182- CAB	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	×	3.15	72.28	20.54	3.01	150.0	±9.6 %
		Y	3.12	72.45	20.92		150.0	
		Z	3.75	76.05	21.92		150.0	
10183- AAA	LTE-FDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	x	2.65	68.65	17.84	3.01	150.0	± 9.6 %
AAA								
		Y	2.59	68.43	17.96		150.0	

### September 29, 2016

10184- CAC	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	×	2.55	67.18	18.26	3.01	150.0	±9.6 %
UAC	GP OIL	Y	2.53	67.22	18.49		150.0	
		Z	2.65	68.34	18.65		150.0	
10185- CAC	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 16- QAM)	x	3.17	72.35	20.58	3.01	150.0	±9.6 %
UNU	(anity)	Y	3.14	72.52	20.96		150.0	1-
		Z	3.78	76.15	21.96		150.0	
10186-	LTE-FDD (SC-FDMA, 1 RB, 3 MHz, 64-	x	2.66	68.72	17.87	3.01	150.0	±9.6 %
AAC	QAM)	Y	2.60	68.50	18.00		150.0	
		Z	2.96	71.02	18.65		150.0	
10187- CAC	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	X	2.56	67.23	18.32	3.01	150.0	±9.6 %
UAC	QF3N)	Y	2.53	67.27	18.56	1	150.0	
		z	2.66	68.40	18.73		150.0	
10188- CAC	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	x	3.26	72.94	20.95	3.01	150.0	±9.6 %
CAC	TO-QAWI)	Y	3.23	73.20	21.37		150.0	
		Z	3.93	76.98	22.42		150.0	
10100	LTE-FDD (SC-FDMA, 1 RB, 1.4 MHz,	X	2.71	69.10	18.15	3.01	150.0	±9.6 %
10189- AAC	LTE-FDD (SC-FDMA, 1 KB, 1.4 MHZ, 64-QAM)	100		68.91	18.30	0.01	150.0	
		Y	2.66	71.52	18.97		150.0	
		Z	3.04		16.48	0.00	150.0	± 9.6 %
10193- CAB	IEEE 802.11n (HT Greenfield, 6.5 Mbps, BPSK)	X	4.56	66.88		0.00	150.0	1 3.0 %
		Y	4.55	67.13	16.66			
		Z	4.47	66.83	16.29	0.00	150.0	1000
10194- CAB	IEEE 802.11n (HT Greenfield, 39 Mbps, 16-QAM)	×	4.72	67.18	16.60	0.00	150.0	± 9.6 %
040 10		Y	4.71	67.41	16.79		150.0	
		Z	4.63	67.10	16.42		150.0	
10195- CAB	IEEE 802.11n (HT Greenfield, 65 Mbps, 64-QAM)	×	4.76	67.21	16.62	0.00	150.0	± 9.6 %
one	01 00 00	Y	4.75	67.44	16.80		150.0	
		Z	4.67	67.13	16.44		150.0	
10196- CAB	IEEE 802.11n (HT Mixed, 6.5 Mbps, BPSK)	x	4.56	66.93	16.49	0.00	150.0	± 9.6 %
CAD	brony	Y	4.55	67.17	16.67		150.0	
		Z	4.47	66.86	16.30		150.0	
10197- CAB	IEEE 802.11n (HT Mixed, 39 Mbps, 16- QAM)	X	4.74	67.20	16.61	0.00	150.0	± 9.6 %
GAD	(GAM)	Y	4.73	67.43	16.80		150.0	
		Z	4.64	67.12	16.43		150.0	
10198- CAB	IEEE 802.11n (HT Mixed, 65 Mbps, 64- QAM)	X	4.76	67.22	16.63	0.00	150.0	± 9.6 %
5/10		Y	4.75	67.45	16.81		150.0	
		Z	4.67	67.14	16.45		150.0	
10219- CAB	IEEE 802.11n (HT Mixed, 7.2 Mbps, BPSK)	X	4.51	66.96	16.46	0.00	150.0	±9.6 %
		Y	4.50	67.21	16.65	-	150.0	
		Z	4.42	66.89	16.27		150.0	
			and the second se	67.16	16.60	0.00	150.0	± 9.6 %
10220- CAB	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16- QAM)	X	4.73					
10220- CAB	IEEE 802.11n (HT Mixed, 43.3 Mbps, 16- QAM)	X Y	4.73	67.39	16.78		150.0	
					16.78 16.42		150.0 150.0	
CAB 10221-	QAM) IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-	Y	4.72	67.39		0.00		± 9.6 %
CAB	QAM)	Y Z X	4.72 4.63 4.77	67.39 67.08 67.14	16.42 16.61	0.00	150.0	± 9.6 %
CAB 10221-	QAM) IEEE 802.11n (HT Mixed, 72.2 Mbps, 64-	Y Z X Y	4.72 4.63 4.77 4.76	67.39 67.08 67.14 67.37	16.42 16.61 16.79	0.00	150.0 150.0 150.0	± 9.6 %
CAB 10221- CAB 10222-	QAM) IEEE 802.11n (HT Mixed, 72.2 Mbps, 64- QAM) IEEE 802.11n (HT Mixed, 15 Mbps,	Y Z X	4.72 4.63 4.77	67.39 67.08 67.14	16.42 16.61	0.00	150.0 150.0	± 9.6 %
CAB 10221- CAB	QAM) IEEE 802.11n (HT Mixed, 72.2 Mbps, 64- QAM)	Y Z X Y Z	4.72 4.63 4.77 4.76 4.68	67.39 67.08 67.14 67.37 67.07	16.42 16.61 16.79 16.43		150.0 150.0 150.0 150.0	

September 29, 2016

10223- CAB	IEEE 802.11n (HT Mixed, 90 Mbps, 16- QAM)	X	5.43	67.54	16.86	0.00	150.0	± 9.6 %
	1. (1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1	Y	5.41	67.68	16.99		150.0	
		Z	5.31	67.40	16.66		150.0	
10224- CAB	IEEE 802.11n (HT Mixed, 150 Mbps, 64- QAM)	X	5.16	67.39	16.69	0.00	150.0	± 9.6 %
-	- Me-mex	Y	5.15	67.58	16.86		150.0	
		Z	5.07	67.29	16.52		150.0	
10225- CAB	UMTS-FDD (HSPA+)	X	2.87	66.79	15.84	0.00	150.0	± 9.6 %
		Y	2.90	67.40	16.19		150.0	
		Z	2.77	66.56	15.46		150.0	
10226- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM)	X	6.75	86.12	24.57	6.02	65.0	±9.6 %
		Y	5.51	83.70	24.23		65.0	
		Z	7.36	88.18	24.95		65.0	
10227- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 64-QAM)	X	6.55	84.30	23.23	6.02	65.0	±9.6 %
		Y	5.57	82.77	23.20		65.0	
		Z	7.03	85.94	23.43		65.0	
10228- CAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK)	×	4.87	82.78	25.35	6.02	65.0	±9.6 %
		Y	4.01	79.77	24.60		65.0	
ne no len		Z	4.44	81.69	24.80		65.0	
10229- CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16- QAM)	x	6.43	85.15	24.14	6.02	65.0	±9.6 %
		Y	5.26	82.75	23.79		65.0	
		Z	6.91	86.96	24.44		65.0	
	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64- QAM)	X	6.20	83.33	22.81	6.02	65.0	± 9.6 %
		Y	5.27	81.77	22.76		65.0	
		Z	6.55	84.76	22.95		65.0	
10231- CAB	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK)	X	4.70	82.07	25.00	6.02	65.0	± 9.6 %
		Y	3.88	79.08	24.24		65.0	
		Z	4.29	80.96	24.43		65.0	
10232- CAB	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM)	X	6.41	85.13	24.13	6.02	65.0	± 9.6 %
		Y	5.25	82.73	23.78		65.0	
		Z	6.89	86.94	24.43		65.0	
10233- CAB	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM)	×	6.18	83.31	22.81	6.02	65.0	± 9.6 %
2012/01/1		Y	5.26	81.74	22.75		65.0	
		Z	6.54	84.73	22.94		65.0	
10234- CAB	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK)	×	4.57	81.42	24.64	6.02	65.0	± 9.6 %
		Y	3.78	78.50	23.89		65.0	
		Z	4.16	80.31	24.06		65.0	
10235- CAB	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16-QAM)	x	6.41	85.15	24.14	6.02	65.0	± 9.6 %
		Y	5.25	82.75	23.79		65.0	
		Z	6.90	86.97	24.44		65.0	
10236- CAB	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-QAM)	×	6.25	83.46	22.85	6.02	65.0	± 9.6 %
		Y	5.31	81.89	22.79		65.0	
		Z	6.62	84.89	22.99		65.0	
10237- CAB	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK)	x	4.70	82.10	25.02	6.02	65.0	±9.6 %
		Y	3.88	79.10	24.25		65.0	
0.000		Z	4.28	80.99	24.45		65.0	
10238- CAB	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16-QAM)	x	6.39	85.10	24.12	6.02	65.0	± 9.6 %
		Y	5.23	82.70	23.77		65.0	
		Z	6.87	86.91	24.42		65.0	

### September 29, 2016

10239- CAB	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64-QAM)	X	6.16	83.27	22.79	6.02	65.0	±9.6 %
		Y	5.24	81.70	22.74		65.0	
		Z	6.51	84.68	22.93		65.0	
10240- CAB	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK)	X	4.69	82.06	25.00	6.02	65.0	±9.6 %
0710	a. orig	Y	3.87	79.07	24.24		65.0	
		Z	4.27	80.95	24.43		65.0	
10241- CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM)	x	6.22	77.11	23.59	6.98	65.0	± 9.6 %
		Y	5.77	75.93	23.22		65.0	
		Z	6.24	77.90	23.77		65.0	
10242- CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM)	x	5.51	74.58	22.40	6.98	65.0	±9.6 %
		Y	5.24	73.95	22.24		65.0	
		Z	5.87	76.65	23.15	_	65.0	
10243- CAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK)	×	4.68	71.62	21.91	6.98	65.0	± 9.6 %
		Y	4.54	71.28	21.85		65.0	
		Z	4.81	72.91	22.39		65.0	and action
10244- CAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	x	3.78	69.70	15.17	3.98	65.0	± 9.6 %
		Y	3.50	69.28	15.08	1	65.0	
		Z	3.41	68.43	14.10		65.0	
10245- CAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	X	3.73	69.26	14.92	3.98	65.0	± 9.6 %
		Y	3.44	68.78	14.78		65.0	
		Z	3.37	68.01	13.85		65.0	
10246- CAB	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	Х	3.79	73.29	17.29	3.98	65.0	± 9.6 %
0/10	or ony	Y	3.43	72.67	17.14		65.0	
		Z	3.13	70.66	15.73		65.0	
10247- CAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM)	X	3.97	70.72	16.89	3.98	65.0	± 9.6 %
01.10		Y	3.63	70.00	16.62		65.0	
		Z	3.55	69.25	15.83		65.0	
10248- CAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM)	×	3.99	70.33	16.69	3.98	65.0	± 9.6 %
		Y	3.64	69.52	16.37		65.0	
		Z	3.58	68.89	15.65		65.0	
10249- CAB	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK)	X	4.84	77.09	19.91	3.98	65.0	±9.6 %
		Y	4.38	76.49	19.88		65.0	
		Z	4.09	74.63	18.58		65.0	
10250- CAB	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM)	X	4.86	73.47	, 19.93	3.98	65.0	± 9.6 %
		Y	4.46	72.59	19.71		65.0	
		Z	4.48	72.29	19.16		65.0	
10251- CAB	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM)	×	4.72	71.74	18.79	3.98	65.0	± 9.6 %
		Y	4.32	70.75	18.45		65.0	
		Z	4.36	70.66	18.04		65.0	
10252- CAB	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK)	X	5.33	77.58	21.22	3.98	65.0	± 9.6 %
		Y	4.80	76.68	21.10		65.0	
		Z	4.73	75.83	20.29		65.0	
10253- CAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM)	X	4.89	71.14	18.94	3.98	65.0	± 9.6 %
		Y	4.51	70.14	18.59		65.0	
		Z	4.59	70.28	18.34		65.0	
10254- CAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM)	X	5.20	71.99	19.64	3.98	65.0	± 9.6 %
CAB		V	4.00	71.02	19.32		65.0	
		Y	4.80	71.02	19.52		00.0	

September 29, 2016

10255- CAB	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK)	X	5.23	74.92	20.40	3.98	65.0	± 9.6 %
		Y	4.77	73.99	20.22		65.0	
		Z	4.80	73.74	19.73		65.0	
10256- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM)	X	2.84	65.90	12.18	3.98	65.0	± 9.6 %
		Y	2.58	65.31	11.90		65.0	
		Z	2.51	64.65	11.02		65.0	
10257- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM)	X	2.81	65.49	11.88	3.98	65.0	± 9.6 %
		Y	2.56	64.86	11.56		65.0	
		Z	2.50	64.29	10.74		65.0	
10258- CAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK)	X	2.76	68.49	14.16	3.98	65.0	± 9.6 %
		Y	2.45	67.63	13.77		65.0	
		Z	2.29	66.25	12.61		65.0	
10259- CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM)	X	4.33	71.85	18.04	3.98	65.0	± 9.6 %
		Y	3.97	71.09	17.80		65.0	
		Z	3.92	70.49	17.09		65.0	
10260- CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM)	×	4.37	71.63	17.94	3.98	65.0	± 9.6 %
		Y	4.01	70.86	17.68		65.0	
1515 A 1940		Z	3.97	70.31	17.01		65.0	
10261- CAB	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK)	X	4.83	76.58	20.18	3.98	65.0	± 9.6 %
		Y	4.36	75.82	20.08		65.0	
		Z	4.20	74.54	19.05		65.0	
	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM)	X	4.85	73.42	19.88	3.98	65.0	± 9.6 %
		Y	4.45	72.54	19.66		65.0	
		Z	4.46	72.24	19.12		65.0	
10263- CAB	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM)	X	4.72	71.71	18.78	3.98	65.0	± 9.6 %
		Y	4.31	70.73	18.44		65.0	
		Z	4.35	70.63	18.03		65.0	
10264- CAB	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK)	X	5.28	77.41	21.12	3.98	65.0	± 9.6 %
		Y	4.76	76.49	21.00		65.0	
		Z	4.69	75.66	20.19		65.0	
10265- CAB	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM)	X	4.98	71.60	19.18	3.98	65.0	± 9.6 %
	e en	Y	4.57	70.54	18.83		65.0	
		Z	4.65	70.66	18.57		65.0	
10266- CAB	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM)	×	5.31	72.51	19.95	3.98	65.0	± 9.6 %
		Y	4.89	71.50	19.64		65.0	
		Z	4.97	71.61	19.36		65.0	
10267- CAB	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK)	x	5.46	75.51	20.44	3.98	65.0	± 9.6 %
		Y	4.98	74.64	20.31		65.0	
		Z	4.99	74.28	19.77		65.0	
10268- CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM)	X	5.63	71.70	19.66	3.98	65.0	± 9.6 %
		Y	5.21	70.69	19.32		65.0	
		Z	5.32	70.92	19.15		65.0	
10269- CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM)	x	5.63	71.35	19.55	3.98	65.0	± 9.6 %
		Y	5.23	70.35	19.21		65.0	
		Z	5.34	70.62	19.06		65.0	
10270- CAB	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	x	5.55	73.38	19.70	3.98	65.0	± 9.6 %
		Y	5.13	72.57	19.54		65.0	
			0.10	1 60-141	10.04		00.0	

September 29, 2016

10274- CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.10)	x	2.69	67.42	15.91	0.00	150.0	±9.6 %
		Y	2.75	68.22	16.36		150.0	
		Z	2.60	67.16	15.51		150.0	
10275- CAB	UMTS-FDD (HSUPA, Subtest 5, 3GPP Rel8.4)	x	1.83	70.40	17.15	0.00	150.0	±9.6 %
477. TUP.		Y	2.03	72.78	18.38		150.0	
		Z	1.68	69.22	16.30		150.0	
10277-	PHS (QPSK)	X	1.85	60.43	5.95	9.03	50.0	±9.6 %
CAA		Y	1.93	60.53	6.11		50.0	
		Z	1.93	60.55	5.62		50.0	
10278- CAA	PHS (QPSK, BW 884MHz, Rolloff 0.5)	X	3.26	67.17	12.22	9.03	50.0	± 9.6 %
CAA		Y	3.32	67.01	12.15		50.0	
		Z	2.98	65.72	11.14		50.0	
40070	DUC (ODCK DW 994MUT Dolloff 0 29)	X	3.37	67.46	12.42	9.03	50.0	±9.6 %
10279- CAA	PHS (QPSK, BW 884MHz, Rolloff 0.38)					5.05		1 0.0 %
		Y	3.42	67.28	12.34		50.0	
		Z	3.07	65.97	11.33	0.00	50.0	1000
10290- AAB	CDMA2000, RC1, SO55, Full Rate	×	2.05	74.04	16.37	0.00	150.0	±9.6 %
		Y	3.52	81.70	19.05		150.0	
		Z	1.51	70.11	14.15		150.0	
10291- AAB	CDMA2000, RC3, SO55, Full Rate	x	1.16	70.91	15.09	0.00	150.0	±9.6 %
		Y	1.98	78.67	18.04		150.0	8
		Z	0.88	67.38	12.85	-	150.0	
10292- AAB	CDMA2000, RC3, SO32, Full Rate	х	3.02	85.41	21.03	0.00	150.0	± 9.6 %
nno		Y	100.00	134.44	33.38	6-7-1	150.0	
		Z	1.64	76.50	17.15		150.0	
10293- AAB	CDMA2000, RC3, SO3, Full Rate	X	33.02	121.00	31.37	0.00	150.0	±9.6 %
7010		Y	100.00	139.57	35.76		150.0	
		Z	8.07	98.75	24.80		150.0	
10295- AAB	CDMA2000, RC1, SO3, 1/8th Rate 25 fr.	X	8.33	82.26	21.81	9.03	50.0	±9.6 %
1010		Y	9.39	83.62	22.13		50.0	
		Z	7.86	80.54	20.63		50.0	
10297- AAA	LTE-FDD (SC-FDMA, 50% RB, 20 MHz, QPSK)	X	2.95	71.17	17,66	0.00	150.0	± 9.6 %
000	Grony	Y	3.11	72.56	18.44		150.0	
		Z	2.77	70.39	17.13		150.0	
10298- AAB	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, QPSK)	X	1.85	70.66	15.62	0.00	150.0	± 9.6 %
1.0.10		Y	2.24	74.04	16.99		150.0	
		Z	1.54	68.34	14.05		150.0	
10299- AAB	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM)	×	2.20	68.15	13.34	0.00	150.0	± 9.6 %
	is security	Y	2.39	69.78	14.08		150.0	
		Z	2.08	67.29	12.27		150.0	
10300- AAB	LTE-FDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM)	X	1.64	63.80	10.42	0.00	150.0	± 9.6 %
		Y	1.59	63.91	10.43		150.0	
		Z	1.54	63.22	9.51		150.0	10.00
10301- AAA	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, QPSK, PUSC)	×	4.46	64.63	17.23	4.17	50.0	± 9.6 %
2802085		Y	4.49	65.06	17.44		50.0	
		Z	4.27	64.20	16.83	1	50.0	
10302- AAA	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, QPSK, PUSC, 3 CTRL symbols)	×	4.98	65.49	18.08	4.96	50.0	± 9.6 %
		_	the second se			-		
		Y	4.97	65.69	18.16		50.0	

September 29, 2016

10303- AAA	IEEE 802.16e WIMAX (31:15, 5ms, 10MHz, 64QAM, PUSC)	×	4.71	65.04	17.86	4.96	50.0	± 9.6 %
		Y	4.70	65.25	17.95		50.0	
10001		Z	4.60	64.94	17.62		50.0	
10304- AAA	IEEE 802.16e WiMAX (29:18, 5ms, 10MHz, 64QAM, PUSC)	×	4.55	65.03	17.43	4.17	50.0	± 9.6 %
		Y	4.54	65.29	17.55		50.0	
		Z	4.44	64.95	17.19		50.0	
10305- AAA	IEEE 802.16e WiMAX (31:15, 10ms, 10MHz, 64QAM, PUSC, 15 symbols)	x	3.99	65.76	18.80	6.02	35.0	± 9.6 %
		Y	4.07	66.47	19.05		35.0	
10000		Z	3.90	65.70	18.47	-	35.0	
10306- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 64QAM, PUSC, 18 symbols)	x	4.39	65.23	18.61	6.02	35.0	± 9.6 %
		Y	4.43	65.72	18.79		35.0	
		Z	4.30	65.19	18.34		35.0	-
10307- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, PUSC, 18 symbols)	x	4.27	65.29	18.54	6.02	35.0	±9.6 %
		Y	4.31	65.81	18.73		35.0	
100		Z	4.17	65.22	18.24		35.0	-
10308- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, PUSC)	X	4.24	65.45	18.66	6.02	35.0	± 9.6 %
		Y	4.28	65.98	18.86		35.0	
		Z	4.14	65.37	18.37		35.0	
10309- AAA	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, 16QAM, AMC 2x3, 18 symbols)	X	4.43	65.42	18.74	6.02	35.0	± 9.6 %
		Y	4.47	65.88	18.91		35.0	
		Z	4.33	65.33	18.45		35.0	
10310- AAA 10MHz, QI	IEEE 802.16e WiMAX (29:18, 10ms, 10MHz, QPSK, AMC 2x3, 18 symbols)	X	4.33	65.26	18.57	6.02	35.0	± 9.6 %
		Y	4.38	65.77	18.77	_	35.0	
		Z	4.24	65.22	18.31		35.0	
10311- AAA	LTE-FDD (SC-FDMA, 100% RB, 15 MHz, QPSK)	X	3.32	70.27	17.19	0.00	150.0	± 9.6 %
		Y	3.49	71.48	17.88		150.0	
		Z	3.14	69.58	16.73		150.0	
10313- AAA	iDEN 1:3	X	2.34	69.26	14.41	6.99	70.0	± 9.6 %
_		Y	2.05	68.81	14.60		70.0	
		Z	1.98	67.44	13.44		70.0	
10314- AAA	IDEN 1:6	X	3.56	75.90	19.90	10.00	30.0	±9.6 %
		Y	4.44	79.56	21.45		30.0	
		Z	3.07	73.30	18.62		30.0	
10315- AAB	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 96pc duty cycle)	X	1.11	64.48	16.04	0.17	150.0	± 9.6 %
		Y	1.13	65.36	16.82		150.0	
		Z	1.07	63.93	15.43		150.0	
10316- AAB	IEEE 802.11g WiFi 2.4 GHz (ERP- OFDM, 6 Mbps, 96pc duty cycle)	X	4.57	66.77	16.49	0.17	150.0	± 9.6 %
		Y	4.56	66.98	16.66		150.0	
		Z	4.48	66.69	16.30		150.0	
10317- AAB	IEEE 802.11a WiFi 5 GHz (OFDM, 6 Mbps, 96pc duty cycle)	×	4.57	66.77	16.49	0.17	150.0	± 9.6 %
		Y	4.56	66.98	16.66		150.0	
		Z	4.48	66.69	16.30		150.0	
10400- AAC	IEEE 802.11ac WiFi (20MHz, 64-QAM, 99pc duty cycle)	x	4.71	67.24	16.60	0.00	150.0	±9.6 %
		Y	4.69	67.45	16.77		150.0	
		Z	4.61	67.14	16.41		150.0	
10401-	IEEE 802.11ac WiFi (40MHz, 64-QAM, 99pc duty cycle)	X	5.46	67.46	16.79	0.00	150.0	±9.6 %
AC								
010		Y	5.44	67.61	16.92		150.0	

### September 29, 2016

10402- AAC	IEEE 802.11ac WiFi (80MHz, 64-QAM, 99pc duty cycle)	x	5.67	67.60	16.72	0.00	150.0	± 9.6 %
		Y	5.66	67.75	16.85		150.0	
		Z	5.58	67.51	16.55		150.0	
10403- AAB	CDMA2000 (1xEV-DO, Rev. 0)	X	2.05	74.04	16.37	0.00	115.0	±9.6 %
AAB		Y	3.52	81.70	19.05		115.0	
-		Z	1.51	70.11	14.15		115.0	
10101	001400000 (4.51/ DO, Dorr A)	X	2.05	74.04	16.37	0.00	115.0	±9.6 %
10404- AAB	CDMA2000 (1xEV-DO, Rev. A)					0.00	21/2 2020/ L	2 0.0 /4
		Y	3.52	81.70	19.05		115.0	
		Z	1.51	70.11	14.15	0.00	115.0	1000
10406- AAB	CDMA2000, RC3, SO32, SCH0, Full Rate	×	100.00	125.83	31.49	0.00	100.0	± 9.6 %
		Y	100.00	130.71	33.64		100.0	(
		Z	100.00	116.24	27.04	-	100.0	
10410- AAA	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	5.91	85.80	20.18	3.23	80.0	±9.6 %
		Y	12.75	99.43	25.10		80.0	
		Z	4.83	82.66	18.51		80.0	
10415- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 99pc duty cycle)	x	1.06	64.06	15.76	0.00	150.0	± 9.6 %
1001	mops, cope daty cycloj	Y	1.09	65.07	16.59		150.0	
		z	1.03	63.60	15.19		150.0	
10416- AAA	IEEE 802.11g WiFi 2.4 GHz (ERP- OFDM, 6 Mbps, 99pc duty cycle)	×	4.56	66.91	16.55	0.00	150.0	± 9.6 %
PV-V1	or bill, o hisps, cope daily of day	Y	4.55	67.15	16.74		150.0	
		Z	4.47	66.85	16.37		150.0	
10417- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 99pc duty cycle)	X	4.56	66.91	16.55	0.00	150.0	± 9.6 %
////	mops, sopedaty cycle)	Y	4.55	67.15	16.74		150.0	
		z	4.47	66.85	16.37	-	150.0	
10418- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 6 Mbps, 99pc duty cycle, Long preambule)	×	4.55	67.10	16.59	0.00	150.0	± 9.6 %
	produtionally	Y	4.55	67.37	16.79		150.0	
		Z	4.47	67.04	16.41		150.0	
10419- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 6 Mbps, 99pc duty cycle, Short preambule)	×	4.57	67.04	16.58	0.00	150.0	±9.6 %
		Y	4.57	67.29	16.77		150.0	
		Z	4.49	66.98	16.40		150.0	
10422- AAA	IEEE 802.11n (HT Greenfield, 7.2 Mbps, BPSK)	X	4.68	67.01	16.58	0.00	150.0	± 9.6 %
		Y	4.68	67.25	16.76		150.0	
		Z	4.59	66.95	16.41	1	150.0	
10423- AAA	IEEE 802.11n (HT Greenfield, 43.3 Mbps, 16-QAM)	X	4.84	67.31	16.69	0.00	150.0	± 9.6 %
		Y	4.83	67.54	16.86		150.0	
		Z	4.74	67.24	16.51	-	150.0	
10424-	IEEE 802.11n (HT Greenfield, 72.2	X	4.76	67.27	16.67	0.00	150.0	± 9.6 %
AAA	Mbps, 64-QAM)	_				0.00	150.0	
		Y	4.75	67.51	16.85	-	and the second se	
10425-	IEEE 802.11n (HT Greenfield, 15 Mbps,	Z X	4.67 5.38	67.19 67.51	16.49 16.82	0.00	150.0 150.0	± 9.6 %
AAA	BPSK)	11	E 07	07.00	40.07		450.0	
		Y	5.37	67.69	16.97		150.0	
10100		Z	5.28	67.41	16.64	0.00	150.0	1004
10426- AAA	IEEE 802.11n (HT Greenfield, 90 Mbps, 16-QAM)	×	5.41	67.63	16.87	0.00	150.0	± 9.6 %
1997 - 1997 - 19		Y	5.40	67.81	17.03		150.0	
		Z	5.30	67.49	16.68	1	150.0	

September 29, 2016

10427- AAA	IEEE 802.11n (HT Greenfield, 150 Mbps, 64-QAM)	X	5.40	67.54	16.83	0.00	150.0	± 9.6 %
		Y	5.38	67.69	16.96		150.0	
		Z	5.28	67.37	16.62		150.0	
10430- AAA	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1)	X	4.43	72.05	18.90	0.00	150.0	± 9.6 %
		Y	4.75	73.92	19.83		150.0	
		Z	4.35	72.14	18.68		150.0	
10431- AAA	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1)	X	4.24	67.60	16.59	0.00	150.0	± 9.6 %
AAA		Y	4.24	67.95	16.81		150.0	
		Z	4.13	67.49	16.35		150.0	
10432- AAA	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1)	X	4.54	67.38	16.64	0.00	150.0	± 9.6 %
		Y	4.53	67.66	16.84		150.0	
		Z	4.43	67.29	16.44		150.0	
10433- AAA	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1)	x	4.78	67.31	16.69	0.00	150.0	±9.6 %
		Y	4.77	67.54	16.87		150.0	-
		Z	4.68	67.23	16.50		150.0	
10434- AAA	W-CDMA (BS Test Model 1, 64 DPCH)	х	4.62	73.24	18.96	0.00	150.0	± 9.6 %
		Y	5.10	75.63	20.05		150.0	
		Z	4.53	73.27	18.67		150.0	
10435- AAA	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	5.51	84.80	19.80	3.23	80.0	± 9.6 %
		Y	11.08	97.33	24.45		80.0	
		Z	4.49	81.69	18.14		80.0	
10447- AAA	LTE-FDD (OFDMA, 5 MHz, E-TM 3.1, Clipping 44%)	X	3.55	67.81	15.93	0.00	150.0	± 9.6 %
1000 California		Y	3.57	68.36	16.20		150.0	
		Z	3.41	67.56	15.53		150.0	
10448- AAA	LTE-FDD (OFDMA, 10 MHz, E-TM 3.1, Clippin 44%)	x	4.08	67.39	16.46	0.00	150.0	±9.6 %
1011010-001		Y	4.09	67.76	16.69		150.0	
		Z	3.98	67.28	16.22		150.0	
10449- AAA	LTE-FDD (OFDMA, 15 MHz, E-TM 3.1, Cliping 44%)	x	4.36	67.22	16.55	0.00	150.0	± 9.6 %
		Y	4.35	67.51	16.76		150.0	
		Z	4.26	67.13	16.34		150.0	
10450- AAA	LTE-FDD (OFDMA, 20 MHz, E-TM 3.1, Clipping 44%)	x	4.55	67.09	16.55	0.00	150.0	± 9.6 %
		Y	4.55	67.34	16.75		150.0	
		Z	4.46	67.01	16.37		150.0	
10451- AAA	W-CDMA (BS Test Model 1, 64 DPCH, Clipping 44%)	×	3.44	68.02	15.52	0.00	150.0	±9.6 %
		Y	3.47	68.59	15.77		150.0	
_		Z	3.27	67.61	15.01		150.0	
10456- AAA	IEEE 802.11ac WiFi (160MHz, 64-QAM, 99pc duty cycle)	x	6.27	68.07	16.96	0.00	150.0	±9.6 %
		Y	6.29	68.25	17.11		150.0	
	· · · · · · · · · · · · · · · · · · ·	Z	6.20	68.03	16.84		150.0	
10457- AAA	UMTS-FDD (DC-HSDPA)	x	3.83	65.55	16.26	0.00	150.0	±9.6 %
		Y	3.83	65.80	16.46		150.0	
		Z	3.78	65.52	16.08		150.0	
10458- AAA	CDMA2000 (1xEV-DO, Rev. B, 2 carriers)	x	3.23	67.20	14.77	0.00	150.0	±9.6 %
		Y	3.21	67.54	14.85		150.0	
		Z	3.03	66.66	14.13		150.0	
							100.0	
10459- AAA	CDMA2000 (1xEV-DO, Rev. B, 3 carriers)	X	4.36	65.63	15.85	0.00	150.0	± 9.6 %
		X Y	4.36	65.63 65.95	15.85	0.00	150.0	±9.6 %

September 29, 2016

10460-	UMTS-FDD (WCDMA, AMR)	x	1.17	73.22	19.18	0.00	150.0	± 9.6 %
AAA		Y	1.57	79.58	22.21		150.0	
		Z	1.00	70.63	17.58		150.0	
10461- AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	2.65	77.54	18.53	3.29	80.0	± 9.6 %
		Y	9.41	97.51	25.55		80.0	
		Z	1.84	73.13	16.45		80.0	
10462- AAA	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	0,74	60.00	7.39	3.23	80.0	±9.6 %
		Y	0.68	60.00	7.80		80.0	
10463-	LTE-TDD (SC-FDMA, 1 RB, 1.4 MHz,	X	0.70	60.00 60.00	6.55 6.77	3.23	80.0 80.0	± 9.6 %
AAA	64-QAM, UL Subframe=2,3,4,7,8,9)	Y	0.70	60.00	7.12	.0.20	80.0	2 0.0 10
		Z	0.74	60.00	5.92		80.0	
10464- AAA	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	1.97	73.34	16.31	3.23	80.0	±9.6 %
////	GI OK, OE OUDITAING-2,0,4,1,0,0	Y	6.17	90.31	22.66		80.0	
		Z	1.35	69.16	14.19		80.0	
10465- AAA	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	x	0.73	60.00	7.31	3.23	80.0	±9.6 %
		Y	0.68	60.00	7.72		80.0	
		Z	0.70	60.00	6.47		80.0	
10466- AAA	LTE-TDD (SC-FDMA, 1 RB, 3 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	×	0.76	60.00	6.72	3.23	80.0	±9.6 %
		Y	0.70	60.00	7.07		80.0	
		Z	0.75	60.00	5.88		80.0	
10467- AAA	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	2.09	74.16	16.66	3.23	80.0	±9.6 %
		Y	7.46	93.01	23.49		80.0	
		Z	1.42	69.79	14.48		80.0	-
10468- AAA	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	×	0.73	60.00	7.33	3.23	80.0	±9.6 %
5175		Y	0.68	60.00	7.74		80.0	1
		Z	0.70	60.00	6.49		80.0	
10469- AAA	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	×	0.76	60.00	6.72	3.23	80.0	± 9.6 %
		Y	0.70	60.00	7.07		80.0	
		Z	0.75	60.00	5.87		80.0	
10470- AAA	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	2.08	74.15	16.65	3.23	80.0	± 9.6 %
		Y	7.52	93.14	23.52	·	80.0	-
		Z	1.41	69.77	14.46		80.0	
10471- AAA	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	×	0.73	60.00	7.31	3.23	80.0	± 9.6 %
		Y	0.68	60.00	7.73		80.0	
10472-	LTE-TDD (SC-FDMA, 1 RB, 10 MHz, 64-	X	0.70	60.00 60.00	6.47 6.70	3.23	80.0 80.0	± 9.6 %
AAA	QAM, UL Subframe=2,3,4,7,8,9)	Y	0.70	60.00	7.05		80.0	-
		Z	0.70	60.00	5.85		80.0	-
10473- AAA	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	2.08	74.11	16.63	3.23	80.0	± 9.6 %
		Y	7.46	93.02	23.48		80.0	
		Z	1.41	69.73	14.44		80.0	
10474- AAA	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	×	0.73	60.00	7.31	3.23	80.0	± 9.6 %
1001		Y	0.68	60.00	7.72		80.0	
		7	0.70	60.00	6.47		80.0	
		Z	0.70	00.00	Q			
10475- AAA	LTE-TDD (SC-FDMA, 1 RB, 15 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	X	0.76	60.00	6.70	3.23	80.0	± 9.6 %
			and the second second second	and the second se	and the further water and the second	3.23	the second se	± 9.6 %

September 29, 2016

10477- AAA	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16- QAM, UL Subframe=2,3,4,7,8,9)	X	0.73	60.00	7.28	3.23	80.0	± 9.6 %
		Y	0.68	60.00	7.70		80.0	
		Z	0.70	60.00	6.44		80.0	
10478- AAA	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 64- QAM, UL Subframe=2,3,4,7,8,9)	X	0.76	60.00	6.69	3.23	80.0	± 9.6 %
		Y	0.70	60.00	7.04		80.0	
		Z	0.75	60.00	5.84		80.0	
10479- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	3.51	75.70	18.88	3.23	80.0	± 9.6 %
1		Y	5.44	83.49	22.01		80.0	
		Z	3.22	74.83	18.15		80.0	
10480- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	2.72	68.98	14.29	3.23	80.0	± 9.6 %
		Y	3.68	73.58	16.30		80.0	
	and the second	Z	2.25	67.06	12.90		80.0	
10481- AAA	LTE-TDD (SC-FDMA, 50% RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	2.26	66.35	12.77	3.23	80.0	± 9.6 %
		Y	2.72	69.29	14.18	-	80.0	
		Z	1.84	64.45	11.34	-	80.0	
10482- AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	2.09	67.92	14.74	2.23	80.0	± 9.6 %
		Y	2.20	69.46	15.46		80.0	
		Z	1.64	65.03	12.95		80.0	
10483- AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	2.23	65.31	12.71	2.23	80.0	± 9.6 %
		Y	2.37	66.69	13.39	-	80.0	
		Z	1.85	63.38	11.28		80.0	
10484- AAA	LTE-TDD (SC-FDMA, 50% RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	х	2.19	64.82	12.48	2.23	80.0	± 9.6 %
		Y	2.27	65.91	13.03		80.0	
		Z	1.83	63.00	11.09		80.0	
10485- AAA	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	2.54	70.24	16.86	2.23	80.0	± 9.6 %
		Y	2.66	71.91	17.75		80.0	
		Z	2.11	67.87	15.46		80.0	
10486- AAA	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	2.57	66.92	14.78	2.23	80.0	± 9.6 %
- 10. O. B.		Y	2.58	67.69	15.18		80.0	
		Z	2.22	65.23	13.58		80.0	
10487- AAA	LTE-TDD (SC-FDMA, 50% RB, 5 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	×	2.58	66.60	14.61	2.23	80.0	± 9.6 %
		Y	2.58	67.26	14.95		80.0	
		Z	2.24	64.98	13.44		80.0	
10488- AAA	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	2.94	70.16	17.68	2.23	80.0	±9.6 %
		Y	2.93	70.95	18.26		80.0	
		Z	2.59	68.53	16.73		80.0	
10489- AAA	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	3.02	67.40	16.44	2.23	80.0	± 9.6 %
		Y	2.98	67.83	16.78		80.0	
		Z	2.78	66.45	15.74		80.0	
10490- AAA	LTE-TDD (SC-FDMA, 50% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	×	3.11	67.31	16.41	2.23	80.0	± 9.6 %
		Y	3.07	67.68	16.72		80.0	
_		Z	2.87	66.41	15.73		80.0	
10491- AAA	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	3.26	69.14	17.41	2.23	80.0	±9.6 %
		Y	3.20	69.55	17.80		80.0	
		Z	2.96	67.96	16.69		80.0	
10492- AAA	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	3.41	66.95	16.60	2.23	80.0	± 9.6 %
		Y	3.33	67.10	16.82		80.0	

September 29, 2016

10493- AAA	LTE-TDD (SC-FDMA, 50% RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	3.48	66.86	16.57	2.23	80.0	± 9.6 %
		Y	3.39	66.99	16.77		80.0	
		Z	3.27	66.18	16.04		80.0	
10494- AAA	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	3.47	70.39	17.80	2.23	80.0	±9.6 %
		Y	3.44	70.95	18.27		80.0	
		Z	3.12	69.00	17.02		80.0	
10495- AAA	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	3.42	67.24	16.78	2.23	80.0	±9.6 %
		Y	3.35	67.39	17.02		80.0	
		Z	3.21	66.48	16.23		80.0	
10496- AAA	LTE-TDD (SC-FDMA, 50% RB, 20 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	3.51	67.06	16.73	2.23	80.0	± 9.6 %
		Y	3.43	67.18	16.94		80.0	
		Z	3.30	66.36	16.21		80.0	
10497- AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	1.42	63.25	11.39	2.23	80.0	± 9.6 %
		Y	1.37	63.53	11.45		80.0	
		Z	1.11	60.91	9.64	0.00	80.0	0.0.0
10498- AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	1.28	60.00	8.56	2.23	80.0	±9.6 %
		Y	1.21	60.00	8.38	1	80.0	e
		Z	1.21	60.00	8.01		80.0	
10499- AAA	LTE-TDD (SC-FDMA, 100% RB, 1.4 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	1.30	60.00	8.41	2.23	80.0	±9.6 %
		Y	1.23	60.00	8.22		80.0	
		Z	1.23	60.00	7.85		80.0	
10500- AAA	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	2.68	70.06	17,15	2.23	80.0	±9.6 %
		Y	2.74	71.29	17.87		80.0	
		Z	2.30	68.10	15.97		80.0	
10501- AAA	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	2.78	67.29	15.50	2.23	80.0	±9.6 %
		Y	2.78	67.98	15.90		80.0	1
		Z	2.48	65.95	14.53		80.0	
10502- AAA	LTE-TDD (SC-FDMA, 100% RB, 3 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	2.84	67.18	15.38	2.23	80.0	± 9.6 %
	-	Y	2.83	67.82	15.75		80.0	
		Z	2.53	65.87	14.42		80.0	
10503- AAA	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	2.91	69.98	17.59	2.23	80.0	± 9.6 %
		Ŷ	2.90	70.74	18.16		80.0	
		Z	2.57	68.37	16.65	0.00	80.0	
10504- AAA	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	3.00	67.31	16.38	2.23	80.0	± 9.6 %
		Y	2.96	67.72	16.72	-	80.0	
		Z	2.77	66.37	15.68	0.00	80.0	1000
10505- AAA	LTE-TDD (SC-FDMA, 100% RB, 5 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	×	3.10	67.22	16.35	2.23	80.0	± 9.6 %
		Y	3.05	67.59	16.66		80.0	
10500	175 TOD (00 FOMA 4000/ DD 40	Z	2.86	66.33	15.67	0.00	80.0	1000
10506- AAA	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	×	3.45	70.27	17.73	2.23	80.0	±9.6 %
		Y	3.41	70.80	18.20		80.0	
10505		Z	3.10	68.88	16.96	0.00	80.0	10.00
10507- AAA	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	3.41	67.19	16.74	2.23	80.0	± 9.6 %
		Y	3.33	67.33	16.97		80.0	
		1 1	0.00	01.00	10.01		00.0	

September 29, 2016

10508- AAA	LTE-TDD (SC-FDMA, 100% RB, 10 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	3.50	66.99	16.69	2.23	80.0	± 9.6 %
		Y	3.42	67.11	16.90		80.0	
		Z	3.29	66.30	16.17		80.0	
10509- AAA	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	X	3.85	69.41	17.39	2.23	80.0	± 9.6 %
		Y	3.80	69.73	17.73		80.0	
		Z	3.56	68.40	16.79		80.0	
10510- AAA	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	×	3.91	67.09	16.83	2.23	80.0	± 9.6 %
		Y	3.81	67.09	16.98		80.0	
		Z	3.70	66.46	16.37		80.0	
10511- AAA	LTE-TDD (SC-FDMA, 100% RB, 15 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	x	3.97	66.90	16.78	2.23	80.0	± 9.6 %
		Y	3.87	66.88	16.92		80.0	
		Z	3.78	66.32	16.35		80.0	
10512- AAA	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, QPSK, UL Subframe=2,3,4,7,8,9)	x	3.94	70.65	17.76	2.23	80.0	± 9.6 %
	1	Y	3.91	71.11	18.18		80.0	-
10511		Z	3.58	69.38	17.06		80.0	
10513- AAA	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 16-QAM, UL Subframe=2,3,4,7,8,9)	x	3.78	67.26	16.90	2.23	80.0	± 9.6 %
		Y	3.69	67.27	17.07		80.0	
10511		Z	3.58	66.56	16.41		80.0	
10514- AAA	LTE-TDD (SC-FDMA, 100% RB, 20 MHz, 64-QAM, UL Subframe=2,3,4,7,8,9)	X	3.83	66.92	16.80	2.23	80.0	± 9.6 %
		Y	3.73	66.91	16.95		80.0	
		Z	3.63	66.29	16.35		80.0	
10515- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 99pc duty cycle)	x	1.03	64.35	15.90	0.00	150.0	± 9.6 %
		Y	1.06	65.48	16.80		150.0	
10510		Z	0.99	63.83	15.29		150.0	
10516- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 99pc duty cycle)	×	1.10	82.16	23.46	0.00	150.0	± 9.6 %
		Y	3.26	105.03	31.90		150.0	
10517-	IEEE 202 11h WIEL2 4 OH- (DOOD 14	Z	0.76	75.15	20.01	0.00	150.0	
AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 99pc duty cycle)	×	0.92	67.48	17.26	0.00	150.0	± 9.6 %
_		Y Z	1.00	69.97	18.89 16.25		150.0	
10518- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 99pc duty cycle)	X	4.55	66.25 67.00	16.54	0.00	150.0 150.0	± 9.6 %
		Y	4.55	67.26	16.73		150.0	
		Z	4.47	66.94	16.35		150.0	
10519- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 99pc duty cycle)	×	4.73	67.21	16.64	0.00	150.0	± 9.6 %
		Y	4.71	67.45	16.82		150.0	
		Z	4.63	67.13	16.45		150.0	
10520- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 99pc duty cycle)	×	4.58	67.18	16.57	0.00	150.0	± 9.6 %
		Y	4.57	67.43	16.76		150.0	
10501		Z	4.48	67.08	16.38		150.0	
10521- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 99pc duty cycle)	X	4.52	67.17	16.56	0.00	150.0	± 9.6 %
		Y	4.51	67.42	16.75		150.0	
10522	IFEE 000 11ab WIFE OUT (OF DUI AA	Z	4.42	67.07	16.36	0.00	150.0	
10522- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 99pc duty cycle)	X	4.58	67.30	16.66	0.00	150.0	± 9.6 %
		Y	4.57	67.55	16.86		150.0	
		Z	4.48	67.20	16.47		150.0	

#### September 29, 2016

10523- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 99pc duty cycle)	X	4.47	67.19	16.53	0.00	150.0	±9.6 %
		Y	4.47	67.48	16.74		150.0	
		Z	4.38	67.13	16.35		150.0	
10524- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 99pc duty cycle)	x	4.52	67.21	16.63	0.00	150.0	±9.6 %
		Y	4.51	67.48	16.83		150.0	
		Z	4.42	67.13	16.44		150.0	
10525- AAA	IEEE 802.11ac WiFi (20MHz, MCS0, 99pc duty cycle)	X	4.52	66.28	16.22	0.00	150.0	±9.6 %
		Y	4.53	66.55	16.43		150.0	
		Z	4.44	66.21	16.05		150.0	
10526- AAA	IEEE 802.11ac WiFi (20MHz, MCS1, 99pc duty cycle)	X	4.68	66.62	16.36	0.00	150.0	±9.6 %
0,000		Y	4.68	66.89	16.56		150.0	
		Z	4.58	66.53	16.17		150.0	
10527- AAA	IEEE 802.11ac WiFi (20MHz, MCS2, 99pc duty cycle)	×	4.61	66.59	16.31	0.00	150.0	±9.6 %
		Y	4.61	66.87	16.52		150.0	
		Z	4.51	66.50	16.12		150.0	
10528- AAA	IEEE 802.11ac WiFi (20MHz, MCS3, 99pc duty cycle)	×	4.62	66.61	16.34	0.00	150.0	±9.6 %
		Y	4.62	66.88	16.54		150.0	
		Z	4.52	66.51	16.15		150.0	
10529- AAA	IEEE 802.11ac WiFi (20MHz, MCS4, 99pc duty cycle)	X	4.62	66.61	16.34	0.00	150.0	±9.6 %
		Y	4.62	66.88	16.54		150.0	
		Z	4.52	66.51	16.15		150.0	
10531- AAA	IEEE 802.11ac WiFi (20MHz, MCS6, 99pc duty cycle)	X	4.60	66.70	16.35	0.00	150.0	± 9.6 %
		Y	4.60	66.97	16.55		150.0	
		Z	4.50	66.58	16.15		150.0	
10532- AAA	IEEE 802.11ac WiFi (20MHz, MCS7, 99pc duty cycle)	X	4.47	66.56	16.29	0.00	150.0	±9.6 %
		Y	4.47	66.84	16.50		150.0	
		Z	4.37	66.44	16.08		150.0	
10533- AAA	IEEE 802.11ac WiFi (20MHz, MCS8, 99pc duty cycle)	X	4.63	66.67	16.34	0.00	150.0	± 9.6 %
1.4.1.0.0		Y	4.63	66.96	16.55		150.0	
		Z	4.53	66.59	16.15		150.0	
10534- AAA	IEEE 802.11ac WiFi (40MHz, MCS0, 99pc duty cycle)	X	5.16	66.62	16.36	0.00	150.0	± 9.6 %
		Y	5,16	66.82	16.53		150.0	
		Z	5.07	66.53	16.19		150.0	
10535- AAA	IEEE 802.11ac WiFi (40MHz, MCS1, 99pc duty cycle)	×	5.23	66.83	_ 16.45	0.00	150.0	±9.6 %
		Y	5.23	67.02	16.62		150.0	
		Z	5.12	66.69	16.27		150.0	
10536- AAA	IEEE 802.11ac WiFi (40MHz, MCS2, 99pc duty cycle)	×	5.10	66.78	16.41	0.00	150.0	± 9.6 %
81925-08		Y	5.11	67.00	16.59		150.0	
		Z	5.01	66.68	16.24		150.0	
10537- AAA	IEEE 802.11ac WiFi (40MHz, MCS3, 99pc duty cycle)	X	5.16	66.73	16.39	0.00	150.0	± 9.6 %
		Y	5.16	66.94	16.56		150.0	
		Z	5.06	66.63	16.22		150.0	
10538- AAA	IEEE 802.11ac WiFi (40MHz, MCS4, 99pc duty cycle)	x	5.24	66.73	16.42	0.00	150.0	± 9.6 %
		Y	5.23	66.92	16.59		150.0	
		Z	5.14	66.62	16.25		150.0	·
10540- AAA	IEEE 802.11ac WiFi (40MHz, MCS6, 99pc duty cycle)	X	5.17	66.74	16.45	0.00	150.0	± 9.6 %
		Y	5.16	66.92	16.61		150.0	

September 29, 2016

10541- AAA	IEEE 802.11ac WiFi (40MHz, MCS7, 99pc duty cycle)	X	5.15	66.61	16.37	0.00	150.0	± 9.6 %
		Y	5.14	66.80	16.53		150.0	
		Z	5.05	66.50	16.19		150.0	
10542- AAA	IEEE 802.11ac WiFi (40MHz, MCS8, 99pc duty cycle)	X	5.30	66.68	16.42	0.00	150.0	±9.6 %
		Y	5.29	66.87	16.58		150.0	
		Z	5.20	66.58	16.25		150.0	
10543- AAA	IEEE 802.11ac WiFi (40MHz, MCS9, 99pc duty cycle)	X	5.37	66.69	16.45	0.00	150.0	± 9.6 %
		Y	5.35	66.87	16.61		150.0	
		Z	5.26	66.59	16.28		150.0	
10544- AAA	IEEE 802.11ac WiFi (80MHz, MCS0, 99pc duty cycle)	X	5.48	66.70	16.33	0.00	150.0	± 9.6 %
		Y	5.48	66.86	16.48		150.0	
		Z	5.40	66.61	16.17	100	150.0	
10545- AAA	IEEE 802.11ac WiFi (80MHz, MCS1, 99pc duty cycle)	X	5.69	67.17	16.52	0.00	150.0	± 9.6 %
		Y	5.69	67.35	16.68		150.0	
		Z	5.58	67.04	16.34		150.0	
10546- AAA	IEEE 802.11ac WiFi (80MHz, MCS2, 99pc duty cycle)	X	5.54	66.88	16.39	0.00	150.0	± 9.6 %
		Y	5.53	67.03	16.53		150.0	
		Z	5.44	66.76	16.21		150.0	
10547-	IEEE 802.11ac WiFi (80MHz, MCS3,	X	5.61	66.94	16.41	0.00	150.0	±9.6 %
AAA	99pc duty cycle)	Ŷ	5.61			0.00	- 2222	1 9.0 %
		Z		67.11	16.56		150.0	
10548- AAA	IEEE 802.11ac WiFi (80MHz, MCS4, 99pc duty cycle)	X	5.51 5.87	66.83 67.91	16.24 16.87	0.00	150.0 150.0	± 9.6 %
	sopo dally official	Y	5.86	68.05	17.00		150.0	
		Z	5.71	67.60	16.60		150.0	
10550- AAA	IEEE 802.11ac WiFi (80MHz, MCS6, 99pc duty cycle)	X	5.58	66.98	16.45	0.00	150.0	±9.6 %
		Y	5.59	67.17	16.61		150.0	
		Z	5.49	66.87	16.28		150.0	
10551- AAA	IEEE 802.11ac WiFi (80MHz, MCS7, 99pc duty cycle)	X	5.57	66.95	16.40	0.00	150.0	± 9.6 %
		Y	5.56	67.09	16.54		150.0	
		Z	5.46	66.79	16.20		150.0	
10552- AAA	IEEE 802.11ac WiFi (80MHz, MCS8, 99pc duty cycle)	x	5.49	66.77	16.31	0.00	150.0	± 9.6 %
		Y	5.49	66.95	16.47		150.0	
		Z	5.41	66.70	16.16		150.0	
10553- AAA	IEEE 802.11ac WiFi (80MHz, MCS9, 99pc duty cycle)	×	5.56	66.77	16.34	0.00	150.0	± 9.6 %
	1 10 55 10 15	Y	5.56	66.93	16.49		150.0	
		Z	5.47	66.68	16.18	_	150.0	
10554- AAA	IEEE 1602.11ac WiFi (160MHz, MCS0, 99pc duty cycle)	×	5.90	67.05	16.41	0.00	150.0	± 9.6 %
	No. Martine Providence	Y	5.90	67.19	16.54		150.0	
		Z	5.81	66.94	16.24		150.0	
10555- AAA	IEEE 1602.11ac WiFi (160MHz, MCS1, 99pc duty cycle)	×	6.03	67.36	16.54	0.00	150.0	± 9.6 %
		Y	6.03	67.50	16.67		150.0	
22222		Z	5.92	67.21	16.36		150.0	
10556- AAA	IEEE 1602.11ac WiFi (160MHz, MCS2, 99pc duty cycle)	X	6.05	67.41	16.56	0.00	150.0	±9.6 %
		Y	6.06	67.57	16.70		150.0	
		Z	5.95	67.29	16.39		150.0	
	IFFF 4000 44 WIFI (400101- 14000	X	6.01	67.28	16.52	0.00	150.0	± 9.6 %
10557- AAA	IEEE 1602.11ac WiFi (160MHz, MCS3, 99pc duty cycle)	<b>^</b>	0.01		100000		1000000	2010/02/02
10557- AAA	99pc duty cycle)	Y	6.00	67.42	16.65		150.0	

#### September 29, 2016

10558- AAA	IEEE 1602.11ac WiFi (160MHz, MCS4, 99pc duty cycle)	×	6.05	67.45	16.61	0.00	150.0	±9.6 %
		Y	6.05	67.58	16.74		150.0	
		Z	5.95	67.30	16.43		150.0	
10560- AAA	IEEE 1602.11ac WiFi (160MHz, MCS6, 99pc duty cycle)	X	6.04	67.28	16.57	0.00	150.0	±9.6 %
		Y	6.04	67.41	16.70		150.0	-
		z	5.94	67.16	16.40		150.0	
10561- AAA	IEEE 1602.11ac WiFi (160MHz, MCS7, 99pc duty cycle)	X	5.98	67.28	16.61	0.00	150.0	±9.6 %
		Y	5.98	67.42	16.74		150.0	
		Z	5.88	67.15	16.43		150.0	
10562- AAA	IEEE 1602.11ac WiFi (160MHz, MCS8, 99pc duty cycle)	×	6.08	67.60	16.76	0.00	150.0	±9.6 %
		Y	6.06	67.70	16.87		150.0	S
		Z	5.96	67.40	16.55		150.0	
10563- AAA	IEEE 1602.11ac WiFi (160MHz, MCS9, 99pc duty cycle)	X	6.19	67.56	16.71	0.00	150.0	±9.6 %
		Y	6.15	67.61	16.79	·	150.0	
		Z	6.04	67.29	16.46		150.0	
10564- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 9 Mbps, 99pc duty cycle)	X	4.86	66.99	16.63	0.46	150.0	± 9.6 %
		Y	4.85	67.18	16.78	_	150.0	
		Z	4.77	66.93	16.45		150.0	
10565- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 12 Mbps, 99pc duty cycle)	X	5.08	67.43	16.95	0.46	150.0	±9.6 %
		Y	5.07	67.64	17.11		150.0	
		Z	4.98	67.36	16.77		150.0	
10566- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 18 Mbps, 99pc duty cycle)	X	4.92	67.27	16.77	0.46	150.0	± 9.6 %
and the second		Y	4.90	67.46	16.92		150.0	
		Z	4.82	67.18	16.58		150.0	
10567- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 24 Mbps, 99pc duty cycle)	X	4.95	67.67	17.13	0.46	150.0	± 9.6 %
		Y	4.94	67.93	17.33		150.0	
		Z	4.85	67.59	16.95		150.0	
10568- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 36 Mbps, 99pc duty cycle)	x	4.82	67.03	16.52	0.46	150.0	± 9.6 %
C 11 19 1 19		Y	4.80	67.19	16.65		150.0	
		Z	4.72	66.92	16.32		150.0	
10569- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 48 Mbps, 99pc duty cycle)	X	4.91	67.80	17.21	0.46	150.0	± 9.6 %
		Y	4.92	68.10	17.44		150.0	
		Z	4.82	67.75	17.05		150.0	
10570- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 54 Mbps, 99pc duty cycle)	×	4.94	67.65	17.14	0.46	150.0	± 9.6 %
		Y	4.93	67.91	17.34		150.0	
		Z	4.84	67.58	16.97		150.0	
10571- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 1 Mbps, 90pc duty cycle)	X	1.15	64.53	16.00	0.46	130.0	± 9.6 %
		Y	1.14	65.11	16.62		130.0	
		Z	1.10	63.89	15.35		130.0	1
10572- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 2 Mbps, 90pc duty cycle)	X	1.16	65.12	16.38	0.46	130.0	± 9.6 %
		Y	1.16	65.82	17.07		130.0	
		Z	1.11	64.41	15.69		130.0	
10573- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 5.5 Mbps, 90pc duty cycle)	X	2.66	93.17	26.84	0.46	130.0	± 9.6 %
		Y	7.16	114.96	34.30		130.0	
		Z	1.37	81.26	22.19		130.0	
10574- AAA	IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps, 90pc duty cycle)	X	1.29	71.56	19.79	0.46	130.0	± 9.6 %
		Y	1.38	74.16	21.48		130.0	

September 29, 2016

10575- AAA	IEEE 802.11g WiFI 2.4 GHz (DSSS- OFDM, 6 Mbps, 90pc duty cycle)	X	4.62	66.66	16.58	0.46	130.0	± 9.6 %
		Y	4.60	66.85	16.73		130.0	
		Z	4.53	66.59	16.38		130.0	
10576- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 9 Mbps, 90pc duty cycle)	×	4.64	66.85	16.65	0.46	130.0	± 9.6 %
		Y	4.63	67.06	16.82	-	130.0	
		Z	4.55	66.78	16.46		130.0	
10577- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 12 Mbps, 90pc duty cycle)	X	4.83	67.12	16.81	0.46	130.0	± 9.6 %
		Y	4.82	67.32	16.98		130.0	
		Z	4.73	67.03	16.62		130.0	
10578- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 18 Mbps, 90pc duty cycle)	X	4.74	67.28	16.93	0.46	130.0	± 9.6 %
_		Y	4.72	67.52	17.12		130.0	
10-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-1-		Z	4.64	67.19	16.73		130.0	
10579- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 24 Mbps, 90pc duty cycle)	X	4.49	66.52	16.20	0.46	130.0	± 9.6 %
		Y	4.46	66.65	16.32		130.0	
		Z	4.39	66.39	15.98		130.0	
10580- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 36 Mbps, 90pc duty cycle)	X	4.54	66.58	16.23	0.46	130.0	± 9.6 %
		Y	4.51	66.71	16.35		130.0	
		Z	4.43	66.45	16.01		130.0	
10581- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 48 Mbps, 90pc duty cycle)	X	4.63	67.32	16.87	0.46	130.0	± 9.6 %
		Y	4.62	67.58	17.07		130.0	
		Z	4.53	67.23	16.68		130.0	
10582- AAA	IEEE 802.11g WiFi 2.4 GHz (DSSS- OFDM, 54 Mbps, 90pc duty cycle)	X	4.43	66.29	15.99	0.46	130.0	±9.6 %
1		Y	4.40	66.39	16.09		130.0	
		Z	4.32	66.15	15.76		130.0	
10583- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 6 Mbps, 90pc duty cycle)	X	4.62	66.66	16.58	0.46	130.0	± 9.6 %
	and the second	Y	4.60	66.85	16.73		130.0	
		Z	4.53	66.59	16.38		130.0	
10584- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 9 Mbps, 90pc duty cycle)	×	4.64	66.85	16.65	0.46	130.0	± 9.6 %
	and the second	Y	4.63	67.06	16.82		130.0	
		Z	4.55	66.78	16.46		130.0	
10585- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 12 Mbps, 90pc duty cycle)	×	4.83	67.12	16.81	0.46	130.0	±9.6 %
		Y	4.82	67.32	16.98		130.0	
		Z	4.73	67.03	16.62		130.0	
10586- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 18 Mbps, 90pc duty cycle)	X	4.74	67.28	16.93	0.46	130.0	±9.6 %
		Y	4.72	67.52	17.12		130.0	
		Z	4.64	67.19	16.73		130.0	
10587- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 24 Mbps, 90pc duty cycle)	X	4.49	66.52	16.20	0.46	130.0	± 9.6 %
		Y	4.46	66.65	16.32		130.0	
		Z	4.39	66.39	15.98		130.0	
10588- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 36 Mbps, 90pc duty cycle)	×	4.54	66.58	16.23	0.46	130.0	±9.6 %
		Y	4.51	66.71	16.35		130.0	
2.000 C		Z	4.43	66.45	16.01		130.0	
10589-	IEEE 802.11a/h WiFi 5 GHz (OFDM, 48 Mbps, 90pc duty cycle)	X	4.63	67.32	16.87	0.46	130.0	±9.6 %
AAA		Y	4.62	67.58	17.07		130.0	
AAA						-		
AAA		Z	4.53	67.23	16.68		130.0	
AAA 10590- AAA	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 90pc duty cycle)		4.53 4.43	67.23 66.29	16.68 15.99	0.46	130.0	±9.6 %
10590-	IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps, 90pc duty cycle)	Z				0.46		± 9.6 %

#### September 29, 2016

10591- AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS0, 90pc duty cycle)	X	4.77	66.73	16.68	0.46	130.0	±9.6 %
		Y	4.75	66.92	16.84		130.0	
		Z	4.68	66.67	16.50		130.0	
10592- AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS1, 90pc duty cycle)	X	4.91	67.06	16.81	0.46	130.0	±9.6 %
	incertector and even	Y	4.89	67.25	16.97		130.0	
		Z	4.81	66.98	16.63		130.0	
10593- AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS2, 90pc duty cycle)	×	4.83	66.95	16.68	0.46	130.0	±9.6 %
	moor, sope day syster	Y	4.81	67.13	16.83		130.0	
		Z	4.73	66.86	16.49		130.0	
10594- AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS3, 90pc duty cycle)	X	4.89	67.12	16.84	0.46	130.0	± 9.6 %
	meest cope any symmetry	Y	4.87	67.32	17.00		130.0	
		Z	4.79	67.04	16.65	17-1-1	130.0	1
10595- AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS4, 90pc duty cycle)	X	4.85	67.08	16.74	0.46	130.0	± 9.6 %
	inde if eepe daily eyerey	Y	4.83	67.28	16.90		130.0	
		Z	4.75	67.00	16.55		130.0	
10596-	IEEE 802.11n (HT Mixed, 20MHz,	X	4.79	67.07	16.74	0.46	130.0	±9.6 %
AAA	MCS5, 90pc duty cycle)	Y	4.77	67.26	16.90		130.0	
		Z	4.68	66.97	16.55		130.0	
10597- AAA	IEEE 802.11n (HT Mixed, 20MHz, MCS6, 90pc duty cycle)	X	4.74	66.96	16.61	0.46	130.0	± 9.6 %
MMM	MCS0, Sope duty cycle)	Y	4.72	67.14	16.76		130.0	
		Z	4.63	66.85	16.41		130.0	
10598-	IEEE 802.11n (HT Mixed, 20MHz,	X	4.03	67.19	16.88	0.46	130.0	± 9.6 %
AAA	MCS7, 90pc duty cycle)				17.06	0.40	130.0	1 0.0 %
_		Y	4.71	67.41	and the second se			
10500		Z	4.62	67.09	16.68	0.40	130.0	+000
10599- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS0, 90pc duty cycle)	X	5.45	67.25	16.90	0.46	130.0	± 9.6 %
		Y	5.44	67.41	17.04		130.0	-
		Z	5.35	67.14	16.71	0.40	130.0	1000
10600- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS1, 90pc duty cycle)	×	5.60	67.73	17.11	0.46	130.0	± 9.6 %
		Y	5,59	67.89	17.25		130.0	
		Z	5.47	67.54	16.89		130.0	
10601- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS2, 90pc duty cycle)	×	5.48	67.44	16.98	0.46	130.0	± 9.6 %
		Y	5.46	67.59	17.12		130.0	
		Z	5.37	67.30	16.79		130.0	
10602- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS3, 90pc duty cycle)	×	5.60	67.56	, 16.96	0.46	130.0	± 9.6 %
		Y	5.59	67.73	17.10		130.0	
		Z	5.50	67.48	16.79		130.0	
10603- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS4, 90pc duty cycle)	×	5.66	67.82	17.23	0.46	130.0	± 9.6 %
		Y	5.66	68.02	17.39		130.0	
		Z	5.57	67.76	17.07		130.0	
10604- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS5, 90pc duty cycle)	×	5.51	67.41	17.01	0.46	130.0	± 9.6 %
		Y	5.53	67.68	17.20	1	130.0	
		Z	5.45	67.41	16.88		130.0	
10605- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS6, 90pc duty cycle)	×	5.59	67.64	17.11	0.46	130.0	± 9.6 %
		Y	5.58	67.78	17.24		130.0	
		Z	5.47	67.46	16.90		130.0	
10606- AAA	IEEE 802.11n (HT Mixed, 40MHz, MCS7, 90pc duty cycle)	X	5.30	66.84	16.57	0.46	130.0	±9.6 %
		Y	5.29	66.99	16.70		130.0	
		Z	5.22	66.77		-	130.0	

September 29, 2016

10607- AAA	IEEE 802.11ac WiFi (20MHz, MCS0, 90pc duty cycle)	×	4.62	66.08	16.32	0.46	130.0	± 9.6 %
		Y	4.61	66.32	16.51		130.0	
		Z	4.53	66.01	16.14		130.0	
10608- AAA	IEEE 802.11ac WiFi (20MHz, MCS1, 90pc duty cycle)	x	4.79	66.47	16.49	0.46	130.0	± 9.6 %
		Y	4.78	66.70	16.67		130.0	
		Z	4.69	66.38	16.30		130.0	
10609- AAA	IEEE 802.11ac WiFi (20MHz, MCS2, 90pc duty cycle)	x	4.68	66.31	16.32	0.46	130.0	± 9.6 %
		Y	4.67	66.53	16.49		130.0	
		Z	4.58	66.21	16.12		130.0	
10610- AAA	IEEE 802.11ac WiFi (20MHz, MCS3, 90pc duty cycle)	x	4.73	66.47	16.48	0.46	130.0	± 9.6 %
		Y	4.72	66.71	16.67		130.0	
		Z	4.63	66.37	16.29		130.0	
10611- AAA	IEEE 802.11ac WiFi (20MHz, MCS4, 90pc duty cycle)	x	4.64	66.27	16.33	0.46	130.0	±9.6 %
_		Y	4.63	66.50	16.50		130.0	
100/-		Z	4.54	66.17	16.13		130.0	
10612- AAA	IEEE 802.11ac WiFi (20MHz, MCS5, 90pc duty cycle)	×	4.65	66.42	16.37	0.46	130.0	±9.6 %
		Y	4.63	66.64	16.54		130.0	
10010		Z	4.54	66.31	16.16		130.0	
10613- AAA	IEEE 802.11ac WiFi (20MHz, MCS6, 90pc duty cycle)	×	4.65	66.28	16.24	0.46	130.0	±9.6 %
		Y	4.63	66.48	16.40		130.0	
10011		Z	4.54	66.15	16.03		130.0	
10614- AAA	IEEE 802.11ac WiFi (20MHz, MCS7, 90pc duty cycle)	×	4.60	66.49	16.49	0.46	130.0	±9.6 %
		Y	4.60	66.74	16.69		130.0	
10010		Z	4.50	66.37	16.28		130.0	
10615- AAA	IEEE 802.11ac WiFi (20MHz, MCS8, 90pc duty cycle)	×	4.64	66.10	16.10	0.46	130.0	±9.6 %
		Y	4.62	66.29	16.25		130.0	
10010		Z	4.54	65.99	15.89		130.0	
10616- AAA	IEEE 802.11ac WiFi (40MHz, MCS0, 90pc duty cycle)	×	5.27	66.50	16.51	0.46	130.0	±9.6 %
_		Y	5.26	66.67	16.66		130.0	
		Z	5.17	66.39	16.33		130.0	
10617- AAA	IEEE 802.11ac WiFi (40MHz, MCS1, 90pc duty cycle)	×	5.35	66.73	16.60	0.46	130.0	±9.6 %
		Y	5.34	66.90	16.75		130.0	
10618- AAA	IEEE 802.11ac WiFi (40MHz, MCS2,	Z X	5.24 5.24	66.59 66.73	16.40 16.61	0.46	130.0 130.0	±9.6 %
	90pc duty cycle)	Y	5.23	66.04	16.70		120.0	
		Z	5.23	66.94	16.78 16.43		130.0	
10619- AAA	IEEE 802.11ac WiFi (40MHz, MCS3, 90pc duty cycle)	X	5.14	66.62 66.51	16.43	0.46	130.0 130.0	±9.6 %
	seebs and along	Y	5.23	66.68	16.58		130.0	
		Z	5.14	66.39	16.25		130.0	
10620- AAA	IEEE 802.11ac WiFi (40MHz, MCS4, 90pc duty cycle)	X	5.33	66.54	16.50	0.46	130.0	±9.6 %
		Y	5.32	66.70	16.64		130.0	
		Z	5.22	66.41	16.31		130.0	
10621- AAA	IEEE 802.11ac WiFi (40MHz, MCS5, 90pc duty cycle)	X	5.34	66.68	16.69	0.46	130.0	± 9.6 %
		Y	5.33	66.87	16.86		130.0	
		Z	5.24	66.56	16.51		130.0	
10622- AAA	IEEE 802.11ac WiFi (40MHz, MCS6, 90pc duty cycle)	X	5.35	66.85	16.77	0.46	130.0	±9.6 %
		Y	5.35	67.07	16.95		130.0	
		Z	5.23	66.67	16.56		130.0	

#### September 29, 2016

10623- AAA	IEEE 802.11ac WiFi (40MHz, MCS7, 90pc duty cycle)	X	5.22	66.35	16.39	0.46	130.0	±9.6 %
		Y	5.20	66.48	16.51		130.0	
		Z	5.11	66.21	16.18		130.0	
10624- AAA	IEEE 802.11ac WiFi (40MHz, MCS8, 90pc duty cycle)	X	5.41	66.56	16.55	0.46	130.0	±9.6 %
0.01	sope day of elof	Y	5.40	66.72	16.70		130.0	
		Z	5.31	66.44	16.37		130.0	
10625-	IEEE 802.11ac WiFi (40MHz, MCS9,	X	5.72	67.36	17.01	0.46	130.0	± 9.6 %
AAA	90pc duty cycle)	Y	5.65	67.35	17.07		130.0	
		Z	5.51	66.93	16.67		130.0	
10626-	IEEE 802.11ac WiFi (80MHz, MCS0, 90pc duty cycle)	X	5.58	66.55	16.46	0.46	130.0	±9.6 %
AAA	sope duty cycle)	Y	5.58	66.69	16.59		130.0	
		Z	5.49	66.45	16.29		130.0	
10007	1555 000 44 WEEL (00MU to MCC4		5.84	67.20	16.75	0.46	130.0	±9.6 %
10627- AAA	IEEE 802.11ac WiFi (80MHz, MCS1, 90pc duty cycle)	×		0.000		0.40	4.010404	1 9.0 %
		Y	5.84	67.37	16.90		130.0	
		Z	5.73	67.04	16.55	0.40	130.0	10.00
10628- AAA	IEEE 802.11ac WiFi (80MHz, MCS2, 90pc duty cycle)	x	5.60	66.61	16.38	0.46	130.0	± 9.6 %
		Y	5.59	66.71	16.50	-	130.0	
		Z	5.50	66.46	16.19		130.0	
10629- AAA	IEEE 802.11ac WiFi (80MHz, MCS3, 90pc duty cycle)	×	5.69	66.69	16.42	0.46	130.0	± 9.6 %
		Y	5.68	66.83	16.55		130.0	
		Z	5.58	66.57	16.24		130.0	
10630- AAA	IEEE 802.11ac WiFi (80MHz, MCS4, 90pc duty cycle)	X	6.13	68.22	17.18	0.46	130.0	±9.6 %
10.03	sopo daty cycley	Y	6.09	68.28	17.27		130.0	
		Z	5.91	67.74	16.83		130.0	
10631- AAA	IEEE 802.11ac WiFi (80MHz, MCS5, 90pc duty cycle)	X	6.00	67.93	17.23	0.46	130.0	±9.6 %
~~~	sope daty cycle)	Y	5.99	68.09	17.38		130.0	-
_		z	5.85	67.68	17.00		130.0	
10632- AAA	IEEE 802.11ac WiFi (80MHz, MCS6, 90pc duty cycle)	X	5.81	67.26	16.92	0.46	130.0	± 9.6 %
~~~	sope daily cycle)	Y	5.82	67.49	17.10		130.0	
		z	5.71	67.16	16.75		130.0	
10633- AAA	IEEE 802.11ac WiFi (80MHz, MCS7, 90pc duty cycle)	X	5.66	66.78	16.50	0.46	130.0	±9.6 %
AAA	sope duty cycle)	Y	5.66	66.93	16.64		130.0	
		Z	5.57	66.67	16.33	-	130.0	
10634- AAA	IEEE 802.11ac WiFi (80MHz, MCS8, 90pc duty cycle)	X	5.65	66.80	16.57	0.46	130.0	± 9.6 %
1001	sops day start	Y	5.64	66.96	16.72		130.0	
		Z	5.55	66.70	16.40		130.0	
10635- AAA	IEEE 802.11ac WiFi (80MHz, MCS9, 90pc duty cycle)	X	5.52	66.10	15.95	0.46	130.0	± 9.6 %
	sops duty cycle)	Y	5.49	66.16	16.03		130.0	
		Z	5.42	65.97	15.76		130.0	-
			0.46		16.54	0.46	130.0	± 9.6 %
10636-	IEEE 1602.11ac WiFi (160MHz, MCS0, 90nc duty cycle)	X	6.01	66.91	10.54		1.00000	1 Sholash 907
10636- AAA	IEEE 1602.11ac WiFi (160MHz, MCS0, 90pc duty cycle)		6.01 6.01	66.91	16.67		130.0	
		×	2 - 544 <b>9</b> 80 - 5-	67.05		0.10		
AAA 10637-	90pc duty cycle) IEEE 1602.11ac WiFi (160MHz, MCS1,	X Y	6.01	10.95500.0	16.67	0.46	130.0	± 9.6 %
AAA	90pc duty cycle)	X Y Z X	6.01 5.92 6.17	67.05 66.81 67.33	16.67 16.37 16.74		130.0 130.0 130.0	± 9.6 %
AAA 10637-	90pc duty cycle) IEEE 1602.11ac WiFi (160MHz, MCS1,	X Y Z X Y	6.01 5.92 6.17 6.17	67.05 66.81 67.33 67.46	16.67 16.37 16.74 16.86		130.0 130.0 130.0 130.0	± 9.6 %
AAA 10637- AAA 10638-	90pc duty cycle) IEEE 1602.11ac WiFi (160MHz, MCS1, 90pc duty cycle) IEEE 1602.11ac WiFi (160MHz, MCS2,	X Y Z X	6.01 5.92 6.17	67.05 66.81 67.33	16.67 16.37 16.74		130.0 130.0 130.0	± 9.6 %
AAA 10637- AAA	90pc duty cycle) IEEE 1602.11ac WiFi (160MHz, MCS1, 90pc duty cycle)	X Y Z X Y Z	6.01 5.92 6.17 6.17 6.06	67.05 66.81 67.33 67.46 67.17	16.67 16.37 16.74 16.86 16.54	0.46	130.0 130.0 130.0 130.0 130.0 130.0	

#### September 29, 2016

	A set ad an a set of a							
10639- AAA	IEEE 1602.11ac WiFi (160MHz, MCS3, 90pc duty cycle)	×	6.13	67.20	16.69	0.46	130.0	±9.6 %
		Y	6.13	67.33	16.81		130.0	
		Z	6.03	67.07	16.51	-	130.0	
10640- AAA	IEEE 1602.11ac WiFi (160MHz, MCS4, 90pc duty cycle)	X	6.13	67.21	16.64	0.46	130.0	± 9.6 %
		Y	6.12	67.31	16.74	-	130.0	
		Z	6.03	67.06	16.44		130.0	
	IEEE 1602.11ac WiFi (160MHz, MCS5, 90pc duty cycle)	X	6.20	67.18	16.64	0.46	130.0	± 9.6 %
		Y	6.20	67.31	16.76		130.0	
		Z	6.10	67.04	16.46		130.0	
10642- AAA	IEEE 1602.11ac WiFi (160MHz, MCS6, 90pc duty cycle)	X	6.22	67.38	16.91	0.46	130.0	± 9.6 %
		Y	6.22	67.52	17.04		130.0	
		Z	6.12	67.26	16.74		130.0	
10643- AAA	IEEE 1602.11ac WiFi (160MHz, MCS7, 90pc duty cycle)	X	6.07	67.09	16.66	0.46	130.0	± 9.6 %
		Y	6.07	67.21	16.78		130.0	
		Z	5.97	66.96	16.48		130.0	
10644- AAA	IEEE 1602.11ac WiFi (160MHz, MCS8, 90pc duty cycle)	X	6.19	67.47	16.87	0.46	130.0	±9.6 %
		Y	6.17	67.53	16.96		130.0	
		Z	6.06	67.25	16.64		130.0	
10645- AAA	IEEE 1602.11ac WiFi (160MHz, MCS9, 90pc duty cycle)	x	6.36	67.63	16.91	0.46	130.0	± 9.6 %
-3111		Y	6.32	67.64	16.97		130.0	
		Z	6.19	67.29	16.63		130.0	
10646- AAB	LTE-TDD (SC-FDMA, 1 RB, 5 MHz, QPSK, UL Subframe=2,7)	X	10.25	97.21	32.85	9.30	60.0	± 9.6 %
10.00		Y	7.85	91.41	30.98		60.0	
		Z	8.65	93.98	31.65		60.0	
10647- AAA	LTE-TDD (SC-FDMA, 1 RB, 20 MHz, QPSK, UL Subframe=2,7)	X	8.96	94.81	32.17	9.30	60.0	± 9.6 %
		Y	6.94	89.26	30.34		60.0	
		Z	7.50	91.40	30.88		60.0	
10648- AAA	CDMA2000 (1x Advanced)	×	0.80	65.94	12.17	0.00	150.0	± 9.6 %
		Y	0.91	68.29	13.16		150.0	
		Z	0.66	63.89	10.52		150.0	

<sup>E</sup> Uncertainty is determined using the max, deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

ac-MR/

Schweizerischer Kalibrierdienst Service suisse d'étalonnage

C Servizio svizzero di taratura

S Swiss Calibration Service

S

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Huawei-SZ (Auden) Client

Certificate No: D835V2-4d126\_Jul15

bject	D835V2 - SN: 4d	126	
alibration procedure(s)	QA CAL-05.v9 Calibration proces	dure for dipole validation kits abo	ve 700 MHz
alibration date:	July 23, 2015		
he measurements and the unce	rtainties with confidence protection the closed laborator	onal standards, which realize the physical uni robability are given on the following pages an ry facility: environment temperature $(22 \pm 3)^{\circ}$	d are part of the certificate.
Calibration Equipment used (M&1	E critical for calibration)		
Liman Clandarda	Line	Cal Date (Catilicate No.)	Scheduled Calibration
	ID #	Cal Date (Certificate No.)	Scheduled Calibration Oct-15
ower meter EPM-442A	GB37480704	07-Oct-14 (No. 217-02020)	Scheduled Calibration Oct-15 Oct-15
ower meter EPM-442A ower sensor HP 8481A	GB37480704 US37292783	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020)	Oct-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A	GB37480704 US37292783 MY41092317	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020)	Oct-15 Oct-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator	GB37480704 US37292783 MY41092317 SN: 5058 (20k)	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131)	Oct-15 Oct-15 Oct-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Type-N mismatch combination	GB37480704 US37292783 MY41092317	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020)	Oct-15 Oct-15 Oct-15 Mar-16
Primary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134)	Oct-15 Oct-15 Oct-15 Mar-16 Mar-16
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14)	Oct-15 Oct-15 Oct-15 Mar-16 Mar-16 Dec-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14)	Oct-15 Oct-15 Oct-15 Mar-16 Mar-16 Dec-15 Aug-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601 ID #	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (in house)	Oct-15 Oct-15 Oct-15 Mar-16 Mar-16 Dec-15 Aug-15 Scheduled Check
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards RF generator R&S SMT-06	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601 ID # 100005 US37390585 S4206	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (in house) 04-Aug-99 (in house check Oct-13) 18-Oct-01 (in house check Oct-14)	Oct-15 Oct-15 Oct-15 Mar-16 Dec-15 Aug-15 Scheduled Check In house check: Oct-16 In house check: Oct-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards RF generator R&S SMT-06 Network Analyzer HP 8753E	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601 ID # 100005 US37390585 S4206 Name	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (In house) 04-Aug-99 (in house check Oct-13) 18-Oct-01 (in house check Oct-14) Function	Oct-15 Oct-15 Oct-15 Mar-16 Mar-16 Dec-15 Aug-15 Scheduled Check In house check: Oct-16
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards RF generator R&S SMT-06	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601 ID # 100005 US37390585 S4206	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (in house) 04-Aug-99 (in house check Oct-13) 18-Oct-01 (in house check Oct-14)	Oct-15 Oct-15 Oct-15 Mar-16 Dec-15 Aug-15 Scheduled Check In house check: Oct-16 In house check: Oct-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards RF generator R&S SMT-06 Vetwork Analyzer HP 8753E	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601 ID # 100005 US37390585 S4206 Name	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (In house) 04-Aug-99 (in house check Oct-13) 18-Oct-01 (in house check Oct-14) Function	Oct-15 Oct-15 Oct-15 Mar-16 Dec-15 Aug-15 Scheduled Check In house check: Oct-16 In house check: Oct-15
Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator Fype-N mismatch combination Reference Probe ES3DV3 DAE4 Secondary Standards RF generator R&S SMT-06 Network Analyzer HP 8753E	GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601 ID # 100005 US37390585 S4206 Name Michael Weber	07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02020) 07-Oct-14 (No. 217-02021) 01-Apr-15 (No. 217-02131) 01-Apr-15 (No. 217-02134) 30-Dec-14 (No. ES3-3205_Dec14) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (In house) 04-Aug-99 (in house check Oct-13) 18-Oct-01 (in house check Oct-14) Function Laboratory Technician	Oct-15 Oct-15 Oct-15 Mar-16 Dec-15 Aug-15 Scheduled Check In house check: Oct-16 In house check: Oct-15

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst

- S Service suisse d'étalonnage C
- Servizio svizzero di taratura
- Swiss Calibration Service S

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

#### Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

# Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010
- d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

### Additional Documentation:

e) DASY4/5 System Handbook

# Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed point exactly below the center marking of the flat phantom section, with the arms oriented . parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole positioned under the liquid filled phantom. The impedance stated is transformed from the measurement at the SMA connector to the feed point. The Return Loss ensures low reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. ٠ No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D835V2-4d126\_Jul15

Page 2 of 8

#### Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.8.8
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	15 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	835 MHz ± 1 MHz	

#### Head TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	41.5	0.90 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	42.4 ± 6 %	0.92 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

#### SAR result with Head TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	2.33 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	9.21 W/kg ± 17.0 % (k≈2)
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL SAR measured	condition 250 mW input power	1.51 W/kg
		1.51 W/kg 5.98 W/kg ± 16.5 % (k=2)

#### Body TSL parameters

The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	55.2	0.97 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	54.9 ± 6 %	1.00 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

## SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAR measured	250 mW input power	2.41 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	9.41 W/kg ± 17.0 % (k=2)
	and the second s	g
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL SAR measured		1.57 W/kg

Certificate No: D835V2-4d126\_Jul15

#### Appendix (Additional assessments outside the scope of SCS 0108)

#### Antenna Parameters with Head TSL

Impedance, transformed to feed point	51.5 Ω - 0.7 jΩ	
Return Loss	- 35.8 dB	

#### Antenna Parameters with Body TSL

Impedance, transformed to feed point	48.0 Ω - 2.5 ]Ω	
Return Loss	- 29.6 dB	

#### **General Antenna Parameters and Design**

Electrical	Delay (one direction)	1.396 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

#### Additional EUT Data

Manufactured by	SPEAG	
Manufactured on	June 29, 2010	

#### DASY5 Validation Report for Head TSL

Date: 22.07.2015

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d126


Communication System: UID 0 - CW; Frequency: 835 MHz Medium parameters used: f = 835 MHz;  $\sigma = 0.92$  S/m;  $\varepsilon_r = 42.4$ ;  $\rho = 1000$  kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:

- Probe: ES3DV3 SN3205; ConvF(6.2, 6.2, 6.2); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

## Dipole Calibration for Head Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:

Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 56.43 V/m; Power Drift = -0.01 dB Peak SAR (extrapolated) = 3.48 W/kg SAR(1 g) = 2.33 W/kg; SAR(10 g) = 1.51 W/kg Maximum value of SAR (measured) = 2.73 W/kg




0 dB = 2.73 W/kg = 4.36 dBW/kg

Certificate No: D835V2-4d126\_Jul15

Page 5 of 8

Impedance Measurement Plot for Head TSL



#### **DASY5 Validation Report for Body TSL**

Date: 23.07.2015

Test Laboratory: SPEAG, Zurich, Switzerland

#### DUT: Dipole 835 MHz; Type: D835V2; Serial: D835V2 - SN: 4d126


Communication System: UID 0 - CW; Frequency: 835 MHz Medium parameters used: f = 835 MHz;  $\sigma = 1$  S/m;  $\epsilon_r = 54.9$ ;  $\rho = 1000$  kg/m<sup>3</sup> Phantom section: Flat Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

#### DASY52 Configuration:

- Probe: ES3DV3 SN3205; ConvF(6.17, 6.17, 6.17); Calibrated: 30.12.2014;
- Sensor-Surface: 3mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 18.08.2014
- Phantom: Flat Phantom 4.9L; Type: QD000P49AA; Serial: 1001
- DASY52 52.8.8(1222); SEMCAD X 14.6.10(7331)

## Dipole Calibration for Body Tissue/Pin=250 mW, d=15mm/Zoom Scan (7x7x7)/Cube 0:


Measurement grid: dx=5mm, dy=5mm, dz=5mm Reference Value = 55.03 V/m; Power Drift = -0.02 dB Peak SAR (extrapolated) = 3.56 W/kg SAR(1 g) = 2.41 W/kg; SAR(10 g) = 1.57 W/kg Maximum value of SAR (measured) = 2.82 W/kg



0 dB = 2.82 W/kg = 4.50 dBW/kg


Page 7 of 8

## Impedance Measurement Plot for Body TSL



## Justification of the extended calibration of Dipole D835V2 SN: 4d126

Per KDB 865664, we have Measured the Impedance and Return Loss as below, and the return loss is <-20dB, with 20% of prior calibration; the real or imaginary parts of the impedance is with 5 ohm of prior calibration. Therefore the verification result should support extended calibration.



## Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





Schweizerischer Kalibrierdienst Service suisse d'étalonnage

C Servizio svizzero di taratura S

Swiss Calibration Service

Accreditation No.: SCS 108

S

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client Huawei-SZ (Auden)

Certificate No: D1900V2-5d143\_Sep14

# CALIBRATION CERTIFICATE

			the second se
Object	D1900V2 - SN:	5d143	
Calibration procedure(s)	QA CAL-05.v9 Calibration proc	cedure for dipole validation kits at	pove 700 MHz
Calibration date:	September 23, ;	2014	
	ertainties with confidence	ational standards, which realize the physical u probability are given on the following pages a	and are part of the certificate.
Calibration Equipment used (M8		ory facility: environment temperature (22 $\pm$ 3)	°C and humidity < 70%,
Calibration Equipment used (M8 Primary Standards	TE critical for calibration)		
Calibration Equipment used (M8	TE critical for calibration)	Cal Date (Certificate No.) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01828) 03-Apr-14 (No. 217-01928) 03-Apr-14 (No. 217-01921) 30-Dec-13 (No. ES3-3205_Dec13) 18-Aug-14 (No. DAE4-601_Aug14)	*C and humidity < 70%. Scheduled Calibration Oct-14 Oct-14 Oct-14 Oct-14 Apr-15 Dec-14 Aug-15
Calibration Equipment used (M8 Primary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator ype-N mismatch combination leference Probe ES3DV3 PAE4 econdary Standards	TE critical for calibration) ID # GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205	Cal Date (Certificate No.) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01828) 03-Apr-14 (No. 217-01918) 03-Apr-14 (No. 217-01921) 30-Dec-13 (No. ES3-3205_Dec13) 18-Aug-14 (No. DAE4-601_Aug14)	Scheduled Calibration Oct-14 Oct-14 Oct-14 Apr-15 Apr-15 Dec-14 Aug-15
Calibration Equipment used (M8 Primary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator ype-N mismatch combination Reference Probe ES3DV3 (AE4	TE critical for calibration) ID # GB37480704 US37292783 MY41092317 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 3205 SN: 601	Cal Date (Certificate No.) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01828) 03-Apr-14 (No. 217-01918) 03-Apr-14 (No. 217-01921) 30-Dec-13 (No. ES3-3205_Dec13)	Scheduled Calibration Oct-14 Oct-14 Oct-14 Apr-15 Apr-15 Dec-14
Calibration Equipment used (M8 Primary Standards Power meter EPM-442A Power sensor HP 8481A Power sensor HP 8481A Reference 20 dB Attenuator ype-N mismatch combination Reference Probe ES3DV3 AE4 econdary Standards F generator R&S SMT-06	TE critical for calibration)           ID #           GB37480704           US37292783           MY41092317           SN: 5058 (20k)           SN: 5058 (20k)           SN: 3205           SN: 601           ID #           100005	Cal Date (Certificate No.) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01827) 09-Oct-13 (No. 217-01828) 03-Apr-14 (No. 217-01918) 03-Apr-14 (No. 217-01921) 30-Dec-13 (No. ES3-3205_Dec13) 18-Aug-14 (No. DAE4-601_Aug14) Check Date (in house) 04-Aug-99 (in house check Oct-13)	Scheduled Calibration Oct-14 Oct-14 Oct-14 Apr-15 Apr-15 Dec-14 Aug-15 Scheduled Check In house check: Oct-16

This calibration certificate shall not be reproduced	except in full without written approval of the laboratory
eerimeate onain not be reproduced	except in full without written approval of the laborator

Katja Pokovic

Issued: September 23, 2014

Certificate No: D1900V2-5d143\_Sep14

Approved by:

Technical Manager

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland





S Schweizerischer Kalibrierdienst

- C Service suisse d'étalonnage
- Servizio svizzero di taratura
- S Swiss Calibration Service

Accredited by the Swiss Accreditation Service (SAS) The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

## Glossary:

TSL	tissue simulating liquid
ConvF	sensitivity in TSL / NORM x,y,z
N/A	not applicable or not measured

## Calibration is Performed According to the Following Standards:

- a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013
- b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300 MHz to 3 GHz)", February 2005
- c) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

## Additional Documentation:

d) DASY4/5 System Handbook

## Methods Applied and Interpretation of Parameters:

- Measurement Conditions: Further details are available from the Validation Report at the end
  of the certificate. All figures stated in the certificate are valid at the frequency indicated.
- Antenna Parameters with TSL: The dipole is mounted with the spacer to position its feed
  point exactly below the center marking of the flat phantom section, with the arms oriented
  parallel to the body axis.
- Feed Point Impedance and Return Loss: These parameters are measured with the dipole
  positioned under the liquid filled phantom. The impedance stated is transformed from the
  measurement at the SMA connector to the feed point. The Return Loss ensures low
  reflected power. No uncertainty required.
- Electrical Delay: One-way delay between the SMA connector and the antenna feed point. No uncertainty required.
- SAR measured: SAR measured at the stated antenna input power.
- SAR normalized: SAR as measured, normalized to an input power of 1 W at the antenna connector.
- SAR for nominal TSL parameters: The measured TSL parameters are used to calculate the nominal SAR result.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: D1900V2-5d143\_Sep14

Accreditation No.: SCS 108

## Measurement Conditions

DASY system configuration, as far as not given on page 1.

DASY Version	DASY5	V52.8.8
Extrapolation	Advanced Extrapolation	
Phantom	Modular Flat Phantom	
Distance Dipole Center - TSL	10 mm	with Spacer
Zoom Scan Resolution	dx, dy, dz = 5 mm	
Frequency	1900 MHz ± 1 MHz	

Head TSL parameters The following parameters and calculations were applied.

te tenering p	Temperature	Permittivity	Conductivity
Nominal Head TSL parameters	22.0 °C	40.0	1.40 mho/m
Measured Head TSL parameters	(22.0 ± 0.2) °C	39.5 ± 6 %	1.37 mho/m ± 6 %
Head TSL temperature change during test	< 0.5 °C		

## SAR result with Head TSL

SAR averaged over 1 cm3 (1 g) of Head TSL	Condition	
SAR measured	250 mW input power	10.1 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	40.8 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Head TSL	condition	
SAR measured	250 mW input power	5.31 W/kg
SAR for nominal Head TSL parameters	normalized to 1W	21.4 W/kg ± 16.5 % (k=2)

Body TSL parameters The following parameters and calculations were applied.

	Temperature	Permittivity	Conductivity
Nominal Body TSL parameters	22.0 °C	53.3	1.52 mho/m
Measured Body TSL parameters	(22.0 ± 0.2) °C	52.6 ± 6 %	1.50 mho/m ± 6 %
Body TSL temperature change during test	< 0.5 °C		

## SAR result with Body TSL

SAR averaged over 1 cm <sup>3</sup> (1 g) of Body TSL	Condition	
SAB measured	250 mW input power	10.0 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	40.2 W/kg ± 17.0 % (k=2)

SAR averaged over 10 cm <sup>3</sup> (10 g) of Body TSL	condition	
SAR measured	250 mW input power	5.32 W/kg
SAR for nominal Body TSL parameters	normalized to 1W	21.3 W/kg ± 16.5 % (k=2)

Certificate No: D1900V2-5d143\_Sep14

# Appendix (Additional assessments outside the scope of SCS108)

## Antenna Parameters with Head TSL

53.1 Ω + 6.3 ]Ω
- 23.3 dB

## Antenna Parameters with Body TSL

Impedance, transformed to feed point	48.9 Ω + 6.2 jΩ
Return Loss	- 24.0 dB

## General Antenna Parameters and Design

Electrical Delay (one direction)	1.195 ns

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals. On some of the dipoles, small end caps are added to the dipole arms in order to improve matching when loaded according to the position as explained in the "Measurement Conditions" paragraph. The SAR data are not affected by this change. The overall dipole length is still according to the Standard.

No excessive force must be applied to the dipole arms, because they might bend or the soldered connections near the feedpoint may be damaged.

## Additional EUT Data

Manufactured by	SPEAG
Manufactured on	March 11, 2011