

APPENDIX C.  
Calibration Certificates

**Schmid & Partner  
Engineering AG**

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

**Calibration Certificate**

**Dosimetric E-Field Probe**

Type:

ET3DV6

Serial Number:

1395

Place of Calibration:

Zurich

Date of Calibration:

August 27, 2002

Calibration Interval:

12 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

D. Vetter

Approved by:

Marie-Kristin

# Probe ET3DV6

**SN:1395**

Manufactured:	October 1, 1999
Last calibration:	July 25, 2001
Recalibrated:	August 27, 2002

Calibrated for System DASY3

**DASY3 - Parameters of Probe: ET3DV6 SN:1395****Sensitivity in Free Space**

NormX	<b>1.69</b> $\mu\text{V}/(\text{V}/\text{m})^2$
NormY	<b>1.73</b> $\mu\text{V}/(\text{V}/\text{m})^2$
NormZ	<b>1.67</b> $\mu\text{V}/(\text{V}/\text{m})^2$

**Diode Compression**

DCP X	<b>94</b>	mV
DCP Y	<b>94</b>	mV
DCP Z	<b>94</b>	mV

**Sensitivity in Tissue Simulating Liquid**

Head	<b>900 MHz</b>	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m
Head	<b>835 MHz</b>	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\%$ mho/m
ConvF X	<b>6.3</b> $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	<b>6.3</b> $\pm 9.5\%$ (k=2)	Alpha	<b>0.42</b>
ConvF Z	<b>6.3</b> $\pm 9.5\%$ (k=2)	Depth	<b>2.57</b>
Head	<b>1800 MHz</b>	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
Head	<b>1900 MHz</b>	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
ConvF X	<b>5.4</b> $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	<b>5.4</b> $\pm 9.5\%$ (k=2)	Alpha	<b>0.61</b>
ConvF Z	<b>5.4</b> $\pm 9.5\%$ (k=2)	Depth	<b>2.26</b>


**Boundary Effect**


Head	900 MHz	Typical SAR gradient: 5 % per mm		
	Probe Tip to Boundary	1 mm	2 mm	
	SAR <sub>ts</sub> [%] Without Correction Algorithm	11.5	6.7	
	SAR <sub>ts</sub> [%] With Correction Algorithm	0.3	0.6	
Head	1800 MHz	Typical SAR gradient: 10 % per mm		
	Probe Tip to Boundary	1 mm	2 mm	
	SAR <sub>ts</sub> [%] Without Correction Algorithm	13.2	8.4	
	SAR <sub>ts</sub> [%] With Correction Algorithm	0.1	0.2	

**Sensor Offset**


Probe Tip to Sensor Center	<b>2.7</b>	mm
Optical Surface Detection	<b>1.4 <math>\pm</math> 0.2</b>	mm

# Receiving Pattern ( $\phi$ ), $\theta = 0^\circ$






### Isotropy Error ( $\phi$ ), $\theta = 0^\circ$





## Frequency Response of E-Field


( TEM-Cell:ifi110, Waveguide R22)



### Dynamic Range $f(\text{SAR}_{\text{brain}})$ ( Waveguide R22 )




## Conversion Factor Assessment



Head	900 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m
Head	835 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\%$ mho/m
	ConvF X	$6.3 \pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	$6.3 \pm 9.5\%$ (k=2)	Alpha 0.42
	ConvF Z	$6.3 \pm 9.5\%$ (k=2)	Depth 2.57
Head	1800 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
Head	1900 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
	ConvF X	$5.4 \pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	$5.4 \pm 9.5\%$ (k=2)	Alpha 0.61
	ConvF Z	$5.4 \pm 9.5\%$ (k=2)	Depth 2.26




## Conversion Factor Assessment





Body	835 MHz	$\epsilon_r = 55.2 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m	
Body	900 MHz	$\epsilon_r = 55.0 \pm 5\%$	$\sigma = 1.05 \pm 5\%$ mho/m	
	ConvF X	$6.2 \pm 9.5\%$ (k=2)	Boundary effect:	
	ConvF Y	$6.2 \pm 9.5\%$ (k=2)	Alpha	0.50
	ConvF Z	$6.2 \pm 9.5\%$ (k=2)	Depth	2.28
Body	1900 MHz	$\epsilon_r = 53.3 \pm 5\%$	$\sigma = 1.52 \pm 5\%$ mho/m	
Body	1800 MHz	$\epsilon_r = 53.3 \pm 5\%$	$\sigma = 1.52 \pm 5\%$ mho/m	
	ConvF X	$4.9 \pm 9.5\%$ (k=2)	Boundary effect:	
	ConvF Y	$4.9 \pm 9.5\%$ (k=2)	Alpha	0.84
	ConvF Z	$4.9 \pm 9.5\%$ (k=2)	Depth	2.01

## Deviation from Isotropy in HSL

Error ( $\theta, \phi$ ),  $f = 900$  MHz



Client **Nokia Salo (TTC)**

CALIBRATION CERTIFICATE			
Object(s)	ET3DV6 - SN 1396		
Calibration procedure(s)	QA CAL-01.v2 Calibration procedure for dosimetric E-field probes		
Calibration date:	January 15, 2003		
Condition of the calibrated item	In Tolerance (according to the specific calibration document)		
This calibration statement documents traceability of M&TE used in the calibration procedures and conformity of the procedures with the ISO/IEC 17025 international standard.			
All calibrations have been conducted in the closed laboratory facility: environment temperature 22 +/- 2 degrees Celsius and humidity < 75%.			
Calibration Equipment used (M&TE critical for calibration)			
Model Type	ID #	Cal Date	Scheduled Calibration
RF generator HP 8684C	US3642U01700	4-Aug-98 (in house check Aug-02)	In house check: Aug-05
Power sensor E4412A	MY41495277	8-Mar-02	Mar-03
Power sensor HP 8481A	MY41092180	18-Sep-02	Sep-03
Power meter EPM E4419B	GB41293874	13-Sep-02	Sep-03
Network Analyzer HP 8753E	US38432426	3-May-00	In house check: May 03
Fluke Process Calibrator Type 702	SN: 6295803	3-Sep-01	Sep-03
Calibrated by:	Name Nico Vetterli	Function Technician	Signature 
Approved by:	Katja Pokovic	Laboratory Director	
Date issued: January 16, 2003			
This calibration certificate is issued as an intermediate solution until the accreditation process (based on ISO/IEC 17025 International Standard) for Calibration Laboratory of Schmid & Partner Engineering AG is completed.			

# Probe ET3DV6

SN:1396

Manufactured:	October 1, 1999
Last calibration:	January 29, 2002
Recalibrated:	January 15, 2003

Calibrated for DASY Systems

(Note: non-compatible with DASY2 system!)

**DASY - Parameters of Probe: ET3DV6 SN:1396****Sensitivity in Free Space****Diode Compression**

NormX	1.72 $\mu\text{V}/(\text{V}/\text{m})^2$	DCP X	93	mV
NormY	1.73 $\mu\text{V}/(\text{V}/\text{m})^2$	DCP Y	93	mV
NormZ	1.84 $\mu\text{V}/(\text{V}/\text{m})^2$	DCP Z	93	mV

**Sensitivity in Tissue Simulating Liquid**


Head	900 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\% \text{ mho/m}$
Head	835 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\% \text{ mho/m}$
ConvF X	6.9 $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	6.9 $\pm 9.5\%$ (k=2)	Alpha	0.35
ConvF Z	6.9 $\pm 9.5\%$ (k=2)	Depth	2.53
Head	1800 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\% \text{ mho/m}$
Head	1900 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\% \text{ mho/m}$
ConvF X	5.6 $\pm 9.5\%$ (k=2)	Boundary effect:	
ConvF Y	5.6 $\pm 9.5\%$ (k=2)	Alpha	0.46
ConvF Z	5.6 $\pm 9.5\%$ (k=2)	Depth	2.71


**Boundary Effect**

Head	900 MHz	Typical SAR gradient: 5 % per mm		
Probe Tip to Boundary		1 mm	2 mm	
SAR <sub>90</sub> [%]	Without Correction Algorithm	9.2	5.2	
SAR <sub>90</sub> [%]	With Correction Algorithm	0.3	0.5	
Head	1800 MHz	Typical SAR gradient: 10 % per mm		
Probe Tip to Boundary		1 mm	2 mm	
SAR <sub>90</sub> [%]	Without Correction Algorithm	13.1	8.9	
SAR <sub>90</sub> [%]	With Correction Algorithm	0.2	0.1	


**Sensor Offset**

Probe Tip to Sensor Center	2.7	mm
Optical Surface Detection	1.5 $\pm$ 0.2	mm

Receiving Pattern ( $\phi$ ),  $\theta = 0^\circ$ 




### Isotropy Error ( $\phi$ ), $\theta = 0^\circ$





## Frequency Response of E-Field

( TEM-Cell:ifi110, Waveguide R22)






**Dynamic Range f(SAR<sub>brain</sub>)**  
( Waveguide R22 )




## Conversion Factor Assessment



Head	900 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m
Head	835 MHz	$\epsilon_r = 41.5 \pm 5\%$	$\sigma = 0.90 \pm 5\%$ mho/m
	ConvF X	<b>6.9</b> $\pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	<b>6.9</b> $\pm 9.5\%$ (k=2)	Alpha <b>0.35</b>
	ConvF Z	<b>6.9</b> $\pm 9.5\%$ (k=2)	Depth <b>2.53</b>
Head	1800 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
Head	1900 MHz	$\epsilon_r = 40.0 \pm 5\%$	$\sigma = 1.40 \pm 5\%$ mho/m
	ConvF X	<b>5.6</b> $\pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	<b>5.6</b> $\pm 9.5\%$ (k=2)	Alpha <b>0.46</b>
	ConvF Z	<b>5.6</b> $\pm 9.5\%$ (k=2)	Depth <b>2.71</b>


## Conversion Factor Assessment



Body	900 MHz	$\epsilon_r = 55.0 \pm 5\%$	$\sigma = 1.05 \pm 5\%$ mho/m
Body	835 MHz	$\epsilon_r = 55.2 \pm 5\%$	$\sigma = 0.97 \pm 5\%$ mho/m
	ConvF X	$6.6 \pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	$6.6 \pm 9.5\%$ (k=2)	Alpha 0.36
	ConvF Z	$6.6 \pm 9.5\%$ (k=2)	Depth 2.57
Body	1800 MHz	$\epsilon_r = 53.3 \pm 5\%$	$\sigma = 1.52 \pm 5\%$ mho/m
Body	1900 MHz	$\epsilon_r = 53.3 \pm 5\%$	$\sigma = 1.52 \pm 5\%$ mho/m
	ConvF X	$5.1 \pm 9.5\%$ (k=2)	Boundary effect:
	ConvF Y	$5.1 \pm 9.5\%$ (k=2)	Alpha 0.53
	ConvF Z	$5.1 \pm 9.5\%$ (k=2)	Depth 2.75

### Deviation from Isotropy in HSL

Error ( $\theta, \phi$ ),  $f = 900$  MHz



# Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

## Calibration Certificate

### 900 MHz System Validation Dipole

Type:

D900V2

Serial Number:

056

Place of Calibration:

Zurich

Date of Calibration:

January 29, 2002

Calibration Interval:

24 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:

*W. K. E. Merian*

Approved by:

*Alain Kappeler*

**Schmid & Partner  
Engineering AG**

---

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

---

**DASY**

**Dipole Validation Kit**

**Type: D900V2**

**Serial: 056**

**Manufactured: September 25, 1999**  
**Calibrated: January 29, 2002**

## 1. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating solution of the following electrical parameters at 900 MHz:

Relative Dielectricity	41.1	± 5%
Conductivity	0.95 mho/m	± 5%

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.48 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 2.50mW ± 3 %. The results are normalized to 1W input power.

## 2. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 1. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm <sup>3</sup> (1 g) of tissue:	11.1 mW/g
averaged over 10 cm <sup>3</sup> (10 g) of tissue:	7.00 mW/g

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well.

### 3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay:	<b>1.348 ns</b>	(one direction)
Transmission factor:	<b>0.986</b>	(voltage transmission, one direction)

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 900 MHz:	$\text{Re}\{Z\} = 50.2 \, \Omega$
	$\text{Im}\{Z\} = -0.7 \, \Omega$
Return Loss at 900 MHz	<b>-42.9 dB</b>

### 4. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with body simulating solution of the following electrical parameters at 900 MHz:

Relative Dielectricity	<b>54.8</b>	$\pm 5\%$
Conductivity	<b>1.03 mho/m</b>	$\pm 5\%$

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 6.17 at 900 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 15mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging. The dipole input power (forward power) was  $250\text{mW} \pm 3\%$ . The results are normalized to 1W input power.



## **5. SAR Measurement**

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 4. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm <sup>3</sup> (1 g) of tissue:	11.7 mW/g
averaged over 10 cm <sup>3</sup> (10 g) of tissue:	7.44 mW/g

## **6. Dipole Impedance and Return Loss**

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 900 MHz:	$\text{Re}\{Z\} = 45.7 \Omega$
	$\text{Im}\{Z\} = -2.4 \Omega$
Return Loss at 900 MHz	-25.7 dB

## **7. Handling**

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

## **8. Design**

The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.


## **9. Power Test**

After long term use with 100W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

01/28/02

Validation Dipole D900V2 SN:056, d = 15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom, Flat Section, Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DM6 - SN1507; ConvF (6.48, 6.48, 6.48) at 900 MHz; IEEE1528 900 MHz;  $\sigma = 0.95$  mholm  $\epsilon_r = 41.1$   $\rho = 1.00$  g/cm<sup>3</sup>  
Cubes (2): Peak: 4.48 mW/g  $\pm 0.01$  dB, SAR (1g): 2.78 mW/g  $\pm 0.02$  dB, SAR (10g): 1.75 mW/g  $\pm 0.02$  dB, (Worst-case extrapolation)  
Penetration depth: 11.5 (10.3, 13.0) [mm]  
Powerdrift: -0.02 dB




18 Jan 2002 11:40:39  
 S41 E V F3 1:58.168 s -726.52 m 242.74 pF 900.000 000 MHz

Det

Cor

Av<sub>2</sub>  
 15

↑




↑

01/29/02

### Validation Dipole D900V2 SN:056, d = 15 mm

Frequency: 900 MHz; Antenna Input Power: 250 [mW]  
SAM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; CorvF(6.17,6.17,6.17) at 900 MHz; Muscle 900 MHz;  $\sigma = 1.03$  mhol/m,  $\epsilon_r = 54.9$ ,  $\rho = 1.00$  g/cm<sup>3</sup>  
Cubes (2): Peak: 4.65 mW/g  $\pm$  0.01 dB, SAR (1g): 2.92 mW/g  $\pm$  0.00 dB, SAR (10g): 1.86 mW/g  $\pm$  0.01 dB, (Worst-case extrapolation)  
Penetration depth: 12.0 (10.7, 13.7) [mm]  
Powerdrift: -0.02 dB



29 Jan 2002 12:13:56  
 S11 1.0 73 1145.685 GHz -2.4453 dB 72.317 pF 900,000 800 MHz

De1

Der

Mag

dB

+



CH2 S11 LOG 5 dB/REF 0 dB 11-25.715 dB 900,000 800 MHz

Der

Mag

dB

+



CENTER 900,000 800 MHz

SPAN 400,000 800 MHz

# Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

## Calibration Certificate

### 1800 MHz System Validation Dipole

Type:

D1800V2

Serial Number:

256

Place of Calibration:

Zurich

Date of Calibration:

January 29, 2002

Calibration Interval:

24 months

Schmid & Partner Engineering AG hereby certifies, that this device has been calibrated on the date indicated above. The calibration was performed in accordance with specifications and procedures of Schmid & Partner Engineering AG.

Wherever applicable, the standards used in the calibration process are traceable to international standards. In all other cases the standards of the Laboratory for EMF and Microwave Electronics at the Swiss Federal Institute of Technology (ETH) in Zurich, Switzerland have been applied.

Calibrated by:



Approved by:



**Schmid & Partner  
Engineering AG**

---

Zeughausstrasse 43, 8004 Zurich, Switzerland, Phone +41 1 245 97 00, Fax +41 1 245 97 79

---

**DASY**

**Dipole Validation Kit**

**Type: D1800V2**

**Serial: 256**

Manufactured: December 23, 1999  
Calibrated: January 29, 2002

## 1. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with head simulating glycol solution of the following electrical parameters at 1800 MHz:

Relative Dielectricity	39.6	± 5%
Conductivity	1.37 mho/m	± 5%

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.31 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging.

The dipole input power (forward power) was 250mW ± 3 %. The results are normalized to 1W input power.

## 2. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 1. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm <sup>3</sup> (1 g) of tissue:	40.0 mW/g
averaged over 10 cm <sup>3</sup> (10 g) of tissue:	20.8 mW/g

Note: If the liquid parameters for validation are slightly different from the ones used for initial calibration, the SAR-values will be different as well. The estimated sensitivities of SAR-values and penetration depths to the liquid parameters are listed in the DASY Application Note 4: 'SAR Sensitivities'.



### 3. Dipole Impedance and Return Loss

The impedance was measured at the SMA-connector with a network analyzer and numerically transformed to the dipole feedpoint. The transformation parameters from the SMA-connector to the dipole feedpoint are:

Electrical delay:	<b>1.180 ns</b>	(one direction)
Transmission factor:	<b>0.980</b>	(voltage transmission, one direction)

The dipole was positioned at the flat phantom sections according to section 1 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 1800 MHz:	$\text{Re}\{Z\} = 47.3 \, \Omega$
	$\text{Im}\{Z\} = -6.6 \, \Omega$
Return Loss at 1800 MHz	<b>-22.7 dB</b>

### 4. Measurement Conditions

The measurements were performed in the flat section of the new generic twin phantom filled with body simulating glycol solution of the following electrical parameters at 1800 MHz:

Relative Dielectricity	<b>53.5</b>	$\pm 5\%$
Conductivity	<b>1.45 mho/m</b>	$\pm 5\%$

The DASY3 System (Software version 3.1d) with a dosimetric E-field probe ET3DV6 (SN:1507, Conversion factor 5.0 at 1800 MHz) was used for the measurements.

The dipole was mounted on the small tripod so that the dipole feedpoint was positioned below the center marking of the flat phantom section and the dipole was oriented parallel to the body axis (the long side of the phantom). The standard measuring distance was 10mm from dipole center to the solution surface. The included distance holder was used during measurements for accurate distance positioning.

The coarse grid with a grid spacing of 20mm was aligned with the dipole. The 5x5x7 fine cube was chosen for cube integration. Probe isotropy errors were cancelled by measuring the SAR with normal and 90° turned probe orientations and averaging. The dipole input power (forward power) was 250mW  $\pm 3\%$ . The results are normalized to 1W input power.

## 5. SAR Measurement

Standard SAR-measurements were performed with the phantom according to the measurement conditions described in section 4. The results have been normalized to a dipole input power of 1W (forward power). The resulting averaged SAR-values are:

averaged over 1 cm<sup>3</sup> (1 g) of tissue:      **39.5 mW/g**

averaged over 10 cm<sup>3</sup> (10 g) of tissue:      **20.4 mW/g**

## 6. Dipole Impedance and Return Loss

The dipole was positioned at the flat phantom sections according to section 4 and the distance holder was in place during impedance measurements.

Feedpoint impedance at 1800 MHz:       **$\text{Re}\{Z\} = 43.2 \Omega$**

**$\text{Im}\{Z\} = -6.3 \Omega$**

Return Loss at 1800 MHz      **-20.1 dB**

## 7. Handling

Do not apply excessive force to the dipole arms, because they might bend. Bending of the dipole arms stresses the soldered connections near the feedpoint leading to a damage of the dipole.

## 8. Design


The dipole is made of standard semirigid coaxial cable. The center conductor of the feeding line is directly connected to the second arm of the dipole. The antenna is therefore short-circuited for DC-signals.

## 9. Power Test

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

# Validation Dipole D1800V2 SN:256, d = 10 mm

Frequency: 1800 MHz, Antenna Input Power: 250 [mW]  
 SAM Phantom, Flat Section, Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
 Probe: ET3DV3 - SN1507, ConvF(5.31,5.31,5.31) at 1800 MHz; IEEE1526 1800 MHz;  $\sigma = 1.37$  mho/m  $\epsilon_r = 39.6$   $\rho = 1.00$  g/cm<sup>3</sup>  
 Cubes (2): Peak: 18.9 mW/g  $\pm$  0.04 dB, SAR (1g): 9.99 mW/g  $\pm$  0.01 dB, SAR (10g): 5.19 mW/g  $\pm$  0.01 dB, (Worst-case extrapolation)  
 Penetration depth: 8.2 (7.7, 9.2) [mm]  
 Powerdrift: -0.02 dB



Case:

15





41

01/29/02

### Validation Dipole D1800V2 SN:256, d = 10 mm

Frequency: 1800 MHz; Antenna Input Power: 250 [mW]  
SJM Phantom; Flat Section; Grid Spacing: Dx = 20.0, Dy = 20.0, Dz = 10.0  
Probe: ET3DV6 - SN1507; ConvF(5.00,5.00,5.00) at 1800 MHz; Muscle 1800 MHz;  $\sigma = 1.45$  mho/m  $\epsilon_r = 53.5$   $\rho = 1.00$  g/cm<sup>3</sup>  
Cubes (2): Peak: 18.7 mW/g  $\pm 0.02$  dB; SAR (1g): 9.87 mW/g  $\pm 0.01$  dB; SAR (10g): 5.11 mW/g  $\pm 0.01$  dB; (Worst-case extrapolation)  
Penetration depth: 8.4 (7.7, 9.8) [mm]  
Powerdrift: -0.01 dB



29 Jan 2002 11:57:01  
 S11 1 u F5 143.175  $\angle$  -5.2773  $\angle$  14.885 pF 1.000.000 000 MHz

DeS

Con

Mag  
16

T

