

11 Measurement Uncertainty

Per KDB 865664 D01 SAR Measurement 100 MHz to 6 GHz, when the highest measured 1-g SAR within a frequency band is < 1.5 W/kg, the extensive SAR measurement uncertainty analysis described in IEEE Std 1528- 2013 is not required in SAR reports submitted for equipment approval.

Report No: RXA1707-0218SAR02R2




ANNEX A: Test Layout





Picture 3: Liquid depth in the head Phantom (5250 MHz, 15.4cm depth)



Picture 4: Liquid depth in the flat Phantom (5250 MHz, 15.3cm depth)



Picture 5: Liquid depth in the head Phantom (5750 MHz, 15.1cm depth)



Picture 6: Liquid depth in the flat Phantom (5750 MHz, 15.0cm depth)



ANNEX B: System Check Results

Plot 1 System Performance Check at 5250 MHz Head TSL DUT: Dipole 5250 MHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1151

Date: 7/13/2017

Communication System: CW; Frequency: 5250 MHz; Duty Cycle: 1:1

Medium parameters used: f = 5250 MHz; σ = 4.75 mho/m; ε_r = 35.96; ρ = 1000 kg/m³

Ambient Temperature: 22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Flat Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(5.66, 5.66, 5.66); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: SAM 2; Type: SAM;

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)

d=10mm, Pin=100mW/Area Scan (61x101x1): Measurement grid: dx=1.000mm, dy=1.000mm

Maximum value of SAR (interpolated) = 9.14 mW/g


d=10mm, Pin=100mW/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 33.6 V/m; Power Drift = -0.095 dB

Peak SAR (extrapolated) = 52.2 W/kg

SAR(1 g) = 7.87 mW/g; SAR(10 g) = 2.25 mW/g

Maximum value of SAR (measured) = 9.73 mW/g





Plot 2 System Performance Check at 5250 MHz Body TSL DUT: Dipole 5250 MHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1151

Date: 7/13/2017

Communication System: CW; Frequency: 5250 MHz; Duty Cycle: 1:1

Medium parameters used: f = 5250 MHz; σ = 5.42 mho/m; ε_r = 46.71; ρ = 1000 kg/m³

Ambient Temperature:22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Flat Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(5.03, 5.03, 5.03); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: ELI v5.0 (30deg probe tilt); Type: QDOVA002AA;

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)


d=10mm, Pin=250mW/Area Scan (61x101x1): Measurement grid: dx=1.000mm, dy=1.000mm Maximum value of SAR (interpolated) = 7.69 mW/g

d=10mm, Pin=250mW/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 36.3 V/m; Power Drift = 0.0277 dB

Peak SAR (extrapolated) = 47.7 W/kg

SAR(1 g) = 7.46 mW/g; SAR(10 g) = 2.26 mW/g Maximum value of SAR (measured) = 7.94 mW/g





Plot 3 System Performance Check at 5750 MHz Head TSL DUT: Dipole 5750 MHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1151

Date: 7/27/2017

Communication System: CW; Frequency: 5750 MHz; Duty Cycle: 1:1

Medium parameters used: f = 5750 MHz; σ = 5.39 mho/m; ε_r = 34.68; ρ = 1000 kg/m³

Ambient Temperature: 22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Flat Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(5.00, 5.00, 5.00); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: SAM 2; Type: SAM;

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)

d=10mm, Pin=100mW/Area Scan (61x101x1): Measurement grid: dx=1.000mm, dy=1.000mm


Maximum value of SAR (interpolated) = 8.27 mW/g

d=10mm, Pin=100mW/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 23.1 V/m; Power Drift = 0.044 dB

Peak SAR (extrapolated) = 23.5 W/kg

SAR(1 g) = 7.65 mW/g; SAR(10 g) = 2.28 mW/gMaximum value of SAR (measured) = 8.96 mW/g





Report No: RXA1707-0218SAR02R2

Plot 4 System Performance Check at 5750 MHz Body TSL DUT: Dipole 5750 MHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1151

Date: 7/27/2017

Communication System: CW; Frequency: 5750 MHz; Duty Cycle: 1:1

Medium parameters used: f = 5750 MHz; σ = 6.07 mho/m; ε_r = 47.73; ρ = 1000 kg/m³

Ambient Temperature:22.3 ℃ Liquid Temperature: 21.5℃

Phantom section: Flat Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(4.52, 4.52, 4.52); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: ELI v5.0 (30deg probe tilt); Type: QDOVA002AA;

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)


d=10mm, Pin=250mW/Area Scan (61x101x1): Measurement grid: dx=1.000mm, dy=1.000mm Maximum value of SAR (interpolated) = 7.87 mW/g

d=10mm, Pin=250mW/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 38 V/m; Power Drift = -0.035 dB

Peak SAR (extrapolated) = 22.8 W/kg

SAR(1 g) = 7.17 mW/g; SAR(10 g) = 1.99 mW/gMaximum value of SAR (measured) = 8.73 mW/g





ANNEX C: Highest Graph Results

Plot 5 802.11n (40M) U-NII-2A Left Tilt CH62

Date: 7/13/2017

Communication System: UID 0, 802.11n(40M) (0); Frequency: 5310 MHz; Duty Cycle: 1:1 Medium parameters used: f = 5310 MHz; $\sigma = 4.831$ S/m; $\epsilon_r = 35.798$; $\rho = 1000$ kg/m³

Ambient Temperature: 22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Left Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(5.66, 5.66, 5.66); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: SAM 2; Type: SAM;

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)

Left Tilt CH62/Area Scan (111x181x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 0.346 W/kg


Left Tilt CH62/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 2.385 V/m; Power Drift = 0.161 dB

Peak SAR (extrapolated) = 2.62 W/kg

SAR(1 g) = 0.373 W/kg; SAR(10 g) = 0.117 W/kg

Maximum value of SAR (measured) = 0.488 W/kg





Plot 6 802.11n (40M) U-NII-2A Back Side CH62 (Distance 10mm)

Date: 7/13/2017

Communication System: UID 0, 802.11n(40M) (0); Frequency: 5310 MHz; Duty Cycle: 1:1 Medium parameters used: f = 5310 MHz; $\sigma = 5.506$ S/m; $\epsilon_r = 46.559$; $\rho = 1000$ kg/m³

Ambient Temperature: 22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Flat Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(5.03, 5.03, 5.03); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: SAM 2; Type: SAM; Serial:

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)

Back Side CH62/Area Scan (111x181x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 0.285 W/kg


Back Side CH62/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 0.6900 V/m; Power Drift = 0.026 dB

Peak SAR (extrapolated) = 0.662 W/kg

SAR(1 g) = 0.230 W/kg; SAR(10 g) = 0.080 W/kg

Maximum value of SAR (measured) = 0.272 W/kg





Plot 7 802.11n (40M) U-NII-3 Left Tilt CH151

Date: 7/27/2017

Communication System: UID 0, 802.11n(40M) (0); Frequency: 5755 MHz; Duty Cycle: 1:1 Medium parameters used: f = 5755 MHz; $\sigma = 5.398$ S/m; $\epsilon_r = 34.667$; $\rho = 1000$ kg/m³

Ambient Temperature: 22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Left Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(5.00, 5.00, 5.00); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: SAM 2; Type: SAM;

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)

Left Tilt CH151/Area Scan (111x181x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 0.0670W/kg


Left Tilt CH151/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 0 V/m; Power Drift = 0.100 dB

Peak SAR (extrapolated) = 0.283 W/kg

SAR(1 g) = 0.065 W/kg; SAR(10 g) = 0.018 W/kg

Maximum value of SAR (measured) = 0.070 W/kg





Plot 8 802.11n (40M) U-NII-3 Top Side CH151 (Distance 10mm)

Date: 7/27/2017

Communication System: UID 0, 802.11n(40M) (0); Frequency: 5755 MHz; Duty Cycle: 1:1 Medium parameters used: f = 5755 MHz; $\sigma = 6.071$ S/m; $\epsilon_r = 47.721$; $\rho = 1000$ kg/m³

Ambient Temperature: 22.3 °C Liquid Temperature: 21.5 °C

Phantom section: Flat Section

DASY5 Configuration:

Sensor-Surface: 1.4mm (Mechanical Surface Detection)

Probe: EX3DV4 - SN3677; ConvF(4.52, 4.52, 4.52); Calibrated: 1/23/2017;

Electronics: DAE4 Sn1317; Calibrated: 8/2/2016

Phantom: SAM 2; Type: SAM; Serial:

Measurement SW: DASY52, Version 52.8 (8); SEMCAD X Version 14.6.10 (7331)

Top Side CH151/Area Scan (51x111x1): Interpolated grid: dx=1.000 mm, dy=1.000 mm

Maximum value of SAR (interpolated) = 0.0951W/kg


Top Side CH151/Zoom Scan (7x7x12)/Cube 0: Measurement grid: dx=4mm, dy=4mm, dz=2mm

Reference Value = 0.8410 V/m; Power Drift = -0.131 dB

Peak SAR (extrapolated) = 0.225 W/k

SAR(1 g) = 0.047 W/kg; SAR(10 g) = 0.013 W/kg

Maximum value of SAR (measured) = 0.061 W/kg



Report No: RXA1707-0218SAR02R2

ANNEX D: Probe Calibration Certificate



E-mail: cttl@chinattl.com

Http://www.chinattl.cn

TA(Shanghai) Client

Certificate No: Z17-97012

CALIBRATION CERTIFICATE

Object

EX3DV4 - SN:3677

Calibration Procedure(s)

FD-Z11-004-01

Calibration Procedures for Dosimetric E-field Probes

Calibration date:

January 23, 2017

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID# | Cal Date(Calibrated by, Certificate No.) | Scheduled Calibration | |
|--|-------------------|---|------------------------------|--|
| Power Meter NRP2 | 101919 | 27-Jun-16 (CTTL, No.J16X04777) | Jun-17 | |
| Power sensor NRP-Z91 | 101547 | 27-Jun-16 (CTTL, No.J16X04777) | Jun-17 | |
| Power sensor NRP-Z91 | 101548 | 27-Jun-16 (CTTL, No.J16X04777) | Jun-17 | |
| Reference10dBAttenuator | 18N50W-10dE | 3 13-Mar-16(CTTL,No.J16X01547) | Mar-18 | |
| Reference20dBAttenuator | 18N50W-20dE | 3 13-Mar-16(CTTL, No.J16X01548) | Mar-18 | |
| Reference Probe EX3DV4 | SN 7433 | 26-Sep-16(SPEAG,No.EX3-7433_Sep16) | Sep-17 | |
| DAE4 | SN 549 | 13-Dec-16(SPEAG, No.DAE4-549_Dec16) | Dec -17 | |
| Secondary Standards SignalGeneratorMG3700A | ID# 6201052605 | Cal Date(Calibrated by, Certificate No.) 27-Jun-16 (CTTL, No.J16X04776) | Scheduled Calibration Jun-17 | |
| Network Analyzer E5071C | | | | |
| | Name | Function | Şignature | |
| Calibrated by: | Yu Zongying | SAR Test Engineer | A TO | |
| Reviewed by: | Qi Dianyuan | SAR Project Leader | Can less | |
| Approved by: | Lu Bingsong | Deputy Director of the laboratory | Fernanda / | |
| | | Issued: Januar | y24, 2017 | |

This calibration certificate shall not be reproduced except in full without written approval of the laboratory.

Certificate No: Z17-97012

Page 1 of 11



In Collaboration with

S D E A G

CALIBRATION LABORATORY

Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2218 Fax: +86-10-62304633-2209 E-mail: cttl@chinattl.com Http://www.chinattl.cn

Glossary:

TSL tissue simulating liquid
NORMx,y,z sensitivity in free space
ConvF sensitivity in TSL / NORMx,y,z
DCP diode compression point

CF crest factor (1/duty_cycle) of the RF signal A,B,C,D modulation dependent linearization parameters

Polarization θ θ rotation around an axis that is in the plane normal to probe axis (at measurement center), i

 θ =0 is normal to probe axis

Connector Angle information used in DASY system to align probe sensor X to the robot coordinate system Calibration is Performed According to the Following Standards:

a) IEEE Std 1528-2013, "IEEE Recommended Practice for Determining the Peak Spatial-Averaged Specific Absorption Rate (SAR) in the Human Head from Wireless Communications Devices: Measurement Techniques", June 2013

 b) IEC 62209-1, "Procedure to measure the Specific Absorption Rate (SAR) for hand-held devices used in close proximity to the ear (frequency range of 300MHz to 3GHz)", February 2005

c) IEC 62209-2, "Procedure to determine the Specific Absorption Rate (SAR) for wireless communication devices used in close proximity to the human body (frequency range of 30 MHz to 6 GHz)", March 2010

d) KDB 865664, "SAR Measurement Requirements for 100 MHz to 6 GHz"

Methods Applied and Interpretation of Parameters:

- NORMx,y,z: Assessed for E-field polarization θ=0 (f≤900MHz in TEM-cell; f>1800MHz: waveguide).
 NORMx,y,z are only intermediate values, i.e., the uncertainties of NORMx,y,z does not effect the E²-field uncertainty inside TSL (see below ConvF).
- NORM(f)x,y,z = NORMx,y,z* frequency_response (see Frequency Response Chart). This
 linearization is implemented in DASY4 software versions later than 4.2. The uncertainty of the
 frequency response is included in the stated uncertainty of ConvF.
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics.
- Ax,y,z; Bx,y,z; Cx,y,z;VRx,y,z:A,B,C are numerical linearization parameters assessed based on the
 data of power sweep for specific modulation signal. The parameters do not depend on frequency nor
 media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- ConvF and Boundary Effect Parameters: Assessed in flat phantom using E-field (or Temperature Transfer Standard for f≤800MHz) and inside waveguide using analytical field distributions based on power measurements for f>800MHz. The same setups are used for assessment of the parameters applied for boundary compensation (alpha, depth) of which typical uncertainty valued are given. These parameters are used in DASY4 software to improve probe accuracy close to the boundary. The sensitivity in TSL corresponds to NORMx,y,z* ConvF whereby the uncertainty corresponds to that given for ConvF. A frequency dependent ConvF is used in DASY version 4.4 and higher which allows extending the validity from±50MHz to±100MHz.
- Spherical isotropy (3D deviation from isotropy): in a field of low gradients realized using a flat phantom exposed by a patch antenna.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the
 probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMx (no uncertainty required).

Certificate No: Z17-97012

Page 2 of 11

Report No: RXA1707-0218SAR02R2





Probe EX3DV4

SN: 3677

Calibrated: January 23, 2017

Calibrated for DASY/EASY Systems

(Note: non-compatible with DASY2 system!)

Certificate No: Z17-97012

Page 3 of 11





DASY/EASY - Parameters of Probe: EX3DV4 - SN: 3677

Basic Calibration Parameters

| | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|---|----------|----------|----------|-----------|
| Norm(µV/(V/m) ²) ^A | 0.39 | 0.44 | 0.38 | ±10.8% |
| DCP(mV) ^B | 97.3 | 102.2 | 101.1 | |

Modulation Calibration Parameters

| UID | Communication System Name | | A dB | B dBõV | С | D dB | VR mV | Unc ^E (k=2) | |
|------|------------------------------|---|---------|-----------|-----|---------|----------|---------------------------|--|
| 0 CW | CW | Х | 0.0 | 0.0 | 1.0 | 0.00 | 180.5 | ±2.0% | |
| | | | Y | 0.0 | 0.0 | 1.0 | | 195.3 | |
| | | Z | 0.0 | 0.0 | 1.0 | | 177.9 | 7 | |

The reported uncertainty of measurement is stated as the standard uncertainty of Measurement multiplied by the coverage factor k=2, which for a normal distribution Corresponds to a coverage probability of approximately 95%.

^B Numerical linearization parameter: uncertainty not required.

Certificate No: Z17-97012

Page 4 of 11

^A The uncertainties of Norm X, Y, Z do not affect the E²-field uncertainty inside TSL (see Page 5 and Page 6).

E Uncertainly is determined using the max. deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Test Report No: RXA1707-0218SAR02R2



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2218 Fax: +86-10-62304633-2209 E-mail: cttl@chinattl.com Http://www.chinattl.cn

DASY/EASY - Parameters of Probe: EX3DV4 - SN: 3677

Calibration Parameter Determined in Head Tissue Simulating Media

| f [MHz] ^C | Relative Permittivity ^F | Conductivity (S/m) F | ConvF X | ConvF Y | ConvF Z | Alpha ^G | Depth ^G (mm) | Unct. (k=2) |
|----------------------|---------------------------------------|-------------------------|---------|---------|---------|--------------------|----------------------------|----------------|
| 750 | 41.9 | 0.89 | 9.58 | 9.58 | 9.58 | 0.30 | 0.75 | ±12% |
| 835 | 41.5 | 0.90 | 9.31 | 9.31 | 9.31 | 0.11 | 1.55 | ±12% |
| 1750 | 40.1 | 1.37 | 8.60 | 8.60 | 8.60 | 0.24 | 1.07 | ±12% |
| 1900 | 40.0 | 1.40 | 8.39 | 8.39 | 8.39 | 0.23 | 1.10 | ±12% |
| 2300 | 39.5 | 1.67 | 8.13 | 8.13 | 8.13 | 0.53 | 0.74 | ±12% |
| 2450 | 39.2 | 1.80 | 7.90 | 7.90 | 7.90 | 0.61 | 0.71 | ±12% |
| 2600 | 39.0 | 1.96 | 7.64 | 7.64 | 7.64 | 0.68 | 0.68 | ±12% |
| 5250 | 35.9 | 4.71 | 5.66 | 5.66 | 5.66 | 0.40 | 1.20 | ±13% |
| 5600 | 35.5 | 5.07 | 4.99 | 4.99 | 4.99 | 0.40 | 1.40 | ±13% |
| 5750 | 35.4 | 5.22 | 5.00 | 5.00 | 5.00 | 0.40 | 1.40 | ±13% |

^c Frequency validity above 300 MHz of ±100MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ±50MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

Certificate No: Z17-97012

Page 5 of 11

F At frequency below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ±5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than \pm 1% for frequencies below 3 GHz and below \pm 2% for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.

Report No: RXA1707-0218SAR02R2



Add: No.51 Xueyuan Road, Haidian District, Beijing, 100191, China Tel: +86-10-62304633-2218 Fax: +86-10-62304633-2209 E-mail: cttl@chinattl.com Http://www.chinattl.cn

DASY/EASY – Parameters of Probe: EX3DV4 – SN: 3677

Calibration Parameter Determined in Body Tissue Simulating Media

| f [MHz] ^C | Relative Permittivity ^F | Conductivity (S/m) ^F | ConvF X | ConvF Y | ConvFZ | Alpha ^G | Depth ^G (mm) | Unct. (k=2) |
|----------------------|---------------------------------------|------------------------------------|---------|---------|--------|--------------------|----------------------------|----------------|
| 750 | 55.5 | 0.96 | 9.99 | 9.99 | 9.99 | 0.30 | 0.95 | ±12% |
| 835 | 55.2 | 0.97 | 9.74 | 9.74 | 9.74 | 0.14 | 1.66 | ±12% |
| 1750 | 53.4 | 1.49 | 8.39 | 8.39 | 8.39 | 0.21 | 1.16 | ±12% |
| 1900 | 53.3 | 1.52 | 7.98 | 7.98 | 7.98 | 0.22 | 1.24 | ±12% |
| 2300 | 52.9 | 1.81 | 7.97 | 7.97 | 7.97 | 0.55 | 0.80 | ±12% |
| 2450 | 52.7 | 1.95 | 7.85 | 7.85 | 7.85 | 0.50 | 0.86 | ±12% |
| 2600 | 52.5 | 2.16 | 7.63 | 7.63 | 7.63 | 0.44 | 0.91 | ±12% |
| 5250 | 48.9 | 5.36 | 5.03 | 5.03 | 5.03 | 0.50 | 1.60 | ±13% |
| 5600 | 48.5 | 5.77 | 4.34 | 4.34 | 4.34 | 0.54 | 1.66 | ±13% |
| 5750 | 48.3 | 5.94 | 4.52 | 4.52 | 4.52 | 0.57 | 1.95 | ±13% |

^c Frequency validity above 300 MHz of ±100MHz only applies for DASY v4.4 and higher (Page 2), else it is restricted to ±50MHz. The uncertainty is the RSS of ConvF uncertainty at calibration frequency and the uncertainty for the indicated frequency band. Frequency validity below 300 MHz is ± 10, 25, 40, 50 and 70 MHz for ConvF assessments at 30, 64, 128, 150 and 220 MHz respectively. Above 5 GHz frequency validity can be extended to ± 110 MHz.

Certificate No: Z17-97012


Page 6 of 11

F At frequency below 3 GHz, the validity of tissue parameters (ϵ and σ) can be relaxed to ±10% if liquid compensation formula is applied to measured SAR values. At frequencies above 3 GHz, the validity of tissue parameters (ϵ and σ) is restricted to ±5%. The uncertainty is the RSS of the ConvF uncertainty for indicated target tissue parameters.

^G Alpha/Depth are determined during calibration. SPEAG warrants that the remaining deviation due to the boundary effect after compensation is always less than ± 1% for frequencies below 3 GHz and below ± 2% for the frequencies between 3-6 GHz at any distance larger than half the probe tip diameter from the boundary.



Frequency Response of E-Field (TEM-Cell: ifi110 EXX, Waveguide: R22)



Uncertainty of Frequency Response of E-field: ±7.5% (k=2)

Certificate No: Z17-97012


Page 7 of 11




Receiving Pattern (Φ), θ=0°

f=600 MHz, TEM

f=1800 MHz, R22





Certificate No: Z17-97012

Page 8 of 11



Dynamic Range f(SAR_{head}) (TEM cell, f = 900 MHz) 10⁶ 105 Input Signal[µV] 104 10² 10° 10°2 10-1 10² 10 103 SAR[mW/cm³] compensated Error[dB] -2 10" SAR[mW/cm not compensated -e- compensated Uncertainty of Linearity Assessment: ±0.9% (k=2)

Certificate No: Z17-97012


Page 9 of 11




Conversion Factor Assessment


f=835 MHz, WGLS R9(H_convF)

f=1750 MHz, WGLS R22(H_convF)





Deviation from Isotropy in Liquid



Uncertainty of Spherical Isotropy Assessment: ±2.8% (K=2)

Certificate No: Z17-97012

Page 10 of 11





DASY/EASY - Parameters of Probe: EX3DV4 - SN: 3677

Other Probe Parameters

| Sensor Arrangement | Triangular |
|---|------------|
| Connector Angle (°) | 117.9 |
| Mechanical Surface Detection Mode | enabled |
| Optical Surface Detection Mode | disable |
| Probe Overall Length | 337mm |
| Probe Body Diameter | 10mm |
| Tip Length | 9mm |
| Tip Diameter | 2.5mm |
| Probe Tip to Sensor X Calibration Point | 1mm |
| Probe Tip to Sensor Y Calibration Point | 1mm |
| Probe Tip to Sensor Z Calibration Point | 1mm |
| Recommended Measurement Distance from Surface | 1.4mm |


Certificate No: Z17-97012

Page 11 of 11



Report No: RXA1707-0218SAR02R2

ANNEX E: D5GHzV2 Dipole Calibration Certificate



E-mail: cttl@chinattl.com

http://www.chinattl.cn



CALIBRATION

Client

TA(Shanghai)

Certificate No:

Z17-97001

CALIBRATION CERTIFICATE

Object

D5GHzV2 - SN: 1151

Calibration Procedure(s)

FD-Z11-003-01

Calibration Procedures for dipole validation kits

Calibration date:

January 5, 2017

This calibration Certificate documents the traceability to national standards, which realize the physical units of measurements(SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate.

All calibrations have been conducted in the closed laboratory facility: environment temperature(22±3)°C and humidity<70%.

Calibration Equipment used (M&TE critical for calibration)

| Primary Standards | ID# | Cal Date(Calibrated by, Certificate No.) | Scheduled Calibration |
|-------------------------|------------|--|-----------------------|
| Power Meter NRP2 | 101919 | 27-Jun-16 (CTTL, No.J16X04777) | Jun-17 |
| Power sensor NRP-Z91 | 101547 | 27-Jun-16 (CTTL, No.J16X04777) | Jun-17 |
| ReferenceProbe EX3DV4 | SN 7433 | 26-Sep-16(SPEAG,No.EX3-7433_Sep16) | Sep-17 |
| DAE4 | SN 771 | 02-Feb-16(CTTL-SPEAG,No.Z16-97011) | Feb-17 |
| | | | |
| Secondary Standards | ID# | Cal Date(Calibrated by, Certificate No.) | Scheduled Calibration |
| Signal Generator E4438C | MY49071430 | 01-Feb-16 (CTTL, No.J16X00893) | Jan-17 |
| NetworkAnalyzer E5071C | MY46110673 | 26-Jan-16 (CTTL, No.J16X00894) | Jan-17 |
| | | | |

Name

Function

Calibrated by:

Zhao Jing

SAR Test Engineer

Reviewed by:

Qi Dianyuan

SAR Project Leader

Approved by:

Lu Bingsong

Deputy Director of the laboratory

Issued: January 7,

This calibration certificate shall not be reproduced except in full without written approval of the laboratory

Certificate No: Z17-97001

Page 1 of 14