

DataBridge - type label sample and location




Figure 1: DataBridge backside incl. label positions

FCC ID: QRI-DATABRIDGE
 This device complies with Part 15 of the FCC Rules.
 Operation is subject to following two conditions: (1) this device may not cause harmful interference, and (2) this device must accept any interference received, including interference that may cause undesired operation.
 IC: 4708A-DATABRIDGE

BOTRONIK
 SE & Co. KG 12359 Berlin
 Woermannkehre 1 Germany

MD DataBridge
SN 627XXXXX
 yyyy-mm-dd
UDI (01)04035479176112(11)yymmdd(21)627XXXXX

IP30

Figure 2: DataBridge type label