

BIOTRONIK SE & Co. KG  
Woermannkehre 1  
12359 Berlin, Germany

Tel +49 (0) 30 68905-1291  
Fax +49 (0) 30 68905-1921  
dirk.koenig@biotronik.com  
www.biotronik.com

Berlin, February 27, 2018

## ICD-4200 SAR-Statement

To whom it may concern,

In FCC KDB 447498 General RF Exposure Guidance D01 v05r01, Section 4.2.4, the FCC provides a power threshold, below which implanted devices excluded from SAR measurement or numerical simulation:

"4.2.4 Transmitters implanted in the body of a user

When the aggregate of the maximum power available at the antenna port and radiating structures of an implanted transmitter, under all operating circumstances, is  $\leq 1.0$  mW, SAR test exclusion may be applied. The maximum available output power requirement and worst case operating conditions must be supported by power measurement results and fully justified in a SAR analysis report, in lieu of SAR measurement or numerical simulation, according to design and implementation requirements of the device."

Since both the radiated output power and conducted output power are well below 1.0 mW, this device is excluded from SAR measurement or numerical simulation.

FCC ID: QRI-ICD4200 is compliant with FCC RF Exposure requirements.

Thank you for your kind attention.

Sincerely yours,

**BIOTRONIK SE & Co. KG**



p.p. Dirk König  
Manager Regulatory Affairs