

BIOTRONIK SE & Co. KG · Postfach 470255 · 12311 Berlin

**Federal Communication Commission  
 Authorization and Evaluation Division  
 Equipment Authorization Branch  
 7435 Oakland Mills Road  
 Columbia, MD 21046**

Tel +49 (0) 30 68905-1291  
 Fax +49 (0) 30 68905-1921  
 dirk.koenig@biotronik.com  
 www.biotronik.com

Berlin, 25 March 2020

**Statement of re-use of original test results of approved radio modules  
 according to KDB 996369 D02 Module Q and A v01 Q1(a)**

Reference of new FCC ID application:

<b>Applicant: BIOTRONIK SE &amp; Co. KG</b>	
FCC ID:	QRI-RENAMICNEO

We, the undersigned, hereby declare that our radio device hosts the following approved radio module(s)

<b>Grantee: Telit Communications S.p.A.</b>			
<b>Type of Module: BlueMod+S50</b>			
FCC ID:	RFR-S50	Report Reference	1-5687/17-01-07-A

The module(s) is/are installed according to the installation instructions of the module manufacturer(s) without modification. The original module test report(s) accurately represent all conducted measured test results under the new conditions. All radiated measurements according to the applicable radio standard(s), the RF-exposure and simultaneous transmission evaluation are performed and reported for the entire host.

Sincerely,



Dirk König  
 Regulatory Affairs Senior Manager  
 BIOTRONIK SE & Co. KG