

Date: November 4, 2021

Element Material Technology
6775 NE Evergreen Parkway, Suite 400
Hillsboro, OR 97124, United States

Subject: Application for Certification of Biotronik Prospera Neuro SCS
External Trial Stimulator with FCC ID: QRI-SCSTS

Dear Sir or Madam:

BIOTRONIK SE & Co. KG, Woermannkehe 1, 12359 Berlin / Germany herein submits original application for Certification of Biotronik Prospera Neuro SCS External Trial Stimulator with Part 15.247 and Part 15.209 transmitters identified by FCC ID number QRI-SCSTS.

External Pulse Generator (EPG), aka Trial Stimulator(TS), which is worn externally by the patient during his/her trial period when used with a Trial Header and may be used to perform intraoperative testing of the leads when used with the Intraoperative Test Cable. The TS communicates with the Patient Remote (PR) for interaction with the patient and the Neuro Service Center. The TS communicates with the Clinician Programmer (CP) for communication with the clinician or sales representative.

Enclosed is a complete Certification Application packages. Please contact me if you require any additional information.

Sincerely,
Gunnar Boersch
Senior Team Manager Regulatory Affairs

BIOTRONIK SE & Co. KG

Woermannkehre 1

12359 Berlin / Germany

Email: Gunnar.Boersch@biotronik.com

Tel + (49)30-689051213

Fax + (49)30-68905961213