

BIOTRONIK SE & Co. KG · Postfach 470255 · 12311 Berlin

Federal Communications Commission  
Authorization and Evaluation Division  
Equipment Authorization Branch  
7435 Oakland Mills Road  
Columbia, MD 21046

BIOTRONIK SE & Co. KG  
Woermannkehre 1  
12359 Berlin, Germany

Tel +49 (0) 30 68905-1282  
Fax +49 (0) 30 68905-1921  
mark.briesemeister@biotronik.com  
www.biotronik.com

Berlin, September 16, 2014

### **BioMonitor 2 SAR Statement**


In FCC KDB 447498 General RF Exposure Guidance D01 v05r02, Section 4.2.4, the FCC provides a power threshold, below which implanted devices excluded from SAR measurement or numerical simulation:

"4.2.4 Transmitters implanted in the body of a user When the aggregate of the maximum power available at the antenna port and radiating structures of an implanted transmitter, under all operating circumstances, is  $\leq 1.0$  mW, SAR test exclusion may be applied. The maximum available output power requirement and worst case operating conditions must be supported by power measurement results and fully justified in a SAR analysis report, in lieu of SAR measurement or numerical simulation, according to design and implementation requirements of the device."

Since both the radiated output power and conducted output power are well below 1.0 mW, this device is excluded from SAR measurement or numerical simulation.

FCC ID: QRIBM2 is compliant with FCC RF Exposure requirements.

Sincerely,



Signature  
Mark Briesemeister  
Junior Manager Regulatory Affairs  
BIOTRONIK SE & CO. KG