

Statement of SAR identity between Tach35 and Tach50


The Lumax 300 (Tach35 4120 Substrate) and the Lumax 600/700 (Tach50 4140 substrate) are using the same mechanical platform. That means the inner assembly, the can, the header and the antenna are identical. Therefore all elements used by the SAR analysis are identical. The result of the Tach35 SAR report is also valid for the Tach50.



BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin

BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin

Berlin, 2011-05-18



Gunnar Gorecki
Diplom Ingenieur
Development Electronic Implants