

To: Federal Communications Commission
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin, Germany

Tel +49 (0) 30 68905-1291
Fax +49 (0) 30 68905-1921
dirk.koenig@biotronik.com
www.biotronik.com

Berlin, April 6, 2016

Statement to the FCC ID: QRIPNP

Labeling Scheme Per 47 CFR 2.926 (e)

"Where it is shown that a permanently affixed nameplate is not desirable or is not feasible, an alternative method of positively identifying the equipment may be used if approved by the commission. The proposed alternative method of identification and the justification for its use must be included with the application for equipment authorization. Note: As an example, a device intended to be implanted within the body of a test animal or person would probably require an alternate method of identification."

Since the DUT is implanted in the human body, the FCC ID will not be applied to outside of the device. Instead, the applicant is proposing the following alternate method of applying the FCC ID number to two locations:

1. The FCC ID, along with the disclosure statement required by 95.1215 (a), is in the user manual.
2. The FCC ID also appears on the package label which is affixed to outside of the shipping box for the Primus Nano Plus device.

Sincerely,

BIOTRONIK SE & Co. KG


Dirk König
Manager Regulatory Affairs