Inventra 7 DR-T

DF4

IS-1
DF4-LLHH

VDE-DDDR

Implantable Cardioverter Defibrillator

REF 399429 Cat. No.

SN 12345678 Serial No.

2015-08 Use By

41-113 °F Stor.Temp.

(01)04035479136246(17)150831(21)12345678

PID 89 Prod.ID

A	V	A	V
DDD-CLS; DDD(R); DDI(R); DDD-ADI(R); VDD(R); VDI(R); D00; **OFF**			
AAI(R)	VVI-CLS; VVI(R); V00		
Burst, Ramp	Burst		
HF burst	Ramp		
	Shock		

Manufacturer:
BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin - Germany
Tel +49 (0) 30 68905-0
Fax +49 (0) 30 6852804
sales@biotronik.com

Distributor:
BIOTRONIK, Inc.
6024 Jean Road
Lake Oswego, OR 97035-5369
(800) 547-0394 (24-hour)
(503) 635-9936 (Fax)

Caution: Federal law (U.S.A.) restricts this device to sale by, or on the order of, a physican (or properly licenced practitioner)

VDE-DDDR

RA IS-1
RV DF4

FCC ID: QRITACH70

OFF

STERILE EO

For single use only, do not resterilize. Refer to technical manual for instructions.
ICD is shipped with all therapy disabled.

A 30009

Inventra 7 DR-T DF4

REF 399429 Cat. No.

2015-08 Use By

SN 12345678 Serial No.

PID 89 Prod.ID

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T DF4 SN 12345678 PID 89

Inventra 7 DR-T

IS-1
DF4-LLHH

VDE-DDDR

DF4

Implantable Cardioverter Defibrillator

REF 399429 Cat. No.

2015-08 Use By

SN 12345678 Serial No.

PID 89 Prod.ID

STERILE EO
Sterilized with ethylene oxide

OFF

For single use only, do not resterilize.
Do not use if packaging is damaged.
Refer to technical manual for instructions.
ICD is shipped with all therapy disabled.

BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin - Germany

(01)04035479136246(17)150831(21)12345678

SN 12345678

SN 12345678

SN 12345678

SN 12345678

Inventra 7 DR-T

IS-1
DF4-LLHH

VDE-DDDR

Implantable Cardioverter Defibrillator

2015-08 Use By

PID 89 Prod.ID

REF 399429 Cat. No.

SN 12345678 Serial No.

DF4

Individual Data Sheet

Parameter	Values at Factory Release	Programmable Parameter Ranges
Bradycardia Therapy Mode	**OFF**	AAI(R); V00; VVI(R); VVI-CLS; VDI(R); VDD(R); D00; DDI(R); DDD(R); DDD-ADI(R); DDD-CLS; OFF
Basic Rate; min-1		30...(5)...100...(10)...160
Pulse Amplitude; V		0.5...(0.25)...4...(0.5)...6; 7.5
Pulse Width; ms		0.4; 0.5...(0.25)...1.5
A ATPs	**OFF**	OFF; BURST; RAMP; HF BURST
V ATPs	**OFF**	OFF, BURST, RAMP
Shock	**OFF**	OFF; DEFIBRILLATION with/without Confirmation
X-Ray Identification	**AH**	

A

For single use only. Refer to Technical Manual for Instructions.

Distributor:

BIOTRONIK, Inc.
6024 Jean Road
Lake Oswego, OR 97035-5369
(800) 547-0394 (24-hour)
(503) 635-9936 (Fax)

BIOTRONIK SE & Co. KG
Woermannkehre 1
12359 Berlin - Germany
Tel +49 (0) 30 68905-0
Fax +49 (0) 30 6852804
sales@biotronik.com