

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID





(01)04035479144623(17)70831(21)12345678



IS-1
IS-1
IS-1
DF-1
DF-1

Implantable Cardioverter-Defibrillator

Caution: Federal law (U.S.A.) restricts this device to sale by, or on the order of, a physician (or properly licenced practitioner)



FCC ID: QRTACHBORAX

417527

Distributor:
BIOTRONIK, Inc.
4024 Jean Road
Lake Oswego, OR 97035-5369
(800) 547-0394 (24-hour)
(503) 635-9936 (Fax)

Manufacturer:
BIOTRONIK SE & Co. KG
Weormannkehe 1
12359 Berlin - Germany
Tel. +49 (0) 30 68905-0
Fax +49 (0) 30 6852804
sales@biotronik.com

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

Use By 2017-08-31

BIOTRONIK

Inlexa 1 HF-T
SN 12345678
PID 89

BIOTRONIK

Inlexa 1 HF-T
SN 12345678
PID 89

BIOTRONIK

Inlexa 1 HF-T
SN 12345678
PID 89

BIOTRONIK


Inlexa 1 HF-T
SN 12345678
PID 89

BIOTRONIK

Inlexa 1 HF-T
SN 12345678
PID 89

BIOTRONIK

Inlexa 1 HF-T
SN 12345678
PID 89



Implantable Cardioverter Defibrillator

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

Use By 2017-08-31

STERILEEO Sterilized with ethylene oxide



For single use only, do not resterilize. Do not use if packaging is damaged. Refer to technical manual for instructions. ICD is shipped with all therapy disabled.

BIOTRONIK SE & Co. KG
Weormannkehe 1
12359 Berlin - Germany



(01)04035479144623(17)70831(21)12345678

SN 12345678
PID 89

SN 12345678
PID 89

SN 12345678
PID 89

SN 12345678
PID 89



AD13FF

Inlexa 1 HF-T
SN 12345678
PID 89

Inlexa 1 HF-T
SN 12345678
PID 89

Inlexa 1 HF-T
SN 12345678
PID 89

Inlexa 1 HF-T
SN 12345678
PID 89

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

IS-1
IS-1
IS-1
DF-1
DF-1

Implantable Cardioverter Defibrillator

Parameter	Values at Factory Release	Programmable Parameter Ranges
Bradycardia Therapy Mode	OFF	AAI(R); VOO; VVI(R); VDI(R); VDD(R); DDO; DDI(R); DDD(R); OFF 30...[5]...100...[10]...160 0.5...[0.25]...4...[0.5]...6; 7.5 0.4; 0.5...[0.25]...1.5 OFF; BURST; RAMP; SCAN OFF; DEFIBRILLATION with/without Confirmation
Basic Rate; min-1		
Pulse Amplitude; V		
Pulse Width; ms		
ATPs	OFF	
Shock	OFF	
X-Ray Identification	HK	

REF 405795 Cat.No. SN 12345678 Serial No. PID 89 Prod.ID

Use By 2017-08-31

For single use only. Refer to Technical Manual for Instructions.

Distributor:
BIOTRONIK, Inc.
6024 Jean Road
Lake Oswego, OR 97035-5369
(800) 547-0394 (24-hour)
(503) 635-9936 (Fax)

BIOTRONIK SE & Co. KG
Weormannkehe 1
12359 Berlin - Germany
Tel. +49 (0) 30 68905-0
Fax +49 (0) 30 6852804
sales@biotronik.com



417527