

## Appendix 2. FCC ID Label design and location

### Product Label Sample with FCC ID Label information

Following is a sample copy of the label that will be placed on the rear cabinet of the product.

The FCC identifier is marked in the product label.

The labelling requirement according to section 15. 19 and information to user according to section 15.21 and 15.105 are described in a prominent location of the user manual (page 5)

### Label Location

The label shown below shall be permanently affixed at a conspicuous location on the device and be readily visible to the user at the time purchase.(Labeling requirements per 2.925)

<b>ADVAN</b> <b>LCD Color Display Monitor</b> Model No. <u>AMM15TK2</u> With respect to electric shock, fire and mechanical hazards only in accordance with UL2601-1 <b>FCC ID QVXAMM15TK2</b> <b>DC,12V --- 3.5A</b>	
⚠ AC/DC Adapter Manufacturer : Ault Inc. Model : MW 160 ⚠ CAUTION - See accompanying documents. ⚠ CAUTION - ELECTRIC SHOCK "To avoid electrical shock, do not open the cabinet. Refer servicing to qualified personnel only."	
CLASSIFIED UL 511J Medical Equipment E215822	UL 2601-1 EN 60601-1 EN 60601-1-2 IEC 60950 <b>FCC CE</b> Tested to comply with FCC Standards For Home or Office Use.
<b>Serial No.</b> <u>                    </u>	<b>Manufactured</b> <u>                    </u>
ADVAN INTERNATIONAL CORPORATION TOLL Free 1-888-786-1688 www.advancorp.com MADE IN KOREA	

