



TEST REPORT

No.I19N02705-HAC RF

For

Yulong Computer Telecommunication Scientific (Shenzhen) Co., Ltd

Smartphone

Model Name: CP3706AS

With

Hardware Version: P1

Software Version: 3706AS.SPRINT.191220.2D

FCC ID: R38YLCP3706AS

Results Summary: M Category = M4

Issued Date: 2020-01-10

Designation Number: CN1210

Note:

The test results in this test report relate only to the devices specified in this report. This report shall not be reproduced except in full without the written approval of SAICT.

Test Laboratory:

Shenzhen Academy of Information and Communications Technology

Building G, Shenzhen International Innovation Center, No.1006 Shennan Road, Futian District, Shenzhen,

Guangdong, P. R. China 518026.

Tel: +86(0)755-33322000, Fax: +86(0)755-33322001

Email: yewu@caict.ac.cn, website: www.cszit.com

©Copyright. All rights reserved by SAICT.



REPORT HISTORY

| Report Number | Revision | Issue Date | Description |
|------------------|----------|------------|---------------------------------|
| I19N02705-HAC RF | Rev.0 | 2020-01-10 | Initial creation of test report |



TABLE OF CONTENT

| 1 SUMMARY OF TEST REPORT | 4 |
|---|---------|
| 1.1 TEST ITEMS 1.2 TEST STANDARDS 1.3 TEST RESULT 1.4 TESTING LOCATION 1.5 PROJECT DATA | |
| 2 CLIENT INFORMATION | 5 |
| 2.1 Applicant Information | |
| 3 EQUIPMENT UNDER TEST (EUT) AND ANCILLARY EQUIPMENT (AE) | 6 |
| 3.1 About EUT 3.2 INTERNAL IDENTIFICATION OF EUT USED DURING THE TEST 3.3 INTERNAL IDENTIFICATION OF AE USED DURING THE TEST 3.4 AIR INTERFACES / BANDS INDICATING OPERATING MODES | 6 6 |
| 4. REFERENCE DOCUMENTS | 7 |
| 5 OPERATIONAL CONDITIONS DURING TEST | 8 |
| 5.1 HAC MEASUREMENT SET-UP 5.2 PROBE SPECIFICATION 5.3 TEST ARCH PHANTOM & PHONE POSITIONER 4.4 ROBOTIC SYSTEM SPECIFICATIONS | 9 10 |
| 6 EUT ARRANGEMENT | 11 |
| 6.1 WD RF EMISSION MEASUREMENTS REFERENCE AND PLANE | 11 |
| 7 SYSTEM VALIDATION | 12 |
| 7.1 Validation Procedure 7.2 Validation Result | |
| 8 MODULATION INTERFERENCE FACTOR (MIF) | 13 |
| 9 EVALUATION FOR LOW-POWER EXEMPTION | 15 |
| 9.1 Product testing threshold 9.2 Conducted power | |
| 10 RF TEST PROCEDURES | 17 |
| 11 MEASUREMENT RESULTS (E-FIELD) | 18 |
| 12 ANSI C 63.19-2011 LIMITS | 19 |
| 13 MEASUREMENT UNCERTAINTY | 20 |
| 14 MAIN TEST INSTRUMENTS | 21 |
| ANNEX A RF EMISSION TEST PLOT | 22 |
| ANNEX B SYSTEM VALIDATION RESULT | 42 |
| ANNEX C PROBE CALIBRATION CERTIFICATE | 45 |
| ANNEX D DIPOLE CALIBRATION CERTIFICATE | 55 |
| ANNEX E UID SPECIFICATION | 70 |



No. I19N02705-HAC RF Page 4 of 91

1 Summary of Test Report

1.1 Test Items

| Description: |
|----------------------|
| Model Name: |
| Applicant's name: |
| Manufacturer's Name: |

Smartphone CP3706AS Yulong Computer Telecommunication Scientific (Shenzhen) Co., Ltd Yulong Computer Telecommunication Scientific (Shenzhen) Co., Ltd

1.2 Test Standards

ANSI C63.19-2011

1.3 Test Result

Please refer to "13. Measurement Results (E-Field)"

1.4 Testing Location

Address: Building G, Shenzhen International Innovation Center, No.1006 Shennan Road, Futian District, Shenzhen, Guangdong, P. R. China

1.5 Project Data

Testing Start Date: Testing End Date: November 24, 2019 November 24, 2019

1.6 Signature

李雨高

Li yongfu (Prepared this test report)

长过我

Zhang Yunzhuan (Reviewed this test report)

老派化

Cao Junfei Deputy Director of the laboratory (Approved this test report)

©Copyright. All rights reserved by SAICT.



2 Client Information

2.1 Applicant Information

| Company Name: | Yulong Computer Telecommunication Scientific (Shenzhen) Co., Ltd |
|----------------|---|
| Address /Post: | Building B, Boton Science Park, Chaguang Road, Xili Town, Nanshan |
| Address /Post. | District, Shenzhen |
| Contact: | Emily zhang |
| Email: | zhangxuzhu@yulong.com |
| Telephone: | +86 15089742056 |
| Fax: | / |

2.2 Manufacturer Information

| Company Name: | Yulong Computer Telecommunication Scientific (Shenzhen) Co., Ltd |
|------------------|---|
| Address /Post: | Building B, Boton Science Park, Chaguang Road, Xili Town, Nanshan |
| Audress / F Usi. | District, Shenzhen |
| Contact: | Emily zhang |
| Email: | zhangxuzhu@yulong.com |
| Telephone: | +86 15089742056 |
| Fax: | / |



3 Equipment under Test (EUT) and Ancillary Equipment (AE)

3.1 About EUT

| Description: | Smartphone |
|------------------------------|--|
| Mode Name: | CP3706AS |
| Operating mode(a): | GSM 850/1900, CDMA BC0/BC1/BC10, WCDMA Band 2/4/5 |
| Operating mode(s): | LTE Band 2/4/5/7/12/13/25/26/41/66/71, BT, Wi-Fi 2.4G/5G |
| Condition of EUT as received | No obvious damage in appearance |

3.2 Internal Identification of EUT used during the test

| EUT ID* | IMEI | HW Version | SW Version |
|---------|-----------------|------------|-------------------------|
| EUT1 | 990015570002444 | P1 | 3706AS.SPRINT.191220.2D |
| EUT2 | 990015570002378 | P1 | 3706AS.SPRINT.191220.2D |

*EUT ID: is used to identify the test sample in the lab internally.

Note: It is performed to test HAC with the EUT 1 & 2.

3.3 Internal Identification of AE used during the test

| AE ID* Description Type | | Туре | Manufacturer |
|-------------------------|---------|------------------|--------------|
| AE1 | Battery | 3706AS_CA406787G | CosMX |

*AE ID: is used to identify the test sample in the lab internally.

3.4 Air Interfaces / Bands Indicating Operating Modes

| Air-interface | Band(MU=) | Turne | C63.19/ | Simultaneous | Name of Voice | Power | |
|---------------|------------------|-------|-------------|---------------|---------------|-----------|--|
| Air-internace | Band(MHz) | Туре | tested | Transmissions | Service | Reduction | |
| GSM | 850 /1900 | VO | Yes | BT,WLAN | CMRS Voice | No | |
| EDGE | 850 /1900 | DT | Yes | BT,WLAN | Google Duo | INO | |
| WCDMA | B2 / B4/ B5 | VO | Yes | BT,WLAN | CMRS Voice | No | |
| VVCDIVIA | HSPA | DT | Yes | BT,WLAN | Google Duo | INO | |
| CDMA | BC0 / BC1 / BC10 | VO | Yes | BT,WLAN | CMRS Voice | No | |
| CDIMA | 1XRTT / EVDO | DT | Yes | BT,WLAN | Google Duo | | |
| | 2/4/5/7/12/13/ | VD | Yes | | VoLTE | | |
| LTE (FDD) | 25/26/66/71 | VD | res | BT,WLAN | Google Duo | No | |
| | 41 | VD | | VoLTE | INO | | |
| LTE (TDD) | 41 | VD | Yes BT,WLAN | | Google Duo | | |
| WLAN | 2.4G | VD | Yes | WWAN | Google Duo | No | |
| WLAN | 5G | VD | Yes | WWAN | Google Duo | No | |
| BT | 2.4G | DT | No | WWAN | NA | No | |

VO: Voice CMRS/PSTN Service Only

VD: Voice CMRS/PSTN and Data Service

DT: Digital Transport

* HAC Rating was not based on concurrent voice and data modes; Non-current mode was found to represent worst case rating for both M and T rating



4. Reference Documents

ANSI C63.19-2011: American National Standard for Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids

FCC KDB 285076 D01v05: Equipment Authorization Guidance for Hearing Aid Compatibility



5 Operational Conditions During Test

5.1 HAC Measurement Set-up

These measurements are performed using the DASY5 NEO automated dosimetric assessment system. It is made by Schmid & Partner Engineering AG (SPEAG) in Zurich, Switzerland. It consists of high precision robotics system (Stäubli), robot controller, Intel Core2 computer, near-field probe, probe alignment sensor. The robot is a six-axis industrial robot performing precise movements. A cell controller system contains the power supply, robot controller, teach pendant (Joystick), and remote control, is used to drive the robot motors. The PC consists of the HP Intel Core2 1.86 GHz computer with Windows XP system and HAC Measurement Software DASY5 NEO, A/D interface card, monitor, mouse, and keyboard. The Stäubli Robot is connected to the cell controller to allow software manipulation of the robot. A data acquisition electronic (DAE) circuit performs the signal amplification, signal multiplexing, AD-conversion, offset measurements, mechanical surface detection, collision detection, etc. is connected to the Electro-optical coupler (EOC). The EOC performs the conversion from the optical into digital electric signal of the DAE and transfers data to the PC plug-in card.




Fig. 1 HAC Test Measurement Set-up

The DAE4 consists of a highly sensitive electrometer-grade preamplifier with auto-zeroing, a channel and gain-switching multiplexer, a fast 16 bit AD-converter and a command decoder and control logic unit. Transmission to the PC-card is accomplished through an optical downlink for data and status information and an optical uplink for commands and clock lines. The mechanical probe mounting device includes two different sensor systems for frontal and sidewise probe contacts. They are also used for mechanical surface detection and probe collision detection. The robot uses its own controller with a built in VME-bus computer.



5.2 Probe Specification

E-Field Probe Description

| Construction | One dipole parallel, two dipoles normal to probe axis | |
|---------------|--|----------|
| | Built-in shielding against static charges | |
| | PEEK enclosure material | E |
| Calibration | In air from 100 MHz to 3.0 GHz (absolute accuracy \pm 6.0%, k=2) | |
| _ | | [ER3DV6] |
| Frequency | 40 MHz to > 6 GHz (can be extended to < 20 MHz) | |
| | Linearity: ± 0.2 dB (100 MHz to 3 GHz) | |
| | | |
| Directivity | \pm 0.2 dB in air (rotation around probe axis) | |
| | \pm 0.4 dB in air (rotation normal to probe axis) | |
| | | |
| Dynamic Range | 2 V/m to > 1000 V/m; Linearity: ± 0.2 dB | |
| | | |
| Dimensions | Overall length: 330 mm (Tip: 16 mm) | |
| | Tip diameter: 8 mm (Body: 12 mm) | |
| | Distance from probe tip to dipole centers: 2.5 mm | |
| | | |
| Application | General near-field measurements up to 6 GHz | |
| | Field component measurements | |
| | Fast automatic scanning in phantoms | |
| | U | |



5.3 Test Arch Phantom & Phone Positioner

The Test Arch phantom should be positioned horizontally on a stable surface. Reference markings on the Phantom allow the complete setup of all predefined phantom positions and measurement grids by manually teaching three points in the robot. It enables easy and well defined positioning of the phone and validation dipoles as well as simple teaching of the robot (Dimensions: 370 x 370 x 370 mm).

The Phone Positioner supports accurate and reliable positioning of any phone with effect on near field $<\pm 0.5$ dB.



Fig. 2 HAC Phantom & Device Holder

4.4 Robotic System Specifications

Specifications

Positioner: Stäubli Unimation Corp. Robot Model: RX160XL Repeatability: ±0.02 mm No. of Axis: 6 Data Acquisition Electronic (DAE) System Cell Controller Processor: Intel Core2 Clock Speed: 1.86 GHz Operating System: Windows XP Data Converter Features:Signal Amplifier, multiplexer, A/D converter, and control logic Software: DASY5 software Connecting Lines: Optical downlink for data and status info. Optical uplink for commands and clock



6 EUT Arrangement

6.1 WD RF Emission Measurements Reference and Plane

Figure 4 illustrates the references and reference plane that shall be used in the WD emissions measurement.

- The grid is 5 cm by 5 cm area that is divided into 9 evenly sized blocks or sub-grids.
- The grid is centered on the audio frequency output transducer of the WD (speaker or T-coil).

• The grid is located by reference to a reference plane. This reference plane is the planar area that contains the highest point in the area of the WD that normally rests against the user's ear

• The measurement plane is located parallel to the reference plane and 15 mm from it, out from the phone. The grid is located in the measurement plane.




Fig. 3 WD reference and plane for RF emission measurements



7 System Validation

7.1 Validation Procedure

Place a dipole antenna meeting the requirements given in ANSI C63.19 in the position normally occupied by the WD. The dipole antenna serves as a known source for an electrical output. Position the E-field probes so that:

• The probes and their cables are parallel to the coaxial feed of the dipole antenna

• The probe cables and the coaxial feed of the dipole antenna approach the measurement area from opposite directions

• The center point of the probe element(s) are 15 mm from the closest surface of the dipole elements.




Fig. 4 Dipole Validation Setup

7.2 Validation Result

| | E-Field Scan | | | | | | |
|------|--------------|-------------|-----------------------|---------------------|------------------------|--------|--|
| Mode | Frequency | Input Power | Measured ¹ | Target ² | Deviation ³ | Limit⁴ | |
| wode | (MHz) | (mW) | Value(dBV/m) | Value(dBV/m) | (%) | (%) | |
| CW | 835 | 100 | 43.62 | 40.72 | 7.1 | ±25 | |
| CW | 1880 | 100 | 39.95 | 39.06 | 2.3 | ±25 | |
| CW | 2600 | 100 | 39.68 | 38.71 | 2.5 | ±25 | |

Notes:

1. Please refer to the attachment for detailed measurement data and plot.

2. Target value is provided by SPEAD in the calibration certificate of specific dipoles.

3. Deviation (%) = 100 * (Measured value minus Target value) divided by Target value.

4. ANSI C63.19 requires values within \pm 25% are acceptable, of which 12% is deviation and 13% is measurement uncertainty. Values independently validated for the dipole actually used in the measurements should be used, when available.



8 Modulation Interference Factor (MIF)

The HAC Standard ANSI C63.19-2011 defines a new scaling using the Modulation Interference Factor (MIF) which replaces the need for the Articulation Weighting Factor (AWF) during the evaluation and is applicable to any modulation scheme.

The Modulation Interference factor (MIF, in dB) is added to the measured average E-field (in dBV/m) and converts it to the RF Audio Interference level (in dBV/m). This level considers the audible amplitude modulation components in the RF E-field. CW fields without amplitude modulation are assumed to not interfere with the hearing aid electronics. Modulations without time slots and low fluctuations at low frequencies have low MIF values, TDMA modulations with narrow transmission and repetition rates of few 100 Hz have high MIF values and give similar classifications as ANSI C63-2007.

Definitions

ER3D, E-field probes have a bandwidth <10 kHz and can therefore not evaluate the RF envelope in the full audio band. DASY52 is therefore using the "indirect" measurement method according to ANSI C63.19-2011 which is the primary method. These near field probes read the averaged E-field measurement. Especially for the new high peak-to-average (PAR) signal types, the probes shall be linearized by probe modulation response (PMR) calibration in order to not overestimate the field reading.

The evaluation method or the MIF is defined in ANSI C63.19-2011 section D.7. An RMS demodulated RF signal is fed to a spectral filter (similar to an A weighting filter) and forwarded to a temporal filter acting as a quasi-peak detector. The averaged output of these filtering is called to a 1 kHz 80% AM signal as reference. MIF measurement requires additional instrumentation and is not well suited for evaluation by the end user with reasonable uncertainty It may alternatively be determined through analysis and simulation, because it is constraint and characteristic for a communication signal. DASY52 uses well defined signals for PMR calibration. The MIF of these signals has been determined by simulation and is automatically applied.

MIF values were not tested by a probe or as specified in the standards but are based on analysis provided by SPEAG for all the air interfaces (GSM, WCDMA, CDMA, LTE). The data included in this report are for the worst case operating modes. The UIDs used are listed below:



| UID | Communication System Name | MIF (dB) |
|-------|---|----------|
| 10021 | GSM-FDD (TDMA, GMSK) | 3.63 |
| 10025 | EDGE-FDD (TDMA, 8PSK, TN 0) | 3.75 |
| 10011 | UMTS-FDD (WCDMA) | -27.23 |
| 10225 | UMTS-FDD (HSPA+) | -20.39 |
| 10295 | CDMA2000 (RC1, SO3, 1/8th Rate 25 fr.) | 3.26 |
| 10403 | CDMA2000 (1xEV-DO, Rev. 0) | -17.67 |
| 10170 | LTE-FDD(SC-FDMA, 1RB, 20MHz, 16-QAM) | -9.76 |
| 10176 | LTE-FDD(SC-FDMA, 1RB, 10MHz, 16-QAM) | -9.76 |
| 10173 | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16QAM) | -1.44 |
| 10061 | IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps) | -2.02 |
| 10069 | IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps) | -3.15 |

A PMR calibrated probe is linearized for the selected waveform over the full dynamic range within the uncertainty specified in its calibration certificate. ER3D, EF3D and EU2D E-field probes have a bandwidth <10 kHz and can therefore not evaluate the RF envelope in the full audio band. DASY52 is therefore using the \indirect" measurement method according to ANSI C63.19-2011 which is the primary method. These near field probes read the averaged E-field measurement. Especially for the new high peak-to-average (PAR) signal types, the probes shall be linearized by PMR calibration in order to not overestimate the field reading.

The MIF measurement uncertainty is estimated as follows, for modulation frequencies from slotted waveforms with fundamental frequency and at least 2 harmonics within 10 kHz:

0.2 dB for MIF -7 to +5 dB, 0.5 dB for MIF -13 to +11 dB 1 dB for MIF > -20 dB



9 Evaluation for low-power exemption

9.1 Product testing threshold

There are two methods for exempting an RF air interface technology from testing. The first method requires evaluation of the MIF for the worst-case operating mode. An RF air interface technology of a device is exempt from testing when its average antenna input power plus its MIF is \leq 17 dBm for any of its operating modes. The second method does not require determination of the MIF. The RF emissions testing exemption shall be applied to an RF air interface technology in a device whose peak antenna input power, averaged over intervals \leq 50 µs20, is \leq 23 dBm. An RF air interface technology that is exempted from testing by either method shall be rated as M4. The first method is used to be exempt from testing for the RF air interface technology in this report.



9.2 Conducted power

| Band | power (dBm) | MIF (dB) | Sum (dBm) | HAC Test |
|-----------------|-------------|----------|-----------|----------|
| GSM 850 | 33.5 | 3.63 | 37.13 | Yes |
| EGPRS 850 | 28 | 3.75 | 31.75 | Yes |
| GSM 1900 | 31 | 3.63 | 34.63 | Yes |
| EGPRS 1900 | 27 | 3.75 | 30.75 | Yes |
| CDMA BC0 | 25 | 3.26 | 28.26 | Yes |
| CDMA BC0 -EVDO | 25 | -17.67 | 7.33 | No |
| CDMA BC1 | 25 | 3.26 | 28.26 | Yes |
| CDMA BC1 -EVDO | 25 | -17.67 | 7.33 | No |
| CDMA BC10 | 25 | 3.26 | 28.26 | Yes |
| CDMA BC10 -EVDO | 25 | -17.67 | 7.33 | No |
| WCDMA B2 | 25.5 | -27.23 | -1.73 | No |
| WCDMA B2 -HSPA | 25 | -20.39 | 4.61 | No |
| WCDMA B4 | 25.5 | -27.23 | -1.73 | No |
| WCDMA B4 -HSPA | 24.5 | -20.39 | 4.11 | No |
| WCDMA B5 | 25.5 | -27.23 | -1.73 | No |
| WCDMA B5 -HSPA | 24.5 | -20.39 | 4.11 | No |
| LTE Band 2 | 24.5 | -9.76 | 14.74 | No |
| LTE Band 4 | 24 | -9.76 | 14.24 | No |
| LTE Band 5 | 24 | -9.76 | 14.24 | No |
| LTE Band 7 | 24.5 | -9.76 | 14.74 | No |
| LTE Band 12 | 24 | -9.76 | 14.24 | No |
| LTE Band 13 | 24 | -9.76 | 14.24 | No |
| LTE Band 25 | 24.5 | -9.76 | 14.74 | No |
| LTE Band 26 | 25.5 | -9.76 | 15.74 | No |
| LTE Band 41 | 24.5 | -1.44 | 23.06 | Yes |
| LTE Band 66 | 24 | -9.76 | 14.24 | No |
| LTE Band 71 | 24 | -9.76 | 14.24 | No |
| WIFI 2.4G | 17.5 | -2.02 | 15.48 | No |
| WIFI 5G | 15.5 | -3.15 | 12.35 | No |

Note:

1. Power = Max turn-up limit.

2. EGPRS data modes are not necessary due the GSM Voice mode is the worst case.



10 RF Test Procedures

The evaluation was performed with the following procedure:

- 1) Confirm proper operation of the field probe, probe measurement system and other instrumentation and the positioning system.
- 2) Position the WD in its intended test position. The gauge block can simplify this positioning.
- 3) Configure the WD normal operation for maximum rated RF output power, at the desired channel and other operating parameters (e.g., test mode), as intended for the test.
- 4) The center sub-grid shall centered on the center of the T-Coil mode axial measurement point or the acoustic output, as appropriate. Locate the field probe at the initial test position in the 50 mm by 50 mm grid, which is contained in the measurement plane. If the field alignment method is used, align the probe for maximum field reception.
- 5) Record the reading.
- 6) Scan the entire 50 mm by 50 mm region in equally spaced increments and record the reading at each measurement point. The distance between measurement points shall be sufficient to assure the identification of the maximum reading.
- 7) Identify the five contiguous sub-grids around the center sub-grid whose maximum reading is the lowest of all available choices. This eliminates the three sub-grids with the maximum readings. Thus, the six areas to be used to determine the WD's highest emissions are identified.
- 8) Identify the maximum field reading within the non-excluded sub-grids identified in Step 7)
- 9) Evaluate the MIF and add to the maximum steady-state rms field-strength reading to obtain the RF audio interference level..
- 10) Compare this RF audio interference level with the categories and record the resulting WD category rating.



11 Measurement Results (E-Field)

| Freq | uency | Measured Value | Power Drift | 0-1 | | | |
|---------|-------------|----------------|-------------|--------------------------|--|--|--|
| MHz | Channel | (dBV/m) | (dB) | Category | | | |
| | GSM 850 | | | | | | |
| 848.8 | 251 | 26.62 | 0.03 | M4 (see Fig A.1) | | | |
| 836.6 | 190 | 29.28 | 0.04 | M4 (see Fig A.2) | | | |
| 824.2 | 128 | 29.40 | 0.01 | M4 (see Fig A.3) | | | |
| | | GSM 19 | 00 | | | | |
| 1909.8 | 810 | 20.76 | 0.01 | M4 (see Fig A.4) | | | |
| 1880 | 661 | 21.08 | 0.10 | M4 (see Fig A.5) | | | |
| 1850.2 | 512 | 21.40 | 0.04 | M4 (see Fig A.6) | | | |
| | | CDMA B | C0 | | | | |
| 848.31 | 777 | 35.13 | -0.03 | M4 (see Fig A.7) | | | |
| 836.52 | 384 | 35.26 | -0.01 | M4 (see Fig A.8) | | | |
| 824.70 | 1013 | 34.77 | -0.02 | M4 (see Fig A.9) | | | |
| | | CDMA B | C1 | | | | |
| 1908.75 | 1175 | 28.00 | 0.05 | M4 (see Fig A.10) | | | |
| 1880 | 600 | 27.72 | 0.08 | M4 (see Fig A.11) | | | |
| 1851.25 | 25 | 27.48 | 0.06 | M4 (see Fig A.12) | | | |
| | | CDMA BO | 210 | | | | |
| 823.1 | 684 | 34.63 | -0.01 | M4 (see Fig A.13) | | | |
| 820.5 | 580 | 34.74 | -0.08 | M4 (see Fig A.14) | | | |
| 817.9 | 476 | 34.96 | -0.07 | M4 (see Fig A.15) | | | |
| | LTE Band 41 | | | | | | |
| 2680 | 41490 | 18.50 | -0.05 | M4 (see Fig A.16) | | | |
| 2636.5 | 41055 | 17.55 | -0.02 | M4 (see Fig A.17) | | | |
| 2593 | 40620 | 17.99 | -0.12 | M4 (see Fig A.18) | | | |
| 2549.5 | 40185 | 19.27 | -0.07 | M4 (see Fig A.19) | | | |
| 2506 | 39750 | 19.09 | -0.07 | M4 (see Fig A.20) | | | |



12 ANSI C 63.19-2011 Limits

WD RF audio interference level categories in logarithmic units

| Emission categories | < 960 MHz | | | |
|---------------------|-------------------|----------|--|--|
| | E-field emissions | | | |
| Category M1 | 50 to 55 | dB (V/m) | | |
| Category M2 | 45 to 50 | dB (V/m) | | |
| Category M3 | 40 to 45 | dB (V/m) | | |
| Category M4 | < 40 | dB (V/m) | | |
| Emission categories | > 960 | MHz | | |
| | E-field er | missions | | |
| Category M1 | 40 to 45 | dB (V/m) | | |
| Category M2 | 35 to 40 | dB (V/m) | | |
| Category M3 | 30 to 35 | dB (V/m) | | |
| Category M4 | < 30 | dB (V/m) | | |



No. I19N02705-HAC RF Page 20 of 91

13 Measurement Uncertainty

| No. | Error source | Туре | Uncert ainty Value (%) | Prob. Dist. | k | C _i E | Standard Uncertainty (%) u_i^{+} (%) E | Degree of freedom V _{eff} or v _i | source |
|------|--|-----------------------|---------------------------------|----------------|------------|---------------------|---|---|--------------|
| 1 | System repeatability | А | 0.24 | Ν | 1 | 1 | 0.24 | 9 | Measurement |
| Meas | surement System | | | | | | | | |
| 2 | Probe Calibration | В | 10.1 | N | 1 | 1 | 10.1 | ∞ | Manufacturer |
| 3 | Axial Isotropy | В | 0.5 | R | $\sqrt{3}$ | 1 | 0.5 | ∞ | Cal report |
| 4 | Sensor Displacement | В | 16.5 | R | $\sqrt{3}$ | 1 | 9.5 | ∞ | Manufacturer |
| 5 | Boundary Effects | В | 2.4 | R | $\sqrt{3}$ | 1 | 1.4 | ∞ | Manufacturer |
| 6 | Linearity | В | 0.6 | R | $\sqrt{3}$ | 1 | 0.35 | ∞ | Cal report |
| 7 | Scaling to Peak Envolope Power | В | 2.0 | R | $\sqrt{3}$ | 1 | 1.2 | ∞ | Standard |
| 8 | System Detection Limit | В | 1.0 | R | $\sqrt{3}$ | 1 | 0.6 | ∞ | Manufacturer |
| 9 | Readout Electronics | В | 0.3 | N | 1 | 1 | 0.3 | ∞ | Manufacturer |
| 10 | Response Time | В | 0.8 | R | $\sqrt{3}$ | 1 | 0.5 | ∞ | Manufacturer |
| 11 | Integration Time | В | 2.6 | R | $\sqrt{3}$ | 1 | 1.5 | ∞ | Manufacturer |
| 12 | RF Ambient Conditions | В | 3.0 | R | $\sqrt{3}$ | 1 | 1.7 | ∞ | Measurement |
| 13 | RF Reflections | В | 12.0 | R | $\sqrt{3}$ | 1 | 6.9 | ∞ | Measurement |
| 14 | Probe Positioner | А | 1.2 | R | $\sqrt{3}$ | 1 | 0.7 | 8 | Manufacturer |
| 15 | Probe Positioning | А | 4.7 | R | $\sqrt{3}$ | 1 | 2.7 | 8 | Manufacturer |
| 16 | Extra. And Interpolation | В | 1.0 | R | $\sqrt{3}$ | 1 | 0.6 | 8 | Manufacturer |
| Test | Sample Related | | | | | | | | |
| 17 | Device Positioning Vertical | В | 4.7 | R | $\sqrt{3}$ | 1 | 2.7 | 8 | Manufacturer |
| 18 | Device Positioning Lateral | В | 1.0 | R | $\sqrt{3}$ | 1 | 0.6 | ∞ | Manufacturer |
| 19 | Device Holder and Phantom | В | 2.4 | R | $\sqrt{3}$ | 1 | 1.4 | 8 | Manufacturer |
| 20 | Power Drift | В | 5.0 | R | $\sqrt{3}$ | 1 | 2.9 | 8 | Measurement |
| Phar | ntom and Setup related | | | | | | | | |
| 21 | Phantom Thickness | В | 2.4 | R | $\sqrt{3}$ | 1 | 1.4 | ∞ | Manufacturer |
| PMF | PMF related | | | | | | | | |
| 22 | Monitor amplitude | В | 3.5 | R | $\sqrt{3}$ | 1 | 2.02 | 8 | Manufacturer |
| 23 | Setup repeatability | А | 2.3 | Ν | 1 | 1 | 2.3 | 9 | Manufacturer |
| 24 | Sensor amplitude | В | 12 | R | $\sqrt{3}$ | 1 | 6.93 | 8 | Manufacturer |
| | Combined standard uncertaint | y(%) | | | | | 18.3 | | |
| | Expanded uncertainty (confidence interval of 95 %) | <i>u</i> _e | $=2u_c$ | Ν | k= | =2 | 36.6 | | |



14 Main Test Instruments

| No. | Name | Туре | Serial Number | Calibration Date | Valid Period |
|-----|---------------------|----------|---------------|------------------|-----------------|
| 01 | Signal Generator | E8257D | MY47461211 | 2019-06-03 | One year |
| 02 | Power meter | E4418B | MY50000366 | 2019 12 14 | Oneveer |
| 03 | Power sensor | E9304A | MY50000188 | 2018-12-14 | One year |
| 04 | Amplifier | VTL5400 | 0404 | / | |
| 05 | HAC Test Arch | N/A | 1150 | / | |
| 06 | DAE | DAE4 | 1527 | 2019-11-11 | One year |
| 07 | E-Field Probe | ER3DV6 | 2424 | 2018-02-23 | Three year |
| 08 | HAC Dipole | CD835V3 | 1165 | 2018-07-19 | Three year |
| 09 | HAC Dipole | CD1880V3 | 1149 | 2018-07-19 | Three year |
| 10 | HAC Dipole | CD2600V3 | 1020 | 2018-10-23 | Three year |
| 11 | BTS | CMU200 | 114544 | 2019-09-02 | One year |
| 12 | BTS | CMU500 | 152499 | 2019-07-18 | One year |

Table 14-1: List of Main Instruments



ANNEX A RF Emission Test Plot

HAC RF E-Field GSM 850 High

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, GSM Frequency: 848.8 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm


Reference Value = 16.71 V/m; Power Drift = 0.03 dB

Applied MIF = 3.63 dB

RF audio interference level = 26.62 dBV/m

Emission category: M4

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 24.65 dBV/m | 25.94 dBV/m | 25.66 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 25.46 dBV/m | 26.62 dBV/m | 26.52 dBV/m |
| | | |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |








HAC RF E-Field GSM 850 Middle

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, GSM Frequency: 836.6 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 21.44 V/m; Power Drift = 0.04 dB Applied MIF = 3.63 dB RF audio interference level = 29.28 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 27.48 dBV/m | 27.68 dBV/m | 26.76 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 29.13 dBV/m | 29.28 dBV/m | 28.68 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 30.67 dBV/m | 30.68 dBV/m | 29.78 dBV/m |








HAC RF E-Field GSM 850 Low

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, GSM Frequency: 824.2 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 21.91 V/m; Power Drift = 0.01 dB Applied MIF = 3.63 dB RF audio interference level = 29.40 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 27.71 dBV/m | 27.82 dBV/m | 26.94 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 29.23 dBV/m | 29.4 dBV/m | 28.83 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 30.76 dBV/m | 30.76 dBV/m | 29.93 dBV/m |







HAC RF E-Field GSM 1900 High


Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, GSM Frequency: 1910 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 7.682 V/m; Power Drift = 0.01 dB Applied MIF = 3.63 dB RF audio interference level = 20.76 dBV/m **Emission category: M4**

MIF scaled E-field

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 19.65 dBV/m | 20.86 dBV/m | 20.15 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 19.52 dBV/m | 20.76 dBV/m | 20.45 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 18.11 dBV/m | 20.19 dBV/m | 20.12 dBV/m |





©Copyright. All rights reserved by SAICT.




HAC RF E-Field GSM 1900 Middle

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, GSM Frequency: 1880 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 8.313 V/m; Power Drift = 0.10 dB Applied MIF = 3.63 dB RF audio interference level = 21.08 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 19.78 dBV/m | 21.06 dBV/m | 20.04 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 19.77 dBV/m | 21.08 dBV/m | 21 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 19.45 dBV/m | 20.93 dBV/m | 20.81 dBV/m |








HAC RF E-Field GSM 1900 Low

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, GSM Frequency: 1850.2 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 8.673 V/m; Power Drift = 0.04 dB Applied MIF = 3.63 dB RF audio interference level = 21.40 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 20.03 dBV/m | 21.09 dBV/m | 20.1 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 20.02 dBV/m | 21.4 dBV/m | 21.33 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 20.34 dBV/m | 21.51 dBV/m | 21.4 dBV/m |








HAC RF E-Field CDMA BC0 High

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 848.31 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device H/Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 48.44 V/m; Power Drift = -0.03 dB Applied MIF = 3.26 dB RF audio interference level = 35.13 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 33.43 dBV/m | 34.83 dBV/m | 34.78 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 33.9 dBV/m | 35.13 dBV/m | 35.09 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 34.62 dBV/m | 35.13 dBV/m | 35.08 dBV/m |







HAC RF E-Field CDMA BC0 Middle

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 836.52 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 48.98 V/m; Power Drift = -0.01 dB Applied MIF = 3.26 dB RF audio interference level = 35.26 dBV/m **Emission category: M4**

MIF scaled E-field

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 33.64 dBV/m | 34.96 dBV/m | 34.9 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 34.08 dBV/m | 35.26 dBV/m | 35.23 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 34.76 dBV/m | 35.28 dBV/m | 35.21 dBV/m |




Fig A.8 HAC RF E-Field CDMA BC0




HAC RF E-Field CDMA BC0 Low

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 824.7 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 46.35 V/m; Power Drift = -0.02 dB Applied MIF = 3.26 dB RF audio interference level = 34.77 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 33.22 dBV/m | 34.48 dBV/m | 34.41 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 33.6 dBV/m | 34.77 dBV/m | 34.71 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 34.25 dBV/m | 34.78 dBV/m | 34.7 dBV/m |








HAC RF E-Field CDMA BC1 High

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 1908.75 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 13.23 V/m; Power Drift = 0.05 dB Applied MIF = 3.26 dB RF audio interference level = 28.00 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M3 | Grid 3 M3 |
|------------------|------------------|------------------|
| 28.6 dBV/m | 30.97 dBV/m | 30.92 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 24.78 dBV/m | 27.85 dBV/m | 28 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 27.64 dBV/m | 27.61 dBV/m | 26.32 dBV/m |








HAC RF E-Field CDMA BC1 Middle

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 1880 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 11.20 V/m; Power Drift = 0.08 dB Applied MIF = 3.26 dB RF audio interference level = 27.72 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M3 | Grid 3 M3 |
|------------------|------------------|------------------|
| 28.63 dBV/m | 30.39 dBV/m | 30.32 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 24.13 dBV/m | 26.49 dBV/m | 26.56 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 27.54 dBV/m | 27.72 dBV/m | 26.85 dBV/m |








HAC RF E-Field CDMA BC1 Low

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 1851.25 MHz Duty Cycle: 1:8.3 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 9.971 V/m; Power Drift = 0.06 dB Applied MIF = 3.26 dB RF audio interference level = 27.48 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M3 | Grid 3 M3 |
|------------------|------------------|------------------|
| 29.02 dBV/m | 30.64 dBV/m | 30.59 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 24.33 dBV/m | 26.59 dBV/m | 26.82 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 27.48 dBV/m | 27.47 dBV/m | 25.95 dBV/m |








HAC RF E-Field CDMA BC10 High

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 823.1 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 46.29 V/m; Power Drift = -0.01 dB Applied MIF = 3.26 dB RF audio interference level = 34.63 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 33.16 dBV/m | 34.39 dBV/m | 34.31 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 33.56 dBV/m | 34.63 dBV/m | 34.54 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 34.27 dBV/m | 34.62 dBV/m | 34.51 dBV/m |








HAC RF E-Field CDMA BC10 Middle

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 820.5 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 46.71 V/m; Power Drift = -0.08 dB Applied MIF = 3.26 dB RF audio interference level = 34.74 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 33.26 dBV/m | 34.5 dBV/m | 34.43 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 33.65 dBV/m | 34.74 dBV/m | 34.66 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 34.35 dBV/m | 34.72 dBV/m | 34.62 dBV/m |








HAC RF E-Field CDMA BC10 Low

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, CDMA Frequency: 817.9 MHz Duty Cycle: 1:1 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 48.01 V/m; Power Drift = -0.07 dB Applied MIF = 3.26 dB RF audio interference level = 34.96 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 33.52 dBV/m | 34.75 dBV/m | 34.65 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 33.87 dBV/m | 34.96 dBV/m | 34.85 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 34.55 dBV/m | 34.93 dBV/m | 34.8 dBV/m |








HAC RF E-Field LTE-Band 41 High

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, LTE_TDD Frequency: 2680 MHz Duty Cycle: 1:1.58 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 6.643 V/m; Power Drift = -0.05 dB Applied MIF = -1.44 dB RF audio interference level = 18.50 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 17.86 dBV/m | 20.32 dBV/m | 20.57 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 15.32 dBV/m | 16.46 dBV/m | 17.83 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 17.54 dBV/m | 18.5 dBV/m | 18.14 dBV/m |








HAC RF E-Field LTE-Band 41 High-2

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, LTE_TDD Frequency: 2636.5 MHz Duty Cycle: 1:1.58 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 6.091 V/m; Power Drift = -0.02 dB Applied MIF = -1.44 dB RF audio interference level = 17.55 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 17.87 dBV/m | 19.85 dBV/m | 20.03 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 15.59 dBV/m | 16.14 dBV/m | 17.17 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 16.95 dBV/m | 17.55 dBV/m | 16.89 dBV/m |







HAC RF E-Field LTE-Band 41 Middle


Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, LTE_TDD Frequency: 2593 MHz Duty Cycle: 1:1.58 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 5.350 V/m; Power Drift = -0.12 dB Applied MIF = -1.44 dB RF audio interference level = 17.99 dBV/m **Emission category: M4**

MIF scaled E-field

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 16.92 dBV/m | 20.02 dBV/m | 20.12 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 13.94 dBV/m | 16.27 dBV/m | 17.1 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 16.66 dBV/m | 17.99 dBV/m | 17.49 dBV/m |







HAC RF E-Field LTE-Band 41 Low-2


Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, LTE_TDD Frequency: 2549.5 MHz Duty Cycle: 1:1.58 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 4.649 V/m; Power Drift = -0.07 dB Applied MIF = -1.44 dB RF audio interference level = 19.27 dBV/m

E-field without scaling

| Grid 1 | Grid 2 | Grid 3 |
|-------------|-------------|-------------|
| 16 dBV/m | 19.86 dBV/m | 19.89 dBV/m |
| Grid 4 | Grid 5 | Grid 6 |
| 13.35 dBV/m | 15.39 dBV/m | 16.08 dBV/m |
| Grid 7 | Grid 8 | Grid 9 |
| 17.9 dBV/m | 19.27 dBV/m | 18.86 dBV/m |








HAC RF E-Field LTE-Band 41 Low

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ S/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Ambient Temperature: 22.0°C Liquid Temperature: 21.5°C Communication System: UID 0, LTE_TDD Frequency: 2506 MHz Duty Cycle: 1:1.58 Probe: ER3DV6 - SN2424 ConvF (1, 1, 1);

E Scan - ER3DV6 - 2011: 15 mm from Probe Center to the Device /Hearing Aid Compatibility

Test (101x101x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 5.432 V/m; Power Drift = -0.07 dB Applied MIF = -1.44 dB RF audio interference level = 19.09 dBV/m **Emission category: M4**

| Grid 1 M4 | Grid 2 M4 | Grid 3 M4 |
|------------------|------------------|------------------|
| 16.05 dBV/m | 19.03 dBV/m | 19.09 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 14.42 dBV/m | 16.17 dBV/m | 16.13 dBV/m |
| Grid 7 M4 | Grid 8 M4 | Grid 9 M4 |
| 18.42 dBV/m | 19.37 dBV/m | 18.97 dBV/m |







ANNEX B System Validation Result


835 MHz

Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: σ = 0 mho/m, ϵ r = 1; ρ = 1000 kg/m3 Communication System: CW; Frequency: 835 MHz; Duty Cycle: 1:1 Probe: ER3DV6 - SN2424; ConvF (1, 1, 1)

E Scan - measurement distance from the probe sensor center to CD835 Dipole = 15mm /Hearing Aid Compatibility Test (41x361x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm Reference Value = 111.5 V/m; Power Drift = 0.03 dB Applied MIF = 0.00 dB RF audio interference level = 43.62 dBV/m **Emission category: M3**

| Grid 1 M3 | Grid 2 M3 | Grid 3 M3 |
|------------------|------------------|------------------|
| 42.95 dBV/m | 43.52 dBV/m | 43.38 dBV/m |
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 38.46 dBV/m | 38.81 dBV/m | 38.76 dBV/m |
| Grid 7 M3 | Grid 8 M3 | Grid 9 M3 |
| 42.12 dBV/m | 43.62 dBV/m | 43.44 dBV/m |



0 dB = 43.62 dBV/m




1880 MHz Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: $\sigma = 0$ mho/m, $\varepsilon_r = 1$; $\rho = 1000$ kg/m³ Communication System: CW; Frequency: 1880 MHz; Duty Cycle: 1:1 Probe: ER3DV6 - SN2424; ConvF (1, 1, 1)

E Scan - measurement distance from the probe sensor center to CD1880 Dipole = 15mm /Hearing Aid Compatibility Test (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm Reference Value = 101.2 V/m; Power Drift = 0.08 dB Applied MIF = 0.00 dB RF audio interference level = 39.95 dBV/m

Emission category: M2

| Grid 1 M2 | Grid 2 M2 | Grid 3 M2 |
|------------------|------------------|------------------|
| 39.50 dBV/m | 39.96 dBV/m | 39.88 dBV/m |
| Grid 4 M2 | Grid 5 M2 | Grid 6 M2 |
| 37.72 dBV/m | 38.05 dBV/m | 38.01 dBV/m |
| Grid 7 M2 | Grid 8 M2 | Grid 9 M2 |
| 39.46 dBV/m | 39.92 dB V/m | 39.85 dBV/m |



0 dB = 39.95 dBV/m



2600 MHz


Date: 2019-11-24 Electronics: DAE4 Sn1527 Medium: Air Medium parameters used: σ = 0 mho/m, ϵ_r = 1; ρ = 1000 kg/m³ Communication System: CW; Frequency: 1880 MHz; Duty Cycle: 1:1 Probe: ER3DV6 - SN2424; ConvF (1, 1, 1)

E Scan - measurement distance from the probe sensor center to CD2600 Dipole = 15mm /Hearing Aid Compatibility Test (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm Reference Value = 98.84 V/m; Power Drift = 0.10 dB Applied MIF = 0.00 dB RF audio interference level = 39.68 dBV/m Emission category: M2

MIF scaled E-field

| Grid 1 M2 | Grid 2 M2 | Grid 3 M2 |
|------------------|------------------|------------------|
| 39.22 dBV/m | 39.48 dBV/m | 39.44 dBV/m |
| Grid 4 M2 | Grid 5 M2 | Grid 6 M2 |
| 38.88 dBV/m | 39.05 dBV/m | 39.00 dBV/m |
| Grid 7 M2 | Grid 8 M2 | Grid 9 M2 |
| 39.52 dBV/m | 39.68 dB V/m | 39.59 dBV/m |



0 dB = 39.68 dBV/m



ANNEX C Probe Calibration Certificate

E_Probe ER3DV6

| Accredited by the Swiss Accredit | tation Ramine (0410) | | creditation No.: SCS 0108 |
|--|---|--|---|
| The Swiss Accreditation Servi | ce is one of the signatories | to the EA | creditation No.: SCS 0108 |
| Multilateral Agreement for the | recognition of calibration of | ortificates | |
| Client CTTL-SZ (Aut | | | ER3-2424 Feb18 |
| erre or trut | actif | Certificate No. | CR3-2424_P0D10 |
| CALIBRATION | CEDTIEICATE | | |
| CALIBITATION | GERTIFICATE | | |
| Object | ER3DV6 - SN:242 | 4 | |
| - class | ERSDV0 - SIN:242 | 19 | |
| | | | |
| Califration procedure(s) | QA CAL-02.vB, QA | | |
| | | lure for E-field probes optimized t | for close near field |
| | evaluations in air | | |
| | | | |
| Calibration date: | February 23, 2018 | 1 | |
| | | | |
| This calibration certificate docu | ments the traceability to nation | sal standards, which realize the physical units | of measurements (SI) |
| The measurements and the unit | certainties with confidence pro | bability are given on the following pages and | are part of the certificate |
| | | | |
| | | | |
| | | | |
| | | facility: environment temperature (22 ± 3)°C (| |
| | | | |
| | ucted in the closed laboratory | | |
| All calibrations have been cond | ucted in the closed laboratory | | |
| All calibrations have been cond Calibration Equipment used (M | ucted in the closed laboratory STE critical for calibration) | facility: environment temperature (22 \pm 3)°C : | and humidity < 70%. |
| All calibrations have been cond | ucted in the closed laboratory | facility: environment temperature (22 ± 3)°C : Cai Date (Certificate No.) | and humidity < 70%. Scheduled Calibration |
| All calibrations have been cond Calibration Equipment used (M Primary Standards | ucted in the closed laboratory STE critical for calibration) | facility: environment temperature (22 ± 3)°C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) | and humidity < 70%. Scheduled Calibration Apr-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP | ID SN: 104778 | facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02521) | and humidey < 70%. Scheduled Calibration Apr.18 Apr.18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 | ATE critical for calibration) | facility: environment temperature (22 ± 3)°C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) | and humidity < 70%. Scheduled Calibration Apr-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Prote ER3DV6 | ID SN: 104778 SN: 103244 SN: 103245 | facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-0252) | Scheduled Calibration Apr-18 Apr-18 Apr-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator | ID SN 104778 SN 103244 SN 103245 SN 103245 SN 55277 (204) | facility: environment temperature (22 ± 3)°C Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02525) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02528) | Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 | ID SN: 104778 SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 55277 (20k) SN: 2328 SN: 2328 SN: 789 | facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) D4-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 2-Aug-17 (No. 2A-785_Aug17) | Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-10 Oct-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards | ID SN 104778 SN 103244 SN 103245 SN 55277 (204) SN 2328 SN 789 ID | facility: environment temperature (22 ± 3)°C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521:02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-0252) 07-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02 | Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Scheduled Check |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Reference Probe ERSDV6 DAE4 Secondary Standards Power motor E4419B | UCted in the closed laboratory 8TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 103245 SN 2328 SN 2328 SN 2328 SN 299 ID SN 0841293874 | facility: environment temperature (22 ± 3)°C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521x02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 06-Apr-16 (In house) 06-Apr-16 (In house) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jup-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards Power motor E44198 Power sensor E4412A | Ucted in the closed laboratory STE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 103245 SN 55277 (204) SN 2328 SN 789 ID SN 0841293874 SN WY41498087 | facility: environment temperature (22 ± 3)°C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 06-Apr-17 (No. CAE4-738, Aug17) Check Date (in house) 06-Apr-16 (in house check Jun-16) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards Power sensor E4419B Power sensor E4412A | Ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103244 SN 103245 SN 55277 (20k) SN 2528 SN 789 ID SN 0841293874 SN 0841293874 SN 4041498087 SN 000110210 | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C | Ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 55277 (20k) SN 2328 SN 789 ID SN 0841293874 SN MY41498087 SN 00110210 SN U53642001700 | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 06-Apr-16 (In house Check Jun-16) 06-Apr-16 (In house Check Jun-16) 06-Apr-16 (In house Check Jun-16) 06-Apr-16 (In house Check Jun-16) 04-Aug-99 (In house Check Jun-16) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards Power sensor E4419B Power sensor E4412A | Ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 55277 (20k) SN 2528 SN 789 ID SN 0841293874 SN 0841293874 SN 090110210 | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C | Ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 55277 (20k) SN 2328 SN 789 ID SN 0841293874 SN MY41498087 SN 00110210 SN U53642001700 | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521/02522) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 06-Apr-16 (In house Check Jun-16) 06-Apr-16 (In house Check Jun-16) 06-Apr-16 (In house Check Jun-16) 06-Apr-16 (In house Check Jun-16) 04-Aug-99 (In house Check Jun-16) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER3DV6 DAE4 Secondary Standards Power meter E44198 Power sensor E4412A Power sensor E4412A RF generator HP 8648C | Ucted in the closed laboratory &TE critical for calibration) ID SN: 104778 SN: 103244 SN: 03245 SN: 03245 SN: 55277 (20x) SN: 2328 SN: 789 ID SN: 0841293874 SN: 0941293874 SN: 0941293874 SN: 041498087 SN: 000110210 SN: US37390585 | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 06-Apr-18 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 18-Oct-01 (In house check Jun-17) Function | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Apr-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator BAE4 Secondary Standards Power sensor E44198 Power sensor E4419A RF generator HP 8648C Network Analyzer HP 8753E | ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 55277 (20k) SN 2528 SN 789 ID SN 0841293874 SN 0841293874 SN 000110210 SN 00542001700 SN US37390565 Name | facility: environment temperature (22 ± 3)°C : Cal Date (Certificate No.) 04-Apr-17 (Ne. 217-02521/02522) 04-Apr-17 (Ne. 217-02521) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02528) 19-Oct-17 (No. 217-02528) 19-Oct-17 (No. 217-02528) 06-Apr-17 (No. CAE-4-798, Aug17) Check Date (in house check Jun-16) 06-Apr-16 (in house check Jun-16) 06-Apr-16 (in house check Jun-16) 04-Aug-99 (in house check Jun-16) 04-Aug-99 (in house check Jun-16) | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator BAE4 Secondary Standards Power sensor E44198 Power sensor E4419A RF generator HP 8648C Network Analyzer HP 8753E | ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 55277 (20k) SN 2528 SN 789 ID SN 0841293874 SN 0841293874 SN 000110210 SN 00542001700 SN US37390565 Name | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 06-Apr-18 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 18-Oct-01 (In house check Jun-17) Function | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator BAE4 Secondary Standards Power sensor E44198 Power sensor E4419A RF generator HP 8648C Network Analyzer HP 8753E | ucted in the closed laboratory &TE critical for calibration) ID SN 104778 SN 103244 SN 103245 SN 55277 (20k) SN 2528 SN 789 ID SN 0841293874 SN 0841293874 SN 000110210 SN 00542001700 SN US37390565 Name | faolity: environment temperature (22 ± 3)*C / Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 07-Apr-17 (No. 217-02525) 06-Apr-18 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 04-Aug-99 (In house check Jun-16) 18-Oct-01 (In house check Jun-17) Function | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER30V6 DAE4 Secondary Standards Power moter E44198 Power sensor E4412A RF generation HP 8646C Network Analyzer HP 8753E Calibrated by | ID SN 104778 SN 104778 SN 103244 SN 103245 SN 103245 SN 2328 SN 2328 SN 2328 SN 789 ID SN 0841293874 SN 009110210 SN 00910210 SN US3642001700 SN US3642001700 SN US36430585 Name Jeton Kastrati | facility: environment temperature (22 ± 3)*C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521302522) 04-Apr-17 (No. 217-02523) 04-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 00-Apr-18 (In house) 06-Apr-16 (In house) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 18-Oct-01 (In house check Oct-17) Function | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER30V6 DAE4 Secondary Standards Power moter E44198 Power sensor E4412A RF generation HP 8646C Network Analyzer HP 8753E Calibrated by | ID SN 104778 SN 104778 SN 103244 SN 103245 SN 103245 SN 2328 SN 2328 SN 2328 SN 789 ID SN 0841293874 SN 009110210 SN 00910210 SN US3642001700 SN US3642001700 SN US36430585 Name Jeton Kastrati | facility: environment temperature (22 ± 3)*C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521002522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 00-Apr-18 (In house) 06-Apr-16 (In house) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 18-Oct-01 (In house check Oct-17) Function | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |
| All calibrations have been cond Calibration Equipment used (M Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Reference 20 dB Attenuator Reference Probe ER30V6 DAE4 Secondary Standards Power moter E44198 Power sensor E4412A RF generation HP 8646C Network Analyzer HP 8753E Calibrated by | ID SN 104778 SN 104778 SN 103244 SN 103245 SN 103245 SN 2328 SN 2328 SN 2328 SN 789 ID SN 0841293874 SN 009110210 SN 00910210 SN US3642001700 SN US3642001700 SN US36430585 Name Jeton Kastrati | facility: environment temperature (22 ± 3)*C : Cal Date (Certificate No.) 04-Apr-17 (No. 217-02521002522) 04-Apr-17 (No. 217-02521) 04-Apr-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 10-Oct-17 (No. 217-02528) 00-Apr-18 (In house) 06-Apr-16 (In house) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 06-Apr-16 (In house check Jun-16) 18-Oct-01 (In house check Oct-17) Function | and humidity < 70%. Scheduled Calibration Apr-18 Apr-18 Apr-18 Apr-18 Oct-18 Aug-18 Scheduled Check In house check: Jun-18 In house check: Jun-18 |



No. I19N02705-HAC RF Page 46 of 91

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 6004 Zurich, Switzerland

S Schweizerischer Kalibrierdienst C Service suisse d'étaionnage Servizio svizzero di taratura Swise Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swee Accreditation Service (SAS) The Swisk Accreditation Service is one of the signatories to the EA Multilateral Agroement for the recognition of calibration certificates

Glossary: NORMX,y,z DCP CF A, B, C, D Polarization ø Polarization 8

sensitivity in free space diode compression point crest factor (1/duty, cycle) of the RF signal modulation dependent linearization parameters or rotation around probe axis 9 rotation around an axis that is in the plane normal to probe axis (at measurement center), i.e., 9 = 0 is normal to probe axis information used in DASY system to align probe sensor X to the robot coordinate system

Connector Angle

Calibration is Performed According to the Following Standards:

- IEEE Std 1309-2005, "IEEE Standard for calibration of electromagnetic field sensors and probes, excluding antennas, from 9 kHz to 40 GHz", December 2005
- b) CTIA Test Plan for Hearing Aid Compatibility, Rev 3.0, November 2013

Methods Applied and Interpretation of Parameters:


- NORMx,y,z: Assessed for E-field polarization & = 0 for XY sensors and & = 90 for Z sensor (f < 900 MHz in TEM-cell; f > 1800 MHz; R22 waveguide).
- NORM(f)x,y,z = NORMx,y,z * frequency_response (see Frequency Response Chart).
- DCPx,y,z: DCP are numerical linearization parameters assessed based on the data of power sweep with CW signal (no uncertainty required). DCP does not depend on frequency nor media.
- PAR: PAR is the Peak to Average Ratio that is not calibrated but determined based on the signal characteristics
- Ax,y,z: Bx,y,z: Cx,y,z; Dx,y,z; VRx,y,z: A, B, C, D are numerical linearization parameters assessed based on the data of power sweep for specific modulation signal. The parameters do not depend on frequency nor media. VR is the maximum calibration range expressed in RMS voltage across the diode.
- Spherical isotropy (3D deviation from isotropy): in a locally homogeneous field realized using an open waveguide setup.
- Sensor Offset: The sensor offset corresponds to the offset of virtual measurement center from the probe tip (on probe axis). No tolerance required.
- Connector Angle: The angle is assessed using the information gained by determining the NORMs (no uncertainty required).

Certificate No. ER3-2424_Feb18

Page 2 of 10



No. I19N02705-HAC RF Page 47 of 91





ER3DV6-SN:2424

February 23, 2018

DASY/EASY - Parameters of Probe: ER3DV6 - SN:2424

Basic Calibration Parameters

| | Sensor X | Sensor Y | Sensor Z | Unc (k=2) |
|--|----------|----------|----------|-----------|
| Norm (µV/(Vim) ²) DCP (mV) ⁶ | 1.46 | 1.51 | 1.82 | ±10.1 % |
| DCP (mV) ⁶ | 100.0 | 98.3 | 100.6 | |

Modulation Calibration Parameters

| VID | Communication System Name | | A dB | B dBõV | c | D dB | VR mV | Unc ^h (k=2) |
|------------------------------------|---|------|---------|-----------|-------|---------|----------|---------------------------|
| 0 | CW | X | 0.0 | 0.0 | 1.0 | 0.00 | 3.691 | 13.5 % |
| | | Y | 0.0 | 0.0 | 1.0 | | 204.8 | |
| | | - 2, | 0.0 | 0.0 | 1.0 | | 200.6 | |
| 10021- GSM-FDD (TDMA, GMSK) DAC | GSM-FDD (TDMA, GMSK) | x | 21.68 | 99.9 | 28.7 | 9.39 | 106.2 | 12,2 % |
| | | Y | 19.41 | 99.7 | 28.8 | | 111.3 | |
| | | Z | 24.71 | 99.5 | 28.2 | | 119.2 | |
| 10061- CAB | IEEE 802.11b W/Fi 2.4 GHz (DSSS, 11 Mbps) | x | 8.35 | 84.6 | 25.4 | 3.60 | 146.9 | ±1.9 % |
| | | Y | 4.81 | 74.8 | .21.7 | | 112.9 | |
| 1.00 | | Z | 6.43 | 78.8 | 22.9 | | 111.9 | |
| 10077- CAB | IEEE 802.11g WiFi 2.4 GHz (DSSS/OFDM, 54 Mbps) | x | 13.28 | 77.7 | 29.3 | 11.00 | 139.0 | #3.8 % |
| | | Y | 11.65 | 73.4 | 26.9 | | 100.8 | |
| 1000 | | Z | 11.41 | 72.1 | 25.6 | | 99.2 | |
| 10172- CAD | LTE-TOD (SC-FDMA, 1 R8, 20 MHz, OPSK) | × | 9.48 | 80.8 | 29.7 | 9.21 | 125.2 | ±3.8 % |
| | | Y | 9.49 | 81.9 | 30.6 | | 134.1 | |
| | | Z | 10.82 | 83.6 | 30.5 | | 136.8 | |
| 10173- CAD | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM) | x | 9.87 | 81.2 | 29.9 | 9.48 | 125.1 | 12.5 % |
| | | Y | 10.11 | 83.1 | 31.3 | | 134.2 | |
| | | Z | 11.30 | 84.2 | 30.8 | | 136.0 | |
| 10295- AAB | CDMA2000, RC1, SO3, 1/8th Rate 25 fr. | х | 16.69 | 99.5 | 40.3 | 12.49 | 96.6 | \$2.5 % |
| | | Y. | 15.42 | 99.3 | 41.1 | | 100.6 | |
| | | Z | 17.91 | 99,9 | -39.8 | | 104.3 | |

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.


* Numerical linearization parameter: uncertainty not required.
* Uncertainty is determined using the mox: deviation from linear response applying rectangular distribution and is expressed for the square of the field value.

Certificate No: ER3-2424_Feb18


Page 4 of 10




No. I19N02705-HAC RF Page 49 of 91



















No. I19N02705-HAC RF Page 53 of 91





ER3DV6 - SN:2424

February 23, 2018

DASY/EASY - Parameters of Probe: ER3DV6 - SN:2424

Other Probe Parameters

| Sensor Arrangement | Rectangular |
|---|-------------|
| Connector Angle (*) | -11.2 |
| Mechanical Surface Detection Mode | enabled |
| Optical Surface Detection Mode | disabled |
| Probe Overall Length | 537 mm |
| Probe Body Diameter | 10 mm |
| Tip Length | 10 mm |
| Tip Diameter | 8 mm |
| Probe Tip to Sensor X Calibration Point | 2.5 mm |
| Probe Tip to Sensor Y Calibration Point | 2.5 mm |
| Probe Tip to Sensor Z Calibration Point | 2.5 mm |

Certificate No: ER3-2424_Feb18

Page 10 of 10



ANNEX D Dipole Calibration Certificate

Dipole 835 MHz

| redited by the Swiss Accreditation swiss Accreditation Service is | | | ccreditation No.: SCS 0108 |
|--|---|--|--|
| Itilateral Agreement for the recog | | | |
| ent CTTL (Auden) | | | : CD835V3-1165_Jul18 |
| Still (manual) | | | |
| ALIBRATION CE | RTIFICATE | | |
| bject | CD835V3 - SN: 1 | 165 | |
| | QA CAL-20.v6 | dure for display in air | |
| | Calibration proce | dure for dipoles in air | |
| Calibration date: | July 19, 2018 | | |
| Contraction of the second s | C (194) () C () () () () () () | | |
| The measurements and the uncerta | inties with confidence p | cruit standards, which realize the physical u obability are given on the following pages a photositics environment hermarature (22 = 3) | ind are part of the certificate. |
| The measurements and the uncerta NI calibrations have been conducte Calibration Equipment used (M&TE | inties with confidence p d in the closed laborator critical for calibration | obability are given on the following pages a y facility: environment temperature (22 ± 3) | nd are part of the certificate. |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards | inties with confidence p d in the closed laborator critical for calibration) | obability are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) | nd are part of the certificate. C and humidity < 70%. Scheduled Calibration |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP | inties with confidence p d in the closed laborator critical for ca8bration) D # SN: 104778 | obability are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) | nd are part of the certificate. °C and humidity < 70%. Scheduled Calibration Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NEIP Power sensor NEIP | Inities with confidence p d in the closed laborator critical for ca8bration) D # SN: 104778 SN: 104778 SN: 103244 | bability are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) | Ind are part of the certificate. C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NEIP Power sensor NRP-Z91 Power sensor NRP-Z91 | Inties with confidence p d in the closed laborator critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 | Ubability are given on the following pages a y facility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-16 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) | rd are part of the certificate. °C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator | In the closed laborator critical for calibration) ID # SN: 104778 SN: 104244 SN: 103245 SN: 5058 (20k) | Obsbility are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) | Ind are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination | Inities with confidence p of in the closed laboration critical for calibration) ID # SN: 104778 SN: 104244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 | ubability are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) | rd are part of the certificate. °C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Prower sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 | Inters with confidence p of in the closed laboration critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 5047.3 SN: 5047.3 | Obblity are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013, Mar18) | nd are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Prower sensor NRP-Z91 Prover sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 | Inities with confidence p of in the closed laboration critical for calibration) ID # SN: 104778 SN: 104244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 | ubability are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) | Indiane part of the curtificate. IC and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 | Inties with confidence p d in the closed laboration critical for celibration) ID # SN: 104778 SN: 104778 SN: 104244 SN: 103245 SN: 5047.2 / 06327 SN: 5047.2 / 06327 SN: 6065 | Obblity are given on the following pages a y tacility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013, Mar18) 30-Dec-17 (No. H3-6065, Dec17) | Indiare part of the curtificate. IC and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards | Inters with confidence p d in the closed laboration critical for calibration) ID # SN: 104778 SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5058 (20k) SN: 5058 (20k) SN: 6065 SN: 781 | Obblity are given on the following pages a y tacility: environment temperature (22 ± 3) Od-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013, Mar18) 30-Dec-17 (No. H3-6065 Dec17) 17-Jan-18 (No. DAE4-781_Jan18) | rd are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B | Indies with confidence pr d in the closed laboration critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103245 SN: 5047.2 / 06327 SN: 6065 SN: 781 ID # | Ubability are given on the following pages a y facility: environment temperature (22 ± 3) 04-Apr-16 (No. 217-02672/02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-M | rd are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 Scheduled Check |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismetich combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A | Indies with confidence pr of in the closed laboration critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 5058 (20k) SN: 781 ID # SN: (3B42420191 | Obsbility are given on the following pages a y facility: environment temperature (22 ± 3) Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. FF3-4013_Mar18) 30-Dec-17 (No. FF3-4013_Mar18) Check Date (In house) 09-Oct-09 (In house check Oct-17) | Indiane part of the certificate. "C and humidity < 70% Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 Scheduled Check In house check: Oct-29 In house check: Oct-20 In house check: Oct-20 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Aftenuator Type-N mismetch combination Probe EF3DV3 Probe H3DV8 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06 | Inties with confidence p of in the closed laboration critical for calibration) ID # SN: 104778 SN: 104778 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 5065 SN: 781 ID # SN: 6065 SN: 781 ID # SN: 084465102 SN: US37295597 SN: 832283/011 | Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02573) 04-Apr-18 (No. 217-02672/02573) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-19 (No. H3-6065, Dect77) 17-Jan-18 (No. OAE4-781_Jan18) Check Date (In house) 09-Oct-09 (In house check Oct-17) 05-Jan-10 (In house check Oct-17) 07-Aug-12 (In house check Oct-17) | rd are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-Z91 Power sensor NRP-Z91 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Aglient 4419B Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06 | Inties with confidence p of in the closed laboration critical for calibration) ID # SN: 104778 SN: 104778 SN: 104245 SN: 5056 (20k) SN: 5047.2 / 06327 SN: 5047.2 / 06327 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 | Obsblitty are given on the following pages a Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 517-02683) 05-Mar-18 (No. EF3-4013, Mar18) 30-Dec-17 (No. H5-6065, Dec17) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) | Indiane part of the certificate. "C and humidity < 70% Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 Scheduled Check In house check: Oct-29 In house check: Oct-20 In house check: Oct-20 |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B | Indies with confidence provide the closed laboration critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5058 (20k) SN: 5058 (20k) SN: 5058 (20k) SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011 SN: US341060477 Name | Obsbility are given on the following pages a y facility: environment temperature (22 ± 3) O4-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-19 (No. EF3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-761_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) 31-Mar-14 (in house check Oct-17) Function | rd are part of the certificate. "C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-16 Signature |
| The measurements and the uncerta All calibrations have been conducte Calibration Equipment used (M&TE Primary Standards Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismetch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Aglient 44198 Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06 | Indies with confidence pr d in the closed laboration critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5038 (20k) SN: 5047 2 / 06327 SN: 6065 SN: 781 ID # SN: 058485102 SN: US37295597 SN: 632283/011 SN: US41060477 | Obsbility are given on the following pages a Call Date (Certilicate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. 277-02683) 05-Dec-17 (No. H3-6065, Dec17) 17-Jan-18 (No. DAE4-761_Jan18) Check Date (In house) 09-Oct-09 (In house check Oct-17) 05-Jan-10 (In house check Oct-17) 07-Aug-12 (In house check Oct-17) 13-Mar-14 (In house check Oct-17) | rd are part of the certificate. "C and humidity < 70% Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mur-10 Dec-18 Jan-19 Scheduled Check In house check: Det-20 In house check: Oct-20 In house check: Oct-20 |

Certificate No: CD835V3-1165_Jul18

Page 1 of 5



No. I19N02705-HAC RF Page 56 of 91

Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



•

S

C

S

Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

References

 ANSI-C63.19-2011 American National Standard, Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.

Methods Applied and Interpretation of Parameters:

- Coordinate System: y-axis is in the direction of the dipole arms, z-axis is from the basis of the antenna (mounted on the table) towards its feed point between the two dipole arms, x-axis is normal to the other axes. In coincidence with the standards [1], the measurement planes (probe sensor center) are selected to be at a distance of 15 mm above the top metal edge of the dipole arms.
- Measurement Conditions: Further details are available from the hardcopies at the end of the certificate. All
 figures stated in the certificate are valid at the frequency indicated. The forward power to the dipole connector
 is set with a calibrated power meter connected and monitored with an auxiliary power meter connected to a
 directional coupler. While the dipole under test is connected, the forward power is adjusted to the same level.
- Antenna Positioning: The dipole is mounted on a HAC Test Arch phantom using the matching dipole positioner with the arms horizontal and the feeding cable coming from the floor. The measurements are performed in a shielded room with absorbers around the setup to reduce the reflections. It is verified before the mounting of the dipole under the Test Arch phantom, that its arms are perfectly in a line. It is installed on the HAC dipole positioner with its arms parallel below the dielectric reference wire and able to move elastically in vertical direction without changing its relative position to the top center of the Test Arch phantom. The vertical distance to the probe is adjusted after dipole mounting with a DASY5 Surface Check job. Before the measurement, the distance between phantom surface and probe tip is verified. The proper measurement distance is selected by choosing the matching section of the HAC Test Arch phantom with the proper device reference point (upper surface of the dipole) and the matching grid reference point (tip of the probe) considering the probe sensor offset. The vertical distance to the probe sensor offset. The vertical distance to the probe is essential for the accuracy.
- Feed Point Impedance and Return Loss: These parameters are measured using a HP 8753E Vector Network Analyzer. The impedance is specified at the SMA connector of the dipole. The influence of reflections was eliminating by applying the averaging function while moving the dipole in the air, at least 70cm away from any obstacles.
- E-field distribution: E field is measured in the x-y-plane with an isotropic ER3D-field probe with 100 mW forward power to the antenna feed point. In accordance with [1], the scan area is 20mm wide, its length exceeds the dipole arm length (180 or 90mm). The sensor center is 15 mm (in z) above the metal top of the dipole arms. Two 3D maxima are available near the end of the dipole arms. Assuming the dipole arms are perfectly in one line, the average of these two maxima (in subgrid 2 and subgrid 8) is determined to compensate for any non-parallelity to the measurement plane as well as the sensor displacement. The E-field value stated as calibration value represents the maximum of the interpolated 3D-E-field, in the plane above the dipole surface.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: CD835V3-1165_Jul18

Page 2 of 5



Measurement Conditions

DASY system configuration, as far as not given on page 1.

| DASY Version | DASY5 | V52.10.1 |
|------------------------------------|-----------------|----------|
| Phantom | HAC Test Arch | |
| Distance Dipole Top - Probe Center | 15 mm | |
| Scan resolution | dx, dy = 5 mm | |
| Frequency | 835 MHz ± 1 MHz | |
| Input power drift | < 0.05 dB | |

Maximum Field values at 835 MHz

| E-field 15 mm above dipole surface | condition | Interpolated maximum |
|------------------------------------|--------------------|--------------------------|
| Maximum measured above high end | 100 mW input power | 108.7 V/m = 40.72 dBV/m |
| Maximum measured above low end | 100 mW input power | 108.6 V/m = 40.72 dBV/m |
| Averaged maximum above arm | 100 mW input power | 108.7 V/m ± 12.8 % (k=2) |

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters

| Frequency | Return Loss | Impedance |
|-----------|-------------|------------------|
| BOO MHz | 16.4 dB | 40.0 Ω - 9.2 jΩ |
| 835 MHz | 25.5 dB | 53.7 Ω + 4.0 jΩ |
| 880 MHz | 17.8 dB | 60.3 Ω - 9.8 jΩ |
| 900 MHz | 16.5 dB | 51.6 Ω - 15.3 jΩ |
| 945 MHz | 21.7 dB | 43.9 Ω + 4.8 jΩ |

3.2 Antenna Design and Handling

The calibration dipole has a symmetric geometry with a built-in two stub matching network, which leads to the enhanced bandwidth.

The dipole is built of standard semirigid coaxial cable. The internal matching line is open ended. The antenna is therefore open for DC signals.


Do not apply force to dipole arms, as they are liable to bend. The soldered connections near the feedpoint may be damaged. After excessive mechanical stress or overheating, check the impedance characteristics to ensure that the internal matching network is not affected.

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Page 3 of 5



Impedance Measurement Plot



Page 4 of 5



DASY5 E-field Result

Date: 19.07.2018

Test Laboratory: SPEAG Lab2

DUT: HAC-Dipole 835 MHz; Type: CD835V3; Serial: CD835V3 - SN: 1165

Communication System: UID 0 - CW ; Frequency: 835 MHz Medium parameters used: $\sigma = 0$ S/m, $z_t = 1$; $\rho = 0$ kg/m³ Phantom section: RF Section Measurement Standard: DASY5 (IEEE/IEC/ANSI C63,19-2011)

DASY52 Configuration:


- Probe: EF3DV3 5N4013; ConvF(1, 1, 1) @ 835 MHz; Calibrated: 05.03.2018
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn781; Calibrated: 17.01.2018
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole E-Field measurement @ 835MHz/E-Scan - 835MHz d=15mm/Hearing Aid Compatibility Test (41x361x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm Reference Value = 130.9 V/m; Power Drift = 0.02 dB Applied MHF = 0.00 dB RF audio interference level = 40.73 dBV/m Emission category: M3

MIF scaled E-field

| | Grid 2 M3 40.72 dBV/m | Grid 3 M3 40.67 dBV/m |
|-------------|--------------------------|--------------------------|
| Grid 4 M4 | Grid 5 M4 | Grid 6 M4 |
| 35.61 dBV/m | 35.96 dBV/m | 35.94 dBV/m |
| Grid 7 M3 | Grid 8 M3 | Grid 9 M3 |
| 40.41 dBV/m | 40.73 dBV/m | 40.67 dBV/m |



0 dB = 108.7 V/m = 40.72 dBV/m

Certificate No: CD835V3-1165_Jul18

Page 5 of 5



Dipole 1880 MHz

| CALIBRATION CERTIFICATE Deject CD1880V3 - SN: 1149 Salbration procedure(s) QA CAL-20,v6 Calibration procedure for dipoles in air Salbration procedure(s) QA CAL-20,v6 Calibration procedure for dipoles in air Salbration date: July 19, 2018 This catbration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with contidence probability are given on the following pages and are part of the continents. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)*C and humidity < 70%. Calibration Equipment used (MATE critical for calibration) Primary Standards D # Power sensor NRP-291 SN: 104778 04-Apr-18 (No: 217-02672) Apr-19 Power sensor NRP-291 SN: 103244 04-Apr-18 (No: 217-02672) Apr-19 Prover sensor NRP-291 SN: 103245 04-Apr-18 (No: 217-02672) Apr-19 Prover sensor NRP-291 SN: 5047-27 (8527 04-Pr-18 (No: 217-02672) Apr-19 Probe FSDOV SN: 6065 30-Dec+7 (No. H3-6065, Dec17) Dec-18 DAE4 SN: 6065 30-Dec+7 (No. H3-6065, Dec17) In house check: Cot-20 Prower sensor HP E4412A SN: 60424201914 </th <th>Engineering AG ughausstrasse 43, 8004 Zurich,</th> <th>Switzerland</th> <th>ilac MRA C S</th> <th>Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service</th> | Engineering AG ughausstrasse 43, 8004 Zurich, | Switzerland | ilac MRA C S | Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service |
|--|--|--|--|---|
| emil CTL (Auden) Certificate No: CD1880V3-1149_Jul18 CALIBRATION CERTIFICATE Autor CD1880V3 - SN: 1149 Autor CD1880V3 - SN: 1149 Calibration procedure(s) OA CAL-20.V6 Calibration procedure for dipoles in air Autor July 19, 2018 Catoration date: July 19, 2018 Catoration Equipment used (M&TE critical for calibration) Certificate No. Scheduled Calibration Processor Sth 104778 Get-Apr-18 (No. 217-02572) Apr-19 Prover meter NRP-291 Sh 103746 O4-Apr-18 (No. 217-02572) Apr-19 Prover meter NRP-291 Sh 103726 O4-Apr-18 (No. 217-02572) Apr-19 Prover meter NRP-291 Sh 103726 O4-Apr-18 (No. 217-02572) Apr-19 Sh 103726 O4-Apr-18 (No. 217-02572) Apr-19 Apr-19 Sh 103727 O4-Apr-18 (No. 217-02572) Apr-19 | 이야지 않는 것이 같은 것이 같은 것이 없는 것이 없는 것이 없다. 것이 없는 것이 없 않이 | | Stat. St. 5 = 201 | creditation No.: SCS 0108 |
| CALIBRATION CERTIFICATE Deject CD1880V3 - SN: 1149 Calibration procedure(s) QA CAL-20, v6 Calibration procedure for dipoles in air Calibration date: July 19, 2018 Calibration certificate documents the traceability to national standarda, which realize the physical units of measurements (SI). The calibration share been conducted in the closed laboratory facility: environment temperature (22 ± 3)*C and humidity < 70%. Calibration Equipment used (M&TE critical for calibration) Calibrate (Certificate No.) Scheduled Calibration Prover meter NRP SN: 104778 O4-Apr-18 (No. 217-02672) Apr-19 Prover relation RP-291 SN: 103244 O4-Apr-18 (No. 217-02672) Apr-19 Prover relation Commission SN: 5056 (20%) O4-Apr-18 (No. 217-02672) Apr-19 Prover relation SNP-291 SN: 103244 O4-Apr-18 (No. 217-02672) Apr-19 Prover relation Commission SN: 5056 (20%) O4-Apr-18 (No. 217-02672) Apr-19 Prover Namanator combinition SN: 5056 (20%) O4-Apr-18 (No. 217-02672) Apr-19 SN: 0055 30-Dec +17 (No. 136-606C, Dec)7) Dec-18 Jan-19 Probe H3DV6 SN: 0055 30-Dec+10 (No. Dase check: Oct-17) In house check: Oct-200 Prover sensor HP | and a state of the | gnition of calibration c | | CD1880V3-1149 Jul18 |
| Calibration procedure(s) CA CAL-20,v6 Calibration procedure for dipoles in air Dailbration date: July 19, 2018 Calibration certificate documents the traceability to national standards, which realize the physical units of measurements (s), the measurements and the uncertainties with confidence probability are given on the following pages and are part of the confidence. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)*C and humidity < 70%. Calibration Equipment used (M&TE entical for calibration) Prower sensor NRP-291 SN: 104778 O4-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-291 SN: 103244 O4-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-291 SN: 103244 O4-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-291 SN: 103244 O4-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-291 SN: 103245 O4-Apr-18 (No. 217-02672) Apr-19 Type-N miamatch combination SN: 50472 / 08327 O4-Apr-18 (No. 217-02673) Apr-19 Poble H3DV% SN: 4013 G5-Mar-18 (No. 217-02673) Apr-19 DAE4 SN: 781 T-Jan-18 (No. 217-02673) Apr-19 Proble H3DV% SN: 60472 / 08327 O4-Apr-18 (No. 217-02673) Apr-19 | And a second second second second | ERTIFICATE | | |
| Calibration procedure for dipoles in air Calibration date: July 19, 2018 This calibration certificate documents the traceability to national standards, which resilize the physical units of measurements (S). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the confidence. All calibration Equipment used (M&TE critical for calibration) Ended to the confidence probability are given on the following pages and are part of the confidence. Primary Standards ID # Cal Date (Certificate No.) Scheduled Calibration Power meter NBP SN: 104778 O4 Apri-18 (No. 217-02672/02673) Apri-19 Power meter NBP-291 SN: 103244 O4 Apri-18 (No. 217-02672/02673) Apri-19 Power ensers NBP-291 SN: 103245 O4 Apri-18 (No. 217-02673) Apri-19 Power sensor NBP-291 SN: 103245 O4 Apri-18 (No. 217-02682) Apri-19 Power sensor NBP-291 SN: 103245 O4 Apri-18 (No. 217-02683) Apri-19 Power sensor NBP-291 SN: 103245 O4 Apri-18 (No. 217-02683) Apri-19 Power sensor NBP-291 SN: 103245 O4 Apri-18 (No. 217-02683) Apri-19 Power sensor NBP-291 SN: 103245 O4 Apri-18 (No. 217-02683) Apri-19 Pobe EF3DV6 SN: 6065 30-Dec-17 (No. H3-6065, Dec-17) Dec-18 DAE4 SN: 781 OF-10 | Object | CD1880V3 - SN: | 1149 | |
| This calibration certificate documents the traceability to national standards, which realize the physical units of measurements (SI). The measurements and the uncertainties with confidence probability are given on the following pages and are part of the certificate. All calibrations have been conducted in the closed laboratory facility: environment temperature (22 ± 3)°C and humidity < 70%. | Calibration procedure(s) | | dure for dipoles in air | |
| Interview Difference Difference <thdifference< th=""> Difference <thdifference< th=""> Difference Difference</thdifference<></thdifference<> | Calibration date: | July 19, 2018 | | |
| Interview Difference Difference <thdifference< th=""> Difference <thdifference< th=""> Difference Difference</thdifference<></thdifference<> | | | | |
| Power sensor NRP-291 SN: 103244 04-Apr-18 (No. 217-02672) Apr-19 Power sensor NRP-291 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Atternuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Type-N mismatch combination SN: 5047 2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Probe EF3DV3 SN: 4013 05-Mar-18 (No. 217-02683) Apr-19 Probe H3DV6 SN: 6065 30-Dec-17 (No. H3-6065_Dec17) Dec-18 DAE4 SN: 781 17-Jan-18 (No. DAE4-781_Jan16) Jan-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power sensor HP 8482A SN: US38485102 06-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US32263/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 RF generator R8S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-20 Calibrated by: Leff Klysner Laboratory Technician Signature | All calibrations have been conducts Calibration Equipment used (M&TE | ed in the closed laborator E critical for calibration) | y facility: environment temperature (22 \pm 3) $^{\circ}\text{C}$ | 2 and humidity $< 70\%_{\rm h}$ |
| Power stensor NRP-291 SN: 103245 04-Apr-18 (No. 217-02673) Apr-19 Reference 20 dB Atternuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Type-N mismatch combination SN: 5047 2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Probe EF30V3 SN: 4013 05-Mar-18 (No. 217-02683) Apr-19 Probe H3DV6 SN: 6065 30-Dec-17 (No. H3-6065_Dec17) Dec-18 DAE4 SN: 781 17-Jan-18 (No. DAE4-781_Jan16) Jan-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power sensor HP 24412A SN: US38485102 06-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US32263/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Aglient E8368A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-20 Name Function Signature Leff Klysner Laboratory Technician Sufficience | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards | ed in the closed laborator E critical for calibration) | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) | and humidity < 70%. Scheduled Calibration |
| Reference 20 dB Atternuator SN: 5058 (20k) 04-Apr-18 (No. 217-02682) Apr-19 Type-N mismatch combination SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Probe EF3DV3 SN: 4013 05-Mar-18 (No. EF3-4013_Mar18) Mar-19 Probe H3DV6 SN: 6065 30-Dec-17 (No. H3-6065_Dec17) Dec-18 DAE4 SN: 781 17-Jan-18 (No. DAE4-781_Jan16) Jan-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power metor Agiliert 4419B SN: GB42420191 09-Oct-09 (in house check Oct-17) In house check: Oct-20 Power sensor HP 24412A SN: US30485102 05-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 9482A SN: US3226597 09-Oct-09 (in house check Oct-17) In house check: Oct-20 RF generator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-18 Calibrated by: Leff Klysiner Laboratory Technician Suffware | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP | ed in the closed laborator E critical for calibration) ID # SN: 104778 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) | and humidity < 70%. Scheduled Calibration Apr-19 |
| Type-N mismatch combination SN: 5047.2 / 06327 04-Apr-18 (No. 217-02683) Apr-19 Probe EF3DV3 SN: 4013 05-Mar-18 (No. EF3-4013_Mar18) Mar-19 Probe H3DV6 SN: 6065 30-Dec-17 (No. H3-6065_Dec17) Dec-18 DAE4 SN: 781 17-Jan-18 (No. DAE4-781_Jan16) Jan-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power metor Agilert 4419B SN: GB42420191 09-Oct-09 (in house check Oct-17) In house check: Oct-20 Power sensor HP 24412A SN: US38485102 05-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 9482A SN: US37295597 09-Oct-09 (in house check Oct-17) In house check: Oct-20 RF generator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-18 Calibrated by: Leff Klysiner Laboratory Technician Suffware | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) | C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 |
| Probe EF3DV3 SN: 4013 05-Mar-18 (No. EF3-4013_Mar18) Mar-19 Probe H3DV6 SN: 6065 30-Dec-17 (No. H3-6065_Dec17) Dec-18 DAE4 SN: 781 17-Jan-18 (No. DAE4-781_Jan16) Jan-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power meter Agilent 4419B SN: GB42420191 09-Oct-09 (in house check Oct-17) In house check: Oct-20 Power sensor HP E4412A SN: US30485102 05-Jan-10 (in house check: Oct-17) In house check: Oct-20 Power sensor HP 9482A SN: US30485102 09-Oct-09 (in house check: Oct-17) In house check: Oct-20 Power sensor HP 9482A SN: US3226597 09-Oct-09 (in house check: Oct-17) In house check: Oct-20 RF generator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check: Oct-17) In house check: Oct-20 Network Analyzer Agilent E8368A SN: US41080477 31-Mar-14 (in house check: Oct-17) In house check: Oct-18 Calibrated by: Leff Klysner Laboratory Technician Suffware | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 | ed in the closed laborator E entical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) | S and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 |
| Note Habitity SN: 781 17-Jan-18 (No. DAE4-781_Jan16) Jan-19 Secondary Standards ID # Check Date (in house) Scheduled Check Power meter Agilent 4419B SN: GB42420191 09-Oct-09 (in house check Oct-17) In house check: Oct-20 Power sensor HP E4412A SN: US38485102 05-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 9482A SN: US3295597 09-Oct-09 (in house check Oct-17) In house check: Oct-20 RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-18 Name Function Signature Calibrated by: Leff Klysner Laboratory Technician Signature | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 5058 (20K) | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) | Sand humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 |
| Secondary Standards ID # Check Date (in house) Scheduled Check Power meter Agilent 4419B SN: GB42420191 09-Oct-09 (in house check Oct-17) In house check: Oct-20 Power sensor HP E4412A SN: US38485102 05-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US3296597 09-Oct-09 (in house check Oct-17) In house check: Oct-20 RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8368A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-18 Name Function Signature Calibrated by: Leff Klysner Laboratory Technician | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 08327 | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) | and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 |
| Secondary Standards SN: GB42420191 O9-Oct-06 (in house check Oct-17) In house check: Oct-20 Power meter Agilent 4419B SN: GB42420191 09-Oct-06 (in house check Oct-17) In house check: Oct-20 Power sensor HP E4412A SN: US38485102 06-Jan-10 (in house check: Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US37296597 09-Oct-08 (in house check: Oct-17) In house check: Oct-20 RF generator R&S SMT-06 SN: 82283/011 27-Aug-12 (in house check: Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check: Oct-17) In house check: Oct-18 Name Function Signature Calibrated by: Leff Klysner Laboratory Technician Sufficient | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) | and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 |
| Power meter Agilent 44198 SN: GB42420191 09-Oct-08 (in house check Oct-17) In house check: Oct-20 Power sensor HP E4412A SN: US38485102 06-Jan-10 (in house check Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US37295597 09-Oct-08 (in house check Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US37295597 09-Oct-08 (in house check Oct-17) In house check: Oct-20 RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-18 Name Function Signature Calibrated by: Leff Klysner Laboratory Technician Sufficient | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17) | and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 |
| Power sensor HP E4412A SN: US38485102 06-Jan-10 (in house check: Oct-17) In house check: Oct-20 Power sensor HP 8482A SN: US37295597 09-Oct-09 (in house check: Oct-17) In house check: Oct-20 RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-20 Name Function Signature Calibrated by: Leff Klysner Laboratory Technician | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 08327 SN: 6065 SN: 681 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. E3-4013_Mar18) 30-Dec-17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-781_Jan18) | Sand humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 |
| Power sensor HP 8482A SN: US37295597 09-Oct-08 (in house check Oct-17) In house check: Oct-20 (in house check Oct-17) RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 (in house check Oct-17) Network Analyzer Agilent E8368A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-20 (in house check Oct-17) Name Function Signature Calibrated by: Leff Klysner Laboratory Technician | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Atternuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mat-18 (No. EF3-4013_Mar18) 30-Dec+17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) | Sand humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check |
| RF penerator R&S SMT-06 SN: 832283/011 27-Aug-12 (in house check Oct-17) In house check: Oct-20 in house check Oct-17) Network Analyzer Agilent E8358A SN: US41080477 31-Mar-14 (in house check Oct-17) In house check: Oct-20 in house check: Oct-18 Name Function Signature Calibrated by: Leff Klysner Laboratory Technician | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Atternuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103244 SN: 103244 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6055 SN: 781 ID # SN: GB42420191 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 30-Dec-17 (No. H5-4013_Mar18) 30-Dec-17 (No. DAE4-781_Jan18) Check Date (In house) 09-Oct-09 (In house check Oct-17) | Sand humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 |
| Calibrated by: Leff Klysner Laboratory Technician Signature | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Atternuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 08327 SN: 5047.2 / 08327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan16) Check Date (In house) 09-Oct-09 (In house check Oct-17) 05-Jan-10 (In house check. Oct-17) | C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 |
| Calibrated by: Leif Klysner Laboratory Technician Sef Tilgan | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Aglient 4419B Power sensor HP E4412A Power sensor HP 8482A | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 06327 SN: 4013 SN: 6065 SN: 781 ID # SN: GB42420191 SN: GB42420191 SN: US38485102 SN: US37295597 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Dec-17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) | C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 |
| Set them | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP 8482A RF generator R&S SMT-06 | ed in the closed laborator E critical for calibration) ID # SN: 104778 SN: 103244 SN: 103245 SN: 103245 SN: 5058 (20K) SN: 5047.2 / 08327 SN: 4013 SN: 6065 SN: 781 ID # SN: G842420191 SN: G842420191 SN: US38485102 SN: US37295597 SN: 832283/011 | y facility: environment temperature (22 ± 3) °C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02673) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 06-Oect-09 (In house check Oct-17) 06-Jan-10 (In house check Oct-17) 09-Oct-09 (In house check Oct-17) 27-Aug-12 (In house check Oct-17) | C and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 |
| Approved by: Katja Pokovic Technical Manager | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP 8482A RF generator R&S SMT-06 | ed in the closed laborator E critical for calibrationy ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047.2 / 08327 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011 SN: US41080477 | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02683) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan18) 30-Dec-17 (No. H3-6065_Dec17) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house check Oct-17) 06-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) 31-Mar-14 (in house check Oct-17) | S and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 |
| | All calibrations have been conducts Calibration Equipment used (M&TE Primary Standards Power meter NRP Power sensor NRP-291 Power sensor NRP-291 Reference 20 dB Attenuator Type-N mismatch combination Probe EF3DV3 Probe H3DV6 DAE4 Secondary Standards Power sensor HP E4412A Power sensor HP 8482A RF penerator R&S SMT-06 Network Analyzer Agilent E8368A | ed in the closed laborator E critical for calibrationy ID # SN: 104778 SN: 103244 SN: 103245 SN: 5058 (20k) SN: 5047 2 / 08327 SN: 6065 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011 SN: US41080477 Name | y facility: environment temperature (22 ± 3)*C Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) 04-Apr-18 (No. 217-02682) 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. 217-02683) 05-Mar-18 (No. DAE4-781_Jan18) 09-Oct-09 (In house check Oct-17) 05-Jan-10 (In house check Oct-17) 06-Jan-10 (In house check Oct-17) 09-Oct-09 (In house check Oct-17) 09-Oct-09 (In house check Oct-17) 09-Oct-09 (In house check Oct-17) 27-Aug-12 (In house check Oct-17) 31-Mar-14 (In house check Oct-17) Function | S and humidity < 70%. Scheduled Calibration Apr-19 Apr-19 Apr-19 Apr-19 Mar-19 Dec-18 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 |

Certificate No: CD1880V3-1149_Jul18

Page 1 of 5



No. I19N02705-HAC RF Page 61 of 91

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



S Sci C Ser S Ser

Schweizerischer Kalibrierdienst Service suisse d'étalonnage

Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

References

[1]

- ANSI-C63.19-2011
- American National Standard, Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.

Methods Applied and Interpretation of Parameters:

- Coordinate System: y-axis is in the direction of the dipole arms, z-axis is from the basis of the antenna (mounted on the table) towards its feed point between the two dipole arms, x-axis is normal to the other axes. In coincidence with the standards [1], the measurement planes (probe sensor center) are selected to be at a distance of 15 mm above the top metal edge of the dipole arms.
- Measurement Conditions: Further details are available from the hardcopies at the end of the certificate. All
 figures stated in the certificate are valid at the frequency indicated. The forward power to the dipole connector
 is set with a calibrated power meter connected and monitored with an auxiliary power meter connected to a
 directional coupler. While the dipole under test is connected, the forward power is adjusted to the same level.
- Antenna Positioning: The dipole is mounted on a HAC Test Arch phantom using the matching dipole positioner with the arms horizontal and the feeding cable coming from the floor. The measurements are performed in a shielded room with absorbers around the setup to reduce the reflections. It is verified before the mounting of the dipole under the Test Arch phantom, that its arms are perfectly in a line. It is installed on the HAC dipole positioner with its arms parallel below the dielectric reference wire and able to move elastically in vertical direction without changing its relative position to the top center of the Test Arch phantom. The vertical distance to the probe is adjusted after dipole mounting with a DASYS Surface Check job. Before the measurement, the distance between phantom surface and probe tip is verified. The proper measurement distance is selected by choosing the matching section of the HAC Test Arch phantom with the proper device reference point (upper surface of the dipole) and the matching grid reference point (tip of the probe) considering the probe sensor offset. The vertical distance to the probe is essential for the accuracy.
- Feed Point Impedance and Return Loss: These parameters are measured using a HP 8753E Vector Network Analyzer. The impedance is specified at the SMA connector of the dipole. The influence of reflections was eliminating by applying the averaging function while moving the dipole in the air, at least 70cm away from any obstacles.
- E-field distribution: E field is measured in the x-y-plane with an isotropic ER3D-field probe with 100 mW forward power to the antenna feed point. In accordance with [1], the scan area is 20mm wide, its length exceeds the dipole arm length (180 or 90mm). The sensor center is 15 mm (in z) above the metal top of the dipole arms. Two 3D maxima are available near the end of the dipole arms. Assuming the dipole arms are perfectly in one line, the average of these two maxima (in subgrid 2 and subgrid 8) is determined to compensate for any non-parallelity to the measurement plane as well as the sensor displacement. The E-field value stated as calibration value represents the maximum of the interpolated 3D-E-field, in the plane above the dipole surface.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: CD1880V3-1149_Jul18

Page 2 of 5



Measurement Conditions

DASY system configuration, as far as not given on page 1.

| DASY Version | DASY5 | V52.10.1 |
|------------------------------------|------------------|----------|
| Phantom | HAC Test Arch | |
| Distance Dipole Top - Probe Center | 15 mm | |
| Scan resolution | dx, dy = 5 mm | |
| Frequency | 1880 MHz ± 1 MHz | |
| Input power drift | < 0.05 dB | |

Maximum Field values at 1880 MHz

| E-field 15 mm above dipole surface | condition | Interpolated maximum |
|------------------------------------|--------------------|-------------------------|
| Maximum measured above high end | 100 mW input power | 89.8 V/m = 39.06 dBV/m |
| Maximum measured above low end | 100 mW input power | 89,3 V/m = 39.02 dBV/m |
| Averaged maximum above arm | 100 mW input power | 89.5 V/m ± 12.8 % (k=2) |

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters

| Frequency | Return Loss | Impedance |
|-----------|-------------|-----------------|
| 1730 MHz | 23.9 dB | 53.9 Ω + 5.4 jΩ |
| 1880 MHz | 22.5 dB | 54.7 Ω + 6.3 jΩ |
| 1900 MHz | 23.4 dB | 55.6 Ω + 4.5 jΩ |
| 1950 MHz | 30.3 dB | 52.9 Ω - 1.3 jΩ |
| 2000 MHz | 21.3 dB | 44.2 Ω + 5.7 JΩ |

3.2 Antenna Design and Handling

The calibration dipole has a symmetric geometry with a built-in two stub matching network, which leads to the enhanced bandwidth.

The dipole is built of standard semirigid coaxial cable. The internal matching line is open ended. The antenna is therefore open for DC signals.

Do not apply force to dipole arms, as they are liable to bend. The soldered connections near the feedpoint may be damaged. After excessive mechanical stress or overheating, check the impedance characteristics to ensure that the internal matching network is not affected.


After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Certificate No: CD1880V3-1149 Jul18

Page 3 of 5



Impedance Measurement Plot



Page 4 of 5



DASY5 E-field Result

Date: 19.07.2018

Test Laboratory: SPEAG Lab2

DUT: HAC Dipole 1880 MHz; Type: CD1880V3; Serial: CD1880V3 - SN: 1149

 $\begin{array}{l} \mbox{Communication System: UID 0 - CW ; Frequency: 1880 MHz \\ \mbox{Medium parameters used: } \sigma = 0 \ S/m, \ t_r = 1; \ \rho = 0 \ kg/m^3 \\ \mbox{Phantom section: RF Section} \\ \mbox{Measurement Standard: DASY5 (IEEE/IEC/ANSI C63,19-2011)} \\ \end{array}$

DASY52 Configuration:


- Probe: EF3DV3 SN4013; ConvF(1, 1, 1) @ 1880 MHz; Calibrated: 05.03.2018
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 5n781; Calibrated: 17.01.2018
- Phantom: HAC Test Arch with AMCC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.1(1476); SEMCAD X 14.6.11(7439)

Dipole E-Field measurement @ 1880MHz/E-Scan - 1880MHz d=15mm/Hearing Aid Compatibility Test (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm

Device Reference Point: 0, 0, -6.3 mm Reference Value = 160,1 V/m; Power Drift = -0.04 dB Applied MIF = 0.00 dB RF audio interference level = 39.06 dBV/m Emission category: M2

MIF scaled E-field

| Grid 1 M2 | Grid 2 M2 | Grid 3 M2 |
|-------------|--------------------------|--------------------------|
| 38.67 dBV/m | 39.06 dBV/m | 39.01 dBV/m |
| Grid 4 M2 | Grid 5 M2 | Grid 6 M2 |
| 36 dBV/m | 36.15 d8V/m | 36.1 dBV/m |
| 50000000000 | Grid 8 M2 39.02 dBV/m | Grid 9 M2 38.91 dBV/m |



Certificate No: CD1880V3-1149_Jul18

Page 5 of 5



No. I19N02705-HAC RF Page 65 of 91

Calibration Laboratory of Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

ILAC-MRA



s

С

S

Schweizerischer Kalibrierdienst Service suisse d'étalonnage Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

Client CTTL-SZ (Auden)

Certificate No: CD2600V3-1020_Oct18

| Object | CD2600V3 - SN: 1020 | | |
|--|---|--|---|
| Calibration procedure(s) | QA CAL-20.v6 Calibration proce | dure for dipoles in air | |
| Calibration date: | October 23, 2018 | 1 | |
| This calibration certificate documer | nts the traceability to nati | onal standards, which realize the physical unit | ts of measurements (SI). |
| | | robability are given on the following pages an | |
| | | | |
| Il calibrations have been conducte | ed in the closed laborator | ry facility: environment temperature (22 \pm 3) ⁺ C | and humidity < 70%. |
| Calibration Equipment used (M&TE | autical for estimation | | |
| | | 0.10.10.10.10.10.10.1 | Webs & And Webbardson |
| Primary Standards Power meter NRP | ID # SN: 104778 | Cal Date (Certificate No.) 04-Apr-18 (No. 217-02672/02673) | Scheduled Calibration |
| ower sensor NRP-Z91 | SN: 103244 | 04-Apr-18 (No. 217-02672/02673) 04-Apr-18 (No. 217-02672) | Apr-19 Apr-19 |
| ower sensor NRP-Z91 | SN: 103245 | 04-Apr-18 (No. 217-02673) | Apr-19 |
| CIMME INSTRUCT FROM SCOTT | 000, 1000040 | 0 | 24012118 |
| alaranan 90 dB Allanuatar | ONLEDED (DOL) | 64 Apr 18 (No. 212 03602) | Arres # G |
| | SN: 5058 (20k) | 04-Apr-18 (No. 217-02682) | Apr-19 |
| ype-N mismatch combination | SN: 5047.2 / 06327 | 04-Apr-18 (No. 217-02683) | Apr-19 |
| Type-N mismatch combination Probe EF3DV3 | SN: 5047.2 / 06327 SN: 4013 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) | Apr-19 Mar-19 |
| Type-N mismatch combination Probe EF3DV3 | SN: 5047.2 / 06327 | 04-Apr-18 (No. 217-02683) | Apr-19 |
| Reference 20 dB Attenuator Type-N mismatoh combination Probe EF3DV3 DAE4 Secondary Standards | SN: 5047.2 / 06327 SN: 4013 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) | Apr-19 Mar-19 |
| Fype-№ miamatoh combination Probe EF3DV3 DAE4 Secondary Standards | SN: 5047.2 / 06327 SN: 4013 SN: 781 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) | Apr-19 Mar-19 Jan-19 Scheduled Check |
| Pype-Ni miamatoh combination Probe EF3DV3 DAE4 Secondary Standards Power meter Agilent 4419B | SN: 5047.2 / 06327 SN: 4013 SN: 781 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) | Apr-19 Mar-19 Jan-19 |
| Pype-Ni miamatoh combination Probe EF3DV3 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A | SN: 5047.2 / 06327 SN: 4013 SN: 781 ID # SN: GB42420191 SN: US38485102 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) | Apr-19 Mar-19 Jan-19 Scheduled Check In house check: Oct-20 |
| Pype-Ni miamatoh combination Probe EF3DV3 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP 8482A | SN: 5047.2 / 06327 SN: 4013 SN: 781 ID # SN: GB42420191 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) | Apr-19 Mar-19 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 |
| Fype-№ mismatoh combination Probe EF3DV3 DAE4 | SN: 5047.2 / 06327 SN: 4013 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) | Apr-19 Mar-19 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 |
| Pype-N mismatoh combination Probe EF3DV3 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06 | SN: 6047.2 / 06327 SN: 4013 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) Check Date (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) | Apr-19 Mar-19 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 |
| Pype-N mismatoh combination Probe EF3DV3 DAE4 Secondary Standards Power meter Agilent 4419B Power sensor HP E4412A Power sensor HP 8482A RF generator R&S SMT-06 | SN: 5047.2 / 06327 SN: 4013 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283011 SN: US41080477 | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) 09-Oct-09 (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) 31-Mar-14 (in house check Oct-18) | Apr-19 Mar-19 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-20 In house check: Oct-19 Signature |
| Pype-N miamatoh combination Probe EF3DV3 DAE4 Secondary Standards Power meter Aglient 44198 Power sensor HP 8482A RF generator R&S SMT-06 Vetwork Analyzer Aglient EB358A | SN: 5047.2 / 06327 SN: 4013 SN: 781 ID # SN: GB42420191 SN: US38485102 SN: US37295597 SN: 832283/011 SN: US41080477 Name | 04-Apr-18 (No. 217-02683) 05-Mar-18 (No. EF3-4013_Mar18) 17-Jan-18 (No. DAE4-781_Jan18) 09-Oct-09 (in house) 09-Oct-09 (in house check Oct-17) 05-Jan-10 (in house check Oct-17) 09-Oct-09 (in house check Oct-17) 27-Aug-12 (in house check Oct-17) 31-Mar-14 (in house check Oct-18) Function | Apr-19 Mar-19 Jan-19 Scheduled Check In house check: Oct-20 In house check: Oct-20 |

Certificate No: CD2600V3-1020 Oct18

Page 1 of 5



No. I19N02705-HAC RF Page 66 of 91

Calibration Laboratory of Schmid & Partner Engineering AG

Zeughausstrasse 43, 8004 Zurich, Switzerland



S Schweizerischer Kalibrierdienst C Service suisse d'étalonnage

Servizio svizzero di taratura Swiss Calibration Service

Accreditation No.: SCS 0108

S

Accredited by the Swiss Accreditation Service (SAS)

The Swiss Accreditation Service is one of the signatories to the EA Multilateral Agreement for the recognition of calibration certificates

References

ANSI-C63.19-2011

American National Standard, Methods of Measurement of Compatibility between Wireless Communications Devices and Hearing Aids.

Methods Applied and Interpretation of Parameters:

- Coordinate System: y-axis is in the direction of the dipole arms. z-axis is from the basis of the antenna (mounted on the table) towards its feed point between the two dipole arms. x-axis is normal to the other axes. In coincidence with the standards [1], the measurement planes (probe sensor center) are selected to be at a distance of 15 mm above the top metal edge of the dipole arms.
- Measurement Conditions: Further details are available from the hardcopies at the end of the certificate. All
 figures stated in the certificate are valid at the frequency indicated. The forward power to the dipole connector
 is set with a calibrated power meter connected and monitored with an auxiliary power meter connected to a
 directional coupler. While the dipole under test is connected, the forward power is adjusted to the same level.
- Antenna Positioning: The dipole is mounted on a HAC Test Arch phantom using the matching dipole positioner with the arms horizontal and the feeding cable coming from the floor. The measurements are performed in a shielded room with absorbers around the setup to reduce the reflections. It is verified before the mounting of the dipole under the Test Arch phantom, that its arms are perfectly in a line. It is installed on the HAC dipole positioner with its arms parallel below the dielectric reference wire and able to move elastically in vertical direction without changing its relative position to the top center of the Test Arch phantom. The vertical distance to the probe is adjusted after dipole mounting with a DASY5 Surface Check job. Before the measurement, the distance between phantom surface and probe tip is verified. The proper measurement distance is selected by choosing the matching section of the HAC Test Arch phantom with the proper device reference point (upper surface of the dipole) and the matching grid reference point (tip of the probe) considering the probe sensor offset. The vertical distance to the probe is essential for the accuracy.
- Feed Point Impedance and Return Loss. These parameters are measured using a HP 8753E Vector Network Analyzer. The impedance is specified at the SMA connector of the dipole. The influence of reflections was eliminating by applying the averaging function while moving the dipole in the air, at least 70cm away from any obstacles.
- E-field distribution: E field is measured in the x-y-plane with an isotropic ER3D-field probe with 100 mW forward power to the antenna feed point. In accordance with [1], the scan area is 20mm wide, its length exceeds the dipole arm length (180 or 90mm). The sensor center is 15 mm (in z) above the metal top of the dipole arms. Two 3D maxima are available near the end of the dipole arms. Assuming the dipole arms are perfectly in one line, the average of these two maxima (in subgrid 2 and subgrid 8) is determined to compensate for any non-parallelity to the measurement plane as well as the sensor displacement. The E-field value stated as calibration value represents the maximum of the interpolated 3D-E-field, in the plane above the dipole surface.

The reported uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%.

Certificate No: CD2600V3-1020_Oct18

Page 2 of 5



Measurement Conditions

DASY system configuration, as far as not given on page 1.

| DASY Version | DASY5 | V52.10.2 |
|------------------------------------|------------------|----------|
| Phantom | HAC Test Arch | |
| Distance Dipole Top - Probe Center | 15 mm | |
| Scan resolution | dx, dy = 5 mm | |
| Frequency | 2600 MHz ± 1 MHz | |
| Input power drift | < 0.05 dB | |

Maximum Field values at 2600 MHz

| E-field 15 mm above dipole surface | condition | Interpolated maximum |
|------------------------------------|--------------------|-------------------------|
| Maximum measured above high end | 100 mW input power | 86.2 V/m = 38.71 dBV/m |
| Maximum measured above low end | 100 mW input power | 85.2 V/m = 38.61 dBV/m |
| Averaged maximum above arm | 100 mW input power | 85.7 V/m ± 12.8 % (k=2) |

Appendix (Additional assessments outside the scope of SCS 0108)

Antenna Parameters

| Frequency | Return Loss | Impedance |
|-----------|-------------|------------------|
| 2450 MHz | 18.6 dB | 42.7 Ω - 8.2 jΩ |
| 2550 MHz | 27.1 dB | 45.9 Ω + 1.2 jΩ |
| 2600 MHz | 32.4 dB | 48.3 Ω + 1.6 jΩ |
| 2650 MHz | 36.6 dB | 51.2 Ω + 1.0 jΩ |
| 2750 MHz | 19.3 dB | 50.9 Ω - 11.0 jΩ |

3.2 Antenna Design and Handling

The calibration dipole has a symmetric geometry with a built-in two stub matching network, which leads to the enhanced bandwidth.

The dipole is built of standard semirigid coaxial cable. The Internal matching line is open ended. The antenna is therefore open for DC signals.


Do not apply force to dipole arms, as they are liable to bend. The soldered connections near the feedpoint may be damaged. After excessive mechanical stress or overheating, check the impedance characteristics to ensure that the internal matching network is not affected.

After long term use with 40W radiated power, only a slight warming of the dipole near the feedpoint can be measured.

Page 3 of 5



Impedance Measurement Plot



Page 4 of 5



No. I19N02705-HAC RF Page 69 of 91

DASY5 E-field Result

Date: 23.10.2018

Test Laboratory: SPEAG Lab2

DUT: HAC Dipole 2600 MHz; Type: CD2600V3; Serial: CD2600V3 - SN: 1020

Communication System: UID 0 - CW : Frequency: 2600 MHz Medium parameters used: $\sigma = 0$ S/m, $s_0 = 1$; $\rho = 0$ kg/m³ Phantom section: RF Section w Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2011)

DASY52 Configuration:


- Probe: EF3DV3 SN4013; ConvF(1, 1, 1) @ 2600 MHz; Calibrated: 05.03.2018
- Sensor-Surface: (Fix Surface)
- Electronics: DAE4 Sn781; Calibrated: 17.01.2018
- Phantom: HAC Test Arch with AMEC; Type: SD HAC P01 BA; Serial: 1070
- DASY52 52.10.2(1495); SEMCAD X 14.6.12(7450)

Dipole E-Field measurement @ 2600MHz/E-Scan - 2600MHz d=15mm/Hearing Aid Compatibility Test (41x181x1): Interpolated grid: dx=0.5000 mm, dy=0.5000 mm Device Reference Point: 0, 0, -6.3 mm

MUE applied T. Rull

Reference Value = 64.09 V/m; Power Drift = 0.01 dB Applied MIF = 0.00 dB RF audio interference level = 38.71 dB V/m Emission category: M2

| Grid 1 M2 | Grid 2 M2 | Grid 3 M2 |
|-------------|--------------------------|--------------------------|
| 38.32 dBV/m | 38.61 dBV/m | 38.53 dBV/m |
| Grid 4 M2 | Grid 5 M2 | Grid 6 M2 |
| 37.96 dBV/m | 38.19 dBV/m | 38.15 dBV/m |
| | Grid 8 M2 38.71 dBV/m | Grid 9 M2 38.63 dBV/m |



Certificate No: CD2600V3-1020_Oct18

Page 5 of 5



ANNEX E UID Specification

Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| Name: | GSM-FDD (TDMA, GMSK) | |
|--|--|--|
| Group: UID: | GSM 10021-DAC | |
| PAR: " MIE: " | 9.39 dB 3.63 dB | |
| Standard Reference: Gategory: Modulation: Frequency Band: | ETSI TS 100 909 V8.9.0 (2005-01) FCC CET KDB 941225, D03 and D04 Periodic pulsed modulation GMSK GSM 450 (459.4 - 457.5 MHz) GSM 460 (479.8 - 466.0 MHz) GSM 710 (698.0 - 718.0 MHz) GSM 750 (747.0 - 763.0 MHz) CSM 850 (524.0 - 845.0 MHz) P-GSM 900 (800.0 - 915.0 MHz) E-GSM 900 (800.0 - 915.0 MHz) DCS 1000 (171.0 - 1755.0 MHz) DCS 1900 (170.0 - 195.0 MHz) PCS 1900 (185.0 - 1915.0 MHz) E-GSM 900 (856.0 - 1915.0 MHz) DCS 1900 (171.0 - 1755.0 MHz) E-GSM 900 (856.0 - 1910.0 MHz) E-GSM 900 (857.0 - 915.0 MHz) | |
| Betailed Specification: | Active Slot: TN0 Data: PN9 continuoue Frame: composed out of 8 Slots Multiframe: 26th (IDLE) Frame set blank | |
| Bandwidth: Integration Time: | Slottype & -timing: Normal burst for GMSK 0.2 MHz 120.0 ms | |

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10021-DAC page 1/2

16.11.2016




Calibration Laboratory of

Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)



UID Specification Sheet

UID 10021-DAC page 2/2

16.11.2016



Calibration Laboratory of Schmid & Partner

Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| Name: | EDGE-FDD (TDMA, 8PSK, TN 0) | |
|--|--|--|
| Group: UID: | GSM 10025-DAC | |
| PAR: [†] MIF: [#] | 12.62 dB 3.75 dB | |
| Standard Reference: Category: Modulation: Frequency Band: | ET5I T5 100 909 V8.9.0 (2005-01) FCC OET KDB 941225, D03 and D04 Periodic pulsed modulation BP9K GSM 450 (450.4 - 457.6 MHz) GSM 450 (476.6 - 486.0 MHz) GSM 710 (598.0 - 716.0 MHz) GSM 750 (747.0 - 783.0 MHz) P-GSM 900 (890.0 - 915.0 MHz) E-GSM 900 (890.0 - 915.0 MHz) R-GSM 900 (890.0 - 915.0 MHz) DCS 1800 (1710.0 + 178.0 MHz) EG 1900 (1850.0 - 910.0 MHz) EG 5M 900 (173.0 - 915.0 MHz) EF-GSM 900 (173.0 - 915.0 MHz) EF-GSM 900 (173.0 - 915.0 MHz) | |
| Detailed Specification: | Active Slot: TN0 Data: PN9 continuous Frame: composed out of 6 Slots Multiframe: 12th (PTCCH) and 28th (IDLE) Frame set blank Slottype 8 - timing: Normal burst for BPSK | |
| Bandwidth | 0.2 MHz | |

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).


UID Specification Sheet

UID 10025-DAC page 1/2


16.11.2016




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







Time Domain

UID Specification Sheet

UID 10025-DAC page 2/2



Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:

UMTS-FDD (WCDMA)

Group: UID:

WCDMA 10011-CAB

2.91 dB

-27.23 dB

PAR: ¹ MIF: ²

| Standard Reference: | 3GPP TS 25.141 Annex A |
|--|--|
| of an official offici | FCC OET KDB 941225 D01 SAR test for 3G devices v02 |
| Category: | Random amplitude modulation |
| Modulation: | QPSK |
| Frequency Band: | Band 1, UTRA/FDD (1920.0-1980.0 MHz, 20000) |
| | Band 2, UTRA/FDD (1850.0-1910.0 MHz, 20001) |
| | Band 3, UTRA/FDD (1710.0-1785.0 MHz, 20002) |
| | Band 4, UTRA/FDD (1710.0-1755.0 MHz, 20003) |
| | Band 5, UTRA/FDD (824.0-849.0 MHz, 20004) |
| | Band 6, UTRA/FDD (830.0-840.0 MHz, 20005) |
| | Band 7, UTRA/FDD (2500.0-2570.0 MHz, 20006) |
| | Band 8, UTRA/FDD (880.0-915.0 MHz, 20007) |
| | Band 9, UTRA/FDD (1749.9-1784.9 MHz, 20008) |
| | Band 10, UTRA/FDD (1710.0-1770.0 MHz, 20009) |
| | Band 11, UTRA/FDD (1427.9-1452.9 MHz, 20010) |
| | Band 12, UTRA/FDD (698.0-716.0 MHz, 20011) |
| | Band 13, UTRA/FDD (777.0-787.0 MHz, 20012) |
| | Band 14, UTRA/FDD (788.0-798.0 MHz, 20013) |
| | Band 19, UTRA/FDD (830.0-845.0 MHz, 20130) |
| | Band 20, UTRA/FDD (832.0-862.0 MHz, 20131) |
| | Band 21, UTRA/FDD (1447.9-1462.9 MHz, 20132) |
| | Band 22, UTRA/FDD (3410.0-3490.0 MHz, 20217) |
| | Band 25, UTRA/FDD (1850.0-1915.0 MHz, 20218) |
| | Band 26, UTRA/FDD (814.0-849.0 MHz, 20219) |
| Detailed Specification: | Dedicated Channel Type: RMC |
| | Bitrate: 12.2 kbps |
| | DPDCH: 60 kbps |
| | DPCCH: 15 kbps |
| | DPCCH/DPDCH power ratio: -5.46 dB |
| Bandwidth: | 5.0 MHz |
| Integration Time: | 100.0 ms |


PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for
 the same communication system (same UID and version).

UID Specification Sheet


UID 10011-CAB page 1/2




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)









UID Specification Sheet

UID 10011-CAB page 2/2



Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name: UMTS-FDD (HSPA+) Group: WCDMA UID: 10225-CAB PAR: 1 5.97 dB MIF: 2 -20.39 dB Standard Reference: 3GPP Rel 7 TS 34.121 FCC OET KDB 941225 D01 SAR test for 3G devices v02 FCC OET KDB 941225 D02 Guidance for 3GPP R6 and R7 HSPA v02v01 Category: Random amplitude modulation Modulation: 16QAM Frequency Band: Band 1, UTRA/FDD (1920.0-1980.0 MHz, 20000) Band 2, UTRA/FDD (1850.0-1910.0 MHz, 20001) Band 3, UTRA/FDD (1710.0-1785.0 MHz, 20002) Band 4, UTRA/FDD (1710.0-1755.0 MHz, 20003) Band 5, UTRA/FDD (824.0-849.0 MHz, 20004) Band 6, UTRA/FDD (830.0-840.0 MHz, 20005) Band 7, UTRA/FDD (2500.0-2570.0 MHz, 20006) Band 8, UTRA/FDD (880.0-915.0 MHz, 20007) Band 9, UTRA/FDD (1749.9-1784.9 MHz, 20008) Band 10, UTRA/FDD (1710.0-1770.0 MHz, 20009) Band 11, UTRA/FDD (1427.9-1452.9 MHz, 20010) Band 12, UTRA/FDD (698.0-716.0 MHz, 20011) Band 13, UTRA/FDD (777.0-787.0 MHz, 20012) Band 14, UTRA/FDD (788.0-798.0 MHz, 20013)

Band 19, UTRA/FDD (830.0-845.0 MHz, 20130) Band 20, UTRA/FDD (832.0-862.0 MHz, 20131) Band 21, UTRA/FDD (1447.9-1462.9 MHz, 20132) Band 22, UTRA/FDD (3410.0-3490.0 MHz, 20217) Band 25, UTRA/FDD (1850.0-1915.0 MHz, 20218) Band 26, UTRA/FDD (814.0-849.0 MHz, 20219) **Detailed Specification:** 12.2 kbps RMC, FRC H-Set 2 CQI value: 2 Sub-test 2 Conditions: DPCCH gain factor (Beta_c) = 6/15 DPDCH gain factor (Beta_d): 15/15 E-DPDCH Settings: Symbol Rate: 2x1960 Mbps Modulation 4PAM Data Type: PN9 Bandwidth: 5.0 MHz Integration Time: 100.0 ms


PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)" Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10225-CAB page 1/2




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)





Time Domain

UID Specification Sheet

UID 10225-CAB page 2/2



Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| Name: | CDMA2000, RC1, SO3, 1/8th Rate 25 fr. |
|-------------------------|---|
| Group: | CDMA2000 |
| UID: | 10295-AAB |
| olo, | 102007010 |
| PAR: 1 | 12.49 dB |
| MIF: ² | 3.26 dB |
| Standard Reference: | 3GPP2 C.S0002-C-1, Chapter 2.1.3.9.2.3 |
| | FCC OET KDB 941225 D01 SAR test for 3G devices (v02) |
| Category: | Random amplitude modulation |
| Modulation: | 64-ary orthogonal |
| Frequency Band: | Band Class 0 (815.0-849.0 MHz, 20220) |
| | Band Class 1 (1850.0-1910.0 MHz, 20040) |
| | Band Class 2 (872.0-915.0 MHz, 20041) |
| | Band Class 3 (887.0-925.0 MHz, 20042) |
| | Band Class 4 (1750.0-1780.0 MHz, 20043) |
| | Band Class 5 (411.7-483.5 MHz, 20044) |
| | Band Class 6 (1920.0-1980.0 MHz, 20045) |
| | Band Class 7 (776.0-794.0 MHz, 20046) |
| | Band Class 8 (1710.0-1785.0 MHz, 20047) |
| | Band Class 9 (880.0-915.0 MHz, 20048) |
| | Band Class 10 (806.0-901.0 MHz, 20049) |
| | Band Class 11 (410.0-462.5 MHz, 20050) |
| | Band Class 12 (870.0-876.0 MHz, 20051) |
| | Band Class 13 (2500.0-2570.0 MHz, 20179) |
| | Band Class 14 (1850.0-1915.0 MHz, 20180) |
| | Band Class 15 (1710.0-1755.0 MHz, 20181) |
| | Band Class 16 (2502.0-2568.0 MHz, 20182) |
| | Band Class 18 (787.0-799.0 MHz, 20184) |
| | Band Class 19 (698.0-716.0 MHz, 20185) |
| | Band Class 20 (1626.5-1660.5 MHz, 20186) |
| Detailed Specification: | Band Class 21 (2000.0-2020.0 MHz, 20187) Radio Configuration 1 (RC1) |
| | Service Option 3 (SO3) |
| | Speech codec: 8k EVRC (Enhanced Voice Rate Codec) |
| Bandwidth: | 1/8th frame rate 1.2 MHz |
| Integration Time: | 500.0 ms |
| megraden mile. | www.witte |


UID Specification Sheet

UID 10295-AAB page 1/2


PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPP)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)









UID Specification Sheet

UID 10295-AAB page 2/2



Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| Name: | CDMA2000 (1xEV-DO, Rev. 0) |
|-------------------------|---|
| Group: | CDMA2000 |
| UID: | 10403-AAB |
| PAR: 1 | 3.76 dB |
| MIF: 2 | -17.67 dB |
| Standard Reference: | 941225 D01 SAR test for 3G devices v02 |
| Category: | Random amplitude modulation |
| Modulation: | BPSK |
| Frequency Band: | Band Class 0 (815.0-849.0 MHz, 20220) |
| | Band Class 1 (1850.0-1910.0 MHz, 20040) |
| | Band Class 2 (872.0-915.0 MHz; 20041) |
| | Band Class 3 (887.0-925.0 MHz, 20042) |
| | Band Class 4 (1750.0-1780.0 MHz, 20043) |
| | Band Class 5 (411.7-483.5 MHz, 20044) |
| | Band Class 6 (1920.0-1980.0 MHz, 20045) |
| | Band Class 7 (776.0-794.0 MHz, 20046) |
| | Band Class 8 (1710.0-1785.0 MHz, 20047) |
| | Band Class 9 (880.0-915.0 MHz, 20048) |
| | Band Class 10 (806.0-901.0 MHz, 20049) |
| | Band Class 11 (410.0-462.5 MHz, 20050) |
| | Band Class 12 (870.0-876.0 MHz, 20051) |
| | Band Class 13 (2500.0-2570.0 MHz, 20179) |
| | Band Class 14 (1850.0-1915.0 MHz, 20180) |
| | Band Class 15 (1710.0-1755.0 MHz, 20181) |
| | Band Class 16 (2502.0-2568.0 MHz, 20182) |
| | Band Class 18 (787.0-799.0 MHz, 20184) |
| | Band Class 19 (698.0-716.0 MHz, 20185) |
| | Band Class 20 (1626.5-1660.5 MHz, 20186) |
| | Band Class 21 (2000.0-2020.0 MHz, 20187) |
| Detailed Specification: | Physical Layer Configuration: Subtype 0 |
| | Reverse Data Channel: 153.6kbps |
| | Forward Traffic Channel: 2-slot version of 307.2kbps, ACK |
| | channel transmitting in all slots |
| | Access Terminal Power Control: "All bits up" |
| Bandwidth: | 1.2 MHz |
| Integration Time: | 100.0 ms |


UID Specification Sheet

UID 10403-AAB page 1/2


PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







Time Domain

UID Specification Sheet

UID 10403-AAB page 2/2



Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:

LTE-FDD (SC-FDMA, 1 R8, 10 MHz, 16-QAM)

Group: UID:

LTE-FDO 10176-CAE

6.52 dB

PAR; 1 MIF; 1

| MIF: # | -9.76 dB |
|--------------------------------|---|
| Standard Referencie: | 3GPP / ETSI TS 136,101 V8.4,0 3GPP / ETSI TS 136,213 V8.4,0 |
| Category: | FCC DET KDB 941225 D05 SAR for LTE Devices v01 Random amplitude modulation |
| Modulation: | 16-QAM |
| Frequency Band: | Band 1, E-UTRA/FDD (1920.0 - 1980.0 MHz) |
| Troughers and some so. | Band 2, E-UTRA/FDD (1850.0 - 1910.0 MHz) |
| | Band 3, E-UTRA/FDD (1710.0 - 1785.0 MHz) |
| | Band 4, E-UTRA/FDD (1710.0 - 1755.0 MHz) |
| | Band 5, E-UTRA/FDD (824.0 - 849.0 MHz) |
| | Band 6, E-UTRA/FDD (830.0 - 840.0 MHz) |
| | Band 7, E-UTRA/FDD (2500.0 - 2570.0 MHz) |
| | Band 8, E-UTRA/FDD (880.0 - 915.0 MHz) |
| | Band 9, E-UTRA/FDD (1749.9 - 1784.9 MHz) |
| | Band 10, E-UTRA/FDD (1710.0 - 1770.0 MHz) |
| | Band 11, E-UTRA/FDD (1427.9 - 1447.9 MHz) |
| | Band 12, E-UTRA/FDD (699.0 - 710.0 MHz) |
| | Band 13, E-UTRA/FDD (777.0 - 787.0 MHz) |
| | Band 14, E-UTRA/FDD (788.0 - 798.0 MHz) |
| | Band 17, E-UTRA/FDD (704.0 - 716.0 MHz) |
| | Band 18, E-UTRA/FDD (815.0 - 830.0 MHz) |
| | Band 19, E-UTRA/FOD (830.0 - 845.0 MHz) |
| | Band 20, E-UTRA/FDD (832.0 - 862.0 MHz) |
| | Band 21, E-UTRA/FDD (1447.9 - 1462.9 MHz) |
| | Band 22, E-UTRA/FDD (3410.0 - 3490.0 MHz) |
| | Band 23, E-UTRA/FDD (2000.0 - 2020.0 MHz) |
| | Band 24, E-UTRA/FDD (1626.5 - 1660.5 MHz) |
| | Band 25, E-UTRA/FDD (1850.0 - 1915.0 MHz) |
| | Band 26 E-UTRA/FDD (814.0 - 849.0 MHz) |
| | Band 27 E-UTRA/FDD (807.0 - 824.0 MHz) |
| | Band 28 E-UTRA/FDD (703.0 - 748.0 MHz) |
| | Band 30, E-UTRA/FDD (2305.0 - 2315.0 MHz) |
| | Band 65, E-UTRA/FDD (1920.0 - 2010.0 MHz) |
| | Band 66, E-UTRA/FDD (1710.0 - 1780.0 MHz) |
| | Band 68, E-UTRA/FDD (698.0 - 728.0 MHz) |
| | Band 70, E-UTRA/FDD (1695.0 - 1710.0 MHz) |
| | Band 71, E-UTRA/FDD (663.0 - 698.0 MHz) |
| | Validation band (0.0 - 6000.0 MHz) |
| Detailed Specification: | Modulation Scheme: SC-FDMA |
| cientose opecinicación. | Number of PUSCHa: 1 |
| | Settings for Subframe #0 to #9: |
| | Modulation Scheme: QPSK |
| | Data Type: UL-SCH |
| | Number RB: 1 |
| | Transport Block Size: 258 |
| | TBS Index: 14 |
| | MCS Index: 15 |
| | Data Type: PNS |
| Bandwidth: | 10.0 MHz |
| Integration Time: | 10.0 ms |
| 27.22 <u>4</u> 141212121212012 | |


PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)" Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for z the same communication system (same UID and version).

UID Specification Sheet


UID 10176-CAE page 1/2




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







UID Specification Sheet

UID 10176-CAE page 2/2



Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| 1 | iame: |
|---|--------|
| ì | irouo: |
| 7 | JID: |

LTE-FDD (SC-FDMA, 1 R8, 20 MHz, 16-QAM)

Band 22, E-UTRA/FDD (3410.0 - 3490.0 MHz) Band 23, E-UTRA/FDD (2000.0 - 2020.0 MHz) Band 25, E-UTRA/FDD (1850.0 - 1915.0 MHz) Band 28 E-UTRA/FDD (703.0 - 748.0 MHz)

3GPP / ETSI TS 136.101 VE.4.0

LTE-FDD 10170-CAD 6.52 dB

-9.76 dB

| PAR: |
|---------|
| r part, |
| |

| | tandar | d Refe | Héndie | |
|---|---------|--------|--------|--|
| c | ategor | γ: | | |
| N | lodulat | ion: | | |

| Property and a second second second | When he had not the rest when he was the |
|-------------------------------------|---|
| | 3GPP / ETSI TS 136,213 V8.4.0 |
| Category: | FCC OET KDB 941225 DOS SAR for LTE Devices v01 Random ampiltude modulation |
| Modulation: | 16-QAM |
| Frequency Band: | Band 1, E-UTRA/FDD (1920.0 - 1980.0 MHz) |
| | Band 2, E-UTRA/FDD (1850:0+1910.0 MHz) |
| | Band 3, E-UTRA/FDD (1710.0 - 1785.0 MHz) |
| | Band 4, E-UTRA/FDD (1710.0 - 1755.0 MHz) |
| | Band 7, E-UTRA/FDD (2500.0 - 2570.0 MHz) |
| | Band 9, E-UTRA/FDD (1748.9 - 1784.9 MHz) |
| | Band 10, E-UTRA/FDD (1710.0 - 1770.0 MHz) |
| | Band 20, E-UTRA/FDD (832.0 - 862.0 MHz) |
| | Band 22, E-UTRA/FOD (3410.0 - 3400.0 MHz) |

Detailed Specification:

Band 65, E-UTRA/FDD (1820.0 - 2010.0 MHz) Band 66, E-UTRA/FDD (1710.0 - 1780.0 MHz) Band 70, E-UTRA/FDD (1695.0 - 1710.0 MHz) Band 71, E-UTRA/FDD (863.0 - 698.0 MHz) Validation band (0.0 - 6000.0 MHz) Modulation Scheme: SC-FDMA Number of PUSCHs: 1 Settings for Subframe #0 to #9: Modulation Scheme: 16QAM Data Type: UL-SCH Number RB: 1 Transport Block Size: 258 TBS Index: 14 MCS Index: 15 Data Type: PN9 20.0 MHz 10.0 ma

Bandwidth: Integration Time:


PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak to Average Power Ratio (PAPR)" z Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10170-CAD page 1/2




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







UID Specification Sheet

UID 10170-CAD page 2/2



Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| Name: | LTE-TDD (SC-FDMA, 1 RB, 20 MHz, 16-QAM) | |
|-------------------------|---|--|
| Group: | LTE-TD0 | |
| UID: | 10173-CAD | |
| PAR: 1 | 9.48 dB | |
| MIE: # | +1.44dB | |
| Standard Reference: | 3GPP / ETSI TS 136.101 V0.4.0 3GPP / ETSI TS 156.213 V6.4.0 | |
| Category: | FCC OET KDB 941225 D05 SAR for LTE Devices v02 Random amplitude modulation | |
| Modulation: | 16-QAM | |
| Frequency Band: | Band 33, E-UTRA/TOD (1900.0 - 1920.0 MHz) Band 35, E-UTRA/TOD (1850.0 - 1910.0 MHz) Band 36, E-UTRA/TDD (1910.0 - 1930.0 MHz) Band 38, E-UTRA/TDD (2570.0 - 2820.0 MHz) Band 38, E-UTRA/TDD (2570.0 - 2820.0 MHz) Band 40, E-UTRA/TDD (2300.0 - 1920.0 MHz) Band 41, E-UTRA/TDD (2400.0 - 1920.0 MHz) Band 42, E-UTRA/TDD (2400.0 - 2800.0 MHz) Band 43, E-UTRA/TDD (2400.0 - 2800.0 MHz) Band 44, E-UTRA/TDD (2400.0 - 2800.0 MHz) Band 44, E-UTRA/TDD (2400.0 - 2800.0 MHz) Band 45, E-UTRA/TDD (2400.0 - 2800.0 MHz) Band 46, E-UTRA/TDD (2600.0 - 2800.0 MHz) Band 46, E-UTRA/TDD (2600.0 - 2800.0 MHz) Band 47, E-UTRA/TDD (2600.0 - 2800.0 MHz) Band 47, E-UTRA/TDD (2600.0 - 2800.0 MHz) Band 46, E-UTRA/TDD (2600.0 - 5925.0 MHz) Band 46, E-UTRA/TDD (2585.0 - 5925.0 MHz) Band 46, E-UTRA/TDD (2580.0 - 3700.0 MHz) Validation band (0.0 - 6000.0 MHz) | |
| Detailed Specification: | Modulation Scheme: SC-FDMA Upfrik-downlink configuration: 1 Special Subframe configuration: 4 Number of Frames: 1 Settings for UL Subframe 2:3,7,8 Number of PUSCHa: 1 Modulation Scheme: 18CAM Allocated RB: 1 Start Number of RB: 50 Data Type: PN9fx | |
| Bandwidth: | 20.0 MHz | |
| Integration Time: | 6.0 ms | |
| | | |


 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).

UID Specification Sheet


UID 10173-CAD page 1/2




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







UID Specification Sheet

UID 10173-CAD page 2/2



Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

Name:

IEEE 802.11b WiFi 2.4 GHz (DSSS, 11 Mbps)

Group: UID: WLAN 10061-CAB

PAR: 1 MIF: 2 3.60 dB -2.02 dB

| IEEE 802.11b-1999 , Part 11, FCC SAR meas for 802 11 a b g v01r02 (248227 D01) |
|---|
| Random amplitude modulation |
| DQPSK |
| WLAN 2.4GHz (2412.0-2484.0 MHz, 20230) |
| Data Rate: 11 Mbps |
| Spreading, Coding: CCK |
| PPDU format: Long Preamble & Heading |
| PSDU Length: 1024 |
| PSDU Data: PN9 20.0 MHz |
| 1.5 ms |
| |

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Ratio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).


UID Specification Sheet

UID 10061-CAB page 1/2


26.11.2014




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







UID Specification Sheet

UID 10061-CAB page 2/2

26.11.2014



Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland

| Name: | IEEE 802.11a/h WiFi 5 GHz (OFDM, 54 Mbps) | |
|---|--|--|
| Group: UID: | WLAN 10069-CAC | |
| PAR: ¹ MIF: ² | 10.56 dB -3.15 dB | |
| Standard Reference: Category: Modulation: | IEEE 802.11a-1999 (P2003), Part 11 IEEE 802.11h-2003, Part 11 FCC SAR meas for 802.11 a b g v01r02 (248227.001) Random amplitude modulation 64-GAM | |
| Frequency Band: | WLAN SGHz (4915.0 - 5825.0 MHz) U-NII-1, U-NII-2A (5170 - 5330 MHz) U-NII-2C Standalone (5440 - 5710 MHz) U-NII-2C Standalone (5490 - 5650 MHz) U-NII-2C (U-NII-3 (5650 - 5635 MHz) U-NII-2C (U-NII-3 (5650 - 5635 MHz) Validation band (0.0 - 6000.0 MHz) | |
| Detailed Specification: | Data Fate: 54 Mbps Coding Rate: 3/4 Coding bits per subcarrier: 6 Codied bits per OFDM symbol: 288 Data bits per OFDM symbol: 216 PSDU Length: 1000 Bytes PSDU Data: PN8 | |
| Bandwidth: Integration Time: | 20.0 MHz .0.3 ms | |

 PAR (0.1%) in accordance with FCC KDB 971168, Section 6.0 "Measurement of the Peak-to-Average Power Batio (PAPR)"
 Modulation Interference Factor (MIF) value valid only in conjunction with advanced probe response linearization calibration for the same communication system (same UID and version).


UID Specification Sheet

UID 10069-CAC page 1/2


23.11.2017




Schmid & Partner Engineering AG Zeughausstrasse 43, 8004 Zurich, Switzerland



Complementary Cumulative Distribution Function (CCDF)







Time Domain

UID Specification Sheet

UID 10069-CAC page 2/2