


REVISIONS

DASH NO.	ECN	REV	DESCRIPTION	DATE	APPROVED

THIS DRAWING (SPECIFICATION) IS THE PROPERTY OF TVM CORPORATION ALL RIGHTS RESERVED

1
—
2
—
3
—
4
—
5
—
6
—
7
—
8
—
9
—
10

$76 \begin{matrix} +0 \\ -0.2 \end{matrix}$



elo TOUCH SOLUTIONS
Elo Touch Solutions, Inc.
2245 B.H. Town Line Rd.
Rochester, NY 14623
www.elotouch.com

Model: ESY15M1 Rating: 12V \pm 6A
E693617 2UWA-0-ZB-MT-NO-WH-MED
This Class A digital apparatus complies with Canadian ICES-003.
Cet appareil numérique de la classe A est conforme à la norme NMB-003 du Canada.
This device complies with Part 15 of the FCC Rules. Operation is subject to the following two conditions: (1) This device may not cause harmful interference, and (2) This device must accept any interference received, including interference that may cause undesired operation. Made in China
Manufacturer: Elo Touch Solutions, Inc
EU importer: Elo Touch Solutions (Belgium) NV, 7414536A0605(7)
Brand name: Elo Touch Solutions

FCC ID: R7QESY15M-RFID
IC: 10757A-ESY15MRFID

6K70 E162681 LISTED
CLASSIFIED
E384528 07EG


Medical Electrical Equipment
ANSI/AAMI ES60601-1
CAN/CSA-C22.2 No. 60601-1
IP21

この装置は、クラス A 情報技術装置です。この装置を家庭環境で使用すると電波妨害を引き起こすことがあります。この場合には使用者が適切な対策を講ずるよう要求されることがあります。 VCCFA

See User's Manual for Power supply information before use. Grounding Reliability can only be achieved when the equipment is connected to an equivalent receptacle marked "Hospital Only" or "Hospital Grade"


$57.6 \begin{matrix} +0 \\ -0.2 \end{matrix}$

E08C789456



Month Code:

MONTH	JAN	FEB	MAR	APR	MAY	JUN	JUL	AUG	SEP	OCT	NOV	DEC
CODE	A	B	C	D	E	F	G	H	I	J	K	L

DASH NO.			NEXT ASSY			USED ON			TVM CONFIDENTIAL			
MATERIAL			CONTRACT NO.						 TVM CORP. Tel.: (02)2771-0693 Fax.: (02)2752-4214			
FINISH			APPROVALS			DATE						
DO NOT SCALE DRAWING			DRAWN S.L.Wu			11.30.2012			B/C Label for ESY15M1-2UWA-0-ZB-MT-NO-WH-MED			
			CHECKED						SIZE: A4	DWG.NO. 7414536A0605		REV A
			ISSUED						SCALE			SHEET 1 OF 1