


REVISIONS			
LTR	DESCRIPTION	DATE	APPROVED
A	ECO 5995: Release to Rev A, Add FCC & IC #		



Notes

1. Material: PM-200 Mylar with OM-200 Compucal Laminate; 2 Mil Matte Clear Mylar Overlay Laminate.
2. Adhesive: L23.
3. Ink Color(s): All content Black on Pantone Cool Gray 1M background unless otherwise specified.
4. Font: Arial, unless otherwise specified.
5. Certification to UL and CSA Standards: LI 1001 or Equivalent, P. O. Number, and Quantity.
6. Attach overlay laminate to top section of label with peel off on bottom section. Overlay laminate boundary shall not interfere with any content.
7. "ABIOMED" and logo shown for reference purposes only. Refer to file AbioLogo286_208.eps for artwork. Scale Artwork to dimension shown.

.XX ± .015"		ABIOMED, Inc.	
APPROVALS	DATE	LABEL, SPECIFICATION SUPRACOR CONSOLE	
DRAWN R.MOSHER	3/6/06		
CHECKED			
RESP ENG			
MFG ENG		SIZE A	DWG NO. 0036-2606
DO NOT SCALE DRAWING	SCALE NONE	CAD FILE 0036-2606-R2.dwg	REV A
		SHEET 1	OF 1