

Subject : Chamber picture question

Dear Leo :

We're very sorry to make this serious mistakes , our careless put the EUT set up picture in 3m compact chamber in EMI test report , this photo. shall be used by CE report for voltage radiated field immunity test (EN 61000-4-3) , PEP perform 3V/m immunity test in 3m compact chamber in His-Chi , If client require 10V/m test , we used Lin-Ko 9mx6mx6m chamber .


we apologize for this mistakes., please refer the correct photos .

Best regards ,

Amy / PEP Testing Laboratory

8.4 Test Setup Photo

* FRONT VIEW *



* REAR VIEW *



9.4 Test Setup Photo

