


**Impedance Measurement Plot for Head TSL**

**DASY5 Validation Report for Body TSL**

Date: 18.06.2012

Test Laboratory: SPEAG, Zurich, Switzerland

**DUT: Dipole 5GHz; Type: D5GHzV2; Serial: D5GHzV2 - SN: 1040**

Communication System: CW; Frequency: 5200 MHz, Frequency: 5500 MHz, Frequency: 5800 MHz  
Medium parameters used:  $f = 5200 \text{ MHz}$ ;  $\sigma = 5.37 \text{ mho/m}$ ;  $\epsilon_r = 47$ ;  $\rho = 1000 \text{ kg/m}^3$ , Medium parameters used:  $f = 5500 \text{ MHz}$ ;  $\sigma = 5.76 \text{ mho/m}$ ;  $\epsilon_r = 46.5$ ;  $\rho = 1000 \text{ kg/m}^3$ , Medium parameters used:  $f = 5800 \text{ MHz}$ ;  $\sigma = 6.16 \text{ mho/m}$ ;  $\epsilon_r = 46$ ;  $\rho = 1000 \text{ kg/m}^3$

Phantom section: Flat Section

Measurement Standard: DASY5 (IEEE/IEC/ANSI C63.19-2007)

DASY52 Configuration:

- Probe: EX3DV4 - SN3503; ConvF(4.91, 4.91, 4.91); Calibrated: 30.12.2011, ConvF(4.43, 4.43, 4.43); Calibrated: 30.12.2011, ConvF(4.38, 4.38, 4.38); Calibrated: 30.12.2011;
- Sensor-Surface: 1.4mm (Mechanical Surface Detection)
- Electronics: DAE4 Sn601; Calibrated: 04.07.2011
- Phantom: Flat Phantom 5.0 (back); Type: QD000P50AA; Serial: 1002
- DASY52 52.8.1(838); SEMCAD X 14.6.5(6469)

**Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5200 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 58.667 V/m; Power Drift = -0.05 dB

Peak SAR (extrapolated) = 29.022 mW/g

**SAR(1 g) = 7.37 mW/g; SAR(10 g) = 2.07 mW/g**

Maximum value of SAR (measured) = 17.2 mW/g

**Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5500 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=1.4mm

Reference Value = 58.708 V/m; Power Drift = -0.02 dB

Peak SAR (extrapolated) = 33.769 mW/g

**SAR(1 g) = 7.87 mW/g; SAR(10 g) = 2.19 mW/g**

Maximum value of SAR (measured) = 19.0 mW/g


**Dipole Calibration for Body Tissue/Pin=100mW, dist=10mm, f=5800 MHz/Zoom Scan, dist=1.4mm (8x8x7)/Cube 0:** Measurement grid: dx=4mm, dy=4mm, dz=1.4mm


Reference Value = 55.529 V/m; Power Drift = 0.02 dB

Peak SAR (extrapolated) = 34.868 mW/g

**SAR(1 g) = 7.44 mW/g; SAR(10 g) = 2.06 mW/g**

Maximum value of SAR (measured) = 18.1 mW/g



**Impedance Measurement Plot for Body TSL**

## ANNEX I DIPOLE QUALIFICATION FOR THE EXTENDED 3-YEAR CALIBRATION INTERVAL

### G1 Dipole750

The information and documentation below are provided to qualify the extended 3-year calibration interval of dipole.

#### G1.1 List of Equipment

No.	Name	Type	Serial Number
01	Network analyzer	E5071C	MY46110673
02	Power meter	NRVD	102083
03	Power sensor	NRV-Z5	100542
04	Signal Generator	E4438C	MY49070393
05	Amplifier	60S1G4	0331848
06	E-field Probe	SPEAG ES3DV3	3149
07	DAE	SPEAG DAE4	771
08	Dipole Validation Kit	SPEAG D750V3	1045

#### G1.2 Results of Impedance, Return-loss and System validation

##### Dipole 750 - Head

		Year		Deviation	Limit
		2011	2012		
Impedance	Real ( $\Omega$ )	54.2	51.7	2.5 $\Omega$	Deviation < 5 $\Omega$
	Imaginary ( $\Omega$ )	-2.3	0.9	3.2 $\Omega$	Deviation < 5 $\Omega$
System validation	Return-loss (dB)	-26.8	-26.7	0.1dB	Deviate < 0.2dB
	10g	1.4	1.43	2.14%	Deviation < 10%
	1g	2.14	2.21	3.27%	Deviation < 10%

##### Dipole 750 - Body

		Year		Deviation	Limit
		2011	2012		
Impedance	Real ( $\Omega$ )	49.5	49.6	0.1 $\Omega$	Deviation < 5 $\Omega$
	Imaginary ( $\Omega$ )	-4.1	0.6	4.7 $\Omega$	Deviation < 5 $\Omega$
System validation	Return-loss (dB)	-27.5	-27.6	0.1dB	Deviate < 0.2dB
	10g	1.45	1.47	1.38%	Deviation < 10%
	1g	2.2	2.25	2.27%	Deviation < 10%

According to the above tables, it is not necessary to recalibration the dipoles in 2012.