



2024-07-25

TUV SUD America Inc. 10 Centennial Drive
Peabody, MA 01960, USA

Dear Sir or Madam,

We, St. Jude Medical 15900 Valley View Court Sylmar, CA 91342, hereby certify the differences described below for all model families listed are valid and true.

Yours sincerely,

A handwritten signature in black ink, appearing to read 'Perry Li'.

Name: Perry Li

Title: Associate Research Fellow



Product Name	Bipolar DF4	Bipolar DF-1
Avant™ HF	CDHFA700D Tested Model (Baseline)	CDHFA700B • Tested Model (Baseline)
Neutrino™ NxT HF	CDHFA800D • No hardware or software difference from baseline. • Different product name and model number.	CDHFA800B • No hardware or software difference from baseline. • Different product name and model number.
Gallant™ HF	CDHFA500D • No hardware or software difference from baseline. • Different product name and model number.	CDHFA500B • No hardware or software difference from baseline. • Different product name and model number.
Neutrino™ NxT HF	CDHFA600D • No hardware or software difference from baseline. • Different product name and model number.	CDHFA600B • No hardware or software difference from baseline. • Different product name and model number.
Entrant™ HF	CDHFA300D • Software change to accommodate different therapy settings. • No change to RF software. • No hardware difference from baseline. • Different product name and model number.	CDHFA300B • Software change to accommodate different therapy settings. • No change to RF software. • No hardware difference from baseline. • Different product name and model number.