

March 7, 2007

Federal Communications commission
7435 Oakland Mills Rd.
Columbia, MD 21036

Re: Application for Grant Certification for FCC ID RIASJMRF


Dear Sir or Madam:

Please find transmitted electronically, an application for certification of implants to be used in a medial implant communications system that operates in the Medical Implant Communications Service (MICS) band from 402 to 405 MHz under Part 95 of the FCC Rules and Regulations. The application contains a separate request for confidentiality of specified documents.

Pursuant to Section 95.1217, the FCC ID number associated with the device and the information required by Section 2.925 of the FCC Rules will be placed in the device manual and on the shipping container label in lieu of being placed directly on the implantable device. The application contains such labeling and proposed placement.

If there are any questions, please feel free to contact me at 818-493-2629.

Sincerely,



Deanna Hughes
Regulatory Affairs
St. Jude Medical


Federal Communication Commission
7435 Oakland Mills Rd.
Columbia, MD 21036

Re: Request for Alternative Method of Identifying for FCC ID: RIASJMRF

To whom it may concern,

As per 47 CFR Part 2.925(f), St. Jude Medical is requesting approval by the FCC of an alternative method of positively identifying the devices under review. Because this is an implantable, sterile product we are requesting approval to place the FCC identifier on the outer shipping box as well as the device user's manual in lieu of placing it on the device itself.

Sincerely,



Deanna L. Hughes
Regulatory Affairs
St. Jude Medical