

April 3, 2009

Federal Communication Commission
7435 Oakland Mills Rd.
Colombia, MD 21036

Re: Application for Grant Certification for FCC ID RIASJMRFB

Dear Sir or Madam,


Please find submitted herewith the application for certification of SJM RF Bradycardia implants to be used in a medical implant communications system that operates in the MICS band from 402 to 405 MHz, under Part 95 of the FCC Rules and Regulations.

The application contains a separate request for permanent and short term confidentiality of specified documents.

Pursuant to Section 95.1217, the FCC ID number associated with the device and the information required by Section 2.925 of the FCC Rules will be placed in the User's manual for the device and on the shipping box label in lieu of being placed directly on the implantable device. The application contains such labeling and placement. Each implant is identified with a unique serial number.

If there are any questions, please feel free to contact me at 408 738 4883 ext 6494.

Sincerely,



Elisabeth E. Neely

Regulatory Affairs

St. Jude Medical