



St. Jude Medical
Cardiac Rhythm Management Division
15900 Valley View Court
Sylmar, CA 91342-3577 USA
Tel 818 362 6822
Fax 818 364 5814

December 21, 2012

Federal Communications Commission
Authorization and Evaluation Division
Equipment Authorization Branch
7435 Oakland Mills Road
Columbia, MD 21046

Re: Correspondence 43067

FCC ID: RIASJMRFB

To Whom it May Concern,

The maximum available power at the antenna port for the active implantable medical devices, FCC ID RIASJMRFB, is ≤ 1.0 mW, under all operating circumstances.

Active implantable medical devices labeled RIASJMRFB qualify for SAR exclusion based on section 4.2.4 of KDB publication 447498.

This statement supersedes the previously submitted SAR reports 60045394 REV01 and 60045397REV01.

If there are any questions regarding this letter, please contact me at 1-818-493-3425 or at ctabion@sjm.com.

Sincerely,

Carolyn Tabion
Sr. Manager, Regulatory Affairs
St. Jude Medical