



St. Jude Medical  
Implantable  
Electronic Systems  
15900 Valley View Court  
Sylmar, CA 91342 USA  
Main 818 362 6822  
Fax 818 364 5814

## Class II permissive change request letter

Date (05/25/2016)

TUV SUD BABT

Octagon House,  
Segensworth Road,  
Fareham,  
Hampshire,  
PO15 5RL

REF: Class II permissive change for FCC ID: RIASJMRFC, IC: 8454A-M3660123 Model 3664

Dear Sir or Madam,

Pursuant to section 2.1043, and IC RSP-100 Sec. 4.4, we CalAmp Wireless Networks. Hereby submit a request for class II permissive change for FCC ID: RIASJMRFC, IC: 8454A-M3660123, Model 3664

Description of the change:

The differences consist of battery capacities (either 5.3 or 7.5 Ah), type of therapy (Tonic or non-Tonic SCS/DBS or Dorsal Root Ganglion therapy (DRG), controlled by software) and method of connecting patient electrodes to the device (SJM, 1.27 mm or DRG header).

Model 3660 Orion IPG, 5.3Ah, SCS, SJM header  
Model 3661 Orion IPG, 5.3Ah, SCS, 1.27mm header  
Model 3662 Orion IPG, 7.5Ah, SCS, SJM header  
Model 3663 Orion IPG, 7.5Ah, SCS, 1.27mm header  
Model 6660 Orion IPG, 5.3Ah, DBS, Tonic, SJM header  
Model 6661 Orion IPG, 5.3Ah, DBS, Tonic, 1.27mm header  
Model 6662 Orion IPG, 7.5Ah, DBS, Tonic, SJM header  
Model 6663 Orion IPG, 7.5Ah, DBS, Tonic, 1.27mm header  
Model 3665 Orion IPG, 5.3Ah, SCS, Tonic, SJM header  
Model 3667 Orion IPG, 7.5Ah, SCS, Tonic, SJM header  
Model 3664 Proclaim™ 5 IPG, 5.3Ah, DRG, DRG header

Sincerely,

Name: Alex Chang  
Company: TUV SUD America