

Declaration Of Equivalence



To whom it may concern :
We have declared below featured for FCC equipment authorization,
Device FCC ID : RNH-EVS2430WI

We, the undersigned, hereby declare that the product, EVS 2430GWi is identical in construction including schematics, PCB layout and electronic components to the model EVS 2430Wi model.

The differences between the modified equipment and original equipment are listed below:

	Main EUT	Rest of multiple models
Product name	Flat Panel Digital X-ray Detector	Flat Panel Digital X-ray Detector
Identification (Model Number)	EVS 2430Wi	EVS 2430GWi
Manufacturer	DRTECH Corporation	DRTECH Corporation
Difference of modified equipment	Scintillator : CsI(Cesium Iodide)	Scintillator : Godox(gadolinium oxysulfide)

Note : Scintillators are not conductive.

DRTECH Corporation

Representative Signature

Kim Young min / Assistant Manager

Printed name and title

2017. 04. 24

Date