

CLASS II PERMISSIVE CHANGE APPLICATION

04/02/2013

FCC ID: ROSTEMPUSIC-2

TO WHOM IT MAY CONCERN

Remote Diagnostic Technologies Ltd. request a Class 2 Permissive Change of the presently authorised equipment under FCC ID: ROSTEMPUSIC-2

The purpose of the request is to allow integration into Remote Diagnostic Technologies Ltd.'s portable host equipment, the Tempus IC 00-1001. A SAR evaluation has been performed to take into account the portable usage condition.

The Tempus IC is a vital signs monitor designed for use in remote locations by trained nonmedical expert personnel. The device monitors a variety of vital signs and transmits the data via existing wired or wireless telecommunications that are available locally in the remote location.

The request is also to permit co-location and simultaneous transmission with ROSTEMPUSIC-BT and P5T-WL54SDIO. EMC tests were performed to ensure continued compliance with the appropriate FCC rule parts.

No parts of the modules (including their antennas) have been modified as a result of their integration into the Tempus IC.

Yours sincerely



Chris Hannan
Programme & Regulatory Affairs Director