TEST REPORT



Dt&C Co., Ltd.

42, Yurim-ro, 154Beon-gil, Cheoin-gu, Yongin-si, Gyeonggi-do, Korea, 17042 Tel: 031-321-2664, Fax: 031-321-1664

1. Report No: DRTFCC2407-0072

2. Customer

• Name (FCC) : SD Biosensor, Inc.

• Address (FCC): C-4th&5th, 16, Deogyeong-daero, 1556beon-gil, Yeongtong-gu, Suwon-si,

Gyeonggi-do, South Korea

3. Use of Report : FCC Original Grant

4. Product Name / Model Name : BTM-FSC-BT982 / FSC-BT982

FCC ID: RPJ-FSC-BT982

5. FCC Regulation(s): Part 15.247

Test Method used: KDB558074 D01v05r02, ANSI C63.10-2013

6. Date of Test: 2024.01.26 ~ 2024.06.13

8. Testing Environment: See appended test report.

9. Test Result: Refer to the attached test result.

The results shown in this test report refer only to the sample(s) tested unless otherwise stated.

This test report is not related to KOLAS accreditation.

Affirmation Tested by

Name: SeokHo Han

Technical Manager

Name: JaeJin Lee

2024.07.10.

Dt&C Co., Ltd.

If this report is required to confirmation of authenticity, please contact to report@dtnc.net

Pages: 2 / 42



Test Report Version

Report No.: DRTFCC2407-0072

Test Report No.	Date	Description	Revised by	Reviewed by
DRTFCC2407-0072	Jul. 10, 2024	Initial issue	SeokHo Han	JaeJin Lee



Table of Contents

Report No.: DRTFCC2407-0072

	General Information	
	1.1. Description of EUT	
	1.2. Declaration by the applicant / manufacturer	
	1.3. Testing Laboratory	5
	1.4. Testing Environment	
	1.5. Measurement Uncertainty	
	1.6. Information about the FHSS characteristics	6
	1.7. Conclusion of worst-case and operation mode	7
	1.8. Test Equipment List	
2.	Antenna Requirement	9
3.	Summary of Test Results	10
4.	Maximum Peak Conducted Output Power	.11
	4.1. Test Setup	. 11
	4.2. Limit	. 11
	4.3. Test Procedure	. 11
	4.4. Test Results	
5.	20 dB BW	14
	5.1. Test Setup	. 14
	5.2. Limit	
	5.3. Test Procedure	. 14
	5.4. Test Results	. 14
6.	Carrier Frequency Separation	17
	6.1. Test Setup	. 17
	6.2. Limit	. 17
	6.3. Test Procedure	. 17
	6.4. Test Results	
	Number of Hopping Channels	
	7.1. Test Setup	
	7.2. Limit	
	7.3. Test Procedure	
	7.4. Test Results	
	Time of Occupancy	
	8.1. Test Setup	
	8.2. Limit	
	8.3. Test Procedure	
	8.4. Test Results	
	Unwanted Emissions	
	9.1. Test Setup	
	9.2. Limit	
	9.3. Test Procedures	
	9.3.1. Test Procedures for Unwanted Emissions(Radiated)	
	9.3.2. Test Procedures for Unwanted Emissions(Conducted)	
	9.4. Test Results	
	9.4.1. Unwanted Emissions(Radiated)	
	9.4.2. Unwanted Emissions(Conducted)	
10). AC Power-Line Conducted Emissions	
	10.1. Test Setup	
	10.2. Limit	
	10.3. Test Procedure	
	10.4. Test Results	
	PPENDIX I	
Α	PPENDIX II	41



1. General Information

1.1. Description of EUT

Equipment Class	Spread Spectrum Transmitter(DSS)	
Product Name	BTM-FSC-BT982	
Model Name	FSC-BT982	
Add Model Name	-	
Firmware Version Identification Number	V945	
EUT Serial Number	No Specified	
Power Supply	DC: 3.3 V	
Frequency Range	2 402 MHz ~ 2 480 MHz	
Max. RF Output Power	-0.86 dBm (0.001 W)	
Modulation Technique (Data rate)	GFSK(1 Mbps)	
Number of Channels	79	
Antenna Specification	Antenna Type: PCB Pattern Antenna Gain : 2.0 dBi (PK)	

1.2. Declaration by the applicant / manufacturer

- NA



1.3. Testing Laboratory

Dt&C Co., Ltd.

The 3 m test site and conducted measurement facility used to collect the radiated data are located at the 42, Yurim-ro, 154beon-gil, Cheoin-gu, Yongin-si, Gyeonggi-do, Korea 17042.

The test site complies with the requirements of Part 2.948 according to ANSI C63.4-2014.

- FCC & IC MRA Designation No.: KR0034

- ISED#: 5740A

<u>www.dtnc.net</u>		
Telephone	:	+ 82-31-321-2664
FAX	:	+ 82-31-321-1664

1.4. Testing Environment

Ambient Condition		
Temperature	+21 °C ~ +24 °C	
 Relative Humidity 	40 % ~ 43 %	

1.5. Measurement Uncertainty

The measurement uncertainties shown below were calculated in accordance with requirements of ANSI C63.4-2014 and ANSI C63.10-2013. All measurement uncertainty values are shown with a coverage factor of k = 2 to indicate a 95 % level of confidence.

Parameter	Measurement uncertainty	
Antenna-port conducted emission	1.0 dB (The confidence level is about 95 %, k = 2)	
AC power-line conducted emission	3.4 dB (The confidence level is about 95 %, k = 2)	
Radiated emission (1 GHz Below)	5.0 dB (The confidence level is about 95 %, k = 2)	
Radiated emission (1 GHz ~ 18 GHz)	4.8 dB (The confidence level is about 95 %, k = 2)	
Radiated emission (18 GHz Above)	5.7 dB (The confidence level is about 95 %, k = 2)	



1.6. Information about the FHSS characteristics

- This Bluetooth module has been tested by a Bluetooth Qualification Lab, and we confirm the following:
 - A) The hopping sequence is pseudorandom
 - Note 1 : Pseudorandom Frequency Hopping Sequence Table as below:

```
Channel: 08, 24, 40, 56, 42, 54, 72, 09, 01, 11, 33, 41, 34, 42, 65, 73, 53, 69, 06, 22, 04, 20, 36, 52, 38, 46, 70, 78, 68, 76, 21, 29, 10, 26, 41, 58, 44, 60, 76, 13, 03, 11, 35, 43, 37, 45, 69, 77, 52, 71, 08, 24, 06, 24, 48, 56, 45, 46, 70, 01, 72, 06, 25, 33, 12, 28, 49, 60, 45, 58, 74, 13, 05, 18, 37, 49 etc
```

FCC ID: RPJ-FSC-BT982

The System receiver have input bandwidths that match the hopping channel bandwidths of their corresponding transmitters and shift frequencies in synchroniztation with the transmit ted signals.

- B) All channels are used equally on average
- C) The receiver input bandwidth equals the transmit bandwidth
- D) The receiver hops in sequenc e with the transmit signal
- 15.247(g): In accordance with the Bluetooth Industry Standard, the system is designed to comply with all of the regulations in Section 15.247 when the transmitter is presented with a continuous data (or information) system.
- 15.247(h): In accordance with the Bluetooth Industry Standard, the system does not coordinate its
 channels selection / hopping sequence with other frequency hopping systems for the express
 purpose of avoiding the simultaneous occupancy of individual hopping frequencies by multiple
 transmitters.
- 15.247(h): The EUT employs Adaptive Frequency Hopping (AFH) which identifies sources of interference namely devices operating in 802.11 WLAN and excludes them from the list of available channels. The process of re-mapping reduces the number of test channels from 79 channels to a minimum number of 20 channels.

This test report is prohibited to copy or reissue in whole or in part without the approval of Dt&C Co., Ltd.

TRF-RF-237(07)210316 Pages: 6 / 42



1.7. Conclusion of worst-case and operation mode

The EUT has three types of modulation (GFSK).

Therefore all applicable requirements were tested with all the modulations.

And packet type was tested at the worst case(DH5).

EUT Operation test setup

Bluetooth tester was used to control the transmit parameters during test.

Tested frequency information,

- Hopping Function : Enable

	Tested Frequency (MHz)	
Hopping Band	2 402 ~ 2 480	

- Hopping Function : Disable

	Tested Frequency (MHz)	
Lowest Channel	2 402	
Middle Channel	2 441	
Highest Channel	2 480	



1.8. Test Equipment List

Spectrum Analyzer Agilent Technologies N9820A 2306/23 24/09/23 U\$47308912 Spectrum Analyzer Agilent Technologies N9820A 2306/23 24/06/23 MY46471622 DC Power Supply Agilent Technologies 6632A 2306/23 24/06/23 3005M0228 DC Power Supply SM techno SDP30-5D 2306/23 24/06/23 3005M0228 Multimeter FLIKE 1794 231/21/5 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 255571 13/18/18 13/18/18 24/12/15 24/12/15 255571 13/18/18 12/18/19 24/12/15 255571 13/18/18 24/12/15 24/12/15 255571 13/18/18 12/18/19 24/12/15 24/12/15 24/12/15 24/12/15 24/12/15 12/18/19 24/12/15 24/12/15 12/18/19 12/18/19 24/12/15 12/18/19 12/18/19 12/18/19 24/12/15 12/18/19 12/18/19 12/18/19 12/18/19 <t< th=""><th>Туре</th><th>Manufacturer</th><th>Model</th><th>Cal.Date (yy/mm/dd)</th><th>Next.Cal.Date (yy/mm/dd)</th><th>S/N</th></t<>	Туре	Manufacturer	Model	Cal.Date (yy/mm/dd)	Next.Cal.Date (yy/mm/dd)	S/N
Spectrum Analyzer	Spectrum Analyzer	Agilent Technologies	N9020A	23/12/15	24/12/15	MY48010133
DC Power Supply Agilent Technologies 68332A 23006/23 24/08/23 US37474175 DC Power Supply SM techno SDP3-05D 23/08/23 24/08/23 305DMQ288 Mullimeter FLIKE 178+ 23/08/23 24/08/23 305DMQ288 Mullimeter TESCOM TC-3000C 23/08/23 24/08/23 3000C000683 Power Splitter Anrisu K241B 23/12/15 24/12/15 255571 Signal Generator Rohde Schwarz SMBV100A 23/12/15 24/12/15 255571 Signal Generator ANRTISU MG3896C 23/12/15 24/12/15 255571 Thermohygrometer BODYCOM BJ5478 23/12/15 24/12/15 120612-2 Horn Antenna ETS-Lin	Spectrum Analyzer	Agilent Technologies	N9020A	23/06/23	24/06/23	US47360812
DC Power Supply	Spectrum Analyzer	Agilent Technologies	N9020A	23/06/23	24/06/23	MY46471622
Multimeter	DC Power Supply	Agilent Technologies	66332A	23/06/23	24/06/23	US37474125
BlueTooth Tester	DC Power Supply	SM techno	SDP30-5D	23/06/23	24/06/23	305DMG288
Power Spillter	Multimeter	FLUKE	17B+	23/12/15	24/12/15	36390701WS
Signal Generator Rohde Schwarz SMBV100A 23/12/15 24/12/15 255571 Signal Generator ANRITSU MG8898C 23/12/15 24/12/15 175501 Thermohygrometer BODYCOM BJ5478 23/12/15 24/12/15 120612-1 Thermohygrometer BODYCOM BJ5478 23/12/15 24/12/15 120612-2 Thermohygrometer BODYCOM BJ5478 23/12/15 24/12/15 120612-2 Loop Antenna ETS-Lindgren 6502 23/11/109 24/11/109 00060496 Hybrid Artenna ETS-Lindgren 6502 23/11/109 24/12/15 3362 Hom Antenna ETS-Lindgren 3117 23/06/23 24/06/23 00/43/278 Hom Antenna ETS-Lindgren 3117 23/06/23 24/06/23 155 PreAmpilifier tsj MLA-1840-J02-45 23/06/23 24/06/23 155 PreAmpilifier ttp 84/17 23/12/15 24/12/15 1865-67 PreAmpilifier ttp 84/17 </td <td>BlueTooth Tester</td> <td>TESCOM</td> <td>TC-3000C</td> <td>23/06/23</td> <td>24/06/23</td> <td>3000C000563</td>	BlueTooth Tester	TESCOM	TC-3000C	23/06/23	24/06/23	3000C000563
Signal Generator	Power Splitter	Anritsu	K241B	23/12/15	24/12/15	1301183
Thermohygrometer	Signal Generator	Rohde Schwarz	SMBV100A	23/12/15	24/12/15	255571
Thermohygrometer BODYCOM BJ5478 23/12/15 24/12/15 120612-2 Thermohygrometer BODYCOM BJ5478 23/06/23 24/06/23 NA Loop Antenna ETS-Lindgren 6502 23/11/19 24/11/19 00060496 Hybrid Antenna Schwarzbeck VULB 9160 23/12/15 24/12/15 3362 Horn Antenna ETS-Lindgren 3117 23/06/23 24/06/23 00143278 Horn Antenna Al-H. Systems Inc. SAS-574 23/06/23 24/06/23 155 PreAmplifier Its] MLA-0118-B01-40 23/12/15 24/12/15 1852267 PreAmplifier Its] MLA-1840-J02-45 23/06/23 24/06/23 16966-10728 PreAmplifier Its] MIA-1840-J02-45 23/06/23 24/06/23 16966-10728 PreAmplifier Its] MIA-1840-J02-45 23/06/23 24/06/23 1696-10728 PreAmplifier Its] MIA-1840-J02-45 23/06/23 24/06/23 169	Signal Generator	ANRITSU	MG3695C	23/12/15	24/12/15	173501
Thermothygrometer BODYCOM BJ5478 23/06/23 24/06/23 N/A	Thermohygrometer	BODYCOM	BJ5478	23/12/15	24/12/15	120612-1
Loop Antenna ETS-Lindgren 6502 23/11/09 24/11/09 00060496 Hybrid Antenna Schwarzbeck VULB 9160 23/12/15 24/12/15 3362 Hom Antenna ETS-Lindgren 3117 23/06/23 24/06/23 00143278 Hom Antenna A.H.Systems Inc. SAS-574 23/06/23 24/06/23 155 PreAmplifier tsj MLA-0118-801-40 23/12/15 24/12/15 185267 PreAmplifier tsj MLA-1840-J02-45 23/06/23 24/06/23 16966-10728 PreAmplifier H.P B447D 23/12/15 24/12/15 2944A07774 High Pass Filter Wainwright Instruments MYKX10-2838-3900-19/06/23 24/06/23 8 High Pass Filter Wainwright Instruments WHNX8.0/26.5-6SS 23/06/23 24/06/23 1 High Pass Filter Wainwright Instruments WHNX8.0/26.5-6SS 23/06/23 24/06/23 3 Attenuator Hefel Shuze SS572.92-10-40 23/06/23 24/06/23 3 Attenuator <td>Thermohygrometer</td> <td>BODYCOM</td> <td>BJ5478</td> <td>23/12/15</td> <td>24/12/15</td> <td>120612-2</td>	Thermohygrometer	BODYCOM	BJ5478	23/12/15	24/12/15	120612-2
Hybrid Antenna	Thermohygrometer	BODYCOM	BJ5478	23/06/23	24/06/23	N/A
Horn Antenna	Loop Antenna	ETS-Lindgren	6502	23/11/09	24/11/09	00060496
Hom Antenna	Hybrid Antenna	Schwarzbeck	VULB 9160	23/12/15	24/12/15	3362
PreAmplifier tsj	Horn Antenna	ETS-Lindgren	3117	23/06/23	24/06/23	00143278
PreAmplifier tsj MLA-1840-J02-45 23/06/23 24/06/23 16966-10728 PreAmplifier H.P. 8447D 23/12/15 24/12/15 2944A07774 High Pass Filter Wainwright Instruments WHKX12-935-1000-1500-40SS 23/06/23 24/06/23 8 High Pass Filter Wainwright Instruments WHKX10-2838-3300-1500-40 23/06/23 24/06/23 1 High Pass Filter Wainwright Instruments WHNX8.0/26.5-6SS 23/06/23 24/06/23 3 Altenuator Hefei Shunze SS5T2.92-10-40 23/06/23 24/06/23 1 Altenuator Aeroflex/Weinschel 56-3 23/06/23 24/06/23 7 Altenuator SMAJK SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 4 <td>Horn Antenna</td> <td>A.H.Systems Inc.</td> <td>SAS-574</td> <td>23/06/23</td> <td>24/06/23</td> <td>155</td>	Horn Antenna	A.H.Systems Inc.	SAS-574	23/06/23	24/06/23	155
PreAmplifier H.P. 8447D 23/12/15 24/12/15 2944A07774 High Pass Filter Wainwright Instruments WHKX12-935-1000-15000-40SS 23/06/23 24/06/23 8 High Pass Filter Wainwright Instruments WHKX10-2838-3300-1800-60SS 23/06/23 24/06/23 1 High Pass Filter Wainwright Instruments WHNX8.0/26.5-6SS 23/06/23 24/06/23 3 Attenuator Hefe Shunze SS572.92-10-40 23/06/23 24/06/23 1 Attenuator Aeroflex/Weinschel 66-3 23/06/23 24/06/23 3 Attenuator SMAJK SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator SMAJK SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator Amylk SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator Amylk SMAJK-2-3 23/06/23 24/06/23 2 Attenuator Amylk SMAJK-2-3 23/06/23 24/06/23 3	PreAmplifier	tsj	MLA-0118-B01-40	23/12/15	24/12/15	1852267
High Pass Filter Wainwright Instruments WHKX12-935-1000-1500-40SS 23/06/23 24/06/23 8 High Pass Filter Wainwright Instruments WHKX10-2838-3300-1600-60SS 23/06/23 24/06/23 1 High Pass Filter Wainwright Instruments WHNX0-06SS 23/06/23 24/06/23 3 Attenuator Hefei Shunze SS5T2.92-10-40 23/06/23 24/06/23 16012202 Attenuator Aeroflex/Weinschel 56-3 23/06/23 24/06/23 3 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor Anritsu ML2496A 23/12/15 24/12/15 1338004 EMI Test Receiver ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-22 23/08/21 24/08/21	PreAmplifier	tsj	MLA-1840-J02-45	23/06/23	24/06/23	16966-10728
High Pass Filter Wainwright Instruments 15000-40SS 23/06/23 24/06/23 1	PreAmplifier	H.P	8447D	23/12/15	24/12/15	2944A07774
High Pass Filter Wainwright Instruments 18000-60SS 23/06/23 24/06/23 1 High Pass Filter Wainwright Instruments WHNX8.0/26.5-6SS 23/06/23 24/06/23 1 Attenuator Hefel Shunze SSF12.92-10-40 23/06/23 24/06/23 16012202 Attenuator Aeroflex/Weinschel 56-3 23/06/23 24/06/23 2 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor MIL2496A 23/12/15 24/12/15 1338004 FULSE LIMITER ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-Z2 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 22/000253 Cable	High Pass Filter	Wainwright Instruments	15000-40SS	23/06/23	24/06/23	8
Attenuator Hefei Shunze SS5T2.92-10-40 23/06/23 24/06/23 16012202 Attenuator Aeroflex/Weinschel 56-3 23/06/23 24/06/23 24/06/23 3 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor Anritsu M2496A 23/12/15 24/12/15 1911481 EMI Test Receiver ROHDE&SCHWARZ ESCIT 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-22 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Dt&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 M-01 Cable Junkosha MWX241 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 M-01 Cable Junkosha MWX241 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241 24/01/03 25/01/03 M-03 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03	High Pass Filter	Wainwright Instruments		23/06/23	24/06/23	1
Attenuator Aeroflex/Weinschel 56-3 23/06/23 24/06/23 Y2370 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor Anritsu ML2496A 23/12/15 24/12/15 1338004 Bandwidth Sensor ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-22 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Di&C Cable 24/01/03 25/01/03 G-2 Cable Di&C Cable </td <td>High Pass Filter</td> <td>Wainwright Instruments</td> <td>WHNX8.0/26.5-6SS</td> <td>23/06/23</td> <td>24/06/23</td> <td>3</td>	High Pass Filter	Wainwright Instruments	WHNX8.0/26.5-6SS	23/06/23	24/06/23	3
Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 3 Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor Anritsu ML2496A MA2411B 23/12/15 24/12/15 1338004 1911481 EMI Test Receiver ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-Z2 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Di&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable HUBER+SUHNER	Attenuator	Hefei Shunze	SS5T2.92-10-40	23/06/23	24/06/23	16012202
Attenuator SMAJK SMAJK-2-3 23/06/23 24/06/23 2 Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor Anritsu ML2496A 23/12/15 24/12/15 1338004 Bandwidth Sensor ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-Z2 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable D1&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable HUBER+SUHNER SUCOFL	Attenuator	Aeroflex/Weinschel	56-3	23/06/23	24/06/23	Y2370
Attenuator Aeroflex/Weinschel 86-10-11 23/06/23 24/06/23 408 Power Meter & Wide Bandwidth Sensor Anritsu ML2496A MA2411B 23/12/15 24/12/15 1338004 1911481 EMI Test Receiver ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-Z2 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable D1&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Junkosha MWX241 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFL	Attenuator	SMAJK	SMAJK-2-3	23/06/23	24/06/23	3
Power Meter & Wide Bandwidth Sensor	Attenuator	SMAJK	SMAJK-2-3	23/06/23	24/06/23	2
Bandwidth Sensor Anritsu MA2411B 23/12/15 24/12/15 1911481 EMI Test Receiver ROHDE&SCHWARZ ESCI7 24/01/29 25/01/29 100910 PULSE LIMITER ROHDE&SCHWARZ ESH3-Z2 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Dt&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-5 Cable OMT YSS21S 24/01/03 25/01/03 M-6 Cable Junkosha MWX241 24/01/03 25/01/03 mmw-1 Cable Junkosha MWX241 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03	Attenuator	Aeroflex/Weinschel	86-10-11	23/06/23	24/06/23	408
PULSE LIMITER ROHDE&SCHWARZ ESH3-Z2 23/08/21 24/08/21 101333 LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Dt&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-4 Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 </td <td>Power Meter & Wide Bandwidth Sensor</td> <td></td> <td>MA2411B</td> <td></td> <td></td> <td>1911481</td>	Power Meter & Wide Bandwidth Sensor		MA2411B			1911481
LISN SCHWARZBECK NSLK 8128 RC 23/10/26 24/10/26 8128 RC-387 Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Dt&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-4 Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 <	EMI Test Receiver	ROHDE&SCHWARZ	ESCI7	24/01/29	25/01/29	100910
Digital Thermo Hygrometer CAS TE-303N 24/02/07 25/02/07 220502531 Cable Dt&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-4 Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-07	PULSE LIMITER	+				
Cable Dt&C Cable 24/01/03 25/01/03 G-2 Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-4 Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-07 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cabl	LISN	SCHWARZBECK	NSLK 8128 RC	23/10/26	24/10/26	8128 RC-387
Cable HUBER+SUHNER SUCOFLEX 100 24/01/03 25/01/03 G-3 Cable Dt&C Cable 24/01/03 25/01/03 G-4 Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-07 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70	Digital Thermo Hygrometer	CAS	TE-303N	24/02/07	25/02/07	220502531
Cable Dt&C Cable 24/01/03 25/01/03 G-4 Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software tsj EMI Measurement NA NA Version V	Cable	Dt&C	Cable	24/01/03	25/01/03	G-2
Cable OMT YSS21S 24/01/03 25/01/03 G-5 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software tsj EMI Measurement NA NA Version Cable FMI Measurement NA NA Version	Cable	HUBER+SUHNER	SUCOFLEX 100	24/01/03	25/01/03	G-3
Cable Junkosha MWX241 24/01/03 25/01/03 mmW-1 Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA NA Version Test Software tsi FMI Measurement NA NA Version	Cable	Dt&C	Cable	24/01/03	25/01/03	G-4
Cable Junkosha MWX241 24/01/03 25/01/03 mmW-4 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA NA Version Test Software tsi FMI Measurement NA NA Version	Cable	OMT	YSS21S	24/01/03	25/01/03	G-5
Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-01 Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA Version Test Software tsi FMI Measurement NA NA Version	Cable	Junkosha	MWX241	24/01/03	25/01/03	mmW-1
Cable HUBER+SUHNER SUCOFLEX100 24/01/03 25/01/03 M-02 Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA Version Test Software tsi FMI Measurement NA NA Version	Cable	Junkosha	MWX241	24/01/03	25/01/03	mmW-4
Cable JUNKOSHA MWX241/B 24/01/03 25/01/03 M-03 Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA Version 2.00.0190 Test Software tsi FMI Measurement NA NA Version	Cable	HUBER+SUHNER	SUCOFLEX100	24/01/03	25/01/03	M-01
Cable JUNKOSHA J12J101757-00 24/01/03 25/01/03 M-07 Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA NA Version 2.00.0190 Test Software tsi FMI Measurement NA NA Version	Cable	HUBER+SUHNER	SUCOFLEX100	24/01/03	25/01/03	M-02
Cable HUBER+SUHNER SUCOFLEX106 24/01/03 25/01/03 M-09 Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA NA Version 2.00.0190 Test Software tsi FMI Measurement NA NA Version	Cable	JUNKOSHA	MWX241/B	24/01/03	25/01/03	M-03
Cable Dt&C Cable 24/01/03 25/01/03 RFC-69 Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA NA Version 2.00.0190 Test Software tsi FMI Measurement NA NA Version	Cable	JUNKOSHA	J12J101757-00	24/01/03	25/01/03	M-07
Cable Radiall TESTPRO3 24/01/03 25/01/03 RFC-70 Test Software (AC Line Conducted) tsj EMI Measurement NA NA Version 2.00.0190 Test Software tsi FMI Measurement NA NA Version	Cable	HUBER+SUHNER	SUCOFLEX106	24/01/03	25/01/03	M-09
Test Software (AC Line Conducted) tsj EMI Measurement NA NA Version 2.00.0190 Test Software tsj EMI Measurement NA NA Version	Cable	Dt&C	Cable	24/01/03	25/01/03	RFC-69
(AC Line Conducted) Test Software tsi EMI Measurement NA NA 2.00.0190 Version	Cable	Radiall	TESTPRO3	24/01/03	25/01/03	RFC-70
I for I FIMI Measurement I NA I NA I	Test Software (AC Line Conducted)	tsj	EMI Measurement	NA	NA	
		tsj	EMI Measurement	NA	NA	

Report No.: DRTFCC2407-0072

Note1: The measurement antennas were calibrated in accordance to the requirements of ANSI C63.5-2017

Note2: The cable is not a regular calibration item, so it has been calibrated by Dt&C itself.

Report No.: DRTFCC2407-0072 FCC ID: RPJ-FSC-BT982

2. Antenna Requirement

An intentional radiator shall be designed to ensure that no antenna other than that furnished by the responsible party shall be used with the device. The use of a permanently attached antenna or of an antenna that uses a unique coupling to the intentional radiator shall be considered sufficient to comply with the provisions.

Conclusion: Comply

The antenna is permanently attached on the device. Therefore this E.U.T complies with the requirement of Part 15.203

Pages: 9 / 42



3. Summary of Test Results

FCC part section(s)	Test Description	st Description Limit (Using in 2 400~ 2 483.5 MHz)		Status Note 1
15.247(a) 15.247(b)	Maximum Peak Conducted Output Power	=< 0.125 W(conducted)		С
	20 dB Bandwidth	NA	_	С
15.247(a)	Carrier Frequency Separation	>= 25 kHz or >= Two thirds of the 20 dB BW, whichever is greater.	Conducted	С
13.247 (a)	Number of Hopping Channels	>= 15 hops		С
	Time of Occupancy	=< 0.4 seconds		С
Unwanted Emissions (Conducted)		The radiated emission to any 100 kHz of out-band shall be at least 20 dB below the highest in-band spectral density.	_	С
15.247(d) 15.205 15.209	Unwanted Emissions (Radiated)	Part 15.209 Limits (Refer to section 9)	Radiated	C Note3
15.207	AC Power-Line Conducted Emissions	Part 15.207 Limits (Refer to section 10)	AC Line Conducted	С
15.203	Antenna Requirement	Part 15.203 (Refer to section 2)	-	С

Report No.: DRTFCC2407-0072

Note 1: C = Comply NC = Not Comply NT = Not Tested NA = Not Applicable

Note 2: For radiated emission tests below 30 MHz were performed on semi-anechoic chamber which is correlated with OATS.

Note 3: This test item was performed in three orthogonal EUT positions and the worst case data was reported.



4. Maximum Peak Conducted Output Power

4.1. Test Setup

Refer to the APPENDIX I.

4.2. Limit

■ FCC Requirements

The maximum peak output power of the intentional radiator shall not exceed the following:

- 1. §15.247(a)(1), Frequency hopping systems shall have hopping channel carrier frequencies separated by a minimum of 25 kHz or the 20 dB bandwidth of the hopping channel, whichever is greater. Alternatively, frequency hopping systems operating in the 2 400 MHz 2 483.5 MHz band may have hopping channel carrier frequencies that are separated by 25 kHz or two-thirds of the 20 dB bandwidth of the hopping channel, whichever is greater, provided the systems operate with an output power no greater than 125 mW.
- 2. §15.247(b)(1), For frequency hopping systems operating in the 2 400 2 483.5 MHz employing at least 75 non-overlapping hopping channels, and all frequency hopping systems in the 5 725 MHz 5 805 MHz band : 1 Watt. For all other frequency hopping systems in the 2 400 MHz 2 483.5 MHz band: 0.125 watts.

4.3. Test Procedure

- 1. The RF output power was measured with a spectrum analyzer connected to the RF Antenna connector (conducted measurement) while EUT was operating in transmit mode at the appropriate center frequency, A spectrum analyzer was used to record the shape of the transmit signal.
- 2. The peak output power of the fundamental frequency was measured with the spectrum analyzer using;

Span = approximately 5 times of the 20 dB bandwidth, centered on a hopping channel

RBW ≥ 20 dB BW

VBW ≥ RBW

Sweep = auto

Detector function = peak

Trace = max hold




4.4. Test Results

Modulation	Tested Channel	Burst Average Output Power		Peak Output Power	
Wodulation		dBm	mW	dBm	mW
	Lowest	-1.01	0.79	-0.86	0.82
<u>GFSK</u>	Middle	-1.36	0.73	-1.22	0.76
	Highest	-2.20	0.60	-1.99	0.63

Note 1: The average output power was tested using an average power meter for reference only.

Note 2: See next pages for actual measured spectrum plots.











Middle Channel & Modulation : GFSK



Peak Output Power

Highest Channel & Modulation : GFSK





5. 20 dB BW

5.1. Test Setup

Refer to the APPENDIX I.

5.2. Limit

Limit: Not Applicable

5.3. Test Procedure

- 1. The 20 dB bandwidth was measured with a spectrum analyzer connected to RF antenna Connector (conducted measurement) while EUT was operating in transmit mode. The analyzer center frequency was set to the EUT carrier frequency, using the analyzer.
- 2. The bandwidth of the fundamental frequency was measured with the spectrum analyzer using below setting:

RBW = 1 % to 5 % of the 20 dB BW

VBW ≥ 3 x RBW

Span = between two times and five times the 20 dB bandwidth

Sweep = auto

Detector function = peak

Trace = max hold

5.4. Test Results

Modulation	Tested Channel	20 dB BW (MHz)
	Lowest	0.933
<u>GFSK</u>	Middle	0.935
	Highest	0.931





Lowest Channel & Modulation : GFSK



20 dB BW

Middle Channel & Modulation : GFSK



TRF-RF-237(07)210316 Pages: 15 / 42



20 dB BW

Highest Channel & Modulation: GFSK





6. Carrier Frequency Separation

6.1. Test Setup

Refer to the APPENDIX I.

6.2. Limit

Limit: ≥ 25 kHz or ≥ Two-Thirds of the 20 dB BW whichever is greater.

6.3. Test Procedure

The carrier frequency separation was measured with a spectrum analyzer connected to the antenna terminal, while EUT had its hopping function enabled.

After the trace being stable, the reading value between the peaks of the adjacent channels using the markerdelta function was recorded as the measurement results.

The spectrum analyzer is set to:

Span = wide enough to capture the peaks of two adjacent channels

RBW = Start with the RBW set to approximately 30 % of the channel spacing; adjust as necessary to

best identify the center of each individual channel.

VBW ≥ RBW Sweep = auto
Detector function = peak Trace = max hold

6.4. Test Results

FH mode

Hopping Mode	Modulation	Peak of reference channel(MHz)	Peak of adjacent Channel(MHz)	Test Result (MHz)
Enable	GFSK	2 440.062	2 441.059	0.997

AFH mode

Hopping Mode	Modulation	Peak of reference channel(MHz)	Peak of adjacent Channel(MHz)	Test Result (MHz)
Enable	GFSK	2441.063	2442.063	1.000

Note 1: See next pages for actual measured spectrum







Carrier Frequency Separation (AFH) Hopping mode : Enable&GFSK



Pages: 18 / 42



7. Number of Hopping Channels

7.1. Test Setup

Refer to the APPENDIX I.

7.2. Limit

Limit: >= 15 hops

7.3. Test Procedure

The number of hopping frequencies was measured with a spectrum analyzer connected to the antenna terminal, while EUT had its hopping function enabled.

To get higher resolution, two frequency ranges for FH mode within the 2 400 MHz ~ 2 483.5 MHz were examined.

The spectrum analyzer is set to:

Span for FH mode = 50 MHz Start Frequency = 2 391.5 MHz, Stop Frequency = 2 441.5 MHz

Start Frequency = 2 441.5 MHz, Stop Frequency = 2 491.5 MHz

FCC ID: RPJ-FSC-BT982

Span for AFH mode = 30 MHz Start Frequency = 2 426.0 MHz, Stop Frequency = 2 456.0 MHz

RBW = To identify clearly the individual channels, set the RBW to less than 30 % of the channel spacing

or the 20 dB bandwidth, whichever is smaller.

VBW ≥ RBW

Sweep = auto

7.4. Test Results

FH mode

Hopping mode	Modulation	Test Result (Total Hops)			
Enable	GFSK	79			


AFH mode

Hopping mode	Modulation	Test Result (Total Hops)
Enable	GFSK	20


Note 1 : See next pages for actual measured spectrum plots.







Number of Hopping Channels 2(FH) <u>Hopping mode : Enable & GFSK</u>







STATUS

TRF-RF-237(07)210316 Pages: 21 / 42

Report No.: DRTFCC2407-0072 FCC ID: RPJ-FSC-BT982

8. Time of Occupancy

8.1. Test Setup

Refer to the APPENDIX I.

8.2. Limit

The maximum permissible time of occupancy is 400 ms within a period of 400 ms multiplied by the number of hopping channels employed.

8.3. Test Procedure

The dwell time was measured with a spectrum analyzer connected to the antenna terminal, while EUT had its hopping function enabled.

The spectrum analyzer is set to:

Center frequency = 2 441 MHz Span = zero

RBW = 1 MHz (RBW shall be ≤ channel spacing and where possible RBW should be set >> 1 / T, where T is the expected dwell time per channel)

VBW ≥ RBW Detector function = peak

Trace = max hold

8.4. Test Results

FH mode

Hopping mode	Packet Type	Number of hopping Channels	Burst On Time (ms)	Test Result (sec)
Enable	DH 5	79	2.835	0.302

AFH mode

Hopping mode	Packet Type	Number of hopping Channels	Burst On Time (ms)	Test Result (sec)
Enable	DH 5	20	2.835	0.151

Note 1 : Dwell Time = 0.4 x Hopping channel x Burst ON time x


((Hopping rate ÷ Time slots) ÷ Hopping channel)

- Time slots for DH5 = 6 slots (TX = 5 slots / RX = 1 slot)
- Hopping Rate = 1 600 for FH mode & 800 for AFH mode

Note 2: See next pages for actual measured spectrum plots.

TRF-RF-237(07)210316 Pages: 22 / 42











9. Unwanted Emissions

9.1. Test Setup

Refer to the APPENDIX I.

9.2. Limit

Part 15.247(d), Part 15.205, Part 15.209

In any 100 kHz bandwidth outside the frequency band in which the spread spectrum or digitally modulated intentional radiator is operating, the radio frequency power that is produced by the intentional radiator shall be at least 20 dB below that in the 100 kHz bandwidth within the band that contains the highest level of the desired power, based on either an RF conducted or a radiated measurement, provided the transmitter demonstrates compliance with the peak conducted power limits. If the transmitter complies with the conducted power limits based on the use of RMS averaging over a time interval, as permitted under paragraph (b)(3) of Part 15.247 the attenuation required under this paragraph shall be 30 dB instead of 20 dB. Attenuation below the general limits specified in §15.209(a) is not required. In addition, radiated emissions which fall in the restricted bands, as defined in §15.205(a), must also comply with the radiated emission limits specified in §15.205(c)).

- Part 15.209: General requirement

Frequency (MHz)	FCC Limit (uV/m)	Measurement Distance (m)
0.009 - 0.490	2 400 / F (kHz)	300
0.490 - 1.705	24 000 / F (kHz)	30
1.705 – 30.0	30	30

Frequency (MHz)	FCC Limit (uV/m)	Measurement Distance (m)
30 ~ 88	100 **	3
88 ~ 216	150 **	3
216 ~ 960	200 **	3
Above 960	500	3

^{**}Except as provided in paragraph (g), fundamental emissions from intentional radiators operating under this section shall not be located in the frequency bands 54-72 MHz, 76-88 MHz, 174-216 MHz or 470-806 MHz. However, operation within these frequency bands is permitted under other sections of this part, e.g., §15.231 and 15.241.

- Part 15.205(a): Restricted band of operation

MHz	MHz	MHz	MHz	GHz	GHz
0.009 ~ 0.110	8.414 25 ~ 8.414 75	108 ~ 121.94	1 300 ~ 1 427	4.5 ~ 5.15	14.47 ~ 14.5
0.495 ~ 0.505	12.29 ~ 12.293	123 ~ 138	1 435 ~ 1 626.5	5.35 ~ 5.46	15.35 ~ 16.2
2.173 5 ~ 2.190 5	12.519 75 ~ 12.520 25	149.9 ~ 150.05	1 645.5 ~ 1 646.5	7.25 ~ 7.75	17.7 ~ 21.4
4.125 ~ 4.128	12.576 75 ~ 12.577 25	156.524 75 ~ 156.525 25	1 660 ~ 1 710	8.025 ~ 8.5	22.01 ~ 23.12
4.177 25 ~ 4.177 75	13.36 ~ 13.41	156.7 ~ 156.9	1 718.8 ~ 1 722.2	9.0 ~ 9.2	23.6 ~ 24.0
4.207 25 ~ 4.207 75	16.42 ~ 16.423	162.012 5 ~ 167.17	2 200 ~ 2 300	9.3 ~ 9.5	31.2 ~ 31.8
6.215 ~ 6.218	16.694 75 ~ 16.695 25	167.72 ~ 173.2	2 310 ~ 2 390	10.6 ~ 12.7	36.43 ~ 36.5
6.267 75 ~ 6.268 25	16.804 25 ~ 16.804 75	240 ~ 285	2 483.5 ~ 2 500	13.25 ~ 13.4	Above 38.6
6.311 75 ~ 6.312 25	25.5 ~ 25.67	322 ~ 335.4	2 655 ~ 2 900		
8.291 ~ 8.294	37.5 ~ 38.25	399.90 ~ 410	3 260 ~ 3 267		
8.362 ~ 8.366	73 ~ 74.6	608 ~ 614	3 332 ~ 3 339		
8.376 25 ~ 8.386 75	74.8 ~ 75.2	960 ~ 1 240	3 345.8 ~ 3 358		
			3 600 ~ 4 400		

TRF-RF-237(07)210316 Pages: 24 / 42



9.3. Test Procedures

9.3.1. Test Procedures for Unwanted Emissions(Radiated)

- 1. The EUT is placed on a non-conductive table. For emission measurements at or below 1 GHz, the table height is 80 cm. For emission measurements above 1 GHz, the table height is 1.5 m. The table was rotated 360 degrees to determine the position of the highest radiation.
- 2. During performing radiated emission below 1 GHz, the EUT was set 3 meters away from the interference receiving antenna, which was mounted on the top of a variable-height antenna tower. During performing radiated emission above 1 GHz, the EUT was set 1 or 3 meter away from the interference-receiving antenna.
- 3. For measurements above 1 GHz absorbers are placed on the floor between the turn table and the antenna mast in such a way so as to maximize the reduction of reflections. For measurements below 1 GHz, the absorbers are removed.
- 4. The antenna is a broadband antenna, and its height is varied from one meter to four meters above the ground to determine the maximum value of the field strength. Both horizontal and vertical polarizations of the antenna are set to make the measurement.
- 5. For each suspected emission, the EUT was arranged to its worst case and then the antenna was tuned to heights from 1 meter to 4 meters and the table was turned from 0 degrees to 360 degrees to find the maximum reading.
- 6. The test-receiver system was set to Peak Detect Function and Specified Bandwidth with Maximum Hold Mode.
- 7. If the emission level of the EUT in peak mode was 10 dB lower than the limit specified, then testing could be stopped and the peak values of the EUT would be reported. Otherwise the emissions that did not have 10 dB margin would be re-tested one by one using peak, quasi-peak or average method as specified and then reported in a data sheet.

Measurement Instrument Setting

- Frequencies less than or equal to 1 000 MHz

 The resolution bandwidth and video bandwidth of test receiver/spectrum analyzer is 120 kHz for Quasipeak detection (QP) at frequency below 1 GHz.
- Frequencies above 1 000 MHz

The resolution bandwidth and video bandwidth of test receiver/spectrum analyzer is 1 MHz for Peak detection and frequency above 1 GHz.

The result of Average measurement is calculated using PK result and duty correction factor.

Report No.: DRTFCC2407-0072 FCC ID: RPJ-FSC-BT982

9.3.2. Test Procedures for Unwanted Emissions(Conducted)

- 1. The transmitter output was connected to the spectrum analyzer.
- 2. The **reference level** of the fundamental frequency was measured with the spectrum analyzer using RBW = 100 kHz, VBW = 300 kHz.
- 3. The conducted spurious emission was tested each ranges were set as below.

Frequency range: 9 kHz ~ 30 MHz

RBW = 100 kHz, VBW = 300 kHz, SWEEP TIME = AUTO, DETECTOR = PEAK, TRACE = MAX HOLD, SWEEP POINT : 40 001

Frequency range: 30 MHz ~ 10 GHz, 10 GHz ~ 25 GHz

RBW = 1 MHz, VBW = 3 MHz, SWEEP TIME = AUTO, DETECTOR = PEAK, TRACE = MAX HOLD, SWEEP POINT : 40 001

LIMIT LINE = 20 dB below of the reference level of above measurement procedure Step 2. (RBW = 100 kHz, VBW = 300 kHz)

If the emission level with above setting was close to the limit (ie, less than 3 dB margin) then zoom scan is required using RBW = 100 kHz, VBW = 300 kHz, SPAN = 100 MHz and BINS = 2 001 to get accurate emission level within 100 kHz BW.

Also the path loss for conducted measurement setup was used as described on the Appendix I of this test report.



9.4. Test Results

9.4.1. Unwanted Emissions(Radiated)

■ Test Notes.

- 1. The radiated emissions below 1 GHz were investigated from 9 kHz and the worst case data was reported.
- 2. Information of Distance Correction Factor

For finding emissions, measurements may be performed at a distance closer than that specified in the regulations. In this case, the distance correction factor is applied to the result.

- Calculation of distance factor

At frequencies below 30 MHz = 40 log(tested distance / specified distance)

At frequencies at or above 30 MHz = 20 log(tested distance / specified distance)

When distance factor is "N/A", the measurements were performed at the specified distance and distance factor is not applied.

3. Sample Calculation.

Margin = Limit - Result / Result = Reading + TF+ DCCF + DCF / TF = AF + CL + HL + AL - AG

Where, TF = Total Factor, AF = Antenna Factor, CL = Cable Loss, AG = Amplifier Gain, HL = High pass filter Loss,


AL = Attenuator Loss, DCCF = Duty Cycle Correction Factor, DCF = Distance Correction Factor

9 kHz ~ 1 GHz Data (Modulation : GFSK)

Lowest Channel

Frequency (MHz)	ANT Pol	EUT Position (Axis)	Detector Mode	Reading (dBuV)	TF (dB/m)	DCCF (dB)	DCF (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)
35.82	V	Х	PK	32.50	-9.38	N/A	N/A	23.12	40.00	16.88
144.46	Н	X	PK	28.00	-6.64	N/A	N/A	21.36	43.50	22.14
800.39	Н	Х	PK	33.40	5.40	N/A	N/A	38.80	46.00	7.20
983.50	V	Χ	PK	25.40	8.70	N/A	N/A	34.10	54.00	19.90

TM1 & Lowest & X & Hor





■ Test Notes.

- 1. The radiated emissions above 1 GHz were investigated up to 25 GHz. And no other spurious and harmonic emissions were found below listed frequencies.
- 2. Information of Distance Correction Factor

For finding emissions, measurements may be performed at a distance closer than that specified in the regulations. In this case, the distance correction factor is applied to the result.

- Calculation of distance factor

At frequencies below 30 MHz = 40 log(tested distance / specified distance)

At frequencies at or above 30 MHz = 20 log(tested distance / specified distance)

When distance factor is "N/A", the measurements were performed at the specified distance and distance factor is not applied.

- 3. DCCF Calculation. (DCCF = Duty Cycle Correction Factor)
 - Time to cycle through all channels = $\Delta t = T$ [ms] X 20 minimum hopping channels , where T = pulse width = 2.835 ms
 - 100 ms / Δt [ms] = H -> Round up to next highest integer, to account for worst case, H' = 100 / (2.835 X 20) = 1.76 = 2
 - The Worst Case Dwell Time = T [ms] x H' = 2.835 ms X 2 = 5.67 ms
 - DCCF = 20 Log(The Worst Case Dwell Time / 100 ms) dB = 20 log(5.67 / 100) = -24.93 dB
- 4. Sample Calculation.

Margin = Limit - Result / Result = Reading + TF+ DCCF + DCF / TF = AF + CL + HL + AL - AG

Where, TF = Total Factor, AF = Antenna Factor, CL = Cable Loss, AG = Amplifier Gain, HL = High pass filter Loss,

AL = Attenuator Loss, DCCF = Duty Cycle Correction Factor, DCF = Distance Correction Factor

1 GHz ~ 25 GHz Data (Modulation: GFSK)

Lowest Channel

Frequency (MHz)	ANT Pol	EUT Position (Axis)	Detector Mode	Reading (dBuV)	TF (dB/m)	DCCF (dB)	DCF (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)
2 389.99	Н	X	PK	50.09	4.48	N/A	N/A	54.57	74.00	19.43
2 389.99	Н	Х	AV	50.09	4.48	-24.93	N/A	29.64	54.00	24.36
4 003.47	Н	Х	PK	55.07	0.62	N/A	N/A	55.69	74.00	18.31
4 003.26	Н	Х	AV	55.07	0.62	-24.93	N/A	30.76	54.00	23.24
4 803.79	Η	Х	PK	50.84	1.64	N/A	N/A	52.48	74.00	21.52
4 803.79	Н	X	AV	50.84	1.64	-24.93	N/A	27.55	54.00	26.45

Middle Channel

Frequency (MHz)	ANT Pol	EUT Position (Axis)	Detector Mode	Reading (dBuV)	TF (dB/m)	DCCF (dB)	DCF (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)
4 068.22	Н	Х	PK	54.75	0.70	N/A	N/A	55.45	74.00	18.55
4 068.27	Τ	Х	AV	54.75	0.70	-24.93	N/A	30.52	54.00	23.48
4 879.67	Η	X	PK	49.78	1.79	N/A	N/A	51.57	74.00	22.43
4 879.67	Н	X	AV	49.78	1.79	-24.93	N/A	26.64	54.00	27.36


Highest Channel

Frequency (MHz)	ANT Pol	EUT Position (Axis)	Detector Mode	Reading (dBuV)	TF (dB/m)	DCCF (dB)	DCF (dB)	Result (dBuV/m)	Limit (dBuV/m)	Margin (dB)
2 483.74	Н	X	PK	60.14	4.98	N/A	N/A	65.12	74.00	8.88
2 483.74	Н	X	AV	60.14	4.98	-24.93	N/A	40.19	54.00	13.81
3 306.86	Н	X	PK	54.25	0.68	N/A	N/A	54.93	74.00	19.07
3 306.55	Н	X	AV	54.25	0.68	-24.93	N/A	30.00	54.00	24.00
4 133.31	Н	X	PK	54.82	0.85	N/A	N/A	55.67	74.00	18.33
4 133.41	Н	X	AV	54.82	0.85	-24.93	N/A	30.74	54.00	23.26
4 959.03	Н	X	PK	48.87	2.52	N/A	N/A	51.39	74.00	22.61
4 959.03	Н	Χ	AV	48.87	2.52	-24.93	N/A	26.46	54.00	27.54




9.4.2. Unwanted Emissions(Conducted)






Low Band-edge <u>Hopping mode & Modulation : GFSK</u>



TRF-RF-237(07)210316 Pages: 29 / 42




Conducted Spurious Emissions <u>Lowest Channel & Modulation : GFSK</u>














Reference for limit

Middle Channel & Modulation : GFSK





Conducted Spurious Emissions <u>Middle Channel & Modulation : GFSK</u>





Conducted Spurious Emissions <u>Middle Channel & Modulation : GFSK</u>










Highest Channel & Modulation: GFSK



High Band-edge


Hopping mode & Modulation : GFSK




TRF-RF-237(07)210316 Pages: 34 / 42



Conducted Spurious Emissions Highest Channel & Modulation : GFSK













10. AC Power-Line Conducted Emissions

10.1. Test Setup

See test photographs for the actual connections between EUT and support equipment.

10.2. Limit

According to §15.207(a) for an intentional radiator that is designed to be connected to the public utility (AC) power line, the radio frequency voltage that is conducted back onto the AC power line on any frequency or frequencies, within the band 150 kHz to 30 MHz, shall not exceed the limits in the following table, as measured using a 50 uH/50 ohm line impedance stabilization network (LISN).

Compliance with the provision of this paragraph shall on the measurement of the radio frequency voltage between each power line and ground at the power terminal. The lower applies at the boundary between the frequency ranges.

Fraguency Pango (MHz)	Conducted Limit (dBuV)				
Frequency Range (MHz)	Quasi-Peak	Average			
0.15 ~ 0.50	66 to 56 *	56 to 46 *			
0.5 ~ 5.0	56	46			
5 ~ 30	60	50			

^{*} Decreases with the logarithm of the frequency

10.3. Test Procedure

Conducted emissions from the EUT were measured according to the ANSI C63.10.

- 1. The test procedure is performed in a 6.5 m × 3.5 m × 3.5 m (L × W × H) shielded room. The EUT along with its peripherals were placed on a 1.0 m (W) × 1.5 m (L) and 0.8 m in height wooden table and the EUT was adjusted to maintain a 0.4 meter space from a vertical reference plane.
- 2. The EUT was connected to power mains through a line impedance stabilization network (LISN) which provides 50 ohm coupling impedance for measuring instrument and the chassis ground was bounded to the horizontal ground plane of shielded room.
- 3. All peripherals were connected to the second LISN and the chassis ground also bounded to the horizontal ground plane of shielded room.
- 4. The excess power cable between the EUT and the LISN was bundled. The power cables of peripherals were unbundled. All connecting cables of EUT and peripherals were moved to find the maximum emission.



10.4. Test Results

AC Power-Line Conducted Emissions (Graph) = Modulation : GFSK

Results of Conducted Emission

Report No.: DRTFCC2407-0072

Date 2024-03-06

Order No. Power Supply Temp/Humi/Atm Test Condition Memo

21 'C / 41 %

1M_2402_QP


Model No. Operator


FSC_BT982 S.H.Han

LIMIT : FCC P15.207 AV FCC P15.207 QP

Lisn Factor 1. NSLK 8128 RC-387_N_23.10.26 2. NSLK 8128 RC-387_L1_23.10.26 Cable Loss
1. C1_LISN TO RECIVER_2023-12-11

Pulse Lmitter
1. PULSE LIMITER_ESH3-Z2_101333_2023.08.21





TRF-RF-237(07)210316 Pages: 38 / 42



AC Power-Line Conducted Emissions (List) = Modulation : <u>GFSK</u>

Results of Conducted Emission

Report No.: DRTFCC2407-0072

Date 2024-03-06

Order No. Power Supply Temp/Humi/Atm

21 'C / 41 % Model No. FSC_BT982 S.H.Han Test Condition Operator

1M_2402_QP Memo

LIMIT : FCC P15.207 AV FCC P15.207 QP

Lisn Factor

1. NSLK 8128 RC-387_N_23.10.26 2. NSLK 8128 RC-387_L1_23.10.26

Cable Loss
1. C1_LISN TO RECIVER_2023-12-11

Pulse Lmitter
1. PULSE LIMITER_ESH3-Z2_101333_2023.08.21


NC	FREQ	READING QP CAV [dBuV][dBu		RESULT QP CAV [dBuV][dBuV]	LIMIT QP CAV [dBuV][dBuV]	QP CAV	PHASE
1	0.26513	22.84 14.4	5 9.89	32.73 24.34	61.27 51.27	28.54 26.93	N
2	0.27672	23.46 16.6	5 9.89	33.35 26.54	60.91 50.91	27.56 24.37	N
3	0.65251	28.44 17.1	8 9.91	38.35 27.09	56.00 46.00	17.65 18.91	N
4	1.29540	16.14 4.8	0 9.93	26.07 14.73	56.00 46.00	29.93 31.27	N
5	5.47300	8.71 -0.7	5 10.03	18.74 9.28	60.00 50.00	41.26 40.72	N
6	25.15800	0.89 -4.0	6 10.34	11.23 6.28	60.00 50.00	48.77 43.72	N
7	0.17250	26.47 13.0	5 9.89	36.36 22.94	64.84 54.84	28.48 31.90	L
8	0.45058	28.02 20.2	6 9.90	37.92 30.16	56.86 46.86	18.94 16.70	L
9	0.64978	34.13 25.9	3 9.91	44.04 35.84	56.00 46.00	11.96 10.16	L
10	1.81460	20.70 11.0	7 9.95	30.65 21.02	56.00 46.00	25.35 24.98	L
11	5.23540	13.88 5.9	9 10.02	23.90 16.01	60.00 50.00	36.10 33.99	L
12	20.93600	0.54 -4.4	7 10.29	10.83 5.82	60.00 50.00	49.17 44.18	L




APPENDIX I

Test set up diagrams

Radiated Measurement



Conducted Measurement



Path loss information

Frequency (GHz)	Path Loss (dB)	Frequency (GHz)	Path Loss (dB)	
0.03	6.53	15	7.83	
1	6.83	20	8.09	
2.402 & 2.441 & 2.480	6.96	25	8.69	
5	7.04	-	-	
10	7.38	-	-	


Note 1: The path loss from EUT to Spectrum analyzer was measured and used for test. Path loss (S/A's correction factor) = Cable A + Power Splitter




APPENDIX II

Unwanted Emissions (Radiated) Test Plot

GFSK & Lowest & X & Hor



GFSK & Highest & X & Hor




TRF-RF-237(07)210316 Pages: 41 / 42



GFSK & Lowest & X & Hor

Detector Mode: AV



TRF-RF-237(07)210316 Pages: 42 / 42