

DRTECH CORPORATION.

DATE: Nov 13, 2019

TIMCO ENGINEERING Inc.

849 N.W. State Road 45 P.O. Box 370 Newberry, Florida 32669

Attn: Director of Certification

SUBJECT: SparkLAN Communications Inc.

FCC ID: RYK-WUBR508N

Class II Permissive Change

Orig. Grant Date : June 11, 2012


This is a description of similarities and differences between the versions of the X-ray Detectors.

Variant models are identical to basic model except for size of exterior and Scintillator type.

	Basic Model	Variant model
Product Name	Flat Panel Digital X-ray Detectors	Flat Panel Digital X-ray Detectors
Model Name	EVS 4343W	EVS 4343WG EVS 3643W EVS 3643WG EVS 4343WP EVS 3643WP
Dimension	460 x 460 x 15	460 x 460 x 15 - EVS 4343WG, EVS 4343WG 460 x 386 x 15 - EVS 3643W, EVS 3643WG, EVS 3643WP
TFT Type	TFT –amorphous Silicon	TFT –amorphous Silicon - EVS 4343WG, EVS 3643W, EVS 3643WG TFT type : TFT – IGZO - EVS 4343WP, EVS 3643WP
Scintillator type	Csl:TI	Csl:TI - EVS 3643W, EVS4343WP, EVS 3643WP GoS - EVS 4343WG, EVS 3643WG

Should you have any questions or comments concerning the above, please contact the undersigned.

Sincerely Yours,

A handwritten signature in black ink, consisting of a stylized 'L' followed by a horizontal line and a vertical line extending downwards.

2019. Nov. 13 | Dongwook, SHU Regulatory Affairs Responsible
DRTECH Corporation