

RF Exposure Evaluation declaration

Product Name : Fanless Medical Grade Box PC
Model No. : xxxMEDPC-6200xxxxxxxxxxx (The x
can be 0 to 9 , a to z , A to Z ,”-“, or
blank for marketing purpose)

Applicant : ONYX HEALTHCARE INC.

Address : 2F., NO.135, LANE 235, PAO CHIAO RD.,
XINDIAN DIST., NEW TAIPEI CITY 231,
TAIWAN (R.O.C.)

Date of Receipt : Aug. 08, 2017

Date of Declaration : Oct. 20, 2017

Report No. : 1780161R-SAUSP03V00

Report Version : V1.0



The test results relate only to the samples tested.

The test results shown in the test report are traceable to the national/international standard through the calibration of the equipment and evaluated measurement uncertainty herein.

This report must not be used to claim product endorsement by TAF or any agency of the government.

The test report shall not be reproduced without the written approval of DEKRA Testing and Certification Co., Ltd.

1. GENERAL INFORMATION

1.1. EUT Description

Product Name	Fanless Medical Grade Box PC
Model No.	xxxMEDPC-6200xxxxxxxxx (The x can be 0 to 9 , a to z , A to Z ,”-“, or blank for marketing purpose)
Trade Name	onyx
FCC ID.	RZ5-MEDPC-6200
Frequency Range	2412-2462MHz for 802.11b/g/n-20BW, 2422-2452MHz for 802.11n-40BW BT/BLE: 2402 – 2480MHz
Number of Channels	802.11b/g/n-20MHz: 11, n-40MHz: 7 BT: 79, BLE: 40CH
Data Speed	802.11b: 1-11Mbps, 802.11g: 6-54Mbps, 802.11n: up to 150Mbps
Channel separation	802.11b/g/n: 5 MHz BT: 1M, BLE: 2M
Type of Modulation	802.11b:DSSS (DBPSK, DQPSK, CCK) 802.11g/n:OFDM (BPSK, QPSK, 16QAM, 64QAM) FHSS: GFSK(1Mbps) / π /4DQPSK(2Mbps) / 8DPSK(3Mbps)
Antenna Type	Dipole Antenna
Channel Control	Auto
Antenna Gain	Refer to the table “Antenna List”

1.2. Antenna List :

No.	Manufacturer	Part No.	Antenna Type	Peak Gain
1	ARISTOTLE	RFA-02-C2M2-M32-3	Dipole Antenna	2.42 dBi for WLAN
2	ARISTOTLE	RFA-02-C2M2-M32-3	Dipole Antenna	2.42 dBi for Bluetooth

2. RF Exposure Evaluation

2.1. Limits

According to FCC 1.1310: The criteria listed in the following table shall be used to evaluate the environment impact of human exposure to radio frequency (RF) radiation as specified in 1.1307(b).

LIMITS FOR MAXIMUM PERMISSIBLE EXPOSURE (MPE)

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (mW/cm ²)	Average Time (Minutes)
(A) Limits for Occupational/ Control Exposures				
300-1500	--	--	F/300	6
1500-100,000	--	--	5	6
(B) Limits for General Population/ Uncontrolled Exposures				
300-1500	--	--	F/1500	30
1500-100,000	--	--	1	30

F= Frequency in MHz

Friis Formula

Friis transmission formula: $P_d = (P_{out} * G) / (4 * \pi * R^2)$

Where

P_d = power density in mW/cm²

P_{out} = output power to antenna in mW

G = gain of antenna in linear scale

π = 3.1416

R = distance between observation point and center of the radiator in cm

P_d is the limit of MPE, 1 mW/cm². If we know the maximum gain of the antenna and the total power input to the antenna, through the calculation, we will know the distance r where the MPE limit is reached.

Simultaneous transmission MPE test exclusion applies when the sum of the MPE ratios for all simultaneously transmitting antennas incorporated in a host device is ≤ 1.0

2.2. Test Result of RF Exposure Evaluation

Product : Fanless Medical Grade Box PC
 Test Item : RF Exposure Evaluation
 Test Site : N/A

Operation Frequency	2412MHz-2462MHz 2422MHz-2452MHz
Maximum Conducted output power	23.57dBm
Antenna gain	2.42dBi

Output Power Into Antenna & RF Exposure Evaluation Distance:

Output Power to Antenna (mW)	Power Density at R = 20 cm (mW/cm ²)
227.51	0.079

Power density is lower than the limit (1 mW/cm²).

Operation Frequency	2402-2480MHz
Maximum Conducted output power	6.7dBm
Antenna gain	2.42dBi

Output Power Into Antenna & RF Exposure Evaluation Distance:

Output Power to Antenna (mW)	Power Density at R = 20 cm (mW/cm ²)
4.67735	0.0016

Power density is lower than the limit (1 mW/cm²).

2.3. calculations for Multi-Transmitter

Mode	Exposure Calculations	result	Limit	Pass/Fail
WLAN	0.079	0.095	1	Pass
BT	0.016			