

---

# FCC Test Report

---

Report No.: AGC02235151201FE03

**FCC ID** : SLR550BT  
**APPLICATION PURPOSE** : Original Equipment  
**PRODUCT DESIGNATION** : Blood Pressure Monitor  
**BRAND NAME** : iHealth  
**MODEL NAME** : KN-550BT  
**CLIENT** : iHealth Labs Inc.  
**DATE OF ISSUE** : Dec.18,2015  
**STANDARD(S)** : FCC Part 15 Rules  
**TEST PROCEDURE(S)**  
**REPORT VERSION** : V1.0

Attestation of Global Compliance (Shenzhen) Co., Ltd



**CAUTION:**

This report shall not be reproduced except in full without the written permission of the test laboratory and shall not be quoted out of context.



**REPORT REVISE RECORD**

<b>Report Version</b>	<b>Revise Time</b>	<b>Issued Date</b>	<b>Valid Version</b>	<b>Notes</b>
V1.0	/	Dec.18,2015	Valid	Original Report

**TABLE OF CONTENTS**

- 1. VERIFICATION OF CONFORMITY..... 5**
- 2. GENERAL INFORMATION..... 6**
  - 2.1. PRODUCT DESCRIPTION ..... 6
  - 2.2. TABLE OF CARRIER FREQUENCYS ..... 6
- 3. MEASUREMENT UNCERTAINTY ..... 7**
- 4. DESCRIPTION OF TEST MODES ..... 7**
- 5. SYSTEM TEST CONFIGURATION..... 9**
  - 5.1. CONFIGURATION OF EUT SYSTEM..... 9
  - 5.2. EQUIPMENT USED IN EUT SYSTEM..... 9
  - 5.3. SUMMARY OF TEST RESULTS ..... 9
- 6. TEST FACILITY ..... 10**
- 7. ALL TEST EQUIPMENT LIST ..... 10**
- 8. RADIATED EMISSION ..... 10**
  - 8.1 TEST LIMIT ..... 12
  - 8.2. MEASUREMENT PROCEDURE ..... 13
  - 8.3. TEST SETUP..... 15
  - 8.4. TEST RESULT..... 17
- 9. BAND EDGE EMISSION ..... 30**
  - 9.1. MEASUREMENT PROCEDURE ..... 30
  - 9.2 TEST SETUP..... 30
  - 9.3 RADIATED TEST RESULT..... 31
- 10 20DB BANDWIDTH ..... 35**
  - 10.1. MEASUREMENT PROCEDURE ..... 35
  - 10.2. TEST SET-UP ..... 35
  - 10.3. LIMITS AND MEASUREMENT RESULTS..... 35
- 11. FCC LINE CONDUCTED EMISSION TEST..... 38**

11.1. LIMITS OF LINE CONDUCTED EMISSION TEST..... 38

11.2. BLOCK DIAGRAM OF LINE CONDUCTED EMISSION TEST ..... 38

11.3. PRELIMINARY PROCEDURE OF LINE CONDUCTED EMISSION TEST..... 39

11.4. FINAL PROCEDURE OF LINE CONDUCTED EMISSION TEST ..... 39

11.5. TEST RESULT OF LINE CONDUCTED EMISSION TEST ..... 39

**APPENDIX A: PHOTOGRAPHS OF TEST SETUP ..... 40**


**APPENDIX B: PHOTOGRAPHS OF EUT ..... 42**

## 1. VERIFICATION OF CONFORMITY

<b>Applicant</b>	iHealth Labs Inc.
<b>Address</b>	719N.Shoreline Blvd,Mountain View,CA94043,USA
<b>Manufacturer</b>	Andon Health Co.Ltd.
<b>Address</b>	No.3 JinPing Street YaAn Road Nankai District Tianjin,China
<b>Product Designation</b>	Blood Pressure Monitor
<b>Brand Name</b>	iHealth
<b>Test Model</b>	KN-550BT
<b>Date of test</b>	Dec.11,2015 to Dec.15,2015
<b>Deviation</b>	None
<b>Condition of Test Sample</b>	Normal
<b>Report Template</b>	AGCRT-US-BR/RF

We hereby certify that:

The above equipment was tested by Dongguan Precise Testing Service Co., Ltd. The test data, data evaluation, test procedures, and equipment configurations shown in this report were made in accordance with the procedures given in ANSI C63.4 (2009) and the energy emitted by the sample EUT tested as described in this report is in compliance with radiated emission limits of FCC Rules Part 15.249.



Tested By \_\_\_\_\_  
Time Huang(Huang Nanhui) Dec.18,2015



Reviewed By \_\_\_\_\_  
Forrest Lei(Lei Yonggang) Dec.18,2015



Approved By \_\_\_\_\_  
Solger Zhang(Zhang Hongyi)  
Authorized Officer Dec.18,2015

## 2. GENERAL INFORMATION

### 2.1. PRODUCT DESCRIPTION

A major technical description of EUT is described as following

<b>Operation Frequency</b>	2.402 GHz to 2.480GHz
<b>RF Output Power</b>	-4.7dBm(Max)
<b>Bluetooth Version</b>	V4.0
<b>Modulation</b>	GFSK
<b>Number of channels</b>	40
<b>Hardware Version</b>	N/A
<b>Software Version</b>	N/A
<b>Antenna Designation</b>	Ceramic Antenna (Met 15.203 Antenna requirement)
<b>Antenna Gain</b>	5dBi
<b>Power Supply</b>	DC6V by battery

### 2.2. TABLE OF CARRIER FREQUENCIES

Frequency Band	Channel Number	Frequency
2400~2483.5MHZ	0	2402MHZ
	1	2404MHZ
	:	:
	19	2440 MHZ
	20	2442 MHZ
	:	:
	38	2478 MHZ
	39	2480 MHZ

### 3. MEASUREMENT UNCERTAINTY

The reported uncertainty of measurement  $y \pm U$ , where expanded uncertainty  $U$  is based on a standard uncertainty multiplied by a coverage factor of  $k=2$ , providing a level of confidence of approximately 95 %.

No.	Item	Uncertainty
1	Conducted Emission Test	$\pm 3.18\text{dB}$
2	All emissions, radiated	$\pm 3.91\text{dB}$
3	Temperature	$\pm 0.5^\circ\text{C}$
4	Humidity	$\pm 2\%$


### 4. DESCRIPTION OF TEST MODES

NO.	TEST MODE DESCRIPTION
1	Low channel GFSK
2	Middle channel GFSK
3	High channel GFSK
4	BT Link

Note:

- All the test modes can be supply by battery, only the result of the worst case was recorded in the report, if no other cases.
- For Radiated Emission, 3axis were chosen for testing for each applicable mode.
- The EUT used new battery when tested.

Software Setting






## 5. SYSTEM TEST CONFIGURATION

### 5.1. CONFIGURATION OF EUT SYSTEM

Configure 1: (Normal hopping)



Configure 2: (Control continuous TX)



### 5.2. EQUIPMENT USED IN EUT SYSTEM

Item	Equipment	Model No.	ID or Specification	Remark
1	Blood Pressure Monitor	iHealth	KN-550BT	EUT
2	Control box	N/A	N/A	A.E
3	PC	Dell	INSPIRON	A.E
6	Temporary Antenna Connector	T10	N/A	A.E

### 5.3. SUMMARY OF TEST RESULTS

FCC RULES	DESCRIPTION OF TEST	RESULT
§15.249	Radiated Emission	Compliant
§15.249	Band Edges	Compliant
§15.207	Conduction Emission	N/A
N/A	BANDWIDTH	Compliant

## 6. TEST FACILITY

<b>Site</b>	Dongguan Precise Testing Service Co., Ltd.
<b>Location</b>	Building D,Baoding Technology Park,Guangming Road2,Dongcheng District, Dongguan, Guangdong, China,
<b>FCC Registration No.</b>	371540
<b>Description</b>	The test site is constructed and calibrated to meet the FCC requirements in documents ANSI C63.4:2009.

## 7. ALL TEST EQUIPMENT LIST

FOR RADIATED EMISSION TEST (BELOW 1GHZ)

Radiated Emission Test Site					
Name of Equipment	Manufacturer	Model Number	Serial Number	Last Calibration	Due Calibration
EMI Test Receiver	Rohde & Schwarz	ESCI	101417	July 4, 2015	July 3, 2016
Trilog Broadband Antenna (25M-1GHz)	SCHWARZBECK	VULB9160	9160-3355	July 4, 2015	July 3, 2016
Signal Amplifier	SCHWARZBECK	BBV 9475	9745-0013	July 4, 2015	July 3, 2016
RF Cable	SCHWARZBECK	AK9515E	96221	July 4, 2015	July 3, 2016
3m Anechoic Chamber	CHENGYU	966	PTS-001	June 6, 2015	June 5, 2016
MULTI-DEVICE Positioning Controller	Max-Full	MF-7802	MF780208339	N/A	N/A
Active loop antenna (9K-30MHz)	Schwarzbeck	FMZB1519	1519-038	June 6, 2015	June 5, 2016
Spectrum analyzer	Agilent	E4407B	MY46185649	June 6, 2015	June 5, 2016
Radiation Cable 1	MXT	RS1	R005	June 6, 2015	June 5, 2016
Radiation Cable 2	MXT	RS1	R006	June 6, 2015	June 5, 2016

## FOR RADIATED EMISSION TEST (1GHZ ABOVE)

Radiated Emission Test Site					
Name of Equipment	Manufacturer	Model Number	Serial Number	Last Calibration	Due Calibration
EMI Test Receiver	Rohde & Schwarz	ESCI	101417	July 4, 2015	July 3, 2016
Horn Antenna (1G-18GHz)	SCHWARZBECK	BBHA9120D	9120D-1246	July 11, 2015	July 10, 2016
Spectrum Analyzer	Agilent	E4411B	MY4511453	July 4, 2015	July 3, 2016
Signal Amplifier	SCHWARZBECK	BBV 9718	9718-269	July 7, 2015	July 6, 2016
RF Cable	SCHWARZBECK	AK9515H	96220	July 8, 2015	July 7, 2016
3m Anechoic Chamber	CHENGYU	966	PTS-001	June 6, 2015	June 5, 2016
MULTI-DEVICE Positioning Controller	Max-Full	MF-7802	MF780208339	N/A	N/A
Horn Ant (18G-40GHz)	Schwarzbeck	BBHA 9170	9170-181	June 6, 2015	June 5, 2016
Radiation Cable 1	MXT	RS1	R005	June 6, 2015	June 5, 2016
Radiation Cable 2	MXT	RS1	R006	June 6, 2015	June 5, 2016

Conducted Emission Test Site					
Name of Equipment	Manufacturer	Model Number	Serial Number	Last Calibration	Due Calibration
EMI Test Receiver	- Rohde & Schwarz	ESCI	101417	July 4, 2015	July 3, 2016
Artificial Mains Network	Narda	L2-16B	000WX31025	July 8, 2015	July 7, 2016
Artificial Mains Network (AUX)	Narda	L2-16B	000WX31026	July 8, 2015	July 7, 2016
RF Cable	SCHWARZBECK	AK9515E	96222	July 4, 2015	July 3, 2016
Shielded Room	CHENGYU	843	PTS-002	June 6,2015	June 5,2016
Conduction Cable	MXT	SE1	S003	June 6,2015	June 5,2016

## 8. RADIATED EMISSION

### 8.1 TEST LIMIT

#### Standard FCC15.249

Fundamental Frequency	Field Strength of Fundamental (millivolts/meter)	Field Strength of Harmonics (microvolts/meter)
900-928MHz	50	500
2400-2483.5MHz	50	500
5725-5875MHz	50	500
24.0-24.25GHz	250	2500

#### Standard FCC 15.209

Frequency (MHz)	Distance Meters	Field Strengths Limit	
		$\mu$ V/m	dB( $\mu$ V)/m
0.009 ~ 0.490	300	2400/F(kHz)	---
0.490 ~ 1.705	30	24000/F(kHz)	---
1.705 ~ 30	30	30	---
30 ~ 88	3	100	40.0
88 ~ 216	3	150	43.5
216 ~ 960	3	200	46.0
960 ~ 1000	3	500	54.0
Above 1000	3	Other:74.0 dB( $\mu$ V)/m (Peak) 54.0 dB( $\mu$ V)/m (Average)	

Remark: (1) Emission level  $\text{dB}\mu\text{V} = 20 \log \text{Emission level } \mu\text{V/m}$   
(2) The smaller limit shall apply at the cross point between two frequency bands.  
(3) Distance is the distance in meters between the measuring instrument, antenna and the closest point of any part of the device or system.

## 8.2. MEASUREMENT PROCEDURE

1. Configure the EUT according to ANSI C63.4. The EUT was placed on the top of the turntable 0.8 meter above ground. The phase center of the receiving antenna mounted on the top of a height-variable antenna tower was placed 3 meters far away from the turntable.
2. Power on the EUT and all the supporting units. The turntable was rotated by 360 degrees to determine the position of the highest radiation.
3. The height of the broadband receiving antenna was varied between one meter and four meters above ground to find the maximum emissions field strength of both horizontal and vertical polarization.
4. For each suspected emissions, the antenna tower was scan (from 1 M to 4 M) and then the turntable was rotated (from 0 degree to 360 degrees) to find the maximum reading.
5. Set the test-receiver system to Peak or CISPR quasi-peak Detect Function with specified bandwidth under Maximum Hold Mode.
6. For emissions above 1GHz, use 1.5MHz VBW and RBW for peak reading. Then 1.5MHz RBW and 10Hz VBW for average reading in spectrum analyzer.

Place the measurement antenna away from each area of the EUT determined to be a source of emissions at the specified measurement distance, while keeping the measurement antenna aimed at the source of emissions at each frequency of significant emissions, with polarization oriented for maximum response. The measurement antenna may have to be higher or lower than the EUT, depending on the radiation pattern of the emission and staying aimed at the emission source for receiving the maximum signal. The final measurement antenna elevation shall be that which maximizes the emissions. The measurement antenna elevation for maximum emissions shall be restricted to a range of heights of from 1 m to 4 m above the ground or reference ground plane.

7. When the radiated emissions limits are expressed in terms of the average value of the emissions, and pulsed operation is employed, the measurement field strength shall be determined by averaging over one complete pulse train, including blanking intervals, as long as the pulse train does not exceed 0.1 seconds. As an alternative (provided the transmitter operates for longer than 0.1 seconds) or in cases where the pulse train exceeds 0.1 seconds, the measured field strength shall be determined from the average absolute voltage during a 0.1 second interval during which the field strength is at its maximum values.
8. If the emissions level of the EUT in peak mode was 3 dB lower than the average limit specified, then testing will be stopped and peak values of EUT will be reported, otherwise, the emissions which do not have 3 dB margin will be repeated one by one using the quasi-peak method for below 1GHz.
9. For testing above 1GHz, the emissions level of the EUT in peak mode was lower than average limit (that means the emissions level in peak mode also complies with the limit in average mode), then testing will be stopped and peak values of EUT will be reported, otherwise, the emissions will be measured in average mode again and reported.
10. In case the emission is lower than 30MHz, loop antenna has to be used for measurement and the recorded data should be QP measured by receiver. High - Low scan is not required in this case.


The following table is the setting of spectrum analyzer and receiver.

<b>Spectrum Parameter</b>	<b>Setting</b>
Start ~Stop Frequency	9KHz~150KHz/RB 200Hz for QP
Start ~Stop Frequency	150KHz~30MHz/RB 9KHz for QP
Start ~Stop Frequency	30MHz~1000MHz/RB 120KHz for QP
Start ~Stop Frequency	1GHz~26.5GHz 1.5MHz/1.5MHz for Peak, 1.5MHz/10Hz for Average


<b>Receiver Parameter</b>	<b>Setting</b>
Start ~Stop Frequency	9KHz~150KHz/RB 200Hz for QP
Start ~Stop Frequency	150KHz~30MHz/RB 9KHz for QP
Start ~Stop Frequency	30MHz~1000MHz/RB 120KHz for QP

8.3. TEST SETUP


Radiated Emission Test-Setup Frequency Below 30MHz



RADIATED EMISSION TEST SETUP 30MHz-1000MHz



RADIATED EMISSION TEST SETUP ABOVE 1000MHz





**8.4. TEST RESULT**


**(Worst modulation: GFSK)**

**RADIATED EMISSION BELOW 30MHZ**

No emission found between lowest internal used/generated frequencies to 30MHz.

**RADIATED EMISSION BELOW 1GHZ**

RADIATED EMISSION TEST- (30MHZ-1GHZ)-LOW CHANNEL-HORIZONTAL



Site: site #1	Polarization: <i>Horizontal</i>	Temperature: 23.3
Limit: FCC Class B 3M Radiation	Power:	Humidity: 56.9 %
EUT: Blood Pressure Monitor	Distance:	
M/N: KN-550BT		
Mode: Low Channel TX		
Note:		


No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna	Table	Comment
		MHz	dBuV	dB/m	dBuV/m	dBuV/m	dB		Height	Degree	
1		41.3167	3.80	11.81	15.61	40.00	-24.39	peak			
2		157.7167	0.16	15.32	15.48	43.50	-28.02	peak			
3		464.8833	0.24	20.75	20.99	46.00	-25.01	peak			
4		751.0333	1.08	26.64	27.72	46.00	-18.28	peak			
5	*	851.2667	1.65	27.34	28.99	46.00	-17.01	peak			
6		985.4500	1.71	29.66	31.37	54.00	-22.63	peak			

**RESULT: PASS**





RADIATED EMISSION TEST- (30MHZ-1GHZ)- MIDDLE CHANNEL -VERTICAL



Site: site #1  
 Limit: FCC Class B 3M Radiation  
 EUT: Blood Pressure Monitor  
 M/N: KN-550BT  
 Mode: Middle Channel TX  
 Note:

Polarization: *Vertical*  
 Power:  
 Distance:


Temperature: 23.3  
 Humidity: 56.9 %

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna	Table	Comment
		MHz	dBuV	dB/m	dBuV/m	dBuV/m	dB		Height	Degree	
									cm	degree	
1		41.3167	11.24	8.81	20.05	40.00	-19.95	peak			
2		169.0333	-0.29	14.76	14.47	43.50	-29.03	peak			
3		350.1000	0.31	18.74	19.05	46.00	-26.95	peak			
4		636.2500	1.07	23.54	24.61	46.00	-21.39	peak			
5		793.0667	3.00	27.22	30.22	46.00	-15.78	peak			
6	*	903.0000	2.12	28.69	30.81	46.00	-15.19	peak			

**RESULT: PASS**

- Note:** 1. Factor=Antenna Factor + Cable loss, Margin=Measurement-Limit.  
 2. The "Factor" value can be calculated automatically by software of measurement system.

RADIATED EMISSION TEST- (30MHZ-1GHZ)-HIGH CHANNEL-HORIZONTAL



Site: site #1  
 Limit: FCC Class B 3M Radiation  
 EUT: Blood Pressure Monitor  
 M/N: KN-550BT  
 Mode: High Channel TX  
 Note:

Polarization: *Horizontal*  
 Power:  
 Distance:

Temperature: 23.3  
 Humidity: 56.9 %


No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dB/m	dBuV/m	dBuV/m	dB		cm	degree	
1		141.5500	0.09	15.21	15.30	43.50	-28.20	peak			
2		332.3167	1.63	17.56	19.19	46.00	-26.81	peak			
3		479.4333	0.98	20.91	21.89	46.00	-24.11	peak			
4		634.6333	1.12	23.81	24.93	46.00	-21.07	peak			
5		751.0333	1.58	26.64	28.22	46.00	-17.78	peak			
6	*	851.2667	2.65	27.34	29.99	46.00	-16.01	peak			

RESULT: PASS



**RADIATED EMISSION ABOVE 1GHZ**

**RADIATED EMISSION TEST- (ABOVE 1GHZ)-LOW CHANNEL-HORIZONTAL**




Site: Conduction Polarization: *Horizontal* Temperature: 26  
 Limit: FCC Class B 3M Radiation above 1GHZ(PK)- Power: Humidity: 60 %  
 EUT: Blood Pressure Monitor Distance: 3m  
 M/N: KN-550BT  
 Mode: Low Channel TX  
 Note:

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dBuV/m	dBuV/m	dBuV/m	dB		cm	degree	
1		2402.000	105.19	-9.68	95.51	114.00	-18.49	peak			
2		3250.000	52.14	-8.12	44.02	74.00	-29.98	peak			
3		4558.333	47.26	-2.96	44.30	74.00	-29.70	peak			
4	*	2402.000	92.95	-9.68	83.27	94.00	-10.73	AVG	100	139	

**RESULT: PASS**

RADIATED EMISSION TEST- (ABOVE 1GHZ)-LOW CHANNEL- VERTICAL




Site: Conduction Polarization: *Vertical* Temperature: 26  
 Limit: FCC Class B 3M Radiation above 1GHZ(PK)- Power: Humidity: 60 %  
 EUT: Blood Pressure Monitor Distance: 3m  
 M/N: KN-550BT  
 Mode: Low Channel TX  
 Note:

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dBuV/m	dBuV/m	dBuV/m	dB		cm	degree	
1		2402.000	105.26	-9.68	95.58	114.00	-18.42	peak			
2		3041.667	52.60	-8.32	44.28	74.00	-29.72	peak			
3		4983.333	46.21	-1.84	44.37	74.00	-29.63	peak			
4	*	2402.000	93.12	-9.68	83.44	94.00	-10.56	AVG	100	288	

RESULT: PASS



RADIATED EMISSION TEST- (ABOVE 1GHZ)-MIDDLE CHANNEL-HORIZONTAL




Site: Conduction Polarization: *Horizontal* Temperature: 26  
 Limit: FCC Class B 3M Radiation above 1GHZ(PK)- Power: Humidity: 60 %  
 EUT: Blood Pressure Monitor Distance: 3m  
 M/N: KN-500BT  
 Mode: Middle Channel TX  
 Note:

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dBuV/m	dBuV/m	dBuV/m	dB		cm	degree	
1		2440.000	104.35	-9.64	94.71	114.00	-19.29	peak			
2		3433.333	52.41	-7.95	44.46	74.00	-29.54	peak			
3		5316.667	47.42	-1.81	45.61	74.00	-28.39	peak			
4	*	2440.000	92.90	-9.64	83.26	94.00	-10.74	AVG	100	129	

RESULT: PASS



RADIATED EMISSION TEST- (ABOVE 1GHZ)-HIGH CHANNEL-HORIZONTAL




Site: Conduction Polarization: *Horizontal* Temperature: 26  
 Limit: FCC Class B 3M Radiation above 1GHZ(PK)- Power: Humidity: 60 %  
 EUT: Blood Pressure Monitor Distance: 3m  
 M/N: KN-550BT  
 Mode: High Channel TX  
 Note:

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dBuV/m	dBuV/m	dBuV/m	dB		cm	degree	
1		2480.000	103.30	-9.59	93.71	114.00	-20.29	peak			
2		3450.000	52.94	-7.94	45.00	74.00	-29.00	peak			
3		5091.667	47.62	-1.80	45.82	74.00	-28.18	peak			
4	*	2480.000	91.32	-9.59	81.73	94.00	-12.27	AVG	100	130	

RESULT: PASS

RADIATED EMISSION TEST- (ABOVE 1GHZ)-HIGH CHANNEL- VERTICAL



Site: Conduction Polarization: *Vertical* Temperature: 26  
 Limit: FCC Class B 3M Radiation above 1GHZ(PK)- Power: Humidity: 60 %  
 EUT: Blood Pressure Monitor Distance: 3m  
 M/N: KN-550BT  
 Mode: High Channel TX  
 Note:

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dBuV/m	dBuV/m	dBuV/m	dB		cm	degree	
1		2480.000	102.92	-9.59	93.33	114.00	-20.67	peak			
2		3058.333	52.79	-8.30	44.49	74.00	-29.51	peak			
3		4591.667	46.89	-2.87	44.02	74.00	-29.98	peak			
4	*	2480.000	91.08	-9.59	81.49	94.00	-12.51	AVG	100	286	

**RESULT: PASS**

**Note:** 6~25GHz at least have 20dB margin. No recording in the test report.  
 Factor=Antenna Factor + Cable loss - Amplifier gain, Margin=Measurement-Limit.  
 The "Factor" value can be calculated automatically by software of measurement system.

**Field strength of the fundamental signal****Peak value**

Frequency	Reading Level	Factor	Measurement	Limit	Over	Antenna
(MHz)	(dBuv)	(dB/m)	(dBuv/m)	(dBuv/m)	(dB)	Polarization
2402	105.19	-9.68	95.51	114	-18.49	Horizontal
2402	105.26	-9.68	95.58	114	-18.42	Vertical
2440	104.35	-9.64	94.71	114	-19.29	Horizontal
2440	104.24	-9.64	94.60	114	-19.40	Vertical
2480	103.30	-9.59	93.71	114	-20.29	Horizontal
2480	102.92	-9.59	93.33	114	-20.67	Vertical

**Average value**

Frequency	Reading Level	Factor	Measurement	Limit	Over	Antenna
(MHz)	(dBuv)	(dB/m)	(dBuv/m)	(dBuv/m)	(dB)	Polarization
2402	92.95	-9.68	83.27	94	-10.73	Horizontal
2402	93.12	-9.68	83.44	94	-10.56	Vertical
2440	92.90	-9.64	83.26	94	-10.74	Horizontal
2440	92.40	-9.64	82.76	94	-11.24	Vertical
2480	91.32	-9.59	81.73	94	-12.27	Horizontal
2480	91.08	-9.59	81.49	94	-12.51	Vertical

## 9. BAND EDGE EMISSION

### 9.1. MEASUREMENT PROCEDURE

1The EUT operates at hopping-off test mode. The lowest or highest channels are tested to verify the largest transmission and spurious emissions power at the continuous transmission mode.


2Max hold the trace of the setp 1,and the EUT operates at hopping-on test mode to verify the largest spurious emissions power.

3Set the spectrum analyzer in the following setting in order to capture the lower and upper band-edges of the emission: (a) PEAK: RBW=VBW=1.5MHz / Sweep=AUTO

(b) AVERAGE: RBW=1.5MHz ; VBW=1/on time(1KHz) / Sweep=AUTO

### 9.2 TEST SETUP


RADIATED EMISSION TEST SETUP



9.3 RADIATED TEST RESULT

(Worst modulation: GFSK)

TEST PLOT OF BAND EDGE FOR LOW CHANNEL-Horizontal



Site: site #1 Polarization: *Horizontal* Temperature: 26  
 Limit: FCC Class B 3M Radiation above 1GHZ(PK) Power: Humidity: 60 %  
 EUT: Blood Pressure Monitor Distance:  
 M/N: KN-550BT  
 Mode: Low Channel TX  
 Note:

No.	Mk	Freq.	Reading	Factor	Measurement	Limit	Over	Detector	Antenna Height	Table Degree	Comment
		MHz	dBuV	dB/m	dBuV/m	dBuV/m	dB		cm	degree	
1		2277.558	29.11	10.19	39.30	74.00	-34.70	peak			
2		2390.000	28.62	10.31	38.93	74.00	-35.07	peak			
3	*	2402.000	85.41	10.32	95.73	74.00	21.73	peak			










**10 20DB BANDWIDTH**

**10.1. MEASUREMENT PROCEDURE**

1. Connect EUT RF output port to the Spectrum Analyzer through an RF attenuator
2. Set the EUT Work on the top, the middle and the bottom operation frequency individually.
3. Set Span = approximately 2 to 3 times the 20 dB bandwidth, centered on a hopping channel  
 RBW ≥ 1% of the 20 dB bandwidth, VBW ≥ RBW; Sweep = auto; Detector function = peak
4. Set SPA Trace 1 Max hold, then View.

**10.2. TEST SET-UP**

(BLOCK DIAGRAM OF CONFIGURATION)




Note: The EUT has been used temporary antenna connector for testing.


**10.3. LIMITS AND MEASUREMENT RESULTS**

LIMITS AND MEASUREMENT RESULTS			
Applicable Limits	Measurement Result		
	Test Data (MHz)		Criteria
N/A	Low Channel	1.246	PASS
	Middle Channel	1.235	PASS
	High Channel	1.242	PASS


TEST PLOT OF BANDWIDTH FOR LOW CHANNEL



TEST PLOT OF BANDWIDTH FOR MIDDLE CHANNEL



TEST PLOT OF BANDWIDTH FOR HIGH CHANNEL



**11. FCC LINE CONDUCTED EMISSION TEST**


**11.1. LIMITS OF LINE CONDUCTED EMISSION TEST**

Frequency	Maximum RF Line Voltage	
	Q.P.( dBuV)	Average( dBuV)
150kHz~500kHz	66-56	56-46
500kHz~5MHz	56	46
5MHz~30MHz	60	50

Note:

1. The lower limit shall apply at the transition frequency.
2. The limit decreases linearly with the logarithm of the frequency in the range 0.15 MHz to 0.50 MHz.

**11.2. BLOCK DIAGRAM OF LINE CONDUCTED EMISSION TEST**



### **11.3. PRELIMINARY PROCEDURE OF LINE CONDUCTED EMISSION TEST**

1. The equipment was set up as per the test configuration to simulate typical actual usage per the user's manual. When the EUT is a tabletop system, a wooden table with a height of 0.8 meters is used and is placed on the ground plane as per ANSI C63.4 (see Test Facility for the dimensions of the ground plane used). When the EUT is a floor-standing equipment, it is placed on the ground plane which has a 3-12 mm non-conductive covering to insulate the EUT from the ground plane.
2. Support equipment, if needed, was placed as per ANSI C63.4.
3. All I/O cables were positioned to simulate typical actual usage as per ANSI C63.4.
4. All support equipments received AC120V/60Hz power from a LISN, if any.
5. The EUT received DC charging voltage by PC or by adapter which received 120V/60Hz power by a LISN.
6. The test program was started. Emissions were measured on each current carrying line of the EUT using a spectrum Analyzer / Receiver connected to the LISN powering the EUT. The LISN has two monitoring points: Line 1 (Hot Side) and Line 2 (Neutral Side). Two scans were taken: one with Line 1 connected to Analyzer / Receiver and Line 2 connected to a 50 ohm load; the second scan had Line 1 connected to a 50 ohm load and Line 2 connected to the Analyzer / Receiver.
7. Analyzer / Receiver scanned from 150 kHz to 30MHz for emissions in each of the test modes.
8. During the above scans, the emissions were maximized by cable manipulation.
9. The test mode(s) were scanned during the preliminary test.

Then, the EUT configuration and cable configuration of the above highest emission level were recorded for reference of final testing.



### **11.4. FINAL PROCEDURE OF LINE CONDUCTED EMISSION TEST**

- EUT and support equipment was set up on the test bench as per step 2 of the preliminary test.
2. A scan was taken on both power lines, Line 1 and Line 2, recording at least the six highest emissions. Emission frequency and amplitude were recorded into a computer in which correction factors were used to calculate the emission level and compare reading to the applicable limit. If EUT emission level was less -2dB to the A.V. limit in Peak mode, then the emission signal was re-checked using Q.P and Average detector.
3. The test data of the worst case condition(s) was reported on the Summary Data page.



### **11.5. TEST RESULT OF LINE CONDUCTED EMISSION TEST**

N/A

**APPENDIX A: PHOTOGRAPHS OF TEST SETUP**  
FCC RADIATED EMISSION TEST SETUP







**APPENDIX B: PHOTOGRAPHS OF EUT**  
TOP VIEW OF EUT



**BOTTOM VIEW OF EUT**



FRONT VIEW OF EUT



BACK VIEW OF EUT







LEFT VIEW OF EUT



RIGHT VIEW OF EUT




OPEN VIEW OF EUT







INTERNAL VIEW OF EUT-1




INTERNAL VIEW OF EUT-2



INTERNAL VIEW OF EUT-3




INTERNAL VIEW OF EUT-4





INTERNAL VIEW OF EUT-5



-----END OF REPORT-----