

RF Exposure Report

Report No.: FCC_RF_SL20060401-STR-006_MPE Rev 2.0

FCC ID: SSH-SYNK4KRX

Test Model: 0240031075

Received Date: 06/24/2020

Test Date: 06/24/2020 – 06/29/2020

Issued Date: 09/01/2020

Applicant: Stryker Endoscopy

Address: 5900 Optical Court, San Jose, CA, 95138, USA

Manufacturer: Stryker Endoscopy

Address: 5900 Optical Court, San Jose, CA, 95138, USA

Issued By: Bureau Veritas Consumer Products Services, Inc.

Lab Address: 775 Montague Expressway, Milpitas, CA 95035

Test Location (1): 775 Montague Expressway, Milpitas, CA 95035

**FCC Registration /
Designation Number:** 540430



This report is for your exclusive use. Any copying or replication of this report to or for any other person or entity, or use of our name or trademark, is permitted only with our prior written permission. This report sets forth our findings solely with respect to the test samples identified herein. The results set forth in this report are not indicative or representative of the quality or characteristics of the lot from which a test sample was taken or any similar or identical product unless specifically and expressly noted. Our report includes all of the tests requested by you and the results thereof based upon the information that you provided to us. You have 60 days from date of issuance of this report to notify us of any material error or omission caused by our negligence, provided, however, that such notice shall be in writing and shall specifically address the issue you wish to raise. A failure to raise such issue within the prescribed time shall constitute your unqualified acceptance of the completeness of this report, the tests conducted and the correctness of the report contents. Unless specific mention, the uncertainty of measurement has been explicitly taken into account to declare the compliance or non-compliance to the specification. The report must not be used by the client to claim product certification, approval, or endorsement by A2LA or any government agencies.

Table of Contents

Release Control Record	3
1 Certificate of Conformity	4
2 RF Exposure	5
2.1 Limits for Maximum Permissible Exposure (MPE).....	5
2.2 MPE Calculation Formula	5
2.3 Classification	5
2.4 Antenna Gain	5
2.5 Calculation Result of Maximum Conducted Power	6
3 Conclusion	6

Release Control Record

Issue No.	Description	Date Issued
FCC_RF_SL20060401-STR-006_MPE	Original Release	08/07/2020
FCC_RF_SL20060401-STR-006_MPE_Rev_1.0	Add series model	08/18/2020
FCC_RF_SL20060401-STR-006_MPE_Rev_2.0	Update Antenna gain and recalculate the result.	09/01/2020

1 Certificate of Conformity

Product: SYNK 4K Wireless Transmitter

Brand: Stryker

Test Model: 0240031065

Series Model: 0240031061

Sample Status: Engineering sample

Applicant: Stryker Endoscopy

Test Date: 06/24/2020 – 06/29/2020


Standards: FCC Part 2 (Section 2.1093)

KDB 447498 D01 General RF Exposure Guidance v06

IEEE C95.1-1992

The above equipment has been tested by **Bureau Veritas Consumer Products Services, Inc., Milpitas Branch**, and found compliance with the requirement of the above standards. The test record, data evaluation & Equipment Under Test (EUT) configurations represented herein are true and accurate accounts of the measurements of the sample's EMC characteristics under the conditions specified in this report.

Prepared by :


Deon Dai / Test Engineer

Date:

09/01/2020

Approved by :


Chen Ge / Engineer Reviewer

Date:

09/01/2020

2 RF Exposure

2.1 Limits for Maximum Permissible Exposure (MPE)

Frequency Range (MHz)	Electric Field Strength (V/m)	Magnetic Field Strength (A/m)	Power Density (mW/cm ²)	Average Time (minutes)
Limits For General Population / Uncontrolled Exposure				
0.3-1.34	614	1.63	(100)*	30
1.34-30	824/f	2.19/f	(180/f ²)*	30
30-300	27.5	0.073	0.2	30
300-1500	f/1500	30
1500-100,000	1.0	30

f = Frequency in MHz; *Plane-wave equivalent power density

2.2 MPE Calculation Formula

$$P_d = (P_{out} \cdot G) / (4 \cdot \pi \cdot r^2)$$

Where

P_d = power density in mW/cm²

P_{out} = output power to antenna in mW

G = gain of antenna in linear scale

π = 3.1416

R = distance between observation point and center of the radiator in cm

2.3 Classification

The antenna of this product, under normal use condition, is at least 20cm away from the body of the user.
So, this device is classified as Mobile Device.

2.4 Antenna Gain

The antenna type is Chip antenna with 2 dBi peak gain. 4x4

MIMO - Completely unCorrelated, Directive Antenna gain

4X4 MIMO	dBi	Numeric gain	
Ant 0	2	1.58	
Ant 1	2	1.58	
Ant 1	2	1.58	
Ant 2	2	1.58	
Directional Antenna Gain		3.25	dBi

2.5 Calculation Result of Maximum Conducted Power

Frequency Band (MHz)	Max Power (dBm)	Max Power (mW)	Turn-Up Tolerance	Antenna Gain (dBi)	Distance (cm)	Power Density (mW/cm ²)	Limit (mW/cm ²)
5190-5230	20.56	113.76	± 1dB	3.25	20	0.0602	1
5270-5310 5510-5710	21.84	152.76	± 1dB	3.25	20	0.0809	1
5775-5795	21.25	133.35	± 1dB	3.25	20	0.0706	1

Note:

1. Determining compliance based on the results of the compliance measurement, not taking into account measurement instrumentation uncertainty.
2. Calculate MPE from condition “1” formulas.

3 Conclusion

Conclusion:

The formula of calculated the MPE is:

$CPD1 / LPD1 + CPD2 / LPD2 + \dots \text{etc.} < 1$

CPD = Calculation power density

LPD = Limit of power density

Maximum = 0.0809 < 1

Therefore the maximum calculations of above situations are less than the “1” limit.

--- END ---