

DASY

Dipole Validation Kit

Type: D835V2

Serial: 406

Manufactured:	January 04, 2003
Last Calibration:	January 11, 2003
Recalibrated:	January 06, 2005

CH1 S11 1 U FS


1: 51.324 Ω 1.4824 Ω 282.56 pF

835.000 000 MHz

PRM
0e1

Cor
Avg
16


↑



CENTER 835.000 000 MHz

SPAN 400.000 000 MHz

[CH1] S11 LOG 5 dB/REF 0 dB 1:-34.182 dB 835.000 000 MHz



CENTER 835.000 000 MHz

SPAN 400.000 000 MHz

