

Information about the Applicant

Company Name	Ventana Medical Systems, Inc.
Grantee	Bill Richards
Address	1910 E. Innovation Park Drive
City, State, Zip	Tucson, AZ 85737
Job Number	vent0023
Model	Symphony Staining System
FCC ID	T652160000
Agent	Northwest EMC, Inc.
Approval Type	Original
Equipment Class	DXT Part 15, Low Power Transceiver, Rx Verified
Rule Part	15.225

Overview

This radio is installed in the Ventana Symphony system, which is a high volume tissue slide staining system. It utilizes RFID technology with a module manufactured by Secura Key, A Division of Soundcraft. The radio transmits at 13.56MHz at a low output power.

Recommendation

All items have been resolved and completed to my satisfaction; therefore I recommend this application for approval.

Signature

5/24/2006

Dave Tolman, TCB Committee