



## **Appendix - Dipole Calibration**

**Validation Dipole 2450MHz**

**P/N: ALS-D-2450-S-2**

**S/N: QTK-319**

# NCL CALIBRATION LABORATORIES

Calibration File No: DC-409-1  
Project Number: QTKB-Dipole Cal-5228

## CERTIFICATE OF CALIBRATION

It is certified that the equipment identified below has been calibrated in the  
**NCL CALIBRATION LABORATORIES** by qualified personnel following recognized  
procedures and using transfer standards traceable to NRC/NIST.

Quietek Validation Dipole

Manufacturer: APREL Laboratories

Part number: ALS-D-2450-S-2

Frequency: 2.45 GHz

Serial No: QTK-319

Customer: Quietek

Calibrated: 15 June 2006  
Released on: 15 June 2006

Released By: \_\_\_\_\_



### **NCL** CALIBRATION LABORATORIES

51 SPECTRUM WAY  
NEPEAN, ONTARIO  
CANADA K2R 1E6

Division of APREL Lab.  
TEL: (613) 820-4888  
FAX: (613) 820-4161

## Calibration Results Summary

The following results relate the Calibrated Dipole and should be used as a quick reference for the user.

### Mechanical Dimensions

**Length:** 53.5 mm

**Height:** 30.4 mm

### Electrical Specification


**SWR:** 1.21 U

**Return Loss:** -20.7 dB

**Impedance:** 47.7  $\Omega$

### System Validation Results

Frequency	1 Gram	10 Gram	Peak
2.45 GHz	48.07	25.65	95.6



## Introduction

This Calibration Report has been produced in line with the SSI Dipole Calibration Procedure SSI-TP-018-ALSAS. The results contained within this report are for Validation Dipole QTK-319. The calibration routine consisted of a three-step process. Step 1 was a mechanical verification of the dipole to ensure that it meets the IEEE/APREL mechanical specifications. Step 2 was an Electrical Calibration for the Validation Dipole, where the SWR, Impedance, and the Return loss were assessed. Step 3 involved a System Validation using the ALSAS-10U, along with QTK E-020 130 MHz to 26 GHz E-Field Probe Serial Number 212.

## References

SSI-TP-018-ALSAS Dipole Calibration Procedure

SSI-TP-016 Tissue Calibration Procedure

IEEE 1528 "Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body Due to Wireless Communications Devices: Experimental Techniques"

## Conditions

Dipole QTK-319 was received for calibration.

**Ambient Temperature of the Laboratory:** 22 °C +/- 0.5°C  
**Temperature of the Tissue:** 20 °C +/- 0.5°C



## Dipole Calibration Results

### Mechanical Verification

IEEE Length	IEEE Height	Measured Length	Measured Height
51.5 mm	30.4 mm	53.5 mm	30.4 mm

### Tissue Validation

Body Tissue 2450 MHz	Measured
Dielectric constant, $\epsilon_r$	52.5
Conductivity, $\sigma$ [S/m]	1.78

Electrical Calibration

Test	Result
S11 R/L	-20.7 dB
SWR	1.21 U
Impedance	47.7 $\Omega$

The Following Graphs are the results as displayed on the Vector Network Analyzer.

S11 Parameter Return Loss

S22 REVERSE REFLECTION

LOG MAGNITUDE

REF = -20.000 dB

5.000 dB/DIV



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm

MARKER 1  
2.451466 GHz  
-20.669 dB

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

SWR

S22 REVERSE REFLECTION

SWR REF=1.500 U 600.000 mU/DIV



CH 1 - S22  
REFERENCE PLANE  
0.0000 mm


MARKER 1  
2.451466 GHz  
1.208 U

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

### Smith Chart Dipole Impedance

S22 REVERSE REFLECTION  
IMPEDANCE



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm

▶ MARKER 1  
2.451466 GHz  
47.685  $\Omega$   
-8.809  $j\Omega$

MARKER TO MAX  
MARKER TO MIN


MARKER READOUT  
FUNCTIONS

2.299474 - 2.599236 GHz



**System Validation Results Using the Electrically Calibrated Dipole**

Frequency	1 Gram	10 Gram	Peak Above Feed Point
2.45 GHz	48.07	25.65	95.6



## **Test Equipment**

The test equipment used during Probe Calibration, manufacturer, model number and, current calibration status are listed and located on the main APREL server R:\NCL\Calibration Equipment\Instrument List





## **Appendix - Dipole Calibration**

**Validation Dipole 5200MHz**

**P/N: ALS-D-5200-S-2**

**S/N: QTK-320**

## NCL CALIBRATION LABORATORIES

Calibration File No: DC-410-1  
Project Number: QTKB-Dipole Cal-5229

# CERTIFICATE OF CALIBRATION

It is certified that the equipment identified below has been calibrated in the  
**NCL CALIBRATION LABORATORIES** by qualified personnel following recognized  
procedures and using transfer standards traceable to NRC/NIST.

Quietek Validation Dipole

Manufacturer: APREL Laboratories

Part number: ALS-D-5200-S-2

Frequency: 5.2 GHz

Serial No: QTK-320

Customer: Quietek

Calibrated: 15 June 2006  
Released on: 15 June 2006

Released By: \_\_\_\_\_



## **NCL** CALIBRATION LABORATORIES

51 SPECTRUM WAY  
NEPEAN, ONTARIO  
CANADA K2R 1E6

Division of APREL Lab.  
TEL: (613) 820-4988  
FAX: (613) 820-4161

### Calibration Results Summary

The following results relate the Calibrated Dipole and should be used as a quick reference for the user.

#### Mechanical Dimensions


**Length:** 23.6 mm  
**Height:** 14.0 mm

#### Electrical Specification

**SWR:** 1.57 U  
**Return Loss:** -13.15 dB  
**Impedance:** 78.2  $\Omega$

#### System Validation Results

Frequency	1 Gram
5200 GHz	58.8



## Introduction

This Calibration Report has been produced in line with the SSI Dipole Calibration Procedure SSI-TP-018. The results contained within this report are for Validation Dipole QTK-320 at 5.2 GHz. The calibration routine consisted of a three-step process. Step 1 was a mechanical verification of the dipole to ensure that it meets the IEEE mechanical specification. Step 2 was an Electrical Calibration for the Validation Dipole, where the SWR, Impedance, and the Return loss were assessed. Step 3 involved a System Validation using the ALSAS-10U, along with QTK E-020 130 MHz to 26 GHz E-Field Probe Serial Number 212.

## References

SSI-TP-018 Dipole Calibration Procedure  
SSI-TP-016 Tissue Calibration Procedure  
IEEE-1528 "Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body Due to Wireless Communications Devices: Experimental Techniques"

## Conditions

Dipole QTK-320 was received for calibration.

**Ambient Temperature of the Laboratory:** 22 °C +/- 0.5°C  
**Temperature of the Tissue:** 21 °C +/- 0.5°C



## Dipole Calibration Results

### Mechanical Verification

APREL Length	APREL Height	Measured Length	Measured Height
23.6 mm	14.0 mm	23.1 mm	14.2 mm

### Tissue Validation

Head Tissue 5200 MHz	Measured
Dielectric constant, $\epsilon_r$	39.94
Conductivity, $\sigma$ [S/m]	5.24



Electrical Calibration

Test	Result
S11 R/L	-13.15 dB
SWR	1.57 U
Impedance	78.2 $\Omega$

The Following Graphs are the results as displayed on the Vector Network Analyzer.



S11 Parameter Return Loss



SWR

S22 REVERSE REFLECTION

SWR REF=1.500 U 50.000 mU/DIV



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm


MARKER 1  
5.199988 GHz  
1.570 U

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

### Smith Chart Dipole Impedance

S22 REVERSE REFLECTION  
IMPEDANCE



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm


▶ MARKER 1  
5.199988 GHz  
78.201  $\Omega$   
-3.155  $j\Omega$

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

**System Validation Results Using a Complex Dipole Model (FDTD calculations)**

Frequency	1 Gram
5.2 GHz	58.8



**Test Equipment**



**NCL Calibration Laboratories**

Division of APREL Laboratories.

---

The test equipment used during Probe Calibration, manufacturer, model number and, current calibration status are listed and located on the main APREL server R:\NCL\Calibration Equipment\Instrument List





## **Appendix - Dipole Calibration**

**Validation Dipole 5800MHz**

**P/N: ALS-D-5800-S-2**

**S/N: QTK-321**

**NCL CALIBRATION LABORATORIES**

Calibration File No: DC-411-1  
Project Number: QTKB-Dipole Cal-5225

**CERTIFICATE OF CALIBRATION**

It is certified that the equipment identified below has been calibrated in the  
**NCL CALIBRATION LABORATORIES** by qualified personnel following recognized  
procedures and using transfer standards traceable to NRC/NIST.

Quietek Validation Dipole

Manufacturer: APREL Laboratories  
Part number: ALS-D-5800-S-2  
Frequency: 5.8 GHz  
Serial No: QTK-321

Customer: Quietek

Calibrated: 15 June 2006  
Released on: 15 June 2006

Released By: \_\_\_\_\_



**NCL CALIBRATION LABORATORIES**

51 SPECTRUM WAY  
NEPEAN, ONTARIO  
CANADA K2R 1E6

Division of APREL Lab.  
TEL: (613) 820-4988  
FAX: (613) 820-4161

### Calibration Results Summary

The following results relate the Calibrated Dipole and should be used as a quick reference for the user.

#### Mechanical Dimensions


**Length:** 21.6 mm  
**Height:** 12.6 mm

#### Electrical Specification

**SWR:** 1.78 U  
**Return Loss:** -11.0 dB  
**Impedance:** 74.8  $\Omega$

#### System Validation Results

Frequency	1 Gram
5800 GHz	57.9



## Introduction

This Calibration Report has been produced in line with the SSI Dipole Calibration Procedure SSI-TP-018. The results contained within this report are for Validation Dipole QTK-321 at 5.8 GHz. The calibration routine consisted of a three-step process. Step 1 was a mechanical verification of the dipole to ensure that it meets the IEEE mechanical specification. Step 2 was an Electrical Calibration for the Validation Dipole, where the SWR, Impedance, and the Return loss were assessed. Step 3 involved a System Validation using the ALSAS-10U, along with QTK E-020 130 MHz to 26 GHz E-Field Probe Serial Number 212.

## References

SSI-TP-018 Dipole Calibration Procedure

SSI-TP-016 Tissue Calibration Procedure

IEEE-1528 "Recommended Practice for Determining the Peak Spatial-Average Specific Absorption Rate (SAR) in the Human Body Due to Wireless Communications Devices: Experimental Techniques"

## Conditions

Dipole QTK-321 was received for calibration.

**Ambient Temperature of the Laboratory:** 22 °C +/- 0.5°C

**Temperature of the Tissue:** 21 °C +/- 0.5°C





## Dipole Calibration Results

### Mechanical Verification

APREL Length	APREL Height	Measured Length	Measured Height
21.6 mm	12.6 mm	21.2 mm	13.1 mm

### Tissue Validation

Head Tissue 5800 MHz	Measured
Dielectric constant, $\epsilon_r$	35.15
Conductivity, $\sigma$ [S/m]	6.4



**Electrical Calibration**

Test	Result
S11 R/L	-11.0 dB
SWR	1.78 U
Impedance	74.8 $\Omega$

The Following Graphs are the results as displayed on the Vector Network Analyzer.


**S11 Parameter Return Loss**

S22 REVERSE REFLECTION

LOG MAGNITUDE

REF = -10.000 dB

1.000 dB/DIV



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm

MARKER 1  
5.799512 GHz  
-11.006 dB

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

SWR

S22 REVERSE REFLECTION

SWR REF=1.900 U 50.000 mU/DIV



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm


MARKER 1  
5.799512 GHz  
1.776 U

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

### Smith Chart Dipole Impedance

S22 REVERSE REFLECTION  
IMPEDANCE



CH 4 - S22  
REFERENCE PLANE  
0.0000 mm


▶ MARKER 1  
5.799512 GHz  
74.849 Ω  
-25.220 jΩ

MARKER TO MAX  
MARKER TO MIN

MARKER READOUT  
FUNCTIONS

System Validation Results Using a Complex Dipole Model (FDTD calculations)

Frequency	1 Gram
5.8 GHz	57.9



## Test Equipment

The test equipment used during Probe Calibration, manufacturer, model number and, current calibration status are listed and located on the main APREL server R:\NCL\Calibration Equipment\Instrument List

