

TRaC Global  
100 Frobisher Business Park  
Leigh Sinton Road  
Malvern  
Worcestershire  
WR14 1BX  
UK

**Widex A/S**  
Nymoellevej 6  
DK-3540 Lynge  
Denmark

Tel.: (+45) 44 35 56 00  
Fax: (+45) 44 35 56 01  
widex@widex.com  
<http://www.widex.com>  
CVR. Nr. 1577 1100

Date: 17 September, 2010

## **FCC Labeling of Widex model C4-9**

FCC ID: TTY-C49

### To Whom It May Concern:


The dimensions of the hearing aid are too small to allow the FCC ID to be printed on the device in a font size that is readable. Consequently, the FCC ID is printed in the user instructions together with the FCC regulatory statements.

Justification is provided by 3 photographs below that show the C4-9 in 3 different angles together with a printout of the FCC ID in font size 4 point that is the minimum readable font size as defined in FCC OET Labelling Guidelines.


Bottom view



Side view



Top view



Thank you for your attention to this matter.

Sincerely,  
*Widex A/S*

Hans-Otto Bindeballe