

## Measurement Report for Device

### Device Under Test Properties

Model, Manufacturer	Dimensions [mm]	Software Version	DUT Type
Device,	70.0 x 50.0 x 8.0	3.0.0.841	5G Verification Source

### Exposure Conditions

Phantom Section	Position, Test Distance [mm]	Frequency [MHz]	Conversion Factor
5G	FRONT, 10.00	10000.0	1.0

### Hardware Setup

Phantom	Medium	Probe, Calibration Date	DAE, Calibration Date
mmWave - 1044	Air -	EUmmWV4 - SN9461_F1-55GHz, 2023-10-12	DAE4 Sn854, 2023-08-17

### Scans Setup

Scan Type	5G Scan
Grid Extents [mm]	120.0 x 120.0
Grid Steps [lambda]	0.25 x 0.25
Sensor Surface [mm]	10.0

### Measurement Results

Date	2023-12-16
Avg. Area [cm <sup>2</sup> ]	4.00
psPDn+ [W/m <sup>2</sup> ]	58.0
psPDtot+ [W/m <sup>2</sup> ]	58.2
H <sub>max</sub> [A/m]	0.427
E <sub>max</sub> [V/m]	160
max <sub>(Stot)</sub> [W/m <sup>2</sup> ]	66.5
Power Drift [dB]	0.08
IPDn	104

