

Receiver

Federal Communication Commission
Equipment Authorization Division, Application
Processing Branch
7435 Oakland Mills Road
Columbia, MD 21048

Certification and Engineering Bureau
Industry Canada
Spectrum Engineering Branch
3701 Carling Avenue, Building 94
Ottawa, Ontario K2H 8S2

Subject: *Limited Modular Approval Statement*
Date: 5th of April, 2017.

FCC Certification Number: U28AUBTEP

Only applicable for IC certification:

ISED Company Number:	1350B	UPN: (Unique Product No)	AUBTEP
HVIN (Hardware Ver. ID):	Aurora BTE power	PMN: (Product Marketing Name)	Aurora BTE power
HMN: (Host Marketing Name)	N/A	FVIN: (Firmware Version Identification Number)	N/A

TO WHOM IT MAY CONCERN

Pursuant to Paragraphs RSP-100 Issue 11 January 2016 Annex D and CFR § 15.212, we herewith declare for our module:

Modular approval requirement	Yes	No *
(a) The radio elements must have the radio frequency circuitry shielded. Physical/discrete and tuning capacitors may be located external to the shield, but must be on the module assembly.		X
*Please provide a detailed explanation if the answer is "No.": There is no complete RF shield attached to the radio module, only the ground layer in its PCB. Therefore we are only requesting a <i>limited modular approval</i> for the module.		
(b) The module shall have buffered modulation/data input(s) (if such inputs are provided) to ensure that the module will comply with the requirements set out in the applicable RSS standard / FCC part 15 under conditions of excessive data rates or over-modulation.	X	
*Please provide a detailed explanation if the answer is "No.":		
(c) The module shall have its own power supply regulation on the module. This is to ensure that the module will comply with the requirements set out in the applicable standard regardless of the design of the power supplying circuitry in the host device which houses the module.	X	
*Please provide a detailed explanation if the answer is "No.":		

(d) The module shall comply with the provisions for external power amplifiers and antennas detailed in this standard. The equipment certification submission shall contain a detailed description of the configuration of highest antenna gain for each type of antenna.	X	
*Please provide a detailed explanation if the answer is "No.":		
(e) The module shall be tested for compliance with the applicable standard in a stand-alone configuration, i.e. the module must not be inside another device during testing.	X	
*Please provide a detailed explanation if the answer is "No.":		
(f) The module shall comply with applicable RSS-102 exposure requirements and any applicable FCC RF exposure requirement which are based on the intended configuration/integration in a host.	X	
*Please provide a detailed explanation if the answer is "No.":		
<i>Only applicable for FCC certification:</i>		
(g) The module must be equipped with either a permanently affixed label or must be capable of electronically displaying its FCC identification number.	(X)	
*Please provide a detailed explanation if the answer is "No.": The module is intended to be used inside Oticon, Bernafon, Sonic and affiliated hearing instruments only, which are too small to carry the FCC and IC label legible without magnification. Therefore the labels are put in the installation guide and/or user manual, but not on the products carrying the module inside. This is possible due to a label exemption granted by Industry Canada.		
<i>Only applicable for FCC certification:</i>		
(i) The modular transmitter complies with all applicable FCC rules. Instructions for maintaining compliance are given in the user instructions.	X	

If you have any questions, please feel free to contact us at the address shown below

Best Regards,

Asif Muhammad,
Head of Regulatory Affairs, QARA
Oticon A/S, Kongebakken 9
2765 Smorum - DK-Denmark
Phone: +45 3913 7100
Email: asmu@oticon.com

INFO for applicant: LMA may be granted when **one or more** of the requirements in the table above cannot be demonstrated. LMA will also be issued in those instances where applicants can demonstrate that they will retain control over the final installation of the device, such that compliance of the end product is assured. In such cases, an operating condition on the LMA for the module must state that the module is only approved for use when installed in devices produced by a specific manufacturer. When LMA is sought, the application for equipment certification must specifically state **how control of the end product**, into which the module will be installed, will be maintained, such that full compliance of the end product is always ensured.